

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP123022\
 Data File : BP013373.D
 Acq On : 30 Dec 2022 15:40
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020732

Quant Time: Dec 30 23:39:09 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP120722.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Dec 28 00:16:35 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.928	152	900745	20.000	ng/ul	0.00
20) Naphthalene-d8	10.740	136	3794308	20.000	ng/ul	0.00
38) Acenaphthene-d10	14.575	164	2276464	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.328	188	4383200	20.000	ng/ul	0.00
79) Chrysene-d12	21.416	240	3782364	20.000	ng/ul	0.00
88) Perylene-d12	23.880	264	4532066	20.000	ng/ul	-0.01
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.317	96	154614	7.339	ng/uL	0.00
4) Pyridine-d5	3.746	84	1193276	19.940	ng/ul	0.00
7) Phenol-d5	7.093	99	1488852	20.293	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.258	67	948818	21.438	ng/ul	0.00
11) 2-Chlorophenol-d4	7.458	132	1165260	20.176	ng/ul	0.00
15) 4-Methylphenol-d8	8.646	113	1182037	19.663	ng/ul	0.00
21) Nitrobenzene-d5	9.099	128	584907	20.981	ng/ul	0.00
24) 2-Nitrophenol-d4	9.822	143	668260	21.292	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.363	165	1252303	20.543	ng/ul	0.00
31) 4-Chloroaniline-d4	10.881	131	1575745	18.310	ng/ul	0.00
46) Dimethylphthalate-d6	13.981	166	3201799	18.301	ng/ul	0.00
49) Acenaphthylene-d8	14.269	160	3862620	20.092	ng/ul	0.00
54) 4-Nitrophenol-d4	14.775	143	512574	16.287	ng/ul	0.00
60) Fluorene-d10	15.569	176	2781285	18.791	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.693	200	541022	19.181	ng/ul	0.00
73) Anthracene-d10	17.428	188	3895253	19.698	ng/ul	0.00
81) Pyrene-d10	19.663	212	4256662	19.606	ng/ul	0.00
92) Benzo(a)pyrene-d12	23.721	264	4391786	19.442	ng/ul	-0.01
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.358	88	161006	7.321	ng/uL	91
5) Pyridine	3.764	79	1185527	19.933	ng/ul	93
6) Benzaldehyde	7.064	77	604670	21.048	ng/ul	94
8) Phenol	7.116	94	1523369	20.536	ng/ul	95
10) Bis(2-Chloroethyl)ether	7.352	93	1242738	20.411	ng/ul	96
12) 2-Chlorophenol	7.487	128	1180411	19.930	ng/ul	93
13) 2-Methylphenol	8.375	108	1140316	19.717	ng/ul	96
14) 2,2'-oxybis(1-Chloropr...	8.463	45	1912367	22.889	ng/ul	98
16) Acetophenone	8.758	105	1894608	20.554	ng/ul#	94
17) N-Nitroso-di-n-propyla...	8.746	70	975637	21.048	ng/ul	92
18) 4-Methylphenol	8.705	108	1259030	19.655	ng/ul	98
19) Hexachloroethane	9.011	117	484609	20.475	ng/ul	100
22) Nitrobenzene	9.140	77	1438938	21.996	ng/ul	96
23) Isophorone	9.669	82	2661558	20.121	ng/ul	98
25) 2-Nitrophenol	9.852	139	713150	20.971	ng/ul	90
26) 2,4-Dimethylphenol	9.910	107	1372724	20.689	ng/ul	97
27) Bis(2-Chloroethoxy)met...	10.152	93	1692913	20.016	ng/ul	99
29) 2,4-Dichlorophenol	10.387	162	1220396	20.437	ng/ul	99
30) Naphthalene	10.793	128	3893547	19.621	ng/ul	99
32) 4-Chloroaniline	10.905	127	1588278	18.650	ng/ul	100
33) Hexachlorobutadiene	11.069	225	868923	21.193	ng/ul	99
34) Caprolactam	11.693	113	176518	8.669	ng/ul	91
35) 4-Chloro-3-methylphenol	12.028	107	1203749	19.753	ng/ul	100
36) 2-Methylnaphthalene	12.399	142	2641884	19.406	ng/ul	100


Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP123022\
 Data File : BP013373.D
 Acq On : 30 Dec 2022 15:40
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020732

Quant Time: Dec 30 23:39:09 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP120722.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Dec 28 00:16:35 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.616	142	2693546	19.240	ng/ul	100
39) 1,2,4,5-Tetrachloroben...	12.763	216	1634823	22.302	ng/ul	99
40) Hexachlorocyclopentadiene	12.740	237	997535	20.912	ng/ul	98
41) 2,4,6-Trichlorophenol	13.004	196	1003360	21.299	ng/ul	100
42) 2,4,5-Trichlorophenol	13.075	196	1078078	21.305	ng/ul	99
43) 1,1'-Biphenyl	13.404	154	3538501	20.697	ng/ul	99
44) 2-Chloronaphthalene	13.451	162	2805993	20.853	ng/ul	100
45) 2-Nitroaniline	13.657	65	732462	22.451	ng/ul	92
47) Dimethylphthalate	14.028	163	3163006	18.240	ng/ul	100
48) 2,6-Dinitrotoluene	14.151	165	638434	19.811	ng/ul	98
50) Acenaphthylene	14.293	152	4119821	19.933	ng/ul	100
51) 3-Nitroaniline	14.481	138	536938	17.645	ng/ul	97
52) Acenaphthene	14.634	153	2820514	19.693	ng/ul	99
53) 2,4-Dinitrophenol	14.693	184	354031	16.072	ng/ul	98
55) 4-Nitrophenol	14.793	109	390578	18.189	ng/ul	93
56) Dibenzofuran	14.969	168	3933381	19.316	ng/ul	100
57) 2,4-Dinitrotoluene	14.940	165	860743	18.388	ng/ul	95
58) 2,3,4,6-Tetrachlorophenol	15.198	232	879094	19.037	ng/ul	97
59) Diethylphthalate	15.393	149	2888503	17.135	ng/ul	99
61) Fluorene	15.622	166	3110234	19.044	ng/ul	100
62) 4-Chlorophenyl-phenyle...	15.616	204	1687480	19.539	ng/ul	99
63) 4-Nitroaniline	15.645	138	444094	16.585	ng/ul	92
66) 4,6-Dinitro-2-methylph...	15.704	198	531290	19.274	ng/ul	97
67) N-Nitrosodiphenylamine	15.828	169	2543974	20.935	ng/ul	100
68) 4-Bromophenyl-phenylether	16.510	248	1018156	21.389	ng/ul	98
69) Hexachlorobenzene	16.628	284	1159488	20.514	ng/ul	99
70) Atrazine	16.787	200	886187	18.714	ng/ul	99
71) Pentachlorophenol	16.975	266	630975	18.434	ng/ul	100
72) Phenanthrene	17.375	178	4477581	19.612	ng/ul	100
74) Anthracene	17.463	178	4501215	19.746	ng/ul	100
75) 1,2,3,4-Tetrachloroben...	13.369	216	1581869	25.295	ng/ul	99
76) Pentachlorobenzene	14.893	250	1483531	23.268	ng/ul	100
77) Carbazole	17.734	167	3768505	19.630	ng/ul	99
78) Di-n-butylphthalate	18.275	149	4187663	17.915	ng/ul	99
80) Fluoranthene	19.334	202	4929288	20.004	ng/ul	100
82) Pyrene	19.692	202	5121191	20.147	ng/ul	97
83) Butylbenzylphthalate	20.551	149	1701649	17.329	ng/ul	93
84) 3,3'-Dichlorobenzidine	21.328	252	1617989	18.936	ng/ul	99
85) Benzo(a)anthracene	21.398	228	5013080	19.965	ng/ul	100
86) Bis(2-ethylhexyl)phtha...	21.310	149	2418523	17.460	ng/ul	99
87) Chrysene	21.457	228	4742193	19.971	ng/ul	99
89) Di-n-octyl phthalate	22.251	149	4117441	16.217	ng/ul	100
90) Benzo(b)fluoranthene	23.127	252	5487537	18.988	ng/ul	100
91) Benzo(k)fluoranthene	23.174	252	5343280	18.645	ng/ul	99
93) Benzo(a)pyrene	23.774	252	4922178	19.561	ng/ul	99
94) Indeno(1,2,3-cd)pyrene	26.457	276	6979359	22.267	ng/ul	98
95) Dibenzo(a,h)anthracene	26.474	278	5784670	22.290	ng/ul	99
96) Benzo(g,h,i)perylene	27.251	276	5678841	22.568	ng/ul	99





(#) = qualifier out of range (m) = manual integration (+) = signals summed



#2
1,4-Dioxane
Concen: 7.321 ng/uL
RT: 3.358 min Scan# 24
Delta R.T. 0.000 min
Lab File: BP013373.D
Acq: 30 Dec 2022 15:40




Instrument :
BNA_P
ClientSampleId :
SSTD020732


Tgt Ion: 88 Resp: 161006
Ion Ratio Lower Upper
88 100
43 32.6 22.4 33.6
58 80.0 58.2 87.4



#5
Pyridine
Concen: 19.933 ng/uL
RT: 3.764 min Scan# 98
Delta R.T. -0.006 min
Lab File: BP013373.D
Acq: 30 Dec 2022 15:40

Tgt Ion: 79 Resp: 1185527
Ion Ratio Lower Upper
79 100
52 82.8 60.7 91.1
51 39.5 28.9 43.3









#6
Benzaldehyde
Concen: 21.048 ng/ul
RT: 7.064 min Scan# 61
Delta R.T. -0.006 min
Lab File: BP013373.D
Acq: 30 Dec 2022 15:40




Instrument :
BNA_P
ClientSampleId :
SSTD020732


Tgt Ion: 77 Resp: 604670
Ion Ratio Lower Upper
77 100
105 94.5 81.5 122.3
106 92.4 76.9 115.3



#8
Phenol
Concen: 20.536 ng/ul
RT: 7.116 min Scan# 668
Delta R.T. -0.006 min
Lab File: BP013373.D
Acq: 30 Dec 2022 15:40

Tgt Ion: 94 Resp: 1523369
Ion Ratio Lower Upper
94 100
65 29.3 21.1 31.7
66 36.6 27.0 40.6









#10
Bis(2-Chloroethyl)ether
Concen: 20.411 ng/ul
RT: 7.352 min Scan# 708
Delta R.T. 0.000 min
Lab File: BP013373.D
Acq: 30 Dec 2022 15:40




Instrument :
BNA_P
ClientSampleId :
SSTD020732


Tgt Ion: 93 Resp: 1242738
Ion Ratio Lower Upper
93 100
63 79.1 59.8 89.8
95 32.5 26.4 39.6



#12
2-Chlorophenol
Concen: 19.930 ng/ul
RT: 7.487 min Scan# 731
Delta R.T. -0.006 min
Lab File: BP013373.D
Acq: 30 Dec 2022 15:40

Tgt Ion: 128 Resp: 1180411
Ion Ratio Lower Upper
128 100
64 51.1 34.9 52.3
130 31.9 25.8 38.6





#13

2-Methylphenol

Concen: 19.717 ng/ul

RT: 8.375 min Scan# 882

Delta R.T. -0.006 min

Lab File: BP013373.D

Acq: 30 Dec 2022 15:40

Instrument :

BNA_P

ClientSampleId :





SSTD020732

Tgt Ion:108 Resp: 1140316

Ion Ratio Lower Upper

108 100

107 89.6 69.1 103.7



#14

2,2'-oxybis(1-Chloropropane)

Concen: 22.889 ng/ul

RT: 8.463 min Scan# 897

Delta R.T. 0.006 min

Lab File: BP013373.D

Acq: 30 Dec 2022 15:40




Tgt Ion: 45 Resp: 1912367


Ion Ratio Lower Upper

45 100

77 13.4 11.0 16.6

79 10.5 9.3 13.9









#16
Acetophenone
Concen: 20.554 ng/ul
RT: 8.758 min Scan# 947
Delta R.T. -0.006 min
Lab File: BP013373.D
Acq: 30 Dec 2022 15:40




Instrument :
BNA_P
ClientSampleId :
SSTD020732


Tgt Ion:105 Resp: 1894608
Ion Ratio Lower Upper
105 100
77 78.0 59.3 88.9
51 30.1 19.6 29.4#



#17
N-Nitroso-di-n-propylamine
Concen: 21.048 ng/ul
RT: 8.746 min Scan# 945
Delta R.T. 0.000 min
Lab File: BP013373.D
Acq: 30 Dec 2022 15:40

Tgt Ion: 70 Resp: 975637
Ion Ratio Lower Upper
70 100
42 65.2 45.5 68.3
101 9.8 7.9 11.9
130 20.9 17.4 26.2









#18
4-Methylphenol
Concen: 19.655 ng/ul
RT: 8.705 min Scan# 911
Delta R.T. -0.006 min
Lab File: BP013373.D
Acq: 30 Dec 2022 15:40




Instrument :
BNA_P
ClientSampleId :
SSTD020732


Tgt Ion:108 Resp: 1259030
Ion Ratio Lower Upper
108 100
107 111.6 87.4 131.2



#19
Hexachloroethane
Concen: 20.475 ng/ul
RT: 9.011 min Scan# 990
Delta R.T. -0.006 min
Lab File: BP013373.D
Acq: 30 Dec 2022 15:40

Tgt Ion:117 Resp: 484609
Ion Ratio Lower Upper
117 100
201 117.1 93.7 140.5
199 72.7 58.6 88.0









#22
Nitrobenzene
Concen: 21.996 ng/ul
RT: 9.140 min Scan# 1012
Delta R.T. -0.006 min
Lab File: BP013373.D
Acq: 30 Dec 2022 15:40




Instrument :
BNA_P
ClientSampleId :
SSTD020732


Tgt Ion: 77 Resp: 1438938
Ion Ratio Lower Upper
77 100
123 44.5 38.3 57.5
65 13.9 10.3 15.5



#23
Isophorone
Concen: 20.121 ng/ul
RT: 9.669 min Scan# 1102
Delta R.T. 0.000 min
Lab File: BP013373.D
Acq: 30 Dec 2022 15:40

Tgt Ion: 82 Resp: 2661558
Ion Ratio Lower Upper
82 100
95 6.8 5.4 8.2
138 17.6 15.3 22.9









#25
2-Nitrophenol
Concen: 20.971 ng/ul
RT: 9.852 min Scan# 1133
Delta R.T. -0.006 min
Lab File: BP013373.D
Acq: 30 Dec 2022 15:40




Instrument :
BNA_P
ClientSampleId :
SSTD020732


Tgt Ion:139 Resp: 713150
Ion Ratio Lower Upper
139 100
65 54.0 36.6 54.8
109 29.3 21.6 32.4



#26
2,4-Dimethylphenol
Concen: 20.689 ng/ul
RT: 9.910 min Scan# 1143
Delta R.T. -0.006 min
Lab File: BP013373.D
Acq: 30 Dec 2022 15:40

Tgt Ion:107 Resp: 1372724
Ion Ratio Lower Upper
107 100
121 50.8 41.8 62.6
122 87.0 72.6 108.8





#27

Bis(2-Chloroethoxy)methane

Concen: 20.016 ng/ul

RT: 10.152 min Scan# 1184

Delta R.T. 0.000 min

Lab File: BP013373.D

Acq: 30 Dec 2022 15:40

Instrument :

BNA_P

ClientSampleId :

SSTD020732





Tgt Ion: 93 Resp: 1692913

Ion Ratio Lower Upper

93 100

95 33.0 26.0 39.0

123 11.5 9.2 13.8



#29

2,4-Dichlorophenol

Concen: 20.437 ng/ul

RT: 10.387 min Scan# 1224

Delta R.T. -0.006 min

Lab File: BP013373.D

Acq: 30 Dec 2022 15:40




Tgt Ion:162 Resp: 1220396


Ion Ratio Lower Upper

162 100

164 65.1 52.5 78.7

98 34.1 26.6 39.8









#30
Naphthalene
Concen: 19.621 ng/ul
RT: 10.793 min Scan# 1293
Delta R.T. -0.006 min
Lab File: BP013373.D
Acq: 30 Dec 2022 15:40




Instrument :
BNA_P
ClientSampleId :
SSTD020732


Tgt Ion:128 Resp: 3893547
Ion Ratio Lower Upper
128 100
129 11.1 8.7 13.1
127 13.3 10.4 15.6



#32
4-Chloroaniline
Concen: 18.650 ng/ul
RT: 10.905 min Scan# 1312
Delta R.T. -0.006 min
Lab File: BP013373.D
Acq: 30 Dec 2022 15:40

Tgt Ion:127 Resp: 1588278
Ion Ratio Lower Upper
127 100
129 32.1 25.6 38.4









#33
Hexachlorobutadiene
Concen: 21.193 ng/ul
RT: 11.069 min Scan# 11
Delta R.T. -0.006 min
Lab File: BP013373.D
Acq: 30 Dec 2022 15:40




Instrument :
BNA_P
ClientSampleId :
SSTD020732


Tgt Ion:225 Resp: 868923
Ion Ratio Lower Upper
225 100
223 62.4 49.2 73.8
227 63.4 50.2 75.2



#34
Caprolactam
Concen: 8.669 ng/ul
RT: 11.693 min Scan# 1446
Delta R.T. 0.000 min
Lab File: BP013373.D
Acq: 30 Dec 2022 15:40

Tgt Ion:113 Resp: 176518
Ion Ratio Lower Upper
113 100
55 186.3 141.9 212.9
56 139.7 99.6 149.4









#35
4-Chloro-3-methylphenol
Concen: 19.753 ng/ul
RT: 12.028 min Scan# 1502
Delta R.T. 0.000 min
Lab File: BP013373.D
Acq: 30 Dec 2022 15:40




Instrument :
BNA_P
ClientSampleId :
SSTD020732


Tgt Ion:107 Resp: 1203749
Ion Ratio Lower Upper
107 100
144 26.8 21.4 32.2
142 82.8 65.8 98.6



#36
2-Methylnaphthalene
Concen: 19.406 ng/ul
RT: 12.399 min Scan# 1566
Delta R.T. -0.006 min
Lab File: BP013373.D
Acq: 30 Dec 2022 15:40

Tgt Ion:142 Resp: 2641884
Ion Ratio Lower Upper
142 100
141 88.7 70.8 106.2









#37
1-Methylnaphthalene
Concen: 19.240 ng/ul
RT: 12.616 min Scan# 1603
Delta R.T. -0.006 min
Lab File: BP013373.D
Acq: 30 Dec 2022 15:40




Instrument :
BNA_P
ClientSampleId :
SSTD020732


Tgt Ion:142 Resp: 2693546
Ion Ratio Lower Upper
142 100
141 91.0 72.7 109.1
116 3.7 2.7 4.1



#39
1,2,4,5-Tetrachlorobenzene
Concen: 22.302 ng/ul
RT: 12.763 min Scan# 1628
Delta R.T. -0.006 min
Lab File: BP013373.D
Acq: 30 Dec 2022 15:40

Tgt Ion:216 Resp: 1634823
Ion Ratio Lower Upper
216 100
214 76.6 61.9 92.9
179 19.1 15.0 22.6
108 14.8 11.0 16.4









#40
Hexachlorocyclopentadiene
Concen: 20.912 ng/ul
RT: 12.740 min Scan# 1624
Delta R.T. -0.006 min
Lab File: BP013373.D
Acq: 30 Dec 2022 15:40




Instrument :
BNA_P
ClientSampleId :
SSTD020732


Tgt Ion:237 Resp: 997535
Ion Ratio Lower Upper
237 100
235 63.5 49.7 74.5
272 13.0 10.8 16.2



#41
2,4,6-Trichlorophenol
Concen: 21.299 ng/ul
RT: 13.004 min Scan# 1669
Delta R.T. -0.006 min
Lab File: BP013373.D
Acq: 30 Dec 2022 15:40

Tgt Ion:196 Resp: 1003360
Ion Ratio Lower Upper
196 100
198 96.2 77.0 115.6
200 31.7 25.3 37.9





#42

2,4,5-Trichlorophenol

Concen: 21.305 ng/ul

RT: 13.075 min Scan# 1

Delta R.T. -0.006 min

Lab File: BP013373.D

Acq: 30 Dec 2022 15:40

Instrument :

BNA_P

ClientSampleId :

SSTD020732





Tgt Ion:196 Resp: 1078078

Ion Ratio Lower Upper

196 100

198 96.1 77.1 115.7

200 30.7 25.4 38.2



#43

1,1'-Biphenyl

Concen: 20.697 ng/ul

RT: 13.404 min Scan# 1737

Delta R.T. -0.006 min

Lab File: BP013373.D

Acq: 30 Dec 2022 15:40




Tgt Ion:154 Resp: 3538501


Ion Ratio Lower Upper

154 100

153 40.1 31.5 47.3

76 12.7 9.6 14.4





#44

2-Chloronaphthalene

Concen: 20.853 ng/ul

RT: 13.451 min Scan# 1745

Delta R.T. 0.000 min

Lab File: BP013373.D

Acq: 30 Dec 2022 15:40

Instrument :

BNA_P

ClientSampleId :

SSTD020732





Tgt Ion:162 Resp: 2805993

Ion Ratio Lower Upper

162 100

164 32.8 26.3 39.5

127 36.2 29.3 43.9



#45

2-Nitroaniline

Concen: 22.451 ng/ul

RT: 13.657 min Scan# 1780

Delta R.T. -0.006 min

Lab File: BP013373.D

Acq: 30 Dec 2022 15:40




Tgt Ion: 65 Resp: 732462


Ion Ratio Lower Upper

65 100

92 66.5 56.4 84.6

138 102.7 90.3 135.5









#47
Dimethylphthalate
Concen: 18.240 ng/ul
RT: 14.028 min Scan# 1843
Delta R.T. 0.000 min
Lab File: BP013373.D
Acq: 30 Dec 2022 15:40




Instrument :
BNA_P
ClientSampleId :
SSTD020732


Tgt Ion:163 Resp: 3163006
Ion Ratio Lower Upper
163 100
194 4.3 3.4 5.0
164 10.2 8.3 12.5



#48
2,6-Dinitrotoluene
Concen: 19.811 ng/ul
RT: 14.151 min Scan# 1864
Delta R.T. -0.006 min
Lab File: BP013373.D
Acq: 30 Dec 2022 15:40

Tgt Ion:165 Resp: 638434
Ion Ratio Lower Upper
165 100
89 55.4 43.0 64.4
121 20.4 16.1 24.1









#50
Acenaphthylene
Concen: 19.933 ng/ul
RT: 14.293 min Scan# 1888
Delta R.T. -0.006 min
Lab File: BP013373.D
Acq: 30 Dec 2022 15:40




Instrument :
BNA_P
ClientSampleId :
SSTD020732


Tgt Ion:152 Resp: 4119821
Ion Ratio Lower Upper
152 100
151 19.9 15.8 23.6
153 13.4 10.6 16.0



#51
3-Nitroaniline
Concen: 17.645 ng/ul
RT: 14.481 min Scan# 1920
Delta R.T. -0.006 min
Lab File: BP013373.D
Acq: 30 Dec 2022 15:40

Tgt Ion:138 Resp: 536938
Ion Ratio Lower Upper
138 100
108 10.2 8.6 13.0
92 113.5 88.4 132.6





#52

Acenaphthene

Concen: 19.693 ng/ul

RT: 14.634 min Scan# 1946

Delta R.T. -0.006 min

Lab File: BP013373.D

Acq: 30 Dec 2022 15:40

Instrument :

BNA_P

ClientSampleId :

SSTD020732





Tgt Ion:153 Resp: 2820514

Ion Ratio Lower Upper

153 100

152 48.0 38.2 57.4

154 90.0 70.8 106.2



#53

2,4-Dinitrophenol

Concen: 16.072 ng/ul

RT: 14.693 min Scan# 1956

Delta R.T. 0.000 min

Lab File: BP013373.D

Acq: 30 Dec 2022 15:40




Tgt Ion:184 Resp: 354031


Ion Ratio Lower Upper

184 100

63 63.0 48.4 72.6

154 58.1 45.5 68.3









#55
4-Nitrophenol
Concen: 18.189 ng/ul
RT: 14.793 min Scan# 1972
Delta R.T. 0.000 min
Lab File: BP013373.D
Acq: 30 Dec 2022 15:40




Instrument :
BNA_P
ClientSampleId :
SSTD020732


Tgt Ion:109 Resp: 390578
Ion Ratio Lower Upper
109 100
139 137.3 120.8 181.2
65 137.7 107.9 161.9



#56
Dibenzofuran
Concen: 19.316 ng/ul
RT: 14.969 min Scan# 2003
Delta R.T. -0.006 min
Lab File: BP013373.D
Acq: 30 Dec 2022 15:40

Tgt Ion:168 Resp: 3933381
Ion Ratio Lower Upper
168 100
139 34.8 27.9 41.9









#57
2,4-Dinitrotoluene
Concen: 18.388 ng/ul
RT: 14.940 min Scan# 1998
Delta R.T. -0.006 min
Lab File: BP013373.D
Acq: 30 Dec 2022 15:40




Instrument :
BNA_P
ClientSampleId :
SSTD020732


Tgt Ion:165 Resp: 860743
Ion Ratio Lower Upper
165 100
63 42.7 31.8 47.6
182 2.6 2.1 3.1



#58
2,3,4,6-Tetrachlorophenol
Concen: 19.037 ng/ul
RT: 15.198 min Scan# 2042
Delta R.T. -0.006 min
Lab File: BP013373.D
Acq: 30 Dec 2022 15:40

Tgt Ion:232 Resp: 879094
Ion Ratio Lower Upper
232 100
131 37.1 31.8 47.6
130 2.1 1.8 2.6
166 27.7 22.9 34.3









#59
Diethylphthalate
Concen: 17.135 ng/ul
RT: 15.393 min Scan# 2075
Delta R.T. 0.000 min
Lab File: BP013373.D
Acq: 30 Dec 2022 15:40




Instrument :
BNA_P
ClientSampleId :
SSTD020732


Tgt Ion:149 Resp: 2888503
Ion Ratio Lower Upper
149 100
177 23.6 18.4 27.6
150 12.2 9.9 14.9



#61
Fluorene
Concen: 19.044 ng/ul
RT: 15.622 min Scan# 2114
Delta R.T. -0.006 min
Lab File: BP013373.D
Acq: 30 Dec 2022 15:40

Tgt Ion:166 Resp: 3110234
Ion Ratio Lower Upper
166 100
165 96.7 77.4 116.2
167 13.1 10.6 16.0









#62
4-Chlorophenyl-phenylether
Concen: 19.539 ng/ul
RT: 15.616 min Scan# 2112
Delta R.T. 0.000 min
Lab File: BP013373.D
Acq: 30 Dec 2022 15:40




Instrument :
BNA_P
ClientSampleId :
SSTD020732


Tgt Ion:204 Resp: 1687480
Ion Ratio Lower Upper
204 100
206 33.3 26.3 39.5
141 49.4 40.4 60.6



#63
4-Nitroaniline
Concen: 16.585 ng/ul
RT: 15.645 min Scan# 2118
Delta R.T. -0.006 min
Lab File: BP013373.D
Acq: 30 Dec 2022 15:40

Tgt Ion:138 Resp: 444094
Ion Ratio Lower Upper
138 100
92 50.9 39.0 58.4
108 78.6 55.1 82.7









#66
4,6-Dinitro-2-methylphenol
Concen: 19.274 ng/ul
RT: 15.704 min Scan# 2128
Delta R.T. -0.006 min
Lab File: BP013373.D
Acq: 30 Dec 2022 15:40




Instrument :
BNA_P
ClientSampleId :
SSTD020732


Tgt Ion:198 Resp: 531290
Ion Ratio Lower Upper
198 100
182 4.0 3.3 4.9
77 22.3 16.6 25.0



#67
N-Nitrosodiphenylamine
Concen: 20.935 ng/ul
RT: 15.828 min Scan# 2149
Delta R.T. -0.006 min
Lab File: BP013373.D
Acq: 30 Dec 2022 15:40

Tgt Ion:169 Resp: 2543974
Ion Ratio Lower Upper
169 100
168 66.2 53.0 79.6
167 35.9 29.1 43.7









#68
4-Bromophenyl-phenylether
Concen: 21.389 ng/ul
RT: 16.510 min Scan# 2265
Delta R.T. -0.006 min
Lab File: BP013373.D
Acq: 30 Dec 2022 15:40




Instrument :
BNA_P
ClientSampleId :
SSTD020732


Tgt Ion:248 Resp: 1018156
Ion Ratio Lower Upper
248 100
250 97.7 77.4 116.0
141 63.7 48.3 72.5



#69
Hexachlorobenzene
Concen: 20.514 ng/ul
RT: 16.628 min Scan# 2285
Delta R.T. -0.006 min
Lab File: BP013373.D
Acq: 30 Dec 2022 15:40

Tgt Ion:284 Resp: 1159488
Ion Ratio Lower Upper
284 100
142 29.4 22.6 34.0
249 30.4 24.6 37.0





#70

Atrazine

Concen: 18.714 ng/ul

RT: 16.787 min Scan# 21

Delta R.T. -0.006 min

Lab File: BP013373.D

Acq: 30 Dec 2022 15:40

Instrument :

BNA_P

ClientSampleId :

SSTD020732





Tgt Ion:200 Resp: 886187

Ion Ratio Lower Upper

200 100

173 24.8 20.5 30.7

215 50.0 39.9 59.9



#71

Pentachlorophenol

Concen: 18.434 ng/ul

RT: 16.975 min Scan# 2344

Delta R.T. -0.006 min

Lab File: BP013373.D

Acq: 30 Dec 2022 15:40




Tgt Ion:266 Resp: 630975


Ion Ratio Lower Upper

266 100

264 63.7 50.9 76.3

268 64.4 51.4 77.2





#72

Phenanthrene

Concen: 19.612 ng/ul

RT: 17.375 min Scan# 2412

Delta R.T. -0.006 min

Lab File: BP013373.D

Acq: 30 Dec 2022 15:40

Instrument :

BNA_P

ClientSampleId :

SSTD020732





Tgt Ion:178 Resp: 4477581

Ion Ratio Lower Upper

178 100

179 15.4 12.5 18.7

176 19.0 15.4 23.0



#74

Anthracene

Concen: 19.746 ng/ul

RT: 17.463 min Scan# 2427

Delta R.T. -0.006 min

Lab File: BP013373.D

Acq: 30 Dec 2022 15:40




Tgt Ion:178 Resp: 4501215


Ion Ratio Lower Upper

178 100

179 15.6 12.5 18.7

176 18.8 14.8 22.2





#75

1,2,3,4-Tetrachlorobenzene

Concen: 25.295 ng/uL

RT: 13.369 min Scan# 1731

Delta R.T. -0.006 min

Lab File: BP013373.D

Acq: 30 Dec 2022 15:40

Instrument :

BNA_P

ClientSampleId :

SSTD020732





Tgt Ion:216 Resp: 1581869

Ion Ratio Lower Upper

216 100

214 79.3 62.2 93.2

218 48.9 38.9 58.3



#76

Pentachlorobenzene

Concen: 23.268 ng/uL

RT: 14.893 min Scan# 1990

Delta R.T. 0.000 min

Lab File: BP013373.D

Acq: 30 Dec 2022 15:40




Tgt Ion:250 Resp: 1483531


Ion Ratio Lower Upper

250 100

248 63.3 50.6 76.0

252 63.8 51.1 76.7









#77
 Carbazole
 Concen: 19.630 ng/ul
 RT: 17.734 min Scan# 2473
 Delta R.T. -0.006 min
 Lab File: BP013373.D
 Acq: 30 Dec 2022 15:40




Instrument :
 BNA_P
 ClientSampleId :
 SSTD020732


Tgt Ion:167 Resp: 3768505
 Ion Ratio Lower Upper
 167 100
 166 21.3 17.2 25.8
 139 12.2 9.4 14.2



#78
 Di-n-butylphthalate
 Concen: 17.915 ng/ul
 RT: 18.275 min Scan# 2565
 Delta R.T. -0.006 min
 Lab File: BP013373.D
 Acq: 30 Dec 2022 15:40

Tgt Ion:149 Resp: 4187663
 Ion Ratio Lower Upper
 149 100
 150 9.1 7.5 11.3
 104 4.9 3.8 5.6









#80
Fluoranthene
Concen: 20.004 ng/ul
RT: 19.334 min Scan# 21
Delta R.T. -0.006 min
Lab File: BP013373.D
Acq: 30 Dec 2022 15:40




Instrument :
BNA_P
ClientSampleId :
SSTD020732


Tgt Ion:202 Resp: 4929288
Ion Ratio Lower Upper
202 100
101 10.2 8.2 12.2
100 7.8 6.1 9.1



#82
Pyrene
Concen: 20.147 ng/ul
RT: 19.692 min Scan# 2806
Delta R.T. -0.006 min
Lab File: BP013373.D
Acq: 30 Dec 2022 15:40

Tgt Ion:202 Resp: 5121191
Ion Ratio Lower Upper
202 100
101 10.7 9.4 14.2
100 8.7 7.8 11.8





#83

Butylbenzylphthalate

Concen: 17.329 ng/ul

RT: 20.551 min Scan# 2952

Delta R.T. -0.012 min

Lab File: BP013373.D

Acq: 30 Dec 2022 15:40

Instrument :

BNA_P

ClientSampleId :

SSTD020732





Tgt Ion:149 Resp: 1701649

Ion Ratio Lower Upper

149 100

91 73.6 53.4 80.0

206 23.1 17.9 26.9



#84

3,3'-Dichlorobenzidine

Concen: 18.936 ng/ul

RT: 21.328 min Scan# 3084

Delta R.T. -0.012 min

Lab File: BP013373.D

Acq: 30 Dec 2022 15:40




Tgt Ion:252 Resp: 1617989


Ion Ratio Lower Upper

252 100

254 65.0 51.8 77.8

126 10.0 7.4 11.2





#85

Benzo(a)anthracene

Concen: 19.965 ng/ul

RT: 21.398 min Scan# 3096

Delta R.T. -0.006 min

Lab File: BP013373.D

Acq: 30 Dec 2022 15:40

Instrument :

BNA_P

ClientSampleId :

SSTD020732





Tgt Ion:228 Resp: 5013080

Ion Ratio Lower Upper

228 100

229 20.2 16.1 24.1

226 27.8 22.0 33.0



#86

Bis(2-ethylhexyl)phthalate

Concen: 17.460 ng/ul

RT: 21.310 min Scan# 3081

Delta R.T. -0.006 min

Lab File: BP013373.D

Acq: 30 Dec 2022 15:40




Tgt Ion:149 Resp: 2418523


Ion Ratio Lower Upper

149 100

167 28.4 22.5 33.7

279 5.2 3.8 5.8





#87

Chrysene

Concen: 19.971 ng/ul

RT: 21.457 min Scan# 3106

Delta R.T. -0.006 min

Lab File: BP013373.D

Acq: 30 Dec 2022 15:40

Instrument :

BNA_P

ClientSampleId :

SSTD020732





Tgt Ion:228 Resp: 4742193

Ion Ratio Lower Upper

228 100

226 30.2 24.5 36.7

229 19.9 16.0 24.0



#89

Di-n-octyl phthalate

Concen: 16.217 ng/ul




RT: 22.251 min Scan# 3241


Delta R.T. -0.012 min

Lab File: BP013373.D

Acq: 30 Dec 2022 15:40

Tgt Ion:149 Resp: 4117441





#90

Benzo(b)fluoranthene

Concen: 18.988 ng/ul

RT: 23.127 min Scan# 3390

Delta R.T. -0.012 min

Lab File: BP013373.D

Acq: 30 Dec 2022 15:40

Instrument :

BNA_P

ClientSampleId :

SSTD020732





Tgt Ion:252 Resp: 5487537

Ion Ratio Lower Upper

252 100

253 22.1 17.7 26.5

125 8.1 6.8 10.2



#91

Benzo(k)fluoranthene

Concen: 18.645 ng/ul

RT: 23.174 min Scan# 3398

Delta R.T. -0.012 min

Lab File: BP013373.D

Acq: 30 Dec 2022 15:40




Tgt Ion:252 Resp: 5343280


Ion Ratio Lower Upper

252 100

253 22.7 17.7 26.5

125 8.3 6.8 10.2





#93

Benzo(a)pyrene

Concen: 19.561 ng/ul

RT: 23.774 min Scan# 3499

Delta R.T. -0.006 min

Lab File: BP013373.D

Acq: 30 Dec 2022 15:40

Instrument :

BNA_P

ClientSampleId :

SSTD020732





Tgt Ion:252 Resp: 4922178

Ion Ratio Lower Upper

252 100

253 21.7 17.5 26.3

125 8.8 7.5 11.3



#94

Indeno(1,2,3-cd)pyrene

Concen: 22.267 ng/ul

RT: 26.457 min Scan# 3956

Delta R.T. -0.018 min

Lab File: BP013373.D

Acq: 30 Dec 2022 15:40




Tgt Ion:276 Resp: 6979359


Ion Ratio Lower Upper

276 100

138 19.7 17.1 25.7

277 25.3 20.0 30.0





#95

Dibenzo(a,h)anthracene

Concen: 22.290 ng/ul

RT: 26.474 min Scan# 3958

Delta R.T. -0.018 min

Lab File: BP013373.D

Acq: 30 Dec 2022 15:40

Instrument :

BNA_P

ClientSampleId :

SSTD020732





Tgt Ion:278 Resp: 5784670

Ion Ratio Lower Upper

278 100

139 13.0 11.1 16.7

279 24.1 19.0 28.6



#96

Benzo(g,h,i)perylene

Concen: 22.568 ng/ul

RT: 27.251 min Scan# 4091

Delta R.T. -0.012 min

Lab File: BP013373.D

Acq: 30 Dec 2022 15:40

Tgt Ion:276 Resp: 5678841

Ion Ratio Lower Upper

276 100

138 18.1 15.4 23.2

277 24.1 19.1 28.7

