

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP123024\
 Data File : BP023591.D
 Acq On : 30 Dec 2024 17:04
 Operator : RC/JU
 Sample : PB165900BL
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 PB165900BL

Quant Time: Dec 30 17:37:49 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP123024.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Dec 30 16:39:09 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.811	152	595440	20.000	ng	0.00
21) Naphthalene-d8	10.587	136	2368454	20.000	ng	0.00
39) Acenaphthene-d10	14.440	164	1459490	20.000	ng	0.00
64) Phenanthrene-d10	17.251	188	2880903	20.000	ng	0.00
76) Chrysene-d12	21.710	240	2971120	20.000	ng	-0.01
86) Perylene-d12	25.139	264	3302464	20.000	ng	-0.02
System Monitoring Compounds						
5) 2-Fluorophenol	5.417	112	5987601	156.949	ng	0.00
7) Phenol-d6	6.969	99	7444981	143.255	ng	0.00
23) Nitrobenzene-d5	8.952	82	4302799	102.520	ng	0.00
42) 2,4,6-Tribromophenol	15.951	330	2467212	146.200	ng	0.00
45) 2-Fluorobiphenyl	13.052	172	9768251	98.220	ng	0.00
79) Terphenyl-d14	19.969	244	15147196	103.696	ng	-0.02
Target Compounds						
						Qvalue
26) 2-Nitrophenol	9.722	139	18	3.175	ng	# 1
32) Benzoic acid	9.846	122	149	6.095	ng	# 66
48) 2-Nitroaniline	13.510	65	297	3.534	ng	# 28
54) 2,4-Dinitrophenol	14.551	184	75	3.846	ng	# 1
57) 2,4-Dinitrotoluene	14.787	165	187	3.898	ng	# 69
65) 4,6-Dinitro-2-methylph...	15.581	198	302	6.544	ng	# 1
80) Butylbenzylphthalate	20.722	149	155	4.101	ng	# 53

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP123024\
Data File : BP023591.D
Acq On : 30 Dec 2024 17:04
Operator : RC/JU
Sample : PB165900BL
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
PB165900BL

Quant Time: Dec 30 17:37:49 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP123024.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Dec 30 16:39:09 2024
Response via : Initial Calibration

