


Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG010219\
 Data File : BG038893.D
 Acq On : 2 Jan 2019 15:02
 Operator : JU/SJ
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\8270-BG121218.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Wed Dec 12 15:26:27 2018



AutoFind: Scans 2508, 2509, 2510; Background Corrected with Scan 2502

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	36.6	23712	PASS
68	69	0.00	2	1.7	338	PASS
69	198	0.00	100	31.1	20151	PASS
70	69	0.00	2	0.4	87	PASS
127	198	10	80	42.1	27272	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	64810	PASS
199	198	5	9	7.3	4715	PASS
275	198	10	60	30.7	19902	PASS
365	198	1	100	4.5	2939	PASS
441	198	0.01	100	13.5	8738	PASS
442	198	50	100	83.0	53762	PASS
443	442	15	24	19.1	10252	PASS