

Data Path : Z:\SVOASRV\HPCHEM1\BNA G\DATA\BG010419\
 Data File : BG038936.D
 Acq On : 4 Jan 2019 19:09
 Operator : JU/SJ
 Sample : K1040-03
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 RBR-200030-RBR-200028-RT-3483

Manual Integrations
 APPROVED

Sohil
 1/7/2019 12:48:16 PM

Quant Time: Jan 04 23:06:30 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG121218.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Dec 12 15:26:27 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.05	152	26313	20.00	ng	-0.01
21) Naphthalene-d8	10.86	136	103337	20.00	ng	-0.02
39) Acenaphthene-d10	14.68	164	74436	20.00	ng	-0.02
64) Phenanthrene-d10	17.43	188	187454	20.00	ng	-0.01
76) Chrysene-d12	21.71	240	190562	20.00	ng	-0.01
87) Perylene-d12	25.00	264	187777	20.00	ng	-0.02
System Monitoring Compounds						
5) 2-Fluorophenol	5.60	112	172599	116.34	ng	0.00
7) Phenol-d6	7.21	99	215705	105.91	ng	0.00
23) Nitrobenzene-d5	9.22	82	165708	81.92	ng	-0.02
42) 2,4,6-Tribromophenol	16.17	330	154179	150.29	ng	-0.02
45) 2-Fluorobiphenyl	13.30	172	446026	82.20	ng	-0.02
79) Terphenyl-d14	20.03	244	852548	97.37	ng	-0.01
Target Compounds						
43) 2,4,6-Trichlorophenol	13.12	196	17711m	10.353	ng	Qvalue
44) 2,4,5-Trichlorophenol	13.21	196	9338	5.155	ng	96
50) Dimethylphthalate	14.14	163	13032	2.144	ng	97
59) 2,3,4,6-Tetrachlorophenol	15.31	232	23372	14.342	ng	# 97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA G\DATA\BG010419\
Data File : BG038936.D
Acq On : 4 Jan 2019 19:09
Operator : JU/SJ
Sample : K1040-03
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
RBR-200030-RBR-200028-RT-3483

Manual Integrations
APPROVED

Sohil
1/7/2019 12:48:16 PM

Quant Time: Jan 04 23:06:30 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG121218.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Dec 12 15:26:27 2018
Response via : Initial Calibration

