

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG010423\
 Data File : BG056199.D
 Acq On : 4 Jan 2023 14:04
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SSTD020583

Quant Time: Jan 05 00:17:43 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG010423.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Jan 05 00:14:25 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.335	152	49079	20.000	ng/ul	0.00
20) Naphthalene-d8	11.167	136	202506	20.000	ng/ul	0.00
38) Acenaphthene-d10	14.945	164	177028	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.683	188	462225	20.000	ng/ul	0.00
79) Chrysene-d12	21.978	240	426881	20.000	ng/ul	0.00
88) Perylene-d12	25.433	264	456668	20.000	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.641	96	8363	7.133	ng/uL	0.00
4) Pyridine-d5	4.070	84	70403	18.417	ng/ul	0.00
7) Phenol-d5	7.466	99	85442	19.192	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.642	67	35840	20.426	ng/ul	0.00
11) 2-Chlorophenol-d4	7.859	132	56976	21.654	ng/ul	0.00
15) 4-Methylphenol-d8	9.029	113	68629	20.483	ng/ul	0.00
21) Nitrobenzene-d5	9.510	128	31903	21.630	ng/ul	0.00
24) 2-Nitrophenol-d4	10.233	143	40966	21.258	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.774	165	82142	19.182	ng/ul	0.00
31) 4-Chloroaniline-d4	11.291	131	95137	19.642	ng/ul	0.00
46) Dimethylphthalate-d6	14.334	166	275066	20.076	ng/ul	0.00
49) Acenaphthylene-d8	14.646	160	297061	19.199	ng/ul	0.00
54) 4-Nitrophenol-d4	15.110	143	42883	20.792	ng/ul	0.00
60) Fluorene-d10	15.926	176	263337	19.472	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	16.032	200	59997	20.192	ng/ul	0.00
73) Anthracene-d10	17.777	188	420790	19.700	ng/ul	0.00
81) Pyrene-d10	20.045	212	497215	21.090	ng/ul	0.00
92) Benzo(a)pyrene-d12	25.192	264	456653	19.674	ng/ul	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.676	88	10264	7.826	ng/uL#	97
5) Pyridine	4.093	79	66421	18.066	ng/ul#	92
6) Benzaldehyde	7.460	77	31483	19.263	ng/ul	96
8) Phenol	7.495	94	83633	18.876	ng/ul	96
10) Bis(2-Chloroethyl)ether	7.736	93	58697	20.126	ng/ul	98
12) 2-Chlorophenol	7.895	128	57523	21.766	ng/ul	97
13) 2-Methylphenol	8.764	108	65995	19.838	ng/ul	100
14) 2,2'-oxybis(1-Chloropr...	8.858	45	11038	29.466	ng/ul#	87
16) Acetophenone	9.164	105	107480	19.366	ng/ul	92
17) N-Nitroso-di-n-propyla...	9.134	70	50437	21.494	ng/ul	96
18) 4-Methylphenol	9.093	108	72989	20.255	ng/ul	99
19) Hexachloroethane	9.434	117	21312	20.056	ng/ul	93
22) Nitrobenzene	9.552	77	72127	19.707	ng/ul	92
23) Isophorone	10.069	82	152529	18.228	ng/ul	97
25) 2-Nitrophenol	10.268	139	40853	21.961	ng/ul	99
26) 2,4-Dimethylphenol	10.310	107	83798	19.038	ng/ul	98
27) Bis(2-Chloroethoxy)met...	10.550	93	85185	19.644	ng/ul	99
29) 2,4-Dichlorophenol	10.803	162	78440	19.179	ng/ul	95
30) Naphthalene	11.220	128	205352	19.462	ng/ul	98
32) 4-Chloroaniline	11.314	127	95752	19.618	ng/ul	98
33) Hexachlorobutadiene	11.496	225	65429	17.071	ng/ul	99
34) Caprolactam	12.060	113	25833	19.756	ng/ul	96
35) 4-Chloro-3-methylphenol	12.407	107	77248	19.660	ng/ul	93
36) 2-Methylnaphthalene	12.801	142	161219	19.058	ng/ul	97


Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG010423\
 Data File : BG056199.D
 Acq On : 4 Jan 2023 14:04
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SSTD020583

Quant Time: Jan 05 00:17:43 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG010423.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Jan 05 00:14:25 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	13.012	142	165270	19.393	ng/ul#	90
39) 1,2,4,5-Tetrachloroben...	13.159	216	132455	18.142	ng/ul	98
40) Hexachlorocyclopentadiene	13.136	237	80048	16.910	ng/ul	95
41) 2,4,6-Trichlorophenol	13.388	196	84114	19.547	ng/ul	96
42) 2,4,5-Trichlorophenol	13.459	196	92112	19.742	ng/ul	99
43) 1,1'-Biphenyl	13.782	154	247264	19.757	ng/ul	97
44) 2-Chloronaphthalene	13.835	162	200768	19.354	ng/ul	100
45) 2-Nitroaniline	14.023	65	39330	27.302	ng/ul	95
47) Dimethylphthalate	14.381	163	272253	20.039	ng/ul	99
48) 2,6-Dinitrotoluene	14.505	165	57632	21.491	ng/ul#	90
50) Acenaphthylene	14.669	152	302392	19.665	ng/ul	97
51) 3-Nitroaniline	14.839	138	43416	21.553	ng/ul	92
52) Acenaphthene	15.010	153	213081	19.726	ng/ul	98
53) 2,4-Dinitrophenol	15.039	184	37911	18.162	ng/ul#	89
55) 4-Nitrophenol	15.127	109	39427	19.650	ng/ul	96
56) Dibenzofuran	15.339	168	328958	19.456	ng/ul	97
57) 2,4-Dinitrotoluene	15.286	165	81680	21.964	ng/ul#	95
58) 2,3,4,6-Tetrachlorophenol	15.556	232	89334	19.370	ng/ul	97
59) Diethylphthalate	15.733	149	253259	20.865	ng/ul	99
61) Fluorene	15.985	166	264178	19.020	ng/ul	98
62) 4-Chlorophenyl-phenyle...	15.968	204	164896	18.275	ng/ul	97
63) 4-Nitroaniline	15.991	138	40793	21.223	ng/ul	97
66) 4,6-Dinitro-2-methylph...	16.050	198	58458	20.520	ng/ul	99
67) N-Nitrosodiphenylamine	16.179	169	241191	20.594	ng/ul	99
68) 4-Bromophenyl-phenylether	16.861	248	113999	19.454	ng/ul	93
69) Hexachlorobenzene	16.984	284	115031	20.282	ng/ul	97
70) Atrazine	17.107	200	110046	19.707	ng/ul	96
71) Pentachlorophenol	17.325	266	69806	19.359	ng/ul	98
72) Phenanthrene	17.724	178	462077	19.692	ng/ul	100
74) Anthracene	17.812	178	469812	19.558	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.758	216	133333	19.017	ng/ul	99
76) Pentachlorobenzene	15.257	250	137829	19.914	ng/ul	96
77) Carbazole	18.077	167	390905	19.904	ng/ul	98
78) Di-n-butylphthalate	18.606	149	409349	22.814	ng/ul	98
80) Fluoranthene	19.710	202	589551	21.266	ng/ul	99
82) Pyrene	20.075	202	592532	21.188	ng/ul	99
83) Butylbenzylphthalate	20.932	149	170747	28.690	ng/ul	98
84) 3,3'-Dichlorobenzidine	21.855	252	206158	21.495	ng/ul	99
85) Benzo(a)anthracene	21.955	228	574211	19.707	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.820	149	243027	26.657	ng/ul	98
87) Chrysene	22.025	228	538533	19.916	ng/ul	99
89) Di-n-octyl phthalate	23.112	149	415117	24.564	ng/ul	100
90) Benzo(b)fluoranthene	24.328	252	601993	19.968	ng/ul	99
91) Benzo(k)fluoranthene	24.399	252	581812	19.777	ng/ul	99
93) Benzo(a)pyrene	25.274	252	518888	19.753	ng/ul	100
94) Indeno(1,2,3-cd)pyrene	29.411	276	666198	19.254	ng/ul	99
95) Dibenzo(a,h)anthracene	29.481	278	559322	19.327	ng/ul	98
96) Benzo(g,h,i)perylene	30.662	276	535602	19.262	ng/ul#	98





(#) = qualifier out of range (m) = manual integration (+) = signals summed



#2
1,4-Dioxane
Concen: 7.826 ng/uL
RT: 3.676 min Scan# 60
Delta R.T. 0.001 min
Lab File: BG056199.D
Acq: 4 Jan 2023 14:04




Instrument :
BNA_G
ClientSampleId :
SSTD020583


Tgt Ion: 88 Resp: 10264
Ion Ratio Lower Upper
88 100
43 6.5 6.8 10.2#
58 25.9 20.0 30.0



#5
Pyridine
Concen: 18.066 ng/uL
RT: 4.093 min Scan# 137
Delta R.T. 0.001 min
Lab File: BG056199.D
Acq: 4 Jan 2023 14:04

Tgt Ion: 79 Resp: 66421
Ion Ratio Lower Upper
79 100
52 18.0 16.6 25.0
51 12.4 13.4 20.2#









#6
Benzaldehyde
Concen: 19.263 ng/ul
RT: 7.460 min Scan# 710
Delta R.T. 0.001 min
Lab File: BG056199.D
Acq: 4 Jan 2023 14:04




Instrument :
BNA_G
ClientSampleId :
SSTD020583


Tgt Ion: 77 Resp: 31483
Ion Ratio Lower Upper
77 100
105 103.5 86.5 129.7
106 104.1 80.8 121.2



#8
Phenol
Concen: 18.876 ng/ul
RT: 7.495 min Scan# 716
Delta R.T. 0.001 min
Lab File: BG056199.D
Acq: 4 Jan 2023 14:04

Tgt Ion: 94 Resp: 83633
Ion Ratio Lower Upper
94 100
65 22.8 15.9 23.9
66 31.9 24.6 37.0









#10
Bis(2-Chloroethyl)ether
Concen: 20.126 ng/ul
RT: 7.736 min Scan# 757
Delta R.T. 0.001 min
Lab File: BG056199.D
Acq: 4 Jan 2023 14:04




Instrument :
BNA_G
ClientSampleId :
SSTD020583


Tgt Ion: 93 Resp: 58697
Ion Ratio Lower Upper
93 100
63 48.4 37.4 56.2
95 33.7 28.1 42.1



#12
2-Chlorophenol
Concen: 21.766 ng/ul
RT: 7.895 min Scan# 784
Delta R.T. 0.001 min
Lab File: BG056199.D
Acq: 4 Jan 2023 14:04

Tgt Ion:128 Resp: 57523
Ion Ratio Lower Upper
128 100
64 33.3 29.0 43.6
130 32.1 25.3 37.9





#13

2-Methylphenol

Concen: 19.838 ng/ul

RT: 8.764 min Scan# 91

Delta R.T. 0.001 min

Lab File: BG056199.D

Acq: 4 Jan 2023 14:04

Instrument :

BNA_G

ClientSampleId :





SSTD020583

Tgt Ion:108 Resp: 65995

Ion Ratio Lower Upper

108 100

107 87.9 70.5 105.7



#14

2,2'-oxybis(1-Chloropropane)

Concen: 29.466 ng/ul

RT: 8.858 min Scan# 948

Delta R.T. 0.001 min

Lab File: BG056199.D

Acq: 4 Jan 2023 14:04




Tgt Ion: 45 Resp: 11038


Ion Ratio Lower Upper

45 100

77 106.8 90.5 135.7

79 73.6 75.8 113.6#









#16
Acetophenone
Concen: 19.366 ng/ul
RT: 9.164 min Scan# 1000
Delta R.T. 0.001 min
Lab File: BG056199.D
Acq: 4 Jan 2023 14:04




Instrument :
BNA_G
ClientSampleId :
SSTD020583


Tgt Ion:105 Resp: 107480
Ion Ratio Lower Upper
105 100
77 73.1 52.8 79.2
51 13.3 10.1 15.1



#17
N-Nitroso-di-n-propylamine
Concen: 21.494 ng/ul
RT: 9.134 min Scan# 995
Delta R.T. 0.001 min
Lab File: BG056199.D
Acq: 4 Jan 2023 14:04

Tgt Ion: 70 Resp: 50437
Ion Ratio Lower Upper
70 100
42 21.8 19.9 29.9
101 11.5 9.8 14.6
130 21.6 18.9 28.3









#18
4-Methylphenol
Concen: 20.255 ng/ul
RT: 9.093 min Scan# 91
Delta R.T. 0.001 min
Lab File: BG056199.D
Acq: 4 Jan 2023 14:04




Instrument :
BNA_G
ClientSampleId :
SSTD020583


Tgt Ion:108 Resp: 72989
Ion Ratio Lower Upper
108 100
107 115.1 91.5 137.3



#19
Hexachloroethane
Concen: 20.056 ng/ul
RT: 9.434 min Scan# 1046
Delta R.T. 0.001 min
Lab File: BG056199.D
Acq: 4 Jan 2023 14:04

Tgt Ion:117 Resp: 21312
Ion Ratio Lower Upper
117 100
201 177.2 132.0 198.0
199 116.0 89.4 134.0









#22
Nitrobenzene
Concen: 19.707 ng/ul
RT: 9.552 min Scan# 1066
Delta R.T. 0.001 min
Lab File: BG056199.D
Acq: 4 Jan 2023 14:04




Instrument :
BNA_G
ClientSampleId :
SSTD020583


Tgt Ion: 77 Resp: 72127
Ion Ratio Lower Upper
77 100
123 46.7 42.2 63.2
65 11.5 10.5 15.7



#23
Isophorone
Concen: 18.228 ng/ul
RT: 10.069 min Scan# 1154
Delta R.T. -0.005 min
Lab File: BG056199.D
Acq: 4 Jan 2023 14:04

Tgt Ion: 82 Resp: 152529
Ion Ratio Lower Upper
82 100
95 7.5 6.9 10.3
138 17.6 15.4 23.2









#25
2-Nitrophenol
Concen: 21.961 ng/ul
RT: 10.268 min Scan# 1188
Delta R.T. 0.001 min
Lab File: BG056199.D
Acq: 4 Jan 2023 14:04




Instrument :
BNA_G
ClientSampleId :
SSTD020583


Tgt Ion:139 Resp: 40853
Ion Ratio Lower Upper
139 100
65 37.9 31.2 46.8
109 38.0 30.1 45.1



#26
2,4-Dimethylphenol
Concen: 19.038 ng/ul
RT: 10.310 min Scan# 1195
Delta R.T. 0.001 min
Lab File: BG056199.D
Acq: 4 Jan 2023 14:04

Tgt Ion:107 Resp: 83798
Ion Ratio Lower Upper
107 100
121 47.6 40.5 60.7
122 82.8 66.8 100.2





#27

Bis(2-Chloroethoxy)methane

Concen: 19.644 ng/ul

RT: 10.550 min Scan# 1236

Delta R.T. 0.007 min

Lab File: BG056199.D

Acq: 4 Jan 2023 14:04

Instrument :

BNA_G

ClientSampleId :

SSTD020583





Tgt Ion: 93 Resp: 85185

Ion Ratio Lower Upper

93 100

95 32.0 25.4 38.0

123 10.1 8.5 12.7



#29

2,4-Dichlorophenol

Concen: 19.179 ng/ul

RT: 10.803 min Scan# 1279

Delta R.T. 0.001 min

Lab File: BG056199.D

Acq: 4 Jan 2023 14:04




Tgt Ion:162 Resp: 78440


Ion Ratio Lower Upper

162 100

164 69.4 52.5 78.7

98 27.3 24.7 37.1









#30
Naphthalene
Concen: 19.462 ng/ul
RT: 11.220 min Scan# 1350
Delta R.T. 0.001 min
Lab File: BG056199.D
Acq: 4 Jan 2023 14:04




Instrument :
BNA_G
ClientSampleId :
SSTD020583


Tgt Ion:128 Resp: 205352
Ion Ratio Lower Upper
128 100
129 11.0 8.5 12.7
127 14.6 10.9 16.3



#32
4-Chloroaniline
Concen: 19.618 ng/ul
RT: 11.314 min Scan# 1366
Delta R.T. 0.001 min
Lab File: BG056199.D
Acq: 4 Jan 2023 14:04

Tgt Ion:127 Resp: 95752
Ion Ratio Lower Upper
127 100
129 30.6 23.5 35.3









#33
Hexachlorobutadiene
Concen: 17.071 ng/ul
RT: 11.496 min Scan# 11
Delta R.T. 0.001 min
Lab File: BG056199.D
Acq: 4 Jan 2023 14:04




Instrument :
BNA_G
ClientSampleId :
SSTD020583


Tgt Ion:225 Resp: 65429
Ion Ratio Lower Upper
225 100
223 62.4 49.8 74.6
227 62.9 51.8 77.6



#34
Caprolactam
Concen: 19.756 ng/ul
RT: 12.060 min Scan# 1493
Delta R.T. 0.001 min
Lab File: BG056199.D
Acq: 4 Jan 2023 14:04

Tgt Ion:113 Resp: 25833
Ion Ratio Lower Upper
113 100
55 52.1 39.3 58.9
56 51.8 39.5 59.3









#35
4-Chloro-3-methylphenol
Concen: 19.660 ng/ul
RT: 12.407 min Scan# 1552
Delta R.T. 0.001 min
Lab File: BG056199.D
Acq: 4 Jan 2023 14:04




Instrument :
BNA_G
ClientSampleId :
SSTD020583


Tgt Ion:107 Resp: 77248
Ion Ratio Lower Upper
107 100
144 28.9 21.0 31.6
142 78.2 67.9 101.9



#36
2-Methylnaphthalene
Concen: 19.058 ng/ul
RT: 12.801 min Scan# 1619
Delta R.T. 0.001 min
Lab File: BG056199.D
Acq: 4 Jan 2023 14:04

Tgt Ion:142 Resp: 161219
Ion Ratio Lower Upper
142 100
141 92.2 71.6 107.4









#37
1-Methylnaphthalene
Concen: 19.393 ng/ul
RT: 13.012 min Scan# 1655
Delta R.T. 0.001 min
Lab File: BG056199.D
Acq: 4 Jan 2023 14:04




Instrument :
BNA_G
ClientSampleId :
SSTD020583


Tgt Ion:142 Resp: 165270
Ion Ratio Lower Upper
142 100
141 99.7 71.7 107.5
116 5.1 3.4 5.0#



#39
1,2,4,5-Tetrachlorobenzene
Concen: 18.142 ng/ul
RT: 13.159 min Scan# 1680
Delta R.T. 0.001 min
Lab File: BG056199.D
Acq: 4 Jan 2023 14:04

Tgt Ion:216 Resp: 132455
Ion Ratio Lower Upper
216 100
214 75.9 62.0 93.0
179 15.4 13.8 20.8
108 10.4 7.9 11.9









#40
Hexachlorocyclopentadiene
Concen: 16.910 ng/ul
RT: 13.136 min Scan# 1676
Delta R.T. 0.001 min
Lab File: BG056199.D
Acq: 4 Jan 2023 14:04




Instrument :
BNA_G
ClientSampleId :
SSTD020583


Tgt Ion:237 Resp: 80048
Ion Ratio Lower Upper
237 100
235 64.7 48.6 72.8
272 11.8 10.2 15.4



#41
2,4,6-Trichlorophenol
Concen: 19.547 ng/ul
RT: 13.388 min Scan# 1719
Delta R.T. 0.001 min
Lab File: BG056199.D
Acq: 4 Jan 2023 14:04

Tgt Ion:196 Resp: 84114
Ion Ratio Lower Upper
196 100
198 99.9 76.2 114.2
200 32.9 25.0 37.4









#42
2,4,5-Trichlorophenol
Concen: 19.742 ng/ul
RT: 13.459 min Scan# 1731
Delta R.T. 0.001 min
Lab File: BG056199.D
Acq: 4 Jan 2023 14:04




Instrument :
BNA_G
ClientSampleId :
SSTD020583


Tgt Ion:196 Resp: 92112
Ion Ratio Lower Upper
196 100
198 98.9 80.2 120.4
200 30.1 23.6 35.4



#43
1,1'-Biphenyl
Concen: 19.757 ng/ul
RT: 13.782 min Scan# 1786
Delta R.T. 0.001 min
Lab File: BG056199.D
Acq: 4 Jan 2023 14:04

Tgt Ion:154 Resp: 247264
Ion Ratio Lower Upper
154 100
153 40.4 34.0 51.0
76 10.0 8.5 12.7









#44
2-Chloronaphthalene
Concen: 19.354 ng/ul
RT: 13.835 min Scan# 1795
Delta R.T. 0.001 min
Lab File: BG056199.D
Acq: 4 Jan 2023 14:04




Instrument :
BNA_G
ClientSampleId :
SSTD020583


Tgt Ion:162 Resp: 200768
Ion Ratio Lower Upper
162 100
164 32.6 25.9 38.9
127 30.6 24.6 36.8



#45
2-Nitroaniline
Concen: 27.302 ng/ul
RT: 14.023 min Scan# 1827
Delta R.T. 0.001 min
Lab File: BG056199.D
Acq: 4 Jan 2023 14:04

Tgt Ion: 65 Resp: 39330
Ion Ratio Lower Upper
65 100
92 91.3 68.2 102.2
138 141.5 109.7 164.5









#47
Dimethylphthalate
Concen: 20.039 ng/ul
RT: 14.381 min Scan# 1888
Delta R.T. 0.001 min
Lab File: BG056199.D
Acq: 4 Jan 2023 14:04




Instrument :
BNA_G
ClientSampleId :
SSTD020583


Tgt Ion:163 Resp: 272253
Ion Ratio Lower Upper
163 100
194 4.6 3.5 5.3
164 10.5 8.6 12.8



#48
2,6-Dinitrotoluene
Concen: 21.491 ng/ul
RT: 14.505 min Scan# 1909
Delta R.T. -0.005 min
Lab File: BG056199.D
Acq: 4 Jan 2023 14:04

Tgt Ion:165 Resp: 57632
Ion Ratio Lower Upper
165 100
89 52.1 36.2 54.4
121 19.5 12.2 18.4#









#50
Acenaphthylene
Concen: 19.665 ng/ul
RT: 14.669 min Scan# 1937
Delta R.T. -0.005 min
Lab File: BG056199.D
Acq: 4 Jan 2023 14:04




Instrument :
BNA_G
ClientSampleId :
SSTD020583


Tgt Ion:152 Resp: 302392
Ion Ratio Lower Upper
152 100
151 21.4 15.4 23.2
153 13.8 10.5 15.7



#51
3-Nitroaniline
Concen: 21.553 ng/ul
RT: 14.839 min Scan# 1966
Delta R.T. 0.001 min
Lab File: BG056199.D
Acq: 4 Jan 2023 14:04

Tgt Ion:138 Resp: 43416
Ion Ratio Lower Upper
138 100
108 10.0 8.3 12.5
92 115.5 85.2 127.8









#52
Acenaphthene
Concen: 19.726 ng/ul
RT: 15.010 min Scan# 1995
Delta R.T. 0.001 min
Lab File: BG056199.D
Acq: 4 Jan 2023 14:04




Instrument :
BNA_G
ClientSampleId :
SSTD020583


Tgt Ion:153 Resp: 213081
Ion Ratio Lower Upper
153 100
152 45.3 38.0 57.0
154 86.4 67.7 101.5



#53
2,4-Dinitrophenol
Concen: 18.162 ng/ul
RT: 15.039 min Scan# 2000
Delta R.T. -0.005 min
Lab File: BG056199.D
Acq: 4 Jan 2023 14:04

Tgt Ion:184 Resp: 37911
Ion Ratio Lower Upper
184 100
63 33.7 22.4 33.6#
154 60.6 42.2 63.4









#55
4-Nitrophenol
Concen: 19.650 ng/ul
RT: 15.127 min Scan# 2015
Delta R.T. 0.001 min
Lab File: BG056199.D
Acq: 4 Jan 2023 14:04




Instrument :
BNA_G
ClientSampleId :
SSTD020583


Tgt Ion:109 Resp: 39427
Ion Ratio Lower Upper
109 100
139 111.8 86.7 130.1
65 81.3 67.9 101.9



#56
Dibenzofuran
Concen: 19.456 ng/ul
RT: 15.339 min Scan# 2051
Delta R.T. 0.001 min
Lab File: BG056199.D
Acq: 4 Jan 2023 14:04

Tgt Ion:168 Resp: 328958
Ion Ratio Lower Upper
168 100
139 32.7 25.0 37.4









#57
2,4-Dinitrotoluene
Concen: 21.964 ng/ul
RT: 15.286 min Scan# 2042
Delta R.T. -0.005 min
Lab File: BG056199.D
Acq: 4 Jan 2023 14:04




Instrument :
BNA_G
ClientSampleId :
SSTD020583


Tgt Ion:165 Resp: 81680
Ion Ratio Lower Upper
165 100
63 28.8 20.9 31.3
182 2.6 2.7 4.1#



#58
2,3,4,6-Tetrachlorophenol
Concen: 19.370 ng/ul
RT: 15.556 min Scan# 2088
Delta R.T. 0.001 min
Lab File: BG056199.D
Acq: 4 Jan 2023 14:04

Tgt Ion:232 Resp: 89334
Ion Ratio Lower Upper
232 100
131 26.6 22.9 34.3
130 1.9 1.7 2.5
166 21.8 17.9 26.9









#59
Diethylphthalate
Concen: 20.865 ng/ul
RT: 15.733 min Scan# 2118
Delta R.T. 0.001 min
Lab File: BG056199.D
Acq: 4 Jan 2023 14:04




Instrument :
BNA_G
ClientSampleId :
SSTD020583


Tgt Ion:149 Resp: 253259
Ion Ratio Lower Upper
149 100
177 26.0 20.2 30.4
150 12.6 9.9 14.9



#61
Fluorene
Concen: 19.020 ng/ul
RT: 15.985 min Scan# 2161
Delta R.T. 0.001 min
Lab File: BG056199.D
Acq: 4 Jan 2023 14:04

Tgt Ion:166 Resp: 264178
Ion Ratio Lower Upper
166 100
165 100.3 78.5 117.7
167 14.1 10.4 15.6









#62
4-Chlorophenyl-phenylether
Concen: 18.275 ng/ul
RT: 15.968 min Scan# 2158
Delta R.T. 0.001 min
Lab File: BG056199.D
Acq: 4 Jan 2023 14:04




Instrument :
BNA_G
ClientSampleId :
SSTD020583


Tgt Ion:204 Resp: 164896
Ion Ratio Lower Upper
204 100
206 32.9 25.0 37.4
141 41.8 31.6 47.4



#63
4-Nitroaniline
Concen: 21.223 ng/ul
RT: 15.991 min Scan# 2162
Delta R.T. 0.001 min
Lab File: BG056199.D
Acq: 4 Jan 2023 14:04

Tgt Ion:138 Resp: 40793
Ion Ratio Lower Upper
138 100
92 47.9 38.8 58.2
108 98.3 75.0 112.4









#66
 4,6-Dinitro-2-methylphenol
 Concen: 20.520 ng/ul
 RT: 16.050 min Scan# 2172
 Delta R.T. 0.001 min
 Lab File: BG056199.D
 Acq: 4 Jan 2023 14:04




Instrument :
 BNA_G
 ClientSampleId :
 SST020583


Tgt Ion:198 Resp: 58458
 Ion Ratio Lower Upper
 198 100
 182 2.7 2.2 3.4
 77 14.8 11.4 17.2



#67
 N-Nitrosodiphenylamine
 Concen: 20.594 ng/ul
 RT: 16.179 min Scan# 2194
 Delta R.T. 0.001 min
 Lab File: BG056199.D
 Acq: 4 Jan 2023 14:04

Tgt Ion:169 Resp: 241191
 Ion Ratio Lower Upper
 169 100
 168 71.2 56.1 84.1
 167 38.3 30.3 45.5









#68
4-Bromophenyl-phenylether
Concen: 19.454 ng/ul
RT: 16.861 min Scan# 21
Delta R.T. 0.001 min
Lab File: BG056199.D
Acq: 4 Jan 2023 14:04




Instrument :
BNA_G
ClientSampleId :
SSTD020583


Tgt Ion:248 Resp: 113999
Ion Ratio Lower Upper
248 100
250 100.7 74.8 112.2
141 50.3 37.0 55.4



#69
Hexachlorobenzene
Concen: 20.282 ng/ul
RT: 16.984 min Scan# 2331
Delta R.T. 0.001 min
Lab File: BG056199.D
Acq: 4 Jan 2023 14:04

Tgt Ion:284 Resp: 115031
Ion Ratio Lower Upper
284 100
142 22.0 18.9 28.3
249 32.2 26.8 40.2









#70
Atrazine
Concen: 19.707 ng/ul
RT: 17.107 min Scan# 21
Delta R.T. -0.005 min
Lab File: BG056199.D
Acq: 4 Jan 2023 14:04




Instrument :
BNA_G
ClientSampleId :
SSTD020583


Tgt Ion:200 Resp: 110046
Ion Ratio Lower Upper
200 100
173 20.5 18.0 27.0
215 48.3 40.8 61.2



#71
Pentachlorophenol
Concen: 19.359 ng/ul
RT: 17.325 min Scan# 2389
Delta R.T. 0.001 min
Lab File: BG056199.D
Acq: 4 Jan 2023 14:04

Tgt Ion:266 Resp: 69806
Ion Ratio Lower Upper
266 100
264 62.7 47.7 71.5
268 67.6 54.3 81.5









#72
Phenanthrene
Concen: 19.692 ng/ul
RT: 17.724 min Scan# 2457
Delta R.T. 0.001 min
Lab File: BG056199.D
Acq: 4 Jan 2023 14:04




Instrument :
BNA_G
ClientSampleId :
SSTD020583


Tgt Ion:178 Resp: 462077
Ion Ratio Lower Upper
178 100
179 15.8 12.4 18.6
176 18.6 15.0 22.6



#74
Anthracene
Concen: 19.558 ng/ul
RT: 17.812 min Scan# 2472
Delta R.T. 0.001 min
Lab File: BG056199.D
Acq: 4 Jan 2023 14:04

Tgt Ion:178 Resp: 469812
Ion Ratio Lower Upper
178 100
179 15.7 13.0 19.6
176 17.8 14.2 21.4









#75
1,2,3,4-Tetrachlorobenzene
Concen: 19.017 ng/uL
RT: 13.758 min Scan# 11
Delta R.T. 0.001 min
Lab File: BG056199.D
Acq: 4 Jan 2023 14:04




Instrument :
BNA_G
ClientSampleId :
SSTD020583


Tgt Ion:216 Resp: 133333
Ion Ratio Lower Upper
216 100
214 78.2 62.4 93.6
218 51.3 39.4 59.0



#76
Pentachlorobenzene
Concen: 19.914 ng/uL
RT: 15.257 min Scan# 2037
Delta R.T. -0.005 min
Lab File: BG056199.D
Acq: 4 Jan 2023 14:04

Tgt Ion:250 Resp: 137829
Ion Ratio Lower Upper
250 100
248 63.7 51.6 77.4
252 63.5 55.2 82.8









#77
 Carbazole
 Concen: 19.904 ng/ul
 RT: 18.077 min Scan# 2517
 Delta R.T. 0.001 min
 Lab File: BG056199.D
 Acq: 4 Jan 2023 14:04




Instrument :
 BNA_G
 ClientSampleId :
 SST020583


Tgt Ion:167 Resp: 390905
 Ion Ratio Lower Upper
 167 100
 166 20.5 17.4 26.0
 139 10.5 8.6 12.8



#78
 Di-n-butylphthalate
 Concen: 22.814 ng/ul
 RT: 18.606 min Scan# 2607
 Delta R.T. 0.001 min
 Lab File: BG056199.D
 Acq: 4 Jan 2023 14:04

Tgt Ion:149 Resp: 409349
 Ion Ratio Lower Upper
 149 100
 150 9.6 7.1 10.7
 104 5.6 4.2 6.4









#80
Fluoranthene
Concen: 21.266 ng/ul
RT: 19.710 min Scan# 21
Delta R.T. 0.001 min
Lab File: BG056199.D
Acq: 4 Jan 2023 14:04




Instrument :
BNA_G
ClientSampleId :
SSTD020583


Tgt Ion:202 Resp: 589551
Ion Ratio Lower Upper
202 100
101 5.6 4.4 6.6
100 4.9 3.8 5.6



#82
Pyrene
Concen: 21.188 ng/ul
RT: 20.075 min Scan# 2857
Delta R.T. 0.001 min
Lab File: BG056199.D
Acq: 4 Jan 2023 14:04

Tgt Ion:202 Resp: 592532
Ion Ratio Lower Upper
202 100
101 6.2 5.2 7.8
100 5.3 4.2 6.2





#83

Butylbenzylphthalate

Concen: 28.690 ng/ul

RT: 20.932 min Scan# 3003

Delta R.T. 0.001 min

Lab File: BG056199.D

Acq: 4 Jan 2023 14:04

Instrument :

BNA_G

ClientSampleId :

SSTD020583





Tgt Ion:149 Resp: 170747

Ion Ratio Lower Upper

149 100

91 65.3 50.0 75.0

206 29.0 23.2 34.8



#84

3,3'-Dichlorobenzidine

Concen: 21.495 ng/ul

RT: 21.855 min Scan# 3160

Delta R.T. 0.001 min

Lab File: BG056199.D

Acq: 4 Jan 2023 14:04




Tgt Ion:252 Resp: 206158


Ion Ratio Lower Upper

252 100

254 63.0 51.1 76.7

126 4.8 3.8 5.6





#85

Benzo(a)anthracene

Concen: 19.707 ng/ul

RT: 21.955 min Scan# 3177

Delta R.T. 0.001 min

Lab File: BG056199.D

Acq: 4 Jan 2023 14:04

Instrument :

BNA_G

ClientSampleId :

SSTD020583





Tgt Ion:228 Resp: 574211

Ion Ratio Lower Upper

228 100

229 19.2 15.9 23.9

226 28.1 22.2 33.2



#86

Bis(2-ethylhexyl)phthalate

Concen: 26.657 ng/ul

RT: 21.820 min Scan# 3154

Delta R.T. -0.005 min

Lab File: BG056199.D

Acq: 4 Jan 2023 14:04




Tgt Ion:149 Resp: 243027


Ion Ratio Lower Upper

149 100

167 30.6 23.7 35.5

279 5.8 4.9 7.3





#87

Chrysene

Concen: 19.916 ng/ul

RT: 22.025 min Scan# 3189

Delta R.T. 0.001 min

Lab File: BG056199.D

Acq: 4 Jan 2023 14:04

Instrument :

BNA_G

ClientSampleId :

SSTD020583





Tgt Ion:228 Resp: 538533

Ion Ratio Lower Upper

228 100

226 31.3 24.5 36.7

229 19.3 15.8 23.8



#89

Di-n-octyl phthalate

Concen: 24.564 ng/ul




RT: 23.112 min Scan# 3374


Delta R.T. 0.001 min

Lab File: BG056199.D

Acq: 4 Jan 2023 14:04

Tgt Ion:149 Resp: 415117





#90

Benzo(b)fluoranthene

Concen: 19.968 ng/ul

RT: 24.328 min Scan# 3581

Delta R.T. 0.001 min

Lab File: BG056199.D

Acq: 4 Jan 2023 14:04

Instrument :

BNA_G

ClientSampleId :

SSTD020583





Tgt Ion:252 Resp: 601993

Ion Ratio Lower Upper

252 100

253 21.4 17.4 26.2

125 3.9 3.0 4.4



#91

Benzo(k)fluoranthene

Concen: 19.777 ng/ul

RT: 24.399 min Scan# 3593

Delta R.T. -0.005 min

Lab File: BG056199.D

Acq: 4 Jan 2023 14:04




Tgt Ion:252 Resp: 581812


Ion Ratio Lower Upper

252 100

253 22.1 18.0 27.0

125 3.4 3.1 4.7





#93

Benzo(a)pyrene

Concen: 19.753 ng/ul

RT: 25.274 min Scan# 31

Delta R.T. 0.001 min

Lab File: BG056199.D

Acq: 4 Jan 2023 14:04

Instrument :

BNA_G

ClientSampleId :

SSTD020583





Tgt Ion:252 Resp: 518888

Ion Ratio Lower Upper

252 100

253 21.6 17.3 25.9

125 4.1 3.2 4.8



#94

Indeno(1,2,3-cd)pyrene

Concen: 19.254 ng/ul

RT: 29.411 min Scan# 4446

Delta R.T. 0.001 min

Lab File: BG056199.D

Acq: 4 Jan 2023 14:04




Tgt Ion:276 Resp: 666198


Ion Ratio Lower Upper

276 100

138 7.8 6.2 9.2

277 25.2 19.8 29.6









#95
 Dibenzo(a,h)anthracene
 Concen: 19.327 ng/ul
 RT: 29.481 min Scan# 4458
 Delta R.T. 0.007 min
 Lab File: BG056199.D
 Acq: 4 Jan 2023 14:04

Instrument :
 BNA_G
 ClientSampleId :
 SSTD020583

Tgt Ion:278 Resp: 559322
 Ion Ratio Lower Upper
 278 100
 139 6.1 5.0 7.4
 279 21.9 18.4 27.6



#96
 Benzo(g,h,i)perylene
 Concen: 19.262 ng/ul
 RT: 30.662 min Scan# 4659
 Delta R.T. -0.005 min
 Lab File: BG056199.D
 Acq: 4 Jan 2023 14:04

Tgt Ion:276 Resp: 535602
 Ion Ratio Lower Upper
 276 100
 138 8.8 5.8 8.8#
 277 24.5 19.1 28.7

