

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG010423\
 Data File : BG056200.D
 Acq On : 4 Jan 2023 14:59
 Operator : CG/JU
 Sample : SST00547
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SST005449

Quant Time: Jan 05 00:18:14 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG010423.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Jan 05 00:14:25 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.335	152	46636	20.000	ng/ul	0.00
20) Naphthalene-d8	11.167	136	184994	20.000	ng/ul	0.00
38) Acenaphthene-d10	14.945	164	164301	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.683	188	433533	20.000	ng/ul	0.00
79) Chrysene-d12	21.978	240	404488	20.000	ng/ul	0.00
88) Perylene-d12	25.433	264	426528	20.000	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.629	96	1841	1.652	ng/uL	0.00
4) Pyridine-d5	4.070	84	15385	4.235	ng/ul	0.00
7) Phenol-d5	7.460	99	18287	4.323	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.636	67	8329	4.995	ng/ul	0.00
11) 2-Chlorophenol-d4	7.853	132	12720	5.088	ng/ul	0.00
15) 4-Methylphenol-d8	9.023	113	15288	4.802	ng/ul	0.00
21) Nitrobenzene-d5	9.504	128	7153	5.309	ng/ul	0.00
24) 2-Nitrophenol-d4	10.233	143	9434	5.359	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.774	165	18859	4.821	ng/ul	0.00
31) 4-Chloroaniline-d4	11.296	131	21596	4.881	ng/ul	0.00
46) Dimethylphthalate-d6	14.334	166	63954	5.029	ng/ul	0.00
49) Acenaphthylene-d8	14.640	160	69117	4.813	ng/ul	0.00
54) 4-Nitrophenol-d4	15.110	143	9512	4.969	ng/ul	0.00
60) Fluorene-d10	15.926	176	59935	4.775	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	16.032	200	13090	4.697	ng/ul	0.00
73) Anthracene-d10	17.683	188	433533	21.640	ng/ul	-0.09
81) Pyrene-d10	20.045	212	118776	5.317	ng/ul	0.00
92) Benzo(a)pyrene-d12	25.192	264	106956	4.934	ng/ul	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.670	88	2191	1.758	ng/uL#	91
5) Pyridine	4.093	79	13542	3.876	ng/ul#	89
6) Benzaldehyde	7.460	77	6854	4.413	ng/ul	95
8) Phenol	7.489	94	18513	4.397	ng/ul	94
10) Bis(2-Chloroethyl)ether	7.736	93	14055	5.072	ng/ul	95
12) 2-Chlorophenol	7.889	128	12148	4.837	ng/ul	96
13) 2-Methylphenol	8.764	108	13194	4.174	ng/ul	93
14) 2,2'-oxybis(1-Chloropr...	8.864	45	2134	5.995	ng/ul#	46
16) Acetophenone	9.158	105	24233	4.595	ng/ul	99
17) N-Nitroso-di-n-propyla...	9.128	70	10771	4.831	ng/ul#	91
18) 4-Methylphenol	9.087	108	15234	4.449	ng/ul	99
19) Hexachloroethane	9.434	117	5035	4.987	ng/ul	88
22) Nitrobenzene	9.546	77	16237	4.856	ng/ul	93
23) Isophorone	10.068	82	34273	4.484	ng/ul	100
25) 2-Nitrophenol	10.268	139	8882	5.227	ng/ul#	90
26) 2,4-Dimethylphenol	10.309	107	17819	4.431	ng/ul	92
27) Bis(2-Chloroethoxy)met...	10.544	93	19174	4.840	ng/ul	95
29) 2,4-Dichlorophenol	10.797	162	17354	4.645	ng/ul	95
30) Naphthalene	11.220	128	46717	4.847	ng/ul	98
32) 4-Chloroaniline	11.314	127	22866	5.128	ng/ul	98
33) Hexachlorobutadiene	11.496	225	14870	4.247	ng/ul	96
34) Caprolactam	12.043	113	2958	2.476	ng/ul#	87
35) 4-Chloro-3-methylphenol	12.401	107	17652	4.918	ng/ul	95
36) 2-Methylnaphthalene	12.795	142	36715	4.751	ng/ul	92

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG010423\
 Data File : BG056200.D
 Acq On : 4 Jan 2023 14:59
 Operator : CG/JU
 Sample : SST00547
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SST005449

Quant Time: Jan 05 00:18:14 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG010423.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Jan 05 00:14:25 2023
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	13.012	142	38470	4.941	ng/ul	97
39) 1,2,4,5-Tetrachloroben...	13.153	216	29908	4.414	ng/ul	96
40) Hexachlorocyclopentadiene	13.135	237	14826	3.375	ng/ul#	86
41) 2,4,6-Trichlorophenol	13.382	196	19163	4.798	ng/ul#	85
42) 2,4,5-Trichlorophenol	13.453	196	20702	4.781	ng/ul	90
43) 1,1'-Biphenyl	13.782	154	58624	5.047	ng/ul	94
44) 2-Chloronaphthalene	13.835	162	46746	4.855	ng/ul	97
45) 2-Nitroaniline	14.023	65	8749	6.544	ng/ul#	83
47) Dimethylphthalate	14.381	163	64078	5.082	ng/ul#	96
48) 2,6-Dinitrotoluene	14.504	165	12875	5.173	ng/ul#	89
50) Acenaphthylene	14.669	152	71574	5.015	ng/ul	97
51) 3-Nitroaniline	14.833	138	10550	5.643	ng/ul	92
52) Acenaphthene	15.010	153	48407	4.828	ng/ul	94
53) 2,4-Dinitrophenol	15.045	184	7657	3.952	ng/ul#	90
55) 4-Nitrophenol	15.121	109	9253	4.969	ng/ul	97
56) Dibenzofuran	15.339	168	77499	4.939	ng/ul	98
57) 2,4-Dinitrotoluene	15.286	165	18871	5.467	ng/ul#	98
58) 2,3,4,6-Tetrachlorophenol	15.556	232	20778	4.854	ng/ul#	87
59) Diethylphthalate	15.727	149	60118	5.336	ng/ul	96
61) Fluorene	15.985	166	61881	4.800	ng/ul	95
62) 4-Chlorophenyl-phenyle...	15.967	204	38313	4.575	ng/ul	99
63) 4-Nitroaniline	15.985	138	8082	4.530	ng/ul	84
66) 4,6-Dinitro-2-methylph...	16.044	198	12283	4.597	ng/ul#	93
67) N-Nitrosodiphenylamine	16.179	169	57662	5.249	ng/ul	97
68) 4-Bromophenyl-phenylether	16.860	248	26916	4.897	ng/ul	99
69) Hexachlorobenzene	16.984	284	26186	4.923	ng/ul	99
70) Atrazine	17.107	200	26689	5.096	ng/ul	95
71) Pentachlorophenol	17.325	266	16691	4.935	ng/ul	97
72) Phenanthrene	17.724	178	110250	5.009	ng/ul	98
74) Anthracene	17.812	178	113848	5.053	ng/ul	96
75) 1,2,3,4-Tetrachloroben...	13.758	216	29619	4.504	ng/ul	96
76) Pentachlorobenzene	15.262	250	32539	5.013	ng/ul	95
77) Carbazole	18.077	167	93236	5.062	ng/ul	99
78) Di-n-butylphthalate	18.605	149	94689	5.626	ng/ul	98
80) Fluoranthene	19.710	202	140565	5.351	ng/ul	99
82) Pyrene	20.074	202	145465	5.489	ng/ul	98
83) Butylbenzylphthalate	20.932	149	40582	7.196	ng/ul	98
84) 3,3'-Dichlorobenzidine	21.855	252	51166	5.630	ng/ul	96
85) Benzo(a)anthracene	21.955	228	141667	5.131	ng/ul	98
86) Bis(2-ethylhexyl)phtha...	21.819	149	60018	6.948	ng/ul	99
87) Chrysene	22.025	228	129577	5.057	ng/ul	100
89) Di-n-octyl phthalate	23.112	149	99820	6.324	ng/ul	100
90) Benzo(b)fluoranthene	24.328	252	139161	4.942	ng/ul	99
91) Benzo(k)fluoranthene	24.328	252	139161	5.065	ng/ul	99
93) Benzo(a)pyrene	25.268	252	125018	5.095	ng/ul	98
94) Indeno(1,2,3-cd)pyrene	29.405	276	139484	4.316	ng/ul	96
95) Dibenzo(a,h)anthracene	29.457	278	53878	1.993	ng/ul	96
96) Benzo(g,h,i)perylene	30.644	276	73520	2.831	ng/ul	94


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG010423\
 Data File : BG056200.D
 Acq On : 4 Jan 2023 14:59
 Operator : CG/JU
 Sample : SSTD00547
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SSTD005449

Quant Time: Jan 05 00:18:14 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG010423.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Jan 05 00:14:25 2023
 Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 1.758 ng/uL
RT: 3.670 min Scan# 61
Delta R.T. -0.005 min
Lab File: BG056200.D
Acq: 4 Jan 2023 14:59




Instrument :
BNA_G
ClientSampleId :
SSTD005449


Tgt Ion: 88 Resp: 2191
Ion Ratio Lower Upper
88 100
43 12.8 6.8 10.2#
58 21.3 20.0 30.0



#5
Pyridine
Concen: 3.876 ng/uL
RT: 4.093 min Scan# 137
Delta R.T. 0.001 min
Lab File: BG056200.D
Acq: 4 Jan 2023 14:59

Tgt Ion: 79 Resp: 13542
Ion Ratio Lower Upper
79 100
52 27.8 16.6 25.0#
51 14.1 13.4 20.2









#6
Benzaldehyde
Concen: 4.413 ng/ul
RT: 7.460 min Scan# 710
Delta R.T. 0.001 min
Lab File: BG056200.D
Acq: 4 Jan 2023 14:59




Instrument :
BNA_G
ClientSampleId :
SSTD005449


Tgt Ion: 77 Resp: 6854
Ion Ratio Lower Upper
77 100
105 114.3 86.5 129.7
106 97.2 80.8 121.2



#8
Phenol
Concen: 4.397 ng/ul
RT: 7.489 min Scan# 715
Delta R.T. -0.005 min
Lab File: BG056200.D
Acq: 4 Jan 2023 14:59

Tgt Ion: 94 Resp: 18513
Ion Ratio Lower Upper
94 100
65 21.1 15.9 23.9
66 35.4 24.6 37.0









#10
 Bis(2-Chloroethyl)ether
 Concen: 5.072 ng/ul
 RT: 7.736 min Scan# 710
 Delta R.T. 0.001 min
 Lab File: BG056200.D
 Acq: 4 Jan 2023 14:59

Instrument :
 BNA_G
 ClientSampleId :
 SSTD005449

Tgt Ion: 93 Resp: 14055




Ion	Ratio	Lower	Upper
93	100		
63	43.5	37.4	56.2
95	31.8	28.1	42.1




#12
 2-Chlorophenol
 Concen: 4.837 ng/ul
 RT: 7.889 min Scan# 783
 Delta R.T. -0.005 min
 Lab File: BG056200.D
 Acq: 4 Jan 2023 14:59

Tgt Ion: 128 Resp: 12148

Ion	Ratio	Lower	Upper
128	100		
64	38.5	29.0	43.6
130	33.9	25.3	37.9









#13
2-Methylphenol
Concen: 4.174 ng/ul
RT: 8.764 min Scan# 911
Delta R.T. 0.001 min
Lab File: BG056200.D
Acq: 4 Jan 2023 14:59




Instrument :
BNA_G
ClientSampleId :
SSTD005449


Tgt Ion:108 Resp: 13194
Ion Ratio Lower Upper
108 100
107 94.3 70.5 105.7



#14
2,2'-oxybis(1-Chloropropane)
Concen: 5.995 ng/ul
RT: 8.864 min Scan# 949
Delta R.T. 0.007 min
Lab File: BG056200.D
Acq: 4 Jan 2023 14:59

Tgt Ion: 45 Resp: 2134
Ion Ratio Lower Upper
45 100
77 46.5 90.5 135.7#
79 51.3 75.8 113.6#









#16
Acetophenone
Concen: 4.595 ng/ul
RT: 9.158 min Scan# 999
Delta R.T. -0.005 min
Lab File: BG056200.D
Acq: 4 Jan 2023 14:59




Instrument :
BNA_G
ClientSampleId :
SSTD005449


Tgt Ion:105 Resp: 24233
Ion Ratio Lower Upper
105 100
77 65.7 52.8 79.2
51 13.4 10.1 15.1



#17
N-Nitroso-di-n-propylamine
Concen: 4.831 ng/ul
RT: 9.128 min Scan# 994
Delta R.T. -0.005 min
Lab File: BG056200.D
Acq: 4 Jan 2023 14:59

Tgt Ion: 70 Resp: 10771
Ion Ratio Lower Upper
70 100
42 22.1 19.9 29.9
101 8.8 9.8 14.6#
130 17.0 18.9 28.3#









#18
4-Methylphenol
Concen: 4.449 ng/ul
RT: 9.087 min Scan# 91
Delta R.T. -0.005 min
Lab File: BG056200.D
Acq: 4 Jan 2023 14:59




Instrument :
BNA_G
ClientSampleId :
SSTD005449


Tgt Ion:108 Resp: 15234
Ion Ratio Lower Upper
108 100
107 113.5 91.5 137.3



#19
Hexachloroethane
Concen: 4.987 ng/ul
RT: 9.434 min Scan# 1046
Delta R.T. 0.001 min
Lab File: BG056200.D
Acq: 4 Jan 2023 14:59

Tgt Ion:117 Resp: 5035
Ion Ratio Lower Upper
117 100
201 151.6 132.0 198.0
199 96.4 89.4 134.0









#22
Nitrobenzene
Concen: 4.856 ng/ul
RT: 9.546 min Scan# 1065
Delta R.T. -0.005 min
Lab File: BG056200.D
Acq: 4 Jan 2023 14:59




Instrument :
BNA_G
ClientSampleId :
SSTD005449


Tgt Ion: 77 Resp: 16237
Ion Ratio Lower Upper
77 100
123 47.0 42.2 63.2
65 12.4 10.5 15.7



#23
Isophorone
Concen: 4.484 ng/ul
RT: 10.068 min Scan# 1154
Delta R.T. -0.005 min
Lab File: BG056200.D
Acq: 4 Jan 2023 14:59

Tgt Ion: 82 Resp: 34273
Ion Ratio Lower Upper
82 100
95 8.4 6.9 10.3
138 19.4 15.4 23.2









#25
2-Nitrophenol
Concen: 5.227 ng/ul
RT: 10.268 min Scan# 1188
Delta R.T. 0.001 min
Lab File: BG056200.D
Acq: 4 Jan 2023 14:59




Instrument :
BNA_G
ClientSampleId :
SSTD005449


Tgt Ion:139 Resp: 8882
Ion Ratio Lower Upper
139 100
65 36.5 31.2 46.8
109 28.1 30.1 45.1#



#26
2,4-Dimethylphenol
Concen: 4.431 ng/ul
RT: 10.309 min Scan# 1195
Delta R.T. 0.001 min
Lab File: BG056200.D
Acq: 4 Jan 2023 14:59

Tgt Ion:107 Resp: 17819
Ion Ratio Lower Upper
107 100
121 46.5 40.5 60.7
122 75.8 66.8 100.2





#27

Bis(2-Chloroethoxy)methane

Concen: 4.840 ng/ul

RT: 10.544 min Scan# 1235

Delta R.T. 0.001 min

Lab File: BG056200.D

Acq: 4 Jan 2023 14:59

Instrument :

BNA_G

ClientSampleId :

SSTD005449





Tgt Ion: 93 Resp: 19174

Ion Ratio Lower Upper

93 100

95 35.2 25.4 38.0

123 9.9 8.5 12.7



#29

2,4-Dichlorophenol

Concen: 4.645 ng/ul

RT: 10.797 min Scan# 1278

Delta R.T. -0.005 min

Lab File: BG056200.D

Acq: 4 Jan 2023 14:59




Tgt Ion:162 Resp: 17354


Ion Ratio Lower Upper

162 100

164 62.4 52.5 78.7

98 26.4 24.7 37.1









#30
Naphthalene
Concen: 4.847 ng/ul
RT: 11.220 min Scan# 1350
Delta R.T. 0.001 min
Lab File: BG056200.D
Acq: 4 Jan 2023 14:59




Instrument :
BNA_G
ClientSampleId :
SSTD005449


Tgt Ion:128 Resp: 46717
Ion Ratio Lower Upper
128 100
129 10.7 8.5 12.7
127 12.4 10.9 16.3



#32
4-Chloroaniline
Concen: 5.128 ng/ul
RT: 11.314 min Scan# 1366
Delta R.T. 0.001 min
Lab File: BG056200.D
Acq: 4 Jan 2023 14:59


Tgt Ion:127 Resp: 22866
Ion Ratio Lower Upper
127 100
129 28.5 23.5 35.3








#33
Hexachlorobutadiene
Concen: 4.247 ng/ul
RT: 11.496 min Scan# 11496
Delta R.T. 0.001 min
Lab File: BG056200.D
Acq: 4 Jan 2023 14:59

Instrument :
BNA_G
ClientSampleId :
SSTD005449







Tgt Ion:225 Resp: 14870
Ion Ratio Lower Upper
225 100
223 64.6 49.8 74.6
227 60.7 51.8 77.6



#34
Caprolactam
Concen: 2.476 ng/ul
RT: 12.043 min Scan# 1490
Delta R.T. -0.017 min
Lab File: BG056200.D
Acq: 4 Jan 2023 14:59

Tgt Ion:113 Resp: 2958
Ion Ratio Lower Upper
113 100
55 64.6 39.3 58.9#
56 46.9 39.5 59.3









#35
 4-Chloro-3-methylphenol
 Concen: 4.918 ng/ul
 RT: 12.401 min Scan# 1551
 Delta R.T. -0.005 min
 Lab File: BG056200.D
 Acq: 4 Jan 2023 14:59




Instrument :
 BNA_G
 ClientSampleId :
 SSTD005449


Tgt Ion:107 Resp: 17652
 Ion Ratio Lower Upper
 107 100
 144 27.2 21.0 31.6
 142 90.9 67.9 101.9



#36
 2-Methylnaphthalene
 Concen: 4.751 ng/ul
 RT: 12.795 min Scan# 1618
 Delta R.T. -0.005 min
 Lab File: BG056200.D
 Acq: 4 Jan 2023 14:59

Tgt Ion:142 Resp: 36715
 Ion Ratio Lower Upper
 142 100
 141 82.3 71.6 107.4





#37

1-Methylnaphthalene

Concen: 4.941 ng/ul

RT: 13.012 min Scan# 1

Delta R.T. 0.001 min

Lab File: BG056200.D

Acq: 4 Jan 2023 14:59

Instrument :

BNA_G

ClientSampleId :

SSTD005449





Tgt Ion:142 Resp: 38470

Ion Ratio Lower Upper

142 100

141 92.8 71.7 107.5

116 4.7 3.4 5.0



#39

1,2,4,5-Tetrachlorobenzene

Concen: 4.414 ng/ul

RT: 13.153 min Scan# 1679

Delta R.T. -0.005 min

Lab File: BG056200.D

Acq: 4 Jan 2023 14:59

Tgt Ion:216 Resp: 29908




Ion Ratio Lower Upper


216 100

214 80.7 62.0 93.0

179 14.2 13.8 20.8

108 9.8 7.9 11.9









#40
Hexachlorocyclopentadiene
Concen: 3.375 ng/ul
RT: 13.135 min Scan# 1676
Delta R.T. 0.001 min
Lab File: BG056200.D
Acq: 4 Jan 2023 14:59




Instrument :
BNA_G
ClientSampleId :
SSTD005449


Tgt Ion:237 Resp: 14826
Ion Ratio Lower Upper
237 100
235 49.4 48.6 72.8
272 10.1 10.2 15.4#



#41
2,4,6-Trichlorophenol
Concen: 4.798 ng/ul
RT: 13.382 min Scan# 1718
Delta R.T. -0.005 min
Lab File: BG056200.D
Acq: 4 Jan 2023 14:59

Tgt Ion:196 Resp: 19163
Ion Ratio Lower Upper
196 100
198 107.9 76.2 114.2
200 20.7 25.0 37.4#









#42
2,4,5-Trichlorophenol
Concen: 4.781 ng/ul
RT: 13.453 min Scan# 1730
Delta R.T. -0.005 min
Lab File: BG056200.D
Acq: 4 Jan 2023 14:59




Instrument :
BNA_G
ClientSampleId :
SSTD005449


Tgt Ion:196 Resp: 20702
Ion Ratio Lower Upper
196 100
198 87.3 80.2 120.4
200 29.0 23.6 35.4



#43
1,1'-Biphenyl
Concen: 5.047 ng/ul
RT: 13.782 min Scan# 1786
Delta R.T. 0.001 min
Lab File: BG056200.D
Acq: 4 Jan 2023 14:59

Tgt Ion:154 Resp: 58624
Ion Ratio Lower Upper
154 100
153 38.7 34.0 51.0
76 8.6 8.5 12.7









#44
2-Chloronaphthalene
Concen: 4.855 ng/ul
RT: 13.835 min Scan# 1795
Delta R.T. 0.001 min
Lab File: BG056200.D
Acq: 4 Jan 2023 14:59




Instrument :
BNA_G
ClientSampleId :
SSTD005449


Tgt Ion:162 Resp: 46746
Ion Ratio Lower Upper
162 100
164 34.1 25.9 38.9
127 28.8 24.6 36.8



#45
2-Nitroaniline
Concen: 6.544 ng/ul
RT: 14.023 min Scan# 1827
Delta R.T. 0.001 min
Lab File: BG056200.D
Acq: 4 Jan 2023 14:59

Tgt Ion: 65 Resp: 8749
Ion Ratio Lower Upper
65 100
92 107.9 68.2 102.2#
138 151.3 109.7 164.5









#47
Dimethylphthalate
Concen: 5.082 ng/ul
RT: 14.381 min Scan# 1888
Delta R.T. 0.001 min
Lab File: BG056200.D
Acq: 4 Jan 2023 14:59




Instrument :
BNA_G
ClientSampleId :
SSTD005449


Tgt Ion:163 Resp: 64078
Ion Ratio Lower Upper
163 100
194 4.3 3.5 5.3
164 8.5 8.6 12.8#



#48
2,6-Dinitrotoluene
Concen: 5.173 ng/ul
RT: 14.504 min Scan# 1909
Delta R.T. -0.005 min
Lab File: BG056200.D
Acq: 4 Jan 2023 14:59

Tgt Ion:165 Resp: 12875
Ion Ratio Lower Upper
165 100
89 51.2 36.2 54.4
121 21.9 12.2 18.4#









#50
Acenaphthylene
Concen: 5.015 ng/ul
RT: 14.669 min Scan# 1937
Delta R.T. -0.005 min
Lab File: BG056200.D
Acq: 4 Jan 2023 14:59




Instrument :
BNA_G
ClientSampleId :
SSTD005449


Tgt Ion:152 Resp: 71574
Ion Ratio Lower Upper
152 100
151 21.1 15.4 23.2
153 12.7 10.5 15.7



#51
3-Nitroaniline
Concen: 5.643 ng/ul
RT: 14.833 min Scan# 1965
Delta R.T. -0.005 min
Lab File: BG056200.D
Acq: 4 Jan 2023 14:59

Tgt Ion:138 Resp: 10550
Ion Ratio Lower Upper
138 100
108 9.4 8.3 12.5
92 115.0 85.2 127.8









#52
Acenaphthene
Concen: 4.828 ng/ul
RT: 15.010 min Scan# 1995
Delta R.T. 0.001 min
Lab File: BG056200.D
Acq: 4 Jan 2023 14:59




Instrument :
BNA_G
ClientSampleId :
SSTD005449


Tgt Ion:153 Resp: 48407
Ion Ratio Lower Upper
153 100
152 49.6 38.0 57.0
154 91.9 67.7 101.5



#53
2,4-Dinitrophenol
Concen: 3.952 ng/ul
RT: 15.045 min Scan# 2001
Delta R.T. 0.001 min
Lab File: BG056200.D
Acq: 4 Jan 2023 14:59

Tgt Ion:184 Resp: 7657
Ion Ratio Lower Upper
184 100
63 37.6 22.4 33.6#
154 48.3 42.2 63.4









#55
4-Nitrophenol
Concen: 4.969 ng/ul
RT: 15.121 min Scan# 2014
Delta R.T. -0.005 min
Lab File: BG056200.D
Acq: 4 Jan 2023 14:59




Instrument :
BNA_G
ClientSampleId :
SSTD005449


Tgt Ion:109 Resp: 9253
Ion Ratio Lower Upper
109 100
139 113.5 86.7 130.1
65 83.9 67.9 101.9



#56
Dibenzofuran
Concen: 4.939 ng/ul
RT: 15.339 min Scan# 2051
Delta R.T. 0.001 min
Lab File: BG056200.D
Acq: 4 Jan 2023 14:59

Tgt Ion:168 Resp: 77499
Ion Ratio Lower Upper
168 100
139 30.1 25.0 37.4









#57
2,4-Dinitrotoluene
Concen: 5.467 ng/ul
RT: 15.286 min Scan# 2042
Delta R.T. -0.005 min
Lab File: BG056200.D
Acq: 4 Jan 2023 14:59




Instrument :
BNA_G
ClientSampleId :
SSTD005449


Tgt Ion:165 Resp: 18871
Ion Ratio Lower Upper
165 100
63 25.0 20.9 31.3
182 1.9 2.7 4.1#



#58
2,3,4,6-Tetrachlorophenol
Concen: 4.854 ng/ul
RT: 15.556 min Scan# 2088
Delta R.T. 0.001 min
Lab File: BG056200.D
Acq: 4 Jan 2023 14:59

Tgt Ion:232 Resp: 20778
Ion Ratio Lower Upper
232 100
131 20.8 22.9 34.3#
130 1.8 1.7 2.5
166 28.7 17.9 26.9#









#59
Diethylphthalate
Concen: 5.336 ng/ul
RT: 15.727 min Scan# 2117
Delta R.T. -0.005 min
Lab File: BG056200.D
Acq: 4 Jan 2023 14:59




Instrument :
BNA_G
ClientSampleId :
SSTD005449


Tgt Ion:149 Resp: 60118
Ion Ratio Lower Upper
149 100
177 27.7 20.2 30.4
150 11.6 9.9 14.9



#61
Fluorene
Concen: 4.800 ng/ul
RT: 15.985 min Scan# 2161
Delta R.T. 0.001 min
Lab File: BG056200.D
Acq: 4 Jan 2023 14:59

Tgt Ion:166 Resp: 61881
Ion Ratio Lower Upper
166 100
165 103.9 78.5 117.7
167 12.5 10.4 15.6





#62

4-Chlorophenyl-phenylether

Concen: 4.575 ng/ul

RT: 15.967 min Scan# 2158

Delta R.T. 0.001 min

Lab File: BG056200.D

Acq: 4 Jan 2023 14:59

Instrument :

BNA_G

ClientSampleId :

SSTD005449





Tgt Ion:204 Resp: 38313

Ion Ratio Lower Upper

204 100

206 31.6 25.0 37.4

141 38.8 31.6 47.4



#63

4-Nitroaniline

Concen: 4.530 ng/ul

RT: 15.985 min Scan# 2161

Delta R.T. -0.005 min

Lab File: BG056200.D

Acq: 4 Jan 2023 14:59




Tgt Ion:138 Resp: 8082


Ion Ratio Lower Upper

138 100

92 40.3 38.8 58.2

108 75.8 75.0 112.4









#66
4,6-Dinitro-2-methylphenol
Concen: 4.597 ng/ul
RT: 16.044 min Scan# 2172
Delta R.T. -0.005 min
Lab File: BG056200.D
Acq: 4 Jan 2023 14:59




Instrument :
BNA_G
ClientSampleId :
SSTD005449


Tgt Ion:198 Resp: 12283
Ion Ratio Lower Upper
198 100
182 4.2 2.2 3.4#
77 17.2 11.4 17.2#



#67
N-Nitrosodiphenylamine
Concen: 5.249 ng/ul
RT: 16.179 min Scan# 2194
Delta R.T. 0.001 min
Lab File: BG056200.D
Acq: 4 Jan 2023 14:59

Tgt Ion:169 Resp: 57662
Ion Ratio Lower Upper
169 100
168 68.4 56.1 84.1
167 35.0 30.3 45.5









#68
4-Bromophenyl-phenylether
Concen: 4.897 ng/ul
RT: 16.860 min Scan# 21
Delta R.T. 0.001 min
Lab File: BG056200.D
Acq: 4 Jan 2023 14:59




Instrument :
BNA_G
ClientSampleId :
SSTD005449


Tgt Ion:248 Resp: 26916
Ion Ratio Lower Upper
248 100
250 92.8 74.8 112.2
141 46.5 37.0 55.4



#69
Hexachlorobenzene
Concen: 4.923 ng/ul
RT: 16.984 min Scan# 2331
Delta R.T. 0.001 min
Lab File: BG056200.D
Acq: 4 Jan 2023 14:59

Tgt Ion:284 Resp: 26186
Ion Ratio Lower Upper
284 100
142 24.4 18.9 28.3
249 33.0 26.8 40.2





#70

Atrazine

Concen: 5.096 ng/ul

RT: 17.107 min Scan# 21

Delta R.T. -0.005 min

Lab File: BG056200.D

Acq: 4 Jan 2023 14:59

Instrument :

BNA_G

ClientSampleId :

SSTD005449





Tgt Ion:200 Resp: 26689

Ion Ratio Lower Upper

200 100

173 20.7 18.0 27.0

215 46.9 40.8 61.2



#71

Pentachlorophenol

Concen: 4.935 ng/ul

RT: 17.325 min Scan# 2389

Delta R.T. 0.001 min

Lab File: BG056200.D

Acq: 4 Jan 2023 14:59




Tgt Ion:266 Resp: 16691


Ion Ratio Lower Upper

266 100

264 59.3 47.7 71.5

268 72.3 54.3 81.5









#72
Phenanthrene
Concen: 5.009 ng/ul
RT: 17.724 min Scan# 2457
Delta R.T. 0.001 min
Lab File: BG056200.D
Acq: 4 Jan 2023 14:59




Instrument :
BNA_G
ClientSampleId :
SSTD005449


Tgt Ion:178 Resp: 110250
Ion Ratio Lower Upper
178 100
179 16.9 12.4 18.6
176 18.4 15.0 22.6



#74
Anthracene
Concen: 5.053 ng/ul
RT: 17.812 min Scan# 2472
Delta R.T. 0.001 min
Lab File: BG056200.D
Acq: 4 Jan 2023 14:59

Tgt Ion:178 Resp: 113848
Ion Ratio Lower Upper
178 100
179 15.0 13.0 19.6
176 15.8 14.2 21.4









#75
1,2,3,4-Tetrachlorobenzene
Concen: 4.504 ng/uL
RT: 13.758 min Scan# 11
Delta R.T. 0.001 min
Lab File: BG056200.D
Acq: 4 Jan 2023 14:59




Instrument :
BNA_G
ClientSampleId :
SSTD005449


Tgt Ion:216 Resp: 29619
Ion Ratio Lower Upper
216 100
214 75.9 62.4 93.6
218 53.2 39.4 59.0



#76
Pentachlorobenzene
Concen: 5.013 ng/uL
RT: 15.262 min Scan# 2038
Delta R.T. 0.001 min
Lab File: BG056200.D
Acq: 4 Jan 2023 14:59

Tgt Ion:250 Resp: 32539
Ion Ratio Lower Upper
250 100
248 57.5 51.6 77.4
252 67.4 55.2 82.8









#77
 Carbazole
 Concen: 5.062 ng/ul
 RT: 18.077 min Scan# 2517
 Delta R.T. 0.001 min
 Lab File: BG056200.D
 Acq: 4 Jan 2023 14:59




Instrument :
 BNA_G
 ClientSampleId :
 SSTD005449


Tgt Ion:167 Resp: 93236
 Ion Ratio Lower Upper
 167 100
 166 22.0 17.4 26.0
 139 10.6 8.6 12.8



#78
 Di-n-butylphthalate
 Concen: 5.626 ng/ul
 RT: 18.605 min Scan# 2607
 Delta R.T. 0.001 min
 Lab File: BG056200.D
 Acq: 4 Jan 2023 14:59

Tgt Ion:149 Resp: 94689
 Ion Ratio Lower Upper
 149 100
 150 9.6 7.1 10.7
 104 4.5 4.2 6.4









#80
Fluoranthene
Concen: 5.351 ng/ul
RT: 19.710 min Scan# 21
Delta R.T. 0.001 min
Lab File: BG056200.D
Acq: 4 Jan 2023 14:59




Instrument :
BNA_G
ClientSampleId :
SSTD005449


Tgt Ion:202 Resp: 140565
Ion Ratio Lower Upper
202 100
101 5.9 4.4 6.6
100 4.4 3.8 5.6



#82
Pyrene
Concen: 5.489 ng/ul
RT: 20.074 min Scan# 2857
Delta R.T. 0.001 min
Lab File: BG056200.D
Acq: 4 Jan 2023 14:59

Tgt Ion:202 Resp: 145465
Ion Ratio Lower Upper
202 100
101 5.8 5.2 7.8
100 6.2 4.2 6.2









#83
 Butylbenzylphthalate
 Concen: 7.196 ng/ul
 RT: 20.932 min Scan# 3003
 Delta R.T. 0.001 min
 Lab File: BG056200.D
 Acq: 4 Jan 2023 14:59




Instrument :
 BNA_G
 ClientSampleId :
 SSTD005449


Tgt Ion:149 Resp: 40582
 Ion Ratio Lower Upper
 149 100
 91 63.3 50.0 75.0
 206 26.9 23.2 34.8



#84
 3,3'-Dichlorobenzidine
 Concen: 5.630 ng/ul
 RT: 21.855 min Scan# 3160
 Delta R.T. 0.001 min
 Lab File: BG056200.D
 Acq: 4 Jan 2023 14:59

Tgt Ion:252 Resp: 51166
 Ion Ratio Lower Upper
 252 100
 254 60.3 51.1 76.7
 126 4.6 3.8 5.6









#85
Benzo(a)anthracene
Concen: 5.131 ng/ul
RT: 21.955 min Scan# 3177
Delta R.T. 0.001 min
Lab File: BG056200.D
Acq: 4 Jan 2023 14:59




Instrument :
BNA_G
ClientSampleId :
SSTD005449


Tgt Ion:228 Resp: 141667
Ion Ratio Lower Upper
228 100
229 19.8 15.9 23.9
226 26.2 22.2 33.2



#86
Bis(2-ethylhexyl)phthalate
Concen: 6.948 ng/ul
RT: 21.819 min Scan# 3154
Delta R.T. -0.005 min
Lab File: BG056200.D
Acq: 4 Jan 2023 14:59

Tgt Ion:149 Resp: 60018
Ion Ratio Lower Upper
149 100
167 29.1 23.7 35.5
279 6.6 4.9 7.3





#87

Chrysene

Concen: 5.057 ng/ul

RT: 22.025 min Scan# 3189

Delta R.T. 0.001 min

Lab File: BG056200.D

Acq: 4 Jan 2023 14:59

Instrument :

BNA_G

ClientSampleId :

SSTD005449





Tgt Ion:228 Resp: 129577

Ion Ratio Lower Upper

228 100

226 30.5 24.5 36.7

229 19.7 15.8 23.8



#89

Di-n-octyl phthalate

Concen: 6.324 ng/ul




RT: 23.112 min Scan# 3374


Delta R.T. 0.001 min

Lab File: BG056200.D

Acq: 4 Jan 2023 14:59

Tgt Ion:149 Resp: 99820









#90
Benzo(b)fluoranthene
Concen: 4.942 ng/ul
RT: 24.328 min Scan# 3581
Delta R.T. 0.001 min
Lab File: BG056200.D
Acq: 4 Jan 2023 14:59




Instrument :
BNA_G
ClientSampleId :
SSTD005449


Tgt Ion:252 Resp: 139161
Ion Ratio Lower Upper
252 100
253 22.2 17.4 26.2
125 4.0 3.0 4.4



#91
Benzo(k)fluoranthene
Concen: 5.065 ng/ul
RT: 24.328 min Scan# 3581
Delta R.T. -0.075 min
Lab File: BG056200.D
Acq: 4 Jan 2023 14:59

Tgt Ion:252 Resp: 139161
Ion Ratio Lower Upper
252 100
253 22.2 18.0 27.0
125 4.0 3.1 4.7









#93
Benzo(a)pyrene
Concen: 5.095 ng/ul
RT: 25.268 min Scan# 31
Delta R.T. -0.005 min
Lab File: BG056200.D
Acq: 4 Jan 2023 14:59




Instrument :
BNA_G
ClientSampleId :
SSTD005449


Tgt Ion:252 Resp: 125018
Ion Ratio Lower Upper
252 100
253 22.8 17.3 25.9
125 3.6 3.2 4.8



#94
Indeno(1,2,3-cd)pyrene
Concen: 4.316 ng/ul
RT: 29.405 min Scan# 4445
Delta R.T. -0.005 min
Lab File: BG056200.D
Acq: 4 Jan 2023 14:59

Tgt Ion:276 Resp: 139484
Ion Ratio Lower Upper
276 100
138 7.4 6.2 9.2
277 21.9 19.8 29.6









#95
Dibenzo(a,h)anthracene
Concen: 1.993 ng/ul
RT: 29.457 min Scan# 4457
Delta R.T. -0.017 min
Lab File: BG056200.D
Acq: 4 Jan 2023 14:59

Instrument :
BNA_G
ClientSampleId :
SSTD005449

Tgt Ion:278 Resp: 53878
Ion Ratio Lower Upper
278 100
139 5.9 5.0 7.4
279 20.6 18.4 27.6



#96
Benzo(g,h,i)perylene
Concen: 2.831 ng/ul
RT: 30.644 min Scan# 4656
Delta R.T. -0.022 min
Lab File: BG056200.D
Acq: 4 Jan 2023 14:59

Tgt Ion:276 Resp: 73520
Ion Ratio Lower Upper
276 100
138 8.6 5.8 8.8
277 20.7 19.1 28.7

