

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG010621\
 Data File : BG047877.D
 Acq On : 6 Jan 2021 15:41
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Jan 06 16:39:21 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\8270-BG121720.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Dec 17 15:55:16 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	100	0.00
2	1,4-Dioxane	40.000	38.757	3.1	123	0.00
3	Pyridine	40.000	40.653	-1.6	116	0.00
4	n-Nitrosodimethylamine	40.000	37.681	5.8	108	0.00
5 S	2-Fluorophenol	80.000	84.163	-5.2	125	0.00
6	Aniline	40.000	43.773	-9.4	131	0.00
7 S	Phenol-d6	80.000	88.799	-11.0	128	0.00
8	2-Chlorophenol	40.000	43.151	-7.9	126	0.00
9	Benzaldehyde	40.000	43.453	-8.6	136	0.00
10 C	Phenol	40.000	45.206	-13.0	130	0.00
11	bis(2-Chloroethyl)ether	40.000	44.403	-11.0	134	0.00
12	1,3-Dichlorobenzene	40.000	42.314	-5.8	124	0.00
13 C	1,4-Dichlorobenzene	40.000	41.262	-3.2	123	0.00
14	1,2-Dichlorobenzene	40.000	42.151	-5.4	127	0.00
15	Benzyl Alcohol	40.000	45.044	-12.6	128	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	46.855	-17.1	141	0.00
17	2-Methylphenol	40.000	44.895	-12.2	134	0.00
18	Hexachloroethane	40.000	41.894	-4.7	120	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	46.165	-15.4	136	0.00
20	3+4-Methylphenols	40.000	46.671	-16.7	138	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	116	0.00
22	Acetophenone	40.000	41.434	-3.6	136	0.00
23 S	Nitrobenzene-d5	80.000	79.221	1.0	130	0.00
24	Nitrobenzene	40.000	38.808	3.0	128	0.00
25	Isophorone	40.000	42.600	-6.5	141	0.00
26 C	2-Nitrophenol	40.000	42.185	-5.5	137	0.00
27	2,4-Dimethylphenol	40.000	41.045	-2.6	138	0.00
28	bis(2-Chloroethoxy)methane	40.000	41.434	-3.6	143	0.00
29 C	2,4-Dichlorophenol	40.000	40.609	-1.5	135	0.00
30	1,2,4-Trichlorobenzene	40.000	40.726	-1.8	129	0.00
31	Naphthalene	40.000	40.376	-0.9	131	0.00
32	Benzoic acid	40.000	25.495	36.3#	153	0.00
33	4-Chloroaniline	40.000	42.663	-6.7	136	0.00
34 C	Hexachlorobutadiene	40.000	38.409	4.0	127	0.00
35	Caprolactam	40.000	44.662	-11.7	151	0.00
36 C	4-Chloro-3-methylphenol	40.000	43.635	-9.1	141	0.00
37	2-Methylnaphthalene	40.000	41.348	-3.4	136	0.00
38	1-Methylnaphthalene	40.000	40.917	-2.3	135	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	123	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	40.342	-0.9	137	0.00
41 P	Hexachlorocyclopentadiene	40.000	45.746	-14.4	141	0.00
42 S	2,4,6-Tribromophenol	80.000	84.715	-5.9	140	0.00
43 C	2,4,6-Trichlorophenol	40.000	41.461	-3.7	140	0.00
44	2,4,5-Trichlorophenol	40.000	42.490	-6.2	141	0.00
45 S	2-Fluorobiphenyl	80.000	81.600	-2.0	137	0.00
46	1,1'-Biphenyl	40.000	40.901	-2.3	138	0.00
47	2-Chloronaphthalene	40.000	40.824	-2.1	133	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG010621\
 Data File : BG047877.D
 Acq On : 6 Jan 2021 15:41
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Jan 06 16:39:21 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\8270-BG121720.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Dec 17 15:55:16 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	42.426	-6.1	141	0.00
49	Acenaphthylene	40.000	41.535	-3.8	138	0.00
50	Dimethylphthalate	40.000	41.403	-3.5	139	0.00
51	2,6-Dinitrotoluene	40.000	41.538	-3.8	141	0.00
52 C	Acenaphthene	40.000	42.800	-7.0	140	0.00
53	3-Nitroaniline	40.000	42.767	-6.9	141	0.00
54 P	2,4-Dinitrophenol	40.000	47.036	-17.6	152	0.00
55	Dibenzofuran	40.000	41.419	-3.5	141	0.00
56 P	4-Nitrophenol	40.000	46.911	-17.3	154	0.00
57	2,4-Dinitrotoluene	40.000	42.802	-7.0	147	0.00
58	Fluorene	40.000	40.714	-1.8	140	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	43.229	-8.1	145	0.00
60	Diethylphthalate	40.000	41.552	-3.9	141	0.00
61	4-Chlorophenyl-phenylether	40.000	41.087	-2.7	139	0.00
62	4-Nitroaniline	40.000	42.525	-6.3	140	0.00
63	Azobenzene	40.000	40.721	-1.8	138	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	116	0.00
65	4,6-Dinitro-2-methylphenol	40.000	43.201	-8.0	139	0.00
66 c	n-Nitrosodiphenylamine	40.000	42.310	-5.8	139	0.00
67	4-Bromophenyl-phenylether	40.000	42.281	-5.7	137	0.00
68	Hexachlorobenzene	40.000	41.453	-3.6	135	0.00
69	Atrazine	40.000	34.465	13.8	114	0.00
70 C	Pentachlorophenol	40.000	44.253	-10.6	147	0.00
71	Phenanthrene	40.000	40.750	-1.9	137	0.00
72	Anthracene	40.000	41.045	-2.6	138	0.00
73	Carbazole	40.000	41.016	-2.5	140	0.00
74	Di-n-butylphthalate	40.000	40.615	-1.5	134	0.00
75 C	Fluoranthene	40.000	37.896	5.3	128	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	102	0.00
77	Benzydine	40.000	32.386	19.0	86	0.00
78	Pyrene	40.000	42.921	-7.3	127	0.00
79 S	Terphenyl-d14	80.000	83.450	-4.3	123	0.00
80	Butylbenzylphthalate	40.000	41.498	-3.7	121	0.00
81	Benzo(a)anthracene	40.000	40.483	-1.2	118	0.00
82	3,3'-Dichlorobenzidine	40.000	40.271	-0.7	115	0.00
83	Chrysene	40.000	40.663	-1.7	120	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	41.412	-3.5	120	0.00
85 c	Di-n-octyl phthalate	40.000	41.291	-3.2	119	0.00
86 I	Perylene-d12	20.000	20.000	0.0	98	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	40.884	-2.2	113	0.01
88	Benzo(b)fluoranthene	40.000	41.353	-3.4	114	0.00
89	Benzo(k)fluoranthene	40.000	40.074	-0.2	113	0.00
90 C	Benzo(a)pyrene	40.000	40.983	-2.5	114	0.00
91	Dibenzo(a,h)anthracene	40.000	41.512	-3.8	115	0.00
92	Benzo(g,h,i)perylene	40.000	41.477	-3.7	115	-0.02

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG010621\
Data File : BG047877.D
Acq On : 6 Jan 2021 15:41
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTDCCC040

Quant Time: Jan 06 16:39:21 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\8270-BG121720.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Dec 17 15:55:16 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 0