

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG010821\
 Data File : BG047896.D
 Acq On : 8 Jan 2021 21:25
 Operator : CG/JU
 Sample : M1044-13
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 TP-2

Quant Time: Jan 09 00:58:26 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\8270-BG121720.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jan 08 11:19:41 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.161	152	42928	20.00	ng	0.00
21) Naphthalene-d8	10.993	136	96993	20.00	ng	# 0.00
39) Acenaphthene-d10	14.794	164	78619	20.00	ng	0.00
64) Phenanthrene-d10	17.532	188	274974	20.00	ng	# 0.00
76) Chrysene-d12	21.804	240	308019	20.00	ng	# 0.00
86) Perylene-d12	25.135	264	223355	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.746	112	299199	120.57	ng	0.00
7) Phenol-d6	7.397	99	415009	114.41	ng	0.00
23) Nitrobenzene-d5	9.330	82	234553	88.95	ng	0.00
42) 2,4,6-Tribromophenol	16.286	330	248782	185.36	ng	0.00
45) 2-Fluorobiphenyl	13.419	172	529376	87.13	ng	0.00
79) Terphenyl-d14	20.123	244	1551648	83.01	ng	0.00
Target Compounds						
10) Phenol	7.438	94	9206	2.80	ng	# 38
50) Dimethylphthalate	14.230	163	41355	6.10	ng	# 98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG010821\
Data File : BG047896.D
Acq On : 8 Jan 2021 21:25
Operator : CG/JU
Sample : M1044-13
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
TP-2

Quant Time: Jan 09 00:58:26 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\8270-BG121720.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jan 08 11:19:41 2021
Response via : Initial Calibration

