

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG010824\
 Data File : BG060291.D
 Acq On : 8 Jan 2024 21:15
 Operator : MA/JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 ICVBG010824

Quant Time: Jan 09 04:10:23 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG010824.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jan 09 04:07:12 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	114	0.00
2	1,4-Dioxane	40.000	39.556	1.1	99	0.00
3	Pyridine	40.000	39.687	0.8	95	0.00
4	n-Nitrosodimethylamine	40.000	40.292	-0.7	101	0.00
5 S	2-Fluorophenol	80.000	81.485	-1.9	124	0.00
6	Aniline	40.000	40.832	-2.1	108	0.00
7 S	Phenol-d6	80.000	82.601	-3.3	101	0.00
8	2-Chlorophenol	40.000	40.305	-0.8	112	0.00
9	Benzaldehyde	40.000	40.092	-0.2	105	0.00
10 C	Phenol	40.000	41.108	-2.8	101	0.00
11	bis(2-Chloroethyl)ether	40.000	40.495	-1.2	114	0.00
12	1,3-Dichlorobenzene	40.000	39.645	0.9	112	0.00
13 C	1,4-Dichlorobenzene	40.000	40.085	-0.2	116	0.00
14	1,2-Dichlorobenzene	40.000	37.673	5.8	96	0.00
15	Benzyl Alcohol	40.000	40.094	-0.2	106	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	37.945	5.1	97	0.00
17	2-Methylphenol	40.000	39.106	2.2	96	0.00
18	Hexachloroethane	40.000	38.753	3.1	99	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	39.063	2.3	93	0.00
20	3+4-Methylphenols	40.000	39.998	0.0	97	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	95	0.00
22	Acetophenone	40.000	40.421	-1.1	95	0.00
23 S	Nitrobenzene-d5	80.000	83.758	-4.7	99	0.00
24	Nitrobenzene	40.000	40.791	-2.0	95	0.00
25	Isophorone	40.000	40.307	-0.8	95	0.00
26 C	2-Nitrophenol	40.000	42.306	-5.8	98	0.00
27	2,4-Dimethylphenol	40.000	40.595	-1.5	94	0.00
28	bis(2-Chloroethoxy)methane	40.000	40.534	-1.3	96	0.00
29 C	2,4-Dichlorophenol	40.000	41.077	-2.7	95	0.00
30	1,2,4-Trichlorobenzene	40.000	40.239	-0.6	95	0.00
31	Naphthalene	40.000	39.424	1.4	95	0.00
32	Benzoic acid	40.000	37.954	5.1	94	0.00
33	4-Chloroaniline	40.000	39.590	1.0	95	0.00
34 C	Hexachlorobutadiene	40.000	38.956	2.6	99	0.00
35	Caprolactam	40.000	39.921	0.2	96	0.00
36 C	4-Chloro-3-methylphenol	40.000	39.874	0.3	95	0.00
37	2-Methylnaphthalene	40.000	38.887	2.8	96	0.00
38	1-Methylnaphthalene	40.000	38.429	3.9	96	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	93	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	38.970	2.6	95	0.00
41 P	Hexachlorocyclopentadiene	40.000	39.890	0.3	90	0.00
42 S	2,4,6-Tribromophenol	80.000	81.299	-1.6	95	0.00
43 C	2,4,6-Trichlorophenol	40.000	40.088	-0.2	95	0.00
44	2,4,5-Trichlorophenol	40.000	41.125	-2.8	97	0.00
45 S	2-Fluorobiphenyl	80.000	77.479	3.2	97	0.00
46	1,1'-Biphenyl	40.000	38.954	2.6	95	0.00
47	2-Chloronaphthalene	40.000	39.422	1.4	95	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG010824\
 Data File : BG060291.D
 Acq On : 8 Jan 2024 21:15
 Operator : MA/JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 ICVBG010824

Quant Time: Jan 09 04:10:23 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG010824.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jan 09 04:07:12 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	41.843	-4.6	96	0.00
49	Acenaphthylene	40.000	40.402	-1.0	95	0.00
50	Dimethylphthalate	40.000	40.096	-0.2	95	0.00
51	2,6-Dinitrotoluene	40.000	42.537	-6.3	96	0.00
52 C	Acenaphthene	40.000	40.342	-0.9	96	0.00
53	3-Nitroaniline	40.000	41.835	-4.6	95	0.00
54 P	2,4-Dinitrophenol	40.000	39.708	0.7	98	0.00
55	Dibenzofuran	40.000	39.897	0.3	95	0.00
56 P	4-Nitrophenol	40.000	44.470	-11.2	94	0.00
57	2,4-Dinitrotoluene	40.000	42.572	-6.4	97	0.00
58	Fluorene	40.000	39.573	1.1	95	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	40.998	-2.5	95	0.00
60	Diethylphthalate	40.000	40.586	-1.5	95	0.00
61	4-Chlorophenyl-phenylether	40.000	39.479	1.3	97	0.00
62	4-Nitroaniline	40.000	41.736	-4.3	95	0.00
63	Azobenzene	40.000	39.686	0.8	94	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	94	0.00
65	4,6-Dinitro-2-methylphenol	40.000	42.699	-6.7	98	0.00
66 c	n-Nitrosodiphenylamine	40.000	40.418	-1.0	94	0.00
67	4-Bromophenyl-phenylether	40.000	40.225	-0.6	94	0.00
68	Hexachlorobenzene	40.000	40.475	-1.2	96	0.00
69	Atrazine	40.000	39.262	1.8	93	0.00
70 C	Pentachlorophenol	40.000	44.367	-10.9	97	0.00
71	Phenanthrene	40.000	40.135	-0.3	95	0.00
72	Anthracene	40.000	39.927	0.2	94	0.00
73	Carbazole	40.000	40.276	-0.7	94	0.00
74	Di-n-butylphthalate	40.000	41.270	-3.2	95	0.00
75 C	Fluoranthene	40.000	39.553	1.1	95	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	93	0.00
77	Benzydine	40.000	34.457	13.9	78	0.00
78	Pyrene	40.000	39.419	1.5	95	0.00
79 S	Terphenyl-d14	80.000	81.722	-2.2	97	0.00
80	Butylbenzylphthalate	40.000	42.420	-6.1	96	0.00
81	Benzo(a)anthracene	40.000	39.709	0.7	94	0.00
82	3,3'-Dichlorobenzidine	40.000	40.511	-1.3	92	0.00
83	Chrysene	40.000	39.813	0.5	95	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	42.162	-5.4	96	0.00
85 c	Di-n-octyl phthalate	40.000	39.099	2.3	88	0.00
86 I	Perylene-d12	20.000	20.000	0.0	84	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	43.865	-9.7	92	0.00
88	Benzo(b)fluoranthene	40.000	40.826	-2.1	84	0.00
89	Benzo(k)fluoranthene	40.000	40.760	-1.9	84	0.00
90 C	Benzo(a)pyrene	40.000	40.282	-0.7	84	0.00
91	Dibenzo(a,h)anthracene	40.000	43.732	-9.3	90	0.00
92	Benzo(g,h,i)perylene	40.000	43.306	-8.3	91	-0.01

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG010824\
Data File : BG060291.D
Acq On : 8 Jan 2024 21:15
Operator : MA/JU
Sample : SSTDICV040
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
ICVBG010824

Quant Time: Jan 09 04:10:23 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG010824.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jan 09 04:07:12 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0