

Data Path : Z:\SVOASRV\HPCHEM1\BNA G\DATA\BG011019\
 Data File : BG039062.D
 Acq On : 10 Jan 2019 21:17
 Operator : JU/SJ
 Sample : K1086-01
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 CF-1819-W

Quant Time: Jan 11 05:44:20 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG010919.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jan 09 14:23:53 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.03	152	33437	20.00	ng	0.00
21) Naphthalene-d8	10.84	136	121369	20.00	ng	0.00
39) Acenaphthene-d10	14.68	164	72529	20.00	ng	0.00
64) Phenanthrene-d10	17.49	188	191607	20.00	ng	0.07
76) Chrysene-d12	21.77	240	181830	20.00	ng	0.07
87) Perylene-d12	25.03	264	134232	20.00	ng	0.05
System Monitoring Compounds						
5) 2-Fluorophenol	5.58	112	69101	39.20	ng	0.00
7) Phenol-d6	7.18	99	54785	21.60	ng	0.00
23) Nitrobenzene-d5	9.20	82	153156	69.05	ng	0.00
42) 2,4,6-Tribromophenol	16.21	330	146276	120.31	ng	0.04
45) 2-Fluorobiphenyl	13.29	172	368045	69.45	ng	0.00
79) Terphenyl-d14	20.11	244	745348	75.83	ng	0.09
Target Compounds						
10) Phenol	7.21	94	13412	5.274	ng	95
50) Dimethylphthalate	14.12	163	34997	5.710	ng	# 78
71) Phenanthrene	17.54	178	407553	40.241	ng	# 58

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA G\DATA\BG011019\
Data File : BG039062.D
Acq On : 10 Jan 2019 21:17
Operator : JU/SJ
Sample : K1086-01
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
CF-1819-W

Quant Time: Jan 11 05:44:20 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG010919.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jan 09 14:23:53 2019
Response via : Initial Calibration

