


Data Path : Z:\SVOASRV\HPCHEM1\BNA G\DATA\BG011019\
 Data File : BG039071.D
 Acq On : 11 Jan 2019 3:10
 Operator : JU/SJ
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\8270-BG010919.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Wed Jan 09 14:23:53 2019



AutoFind: Scans 2505, 2506, 2507; Background Corrected with Scan 2499

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	31.8	14390	PASS
68	69	0.00	2	0.6	71	PASS
69	198	0.00	100	28.0	12647	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	40.0	18095	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	45208	PASS
199	198	5	9	6.8	3074	PASS
275	198	10	60	30.8	13932	PASS
365	198	1	100	3.6	1613	PASS
441	198	0.01	100	12.4	5614	PASS
442	198	50	100	80.4	36328	PASS
443	442	15	24	19.8	7192	PASS