

Data Path : Z:\SVOASRV\HPCHEM1\BNA G\DATA\BG011419\
 Data File : BG039126.D
 Acq On : 14 Jan 2019 18:30
 Operator : JU/SJ
 Sample : K1127-02
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 COMP-1

Quant Time: Jan 14 23:10:42 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG010919.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jan 09 14:23:53 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.03	152	29945	20.00	ng	0.00
21) Naphthalene-d8	10.85	136	118719	20.00	ng	0.00
39) Acenaphthene-d10	14.67	164	67751	20.00	ng	0.00
64) Phenanthrene-d10	17.42	188	153831	20.00	ng	0.00
76) Chrysene-d12	21.69	240	166900	20.00	ng	-0.01
87) Perylene-d12	24.97	264	183413	20.00	ng	-0.02
System Monitoring Compounds						
5) 2-Fluorophenol	5.59	112	66485	42.12	ng	0.00
7) Phenol-d6	7.20	99	63693	28.04	ng	0.01
23) Nitrobenzene-d5	9.21	82	102525	47.26	ng	0.00
42) 2,4,6-Tribromophenol	16.16	330	90848	79.99	ng	0.00
45) 2-Fluorobiphenyl	13.29	172	272653	55.08	ng	0.00
79) Terphenyl-d14	20.02	244	482245	53.45	ng	0.00
Target Compounds						
10) Phenol	7.23	94	42410	18.620	ng	Qvalue 96
15) Benzyl Alcohol	8.29	79	50473	29.502	ng	99
17) 2-Methylphenol	8.48	107	14253	9.012	ng	98
20) 3+4-Methylphenols	8.81	107	113386	50.286	ng	97
27) 2,4-Dimethylphenol	10.02	122	74073	44.388	ng	92
31) Naphthalene	10.89	128	48112	8.080	ng	96
37) 2-Methylnaphthalene	12.50	142	11905	2.667	ng	# 90
50) Dimethylphthalate	14.13	163	36650	6.401	ng	# 95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA G\DATA\BG011419\
Data File : BG039126.D
Acq On : 14 Jan 2019 18:30
Operator : JU/SJ
Sample : K1127-02
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
COMP-1

Quant Time: Jan 14 23:10:42 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG010919.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jan 09 14:23:53 2019
Response via : Initial Calibration

