




Data Path : Z:\svoasrv\HPCHEM1\BNA G\Data\BG011619\
 Data File : BG039152.D
 Acq On : 16 Jan 2019 10:59
 Operator : JU/SJ
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Quant Time: Jan 16 14:23:53 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG010919.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jan 09 14:23:53 2019
 Response via : Initial Calibration



TIC: BG039152.D

(77) Benzidine




19.645min (-0.014) 9749.77ng

response 193537

Ion	Exp%	Act%
184.00	100	100
185.00	15.10	14.80
183.00	11.90	13.08
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA G\Data\BG011619\
Data File : BG039152.D
Acq On : 16 Jan 2019 10:59
Operator : JU/SJ
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Jan 16 14:23:53 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG010919.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jan 09 14:23:53 2019
Response via : Initial Calibration



(70) Pentachlorophenol (C)

17.060min (-0.007) 0.00ng

response 22357

Ion	Exp%	Act%
265.70	100	100
268.00	65.00	65.10
264.00	68.20	61.72
0.00	0.00	0.00