



Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG011722\
 Data File : BG052016.D
 Acq On : 17 Jan 2022 9:41
 Operator : CG/JU
 Sample : DFTPP455
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 DFTPP455

Integration File: LSCINT.P

Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG010622.M
 Title : SVOA CALIBRATION
 Last Update : Thu Jan 06 14:43:02 2022



AutoFind: Scans 2482, 2483, 2484; Background Corrected with Scan 2476

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	41.9	37501	PASS
68	69	0.00	2	0.0	0	PASS
69	198	0.00	100	42.6	38088	PASS
70	69	0.00	2	0.7	269	PASS
127	198	10	80	47.3	42285	PASS
197	198	0.00	2	0.2	215	PASS
198	198	100	100	100.0	89411	PASS
199	198	5	9	7.2	6457	PASS
275	198	10	60	32.4	29005	PASS
365	198	1	100	3.8	3391	PASS
441	198	0.01	100	13.1	11724	PASS
442	198	50	100	78.8	70440	PASS
443	442	15	24	20.0	14062	PASS

m/z	Abundance
39.05	200.0
40.15	246.0
44.15	650.0
57.15	168.0
154.95	161.0
207.05	264.0

Instrument :
BNA_G
ClientSampleId :
DFTPP455

m/z	Abundance
38.15	673.0
39.15	2784.0
40.05	353.0
40.95	154.0
44.05	634.0
50.15	9288.0
51.15	36864.0
52.15	1764.0
55.25	270.0
56.15	947.0
57.05	2514.0
61.05	628.0
62.05	558.0
63.15	2058.0
64.15	320.0
65.05	841.0
69.05	36264.0
70.05	224.0
73.05	686.0
74.15	3595.0
75.15	6384.0
76.15	2319.0
77.15	40296.0
78.15	3151.0
79.15	2449.0
80.05	2797.0
81.15	3393.0
82.05	787.0
83.05	733.0
84.05	159.0
85.15	513.0
86.15	1043.0
86.95	360.0
88.05	183.0
91.05	494.0
92.05	556.0
93.05	4181.0
94.05	228.0
96.15	230.0
97.05	205.0
98.15	3715.0
99.05	2622.0
99.95	158.0
101.15	1516.0
103.05	759.0
104.05	995.0
105.15	1228.0
107.05	12475.0
108.05	1988.0
110.05	19592.0
111.05	3581.0
112.05	468.0
112.95	213.0
116.05	773.0
117.05	12185.0
118.15	747.0
119.95	252.0
122.05	975.0
123.05	1156.0
124.05	568.0
125.15	473.0
127.05	37504.0
128.15	3185.0
129.05	16327.0
130.05	1564.0
131.15	247.0
132.35	226.0
133.95	574.0
135.05	1391.0
136.05	801.0
137.05	643.0
140.05	212.0
141.05	2446.0
142.05	719.0
143.05	371.0
143.95	185.0
146.05	438.0
147.15	1481.0
148.05	2876.0
149.05	457.0
150.95	229.0
151.85	189.0
152.95	560.0
154.15	518.0
155.05	1231.0
156.15	1588.0
156.85	212.0
158.05	299.0
159.05	448.0
159.95	710.0
161.05	1256.0
161.95	636.0

Instrument :
BNA_G
ClientSampleId :
DFTPP455

163.85	158.0
165.05	804.0
166.15	804.0
167.05	4056.0
168.05	2029.0
169.15	320.0
170.95	165.0
171.95	386.0
173.05	644.0
174.05	661.0
175.15	1621.0
176.05	666.0
176.95	910.0
178.05	290.0
179.05	3267.0
180.05	2051.0
181.05	1196.0
182.05	308.0
184.05	172.0
185.05	1556.0
186.05	11419.0
187.05	3238.0
188.05	470.0
189.05	718.0
189.85	200.0
191.05	494.0
192.05	1232.0
193.05	1009.0
194.05	239.0
194.85	231.0
195.15	210.0
196.05	2217.0
198.05	75656.0
199.05	5520.0
200.05	546.0
201.45	531.0
202.95	638.0
204.15	3199.0
205.15	5271.0
206.15	23016.0
207.15	3260.0
208.15	879.0
209.05	311.0
210.25	356.0
211.05	876.0
211.65	278.0
215.05	332.0
216.05	628.0
217.05	6113.0
218.05	782.0
221.05	2845.0
222.05	933.0
223.05	1402.0
224.05	12971.0
225.05	3281.0
226.05	566.0
227.05	5755.0
227.95	943.0
228.95	890.0
230.05	221.0
231.05	601.0
234.05	475.0
235.05	523.0
235.95	197.0
237.05	490.0
239.05	282.0
241.15	513.0
241.95	524.0
243.15	734.0
244.05	9188.0
245.05	1337.0
246.05	2170.0
247.05	573.0
249.05	386.0
250.95	206.0
253.15	169.0
255.05	50712.0
256.05	7943.0
257.05	733.0
258.05	3438.0
259.05	447.0
260.05	246.0
265.05	1143.0
266.05	424.0
272.25	180.0
273.05	1334.0
274.05	4219.0
275.05	23040.0
276.05	3023.0
277.05	1611.0
278.05	371.0
282.95	301.0
283.95	200.0
284.95	334.0
292.15	188.0
293.05	321.0

Instrument :
BNA_G
ClientSampleId :
DFTPP455

294.95	196.0
296.05	6391.0
297.05	1236.0
301.15	184.0
303.05	958.0
314.05	181.0
314.95	511.0
316.15	318.0
321.95	270.0
323.05	2226.0
323.95	420.0
326.95	326.0
328.05	168.0
328.25	167.0
333.05	226.0
334.05	966.0
335.15	406.0
340.95	335.0
346.05	410.0
352.05	451.0
352.95	424.0
354.05	747.0
365.05	2358.0
365.85	229.0
372.05	1079.0
383.05	236.0
402.05	481.0
403.05	666.0
403.95	200.0
421.05	381.0
421.95	639.0
422.95	3063.0
424.05	406.0
440.95	7559.0
442.05	48872.0
443.05	9526.0
444.05	1209.0

Instrument :
BNA_G
ClientSampleId :
DFTPP455

m/z	Abundance
37.15	227.0
38.15	553.0
39.15	3343.0
40.15	217.0
41.15	220.0
44.15	557.0
50.15	12146.0
51.15	44536.0
52.15	2112.0
55.15	375.0
56.15	1284.0
57.15	2969.0
58.05	154.0
61.15	535.0
62.15	747.0
63.15	1747.0
64.05	278.0
65.15	943.0
69.05	45872.0
70.05	231.0
73.25	793.0
74.15	4622.0
75.05	7517.0
76.15	2601.0
77.15	51600.0
78.15	3359.0
79.15	4145.0
80.15	2249.0
81.05	4075.0
82.15	828.0
83.05	1072.0
85.05	732.0
86.05	1136.0
87.15	496.0
91.05	860.0
92.15	1071.0
93.15	5827.0
94.15	369.0
95.05	170.0
95.95	244.0
97.05	260.0
98.05	5092.0
99.05	3086.0
99.95	214.0
101.05	1663.0
102.95	581.0
104.05	1711.0
105.05	1458.0
105.95	344.0
107.05	16264.0
108.05	2374.0
108.95	519.0
110.05	25720.0
111.15	4286.0
112.15	574.0
113.05	249.0
115.05	183.0
116.15	1019.0
117.15	14496.0
118.05	997.0
119.35	208.0
122.05	1520.0
123.05	1920.0
124.05	898.0
124.95	959.0
127.05	50248.0
128.15	3620.0
129.05	21488.0
130.15	1655.0
131.05	520.0
133.15	184.0
134.05	531.0
135.05	1760.0
135.95	580.0
137.15	820.0
139.15	236.0
139.95	482.0
141.15	2991.0
142.15	854.0
143.05	835.0
143.95	167.0
145.15	154.0
146.15	532.0
147.05	1630.0
148.05	3291.0
149.15	783.0
150.05	182.0
151.25	284.0
152.95	1042.0
154.05	737.0
155.05	1982.0
156.05	2244.0

Instrument :
BNA_G
ClientSampleId :
DFTPP455

157.15	375.0
157.95	561.0
159.05	416.0
160.05	918.0
161.15	1216.0
161.95	337.0
165.05	766.0
166.15	991.0
167.15	4934.0
168.05	2545.0
169.05	646.0
170.05	195.0
170.95	317.0
171.95	390.0
173.15	745.0
174.15	1002.0
175.15	2341.0
176.05	707.0
176.95	1011.0
177.95	392.0
179.05	4270.0
180.05	2882.0
181.05	1267.0
182.05	210.0
184.05	454.0
185.05	2101.0
186.05	15112.0
187.15	5259.0
188.15	677.0
189.05	1110.0
189.95	163.0
191.05	542.0
192.05	1510.0
193.05	1567.0
194.05	451.0
195.05	386.0
196.15	3392.0
196.75	645.0
198.05	108000.0
199.05	7666.0
200.05	538.0
201.55	606.0
203.05	1014.0
204.05	4502.0
205.05	8059.0
206.05	30904.0
207.15	3970.0
208.05	991.0
209.05	408.0
210.25	584.0
210.95	1147.0
212.05	301.0
215.15	368.0
216.15	701.0
217.05	9175.0
218.05	966.0
219.85	195.0
221.05	3892.0
223.05	1972.0
224.05	17912.0
225.05	4243.0
226.15	479.0
227.05	8039.0
228.05	1341.0
229.05	1580.0
230.15	255.0
231.15	432.0
233.15	192.0
234.05	541.0
234.95	622.0
235.95	373.0
236.95	585.0
238.95	327.0
239.95	199.0
241.15	454.0
241.95	646.0
243.05	987.0
244.05	13408.0
245.05	1881.0
246.05	3284.0
247.05	409.0
247.95	220.0
248.95	613.0
251.15	192.0
252.15	159.0
253.05	358.0
255.05	73696.0
256.05	10747.0
256.95	698.0
258.05	4715.0
259.05	791.0
259.75	163.0
260.05	185.0
265.05	1737.0
265.95	251.0
270.15	185.0

Instrument :
BNA_G
ClientSampleId :
DFTPP455

272.05	258.0
273.05	2077.0
274.05	6096.0
275.05	33944.0
276.05	4532.0
277.05	2894.0
278.05	386.0
282.25	200.0
282.95	455.0
284.05	176.0
284.85	492.0
289.05	291.0
290.85	162.0
293.05	552.0
294.15	287.0
294.95	171.0
296.05	9337.0
297.05	1095.0
297.95	219.0
301.95	219.0
303.05	1221.0
304.05	292.0
308.05	168.0
309.85	191.0
314.15	618.0
315.05	1029.0
316.05	667.0
321.05	320.0
323.05	3094.0
324.15	688.0
326.95	781.0
328.15	244.0
333.15	286.0
334.05	2045.0
334.95	437.0
341.05	404.0
342.15	156.0
345.95	798.0
351.95	750.0
353.05	611.0
354.05	997.0
355.15	173.0
365.05	3650.0
366.05	475.0
369.95	167.0
370.95	241.0
372.05	1688.0
373.05	402.0
382.95	461.0
389.85	214.0
390.95	161.0
391.25	166.0
392.15	193.0
401.95	708.0
403.05	1086.0
404.05	340.0
421.05	801.0
422.05	810.0
423.05	4643.0
424.05	945.0
441.05	13076.0
442.05	81912.0
443.05	16512.0
444.05	1927.0

Instrument :
BNA_G
ClientSampleId :
DFTPP455

m/z	Abundance
37.05	167.0
38.15	441.0
39.15	2499.0
40.05	359.0
41.05	154.0
43.05	267.0
44.05	562.0
49.25	177.0
50.15	8374.0
51.15	31104.0
52.15	1468.0
55.15	336.0
56.15	1034.0
57.15	2438.0
58.15	180.0
61.05	365.0
62.05	553.0
63.15	1343.0
64.15	172.0
65.15	741.0
69.05	32128.0
70.15	353.0
73.05	390.0
74.15	3904.0
75.05	5355.0
76.15	1576.0
77.15	36320.0
78.15	2526.0
79.15	2436.0
80.05	1863.0
81.05	2542.0
82.05	701.0
83.15	700.0
85.15	468.0
86.05	885.0
87.15	667.0
88.05	169.0
91.05	892.0
92.15	604.0
93.05	4896.0
94.15	308.0
95.25	264.0
96.05	238.0
98.15	3512.0
99.05	2490.0
100.15	475.0
101.05	1565.0
103.05	632.0
104.15	1075.0
105.15	968.0
107.05	12493.0
108.15	1994.0
110.05	18376.0
111.05	3316.0
112.05	342.0
116.15	852.0
117.05	10844.0
117.95	662.0
122.05	1107.0
123.05	1174.0
124.05	607.0
125.15	445.0
127.05	39104.0
128.15	2768.0
129.05	16592.0
130.15	1515.0
134.05	462.0
135.05	1085.0
136.05	672.0
137.15	680.0
139.15	214.0
140.15	273.0
141.05	2594.0
142.15	631.0
143.05	507.0
144.05	178.0
145.15	166.0
146.15	302.0
147.05	1177.0
148.05	2774.0
149.05	612.0
150.05	180.0
151.15	368.0
153.05	849.0
154.15	532.0
155.05	1312.0
156.05	1707.0
157.05	414.0
158.05	509.0
158.95	399.0
160.05	692.0
161.05	1304.0

Instrument :
BNA_G
ClientSampleId :
DFTPP455

162.05	288.0
164.35	183.0
165.05	818.0
165.95	787.0
167.05	3649.0
168.05	2407.0
169.05	484.0
171.15	237.0
172.05	352.0
173.15	677.0
174.15	755.0
175.15	1523.0
176.05	477.0
177.05	813.0
177.85	272.0
179.05	4009.0
180.05	2549.0
181.05	1093.0
184.05	364.0
185.05	1854.0
186.05	13790.0
187.15	3571.0
188.15	420.0
189.05	821.0
190.15	231.0
191.05	457.0
192.05	1005.0
193.05	1176.0
194.05	328.0
195.05	221.0
196.05	2724.0
198.05	84576.0
199.05	6185.0
200.05	634.0
201.25	439.0
202.05	183.0
203.15	611.0
204.05	3384.0
205.05	6000.0
206.05	25368.0
207.15	3352.0
208.05	909.0
209.15	297.0
209.95	316.0
211.05	853.0
211.85	294.0
215.05	403.0
215.95	506.0
217.05	6614.0
218.05	812.0
219.05	160.0
220.15	284.0
221.05	3941.0
221.75	742.0
223.05	1277.0
224.05	14627.0
225.05	3854.0
226.05	500.0
227.05	6164.0
228.05	921.0
229.05	1331.0
231.05	538.0
232.75	206.0
233.95	365.0
235.05	587.0
236.05	450.0
237.05	399.0
238.15	198.0
238.95	269.0
240.15	229.0
241.05	589.0
242.15	732.0
243.15	873.0
244.05	11985.0
245.05	1967.0
246.05	2696.0
247.05	544.0
248.15	159.0
248.95	399.0
249.85	154.0
251.05	211.0
252.95	161.0
255.05	66768.0
256.05	9563.0
257.05	745.0
258.05	3662.0
259.05	550.0
261.15	162.0
265.05	1451.0
265.95	326.0
267.15	153.0
270.75	163.0
271.95	277.0
273.15	2014.0
274.05	5118.0
275.05	30032.0

Instrument :
BNA_G
ClientSampleId :
DFTPP455

276.05	3637.0
277.05	2976.0
278.15	290.0
283.05	356.0
283.95	241.0
284.95	219.0
289.05	190.0
291.75	216.0
293.05	691.0
294.95	248.0
296.05	8622.0
297.05	822.0
302.05	229.0
303.05	979.0
304.05	344.0
314.05	543.0
315.05	918.0
316.05	541.0
320.95	223.0
321.95	217.0
323.05	2275.0
324.05	489.0
326.95	466.0
327.85	171.0
332.15	203.0
332.95	286.0
334.05	1677.0
335.05	547.0
340.95	330.0
345.95	513.0
352.05	1054.0
352.95	496.0
354.05	683.0
365.05	4164.0
366.05	639.0
369.85	171.0
371.05	195.0
372.05	1340.0
373.05	445.0
382.95	412.0
402.05	284.0
403.05	914.0
404.05	318.0
421.05	603.0
422.05	733.0
423.05	4466.0
424.05	1155.0
441.05	14537.0
442.05	80536.0
443.05	16147.0
444.05	1430.0

Instrument :
BNA_G
ClientSampleId :
DFTPP455