




Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG011823\
Data File : BG056329.D
Acq On : 18 Jan 2023 15:18
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Jan 19 04:51:12 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG122722.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jan 19 00:31:30 2023
Response via : Initial Calibration



TIC: BG056329.D\data.ms

(70) Pentachlorophenol (C)




17.306min (-0.002) 23212.89 ng

response 120329

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	66.20	65.85
264.00	64.30	62.51
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG011823\
 Data File : BG056329.D
 Acq On : 18 Jan 2023 15:18
 Operator : CG/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Quant Time: Jan 19 04:51:12 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG122722.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jan 19 00:31:30 2023
 Response via : Initial Calibration



TIC: BG056329.D\data.ms

(77) Benzidine**19.861min (+ 0.004) 99824.65 ng****response 661176**

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	15.20	13.99
183.00	11.70	12.13
0.00	0.00	0.00