



Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG011823\
Data File : BG056323.D
Acq On : 18 Jan 2023 10:58
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
DFTPP

Integration File: rteint.p

Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG122722.M
Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
Last Update : Thu Jan 19 00:31:30 2023






AutoFind: Scans 2513, 2514, 2515; Background Corrected with Scan 2506

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	15.7	21765	PASS
68	69	0.00	2	0.8	246	PASS
69	198	0.00	100	22.8	31717	PASS
70	69	0.00	2	0.7	230	PASS
127	198	10	80	35.1	48803	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	138957	PASS
199	198	5	9	6.9	9525	PASS
275	198	10	60	31.4	43675	PASS
365	198	1	100	4.0	5607	PASS
441	198	0.01	100	11.0	15314	PASS
442	442	50	100	100.0	94715	PASS
443	442	15	24	20.6	19557	PASS

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG011823\
Data File : BG056323.D
Acq On : 18 Jan 2023 10:58
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
DFTPP

Quant Time: Jan 19 04:49:13 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG122722.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jan 19 00:31:30 2023
Response via : Initial Calibration



TIC: BG056323.D\data.ms

(70) Pentachlorophenol (C)

17.306min (-0.001) 381012.95 ng




response 187770

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	66.20	66.90
264.00	64.30	64.60
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG011823\
Data File : BG056323.D
Acq On : 18 Jan 2023 10:58
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
DFTPP

Quant Time: Jan 19 04:49:13 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG122722.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jan 19 00:31:30 2023
Response via : Initial Calibration



TIC: BG056323.D\data.ms

(77) Benzidine

19.862min (+ 0.005) 17250.87 ng

response 905608

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	15.20	14.85
183.00	11.70	12.31
0.00	0.00	0.00

DDT Breakdown

Date	Instrument Name	DFTPP Data File
1/18/2023	BNA_G	<u>BG056323.D</u>
Compound Name	Response	Retention Time
DDT	448261	21.096
DDD	13220	20.643
DDE	1813	20.138
SUM(DDD+DDE)	SUM(DDT+DDD+DDE)	% Breakdown Of DDT
15033	463294	3.24

Instrument :
BNA_G
ClientSampleId :
DFTPP