

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG011922\
 Data File : BG052067.D
 Acq On : 19 Jan 2022 10:20
 Operator : CG/JU
 Sample : PB142080BL
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SBLK080

Quant Time: Jan 19 23:58:38 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG010622.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Jan 06 14:43:02 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.243	152	29188	20.000	ng/u1	-0.01
20) Naphthalene-d8	11.086	136	127294	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.887	164	95369	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.636	188	220262	20.000	ng/u1	# 0.00
79) Chrysene-d12	21.960	240	203394	20.000	ng/u1	# 0.00
88) Perylene-d12	25.449	264	200777	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.573	96	3723	4.558	ng/uL	0.00
4) Pyridine-d5	3.996	84	52734	20.768	ng/u1	0.00
7) Phenol-d5	7.385	99	70716	24.754	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.550	67	41224	22.732	ng/u1	0.00
11) 2-Chlorophenol-d4	7.773	132	48749	25.774	ng/u1	0.00
15) 4-Methylphenol-d8	8.942	113	57142	25.731	ng/u1	0.00
21) Nitrobenzene-d5	9.412	128	26259	26.106	ng/u1	0.00
24) 2-Nitrophenol-d4	10.146	143	31055	27.916	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.698	165	59435	27.334	ng/u1	0.00
31) 4-Chloroaniline-d4	11.209	131	76794	25.360	ng/u1	0.00
46) Dimethylphthalate-d6	14.276	166	198298	25.155	ng/u1	0.00
49) Acenaphthylene-d8	14.581	160	248491	26.096	ng/u1	0.00
54) 4-Nitrophenol-d4	15.069	143	26192	19.873	ng/u1	0.00
60) Fluorene-d10	15.880	176	174305	25.361	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.985	200	25552	18.389	ng/u1	0.00
73) Anthracene-d10	17.736	188	286747	27.443	ng/u1	0.00
81) Pyrene-d10	20.015	212	325401	27.798	ng/u1	0.00
92) Benzo(a)pyrene-d12	25.208	264	287884	27.236	ng/u1	0.01

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG011922\
Data File : BG052067.D
Acq On : 19 Jan 2022 10:20
Operator : CG/JU
Sample : PB142080BL
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
SBLK080

Quant Time: Jan 19 23:58:38 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG010622.M
Quant Title : SVOA CALIBRATION
QLast Update : Thu Jan 06 14:43:02 2022
Response via : Initial Calibration

