


Data Path : Z:\SVOASRV\HPCHEM1\BNA_G\DATA\BG012220\
 Data File : BG044154.D
 Acq On : 22 Jan 2020 11:48
 Operator : JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\8270-BG123019.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Tue Dec 31 13:32:23 2019



AutoFind: Scans 2496, 2497, 2498; Background Corrected with Scan 2490

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	35.9	27325	PASS
68	69	0.00	2	0.6	167	PASS
69	198	0.00	100	34.7	26360	PASS
70	69	0.00	2	0.4	117	PASS
127	198	10	80	45.1	34325	PASS
197	198	0.00	2	0.3	192	PASS
198	198	100	100	100.0	76037	PASS
199	198	5	9	6.8	5174	PASS
275	198	10	60	25.6	19452	PASS
365	198	1	100	3.3	2492	PASS
441	198	0.01	100	11.5	8732	PASS
442	198	50	100	70.2	53413	PASS
443	442	15	24	20.3	10838	PASS