

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG012622\
 Data File : BG052202.D
 Acq On : 28 Jan 2022 2:08
 Operator : CG/JU
 Sample : N1302-02
 Misc :
 ALS Vial : 61 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 C0F32

Quant Time: Jan 28 02:55:19 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG010622.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Jan 26 23:58:52 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.239	152	24091	20.000	ng/u1	0.00
20) Naphthalene-d8	11.077	136	101425	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.883	164	73571	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.632	188	179194	20.000	ng/u1	# 0.00
79) Chrysene-d12	21.950	240	185623	20.000	ng/u1	# 0.00
88) Perylene-d12	25.428	264	186444	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.563	96	2593	3.846	ng/uL	0.00
4) Pyridine-d5	3.998	84	20875	9.960	ng/u1	0.00
7) Phenol-d5	7.382	99	13372	5.671	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.540	67	43259	28.901	ng/u1	0.00
11) 2-Chlorophenol-d4	7.769	132	38596	24.724	ng/u1	0.00
15) 4-Methylphenol-d8	8.938	113	25316	13.812	ng/u1	0.00
21) Nitrobenzene-d5	9.408	128	25907	32.326	ng/u1	0.00
24) 2-Nitrophenol-d4	10.143	143	28708	32.388	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.695	165	51445	29.694	ng/u1	0.00
31) 4-Chloroaniline-d4	11.200	131	64984	26.934	ng/u1	0.00
46) Dimethylphthalate-d6	14.272	166	200692	33.002	ng/u1	0.00
49) Acenaphthylene-d8	14.578	160	248464	33.824	ng/u1	0.00
54) 4-Nitrophenol-d4	15.065	143	4805	4.726	ng/u1	0.00
60) Fluorene-d10	15.870	176	181204	34.176	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.982	200	27097	23.970	ng/u1	0.00
73) Anthracene-d10	17.726	188	330055	38.827	ng/u1	0.00
81) Pyrene-d10	20.006	212	402750	37.700	ng/u1	0.00
92) Benzo(a)pyrene-d12	25.187	264	394611	40.204	ng/u1	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG012622\
Data File : BG052202.D
Acq On : 28 Jan 2022 2:08
Operator : CG/JU
Sample : N1302-02
Misc :
ALS Vial : 61 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
C0F32

Quant Time: Jan 28 02:55:19 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG010622.M
Quant Title : SVOA CALIBRATION
QLast Update : Wed Jan 26 23:58:52 2022
Response via : Initial Calibration

