

Data Path : Z:\SVOASRV\HPCHEM1\BNA G\DATA\BG012820\
 Data File : BG044231.D
 Acq On : 28 Jan 2020 15:28
 Operator : CG/JU
 Sample : L1242-10RX
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 20200063-AMENDED-COMP-1RX

Manual Integrations
 APPROVED

mohammad
 1/30/2020 10:17:43 AM

Quant Time: Jan 28 16:16:26 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG123019.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jan 28 15:57:24 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.93	152	90706	20.00	ng	0.00
21) Naphthalene-d8	10.74	136	412834	20.00	ng	0.00
39) Acenaphthene-d10	14.57	164	284856	20.00	ng	0.00
64) Phenanthrene-d10	17.32	188	543484	20.00	ng	0.00
76) Chrysene-d12	21.56	240	472866	20.00	ng	0.00
87) Perylene-d12	24.67	264	549851	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.51	112	14396	2.91	ng	0.00
7) Phenol-d6	7.10	99	179679	26.02	ng	0.00
23) Nitrobenzene-d5	9.09	82	674418	87.39	ng	0.00
42) 2,4,6-Tribromophenol	16.07	330	8979	3.01	ng	0.01
45) 2-Fluorobiphenyl	13.19	172	1338856	85.16	ng	0.00
79) Terphenyl-d14	19.93	244	1797145	94.24	ng	0.00
Target Compounds						
10) Phenol	7.13	94	40006	5.623	ng	94
50) Dimethylphthalate	14.02	163	320020	15.822	ng	100
71) Phenanthrene	17.36	178	114046	4.375	ng	95
75) Fluoranthene	19.37	202	481256	16.143	ng	96
78) Pyrene	19.72	202	429575	13.917	ng	96
81) Benzo(a)anthracene	21.54	228	251536	8.579	ng	96
83) Chrysene	21.60	228	259869	9.327	ng	95
84) Bis(2-ethylhexyl)phthalate	21.46	149	48617	2.398	ng	95
86) Indeno(1,2,3-cd)pyrene	28.18	276	96612	2.552	ng	# 92
88) Benzo(b)fluoranthene	23.69	252	246878	7.595	ng	96
89) Benzo(k)fluoranthene	23.74	252	109641m	3.547	ng	
90) Benzo(a)pyrene	24.52	252	171576	5.592	ng	# 96
92) Benzo(a,h,i)perylene	29.28	276	87309	2.853	ng	# 93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA G\DATA\BG012820\
Data File : BG044231.D
Acq On : 28 Jan 2020 15:28
Operator : CG/JU
Sample : L1242-10RX
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
20200063-AMENDED-COMP-1RX

Manual Integrations
APPROVED

mohammad
1/30/2020 10:17:43 AM

Quant Time: Jan 28 16:16:26 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG123019.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jan 28 15:57:24 2020
Response via : Initial Calibration

