

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG020122\
 Data File : BG052223.D
 Acq On : 1 Feb 2022 10:52
 Operator : CG/JU
 Sample : PB142372BL
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SBLK372

Quant Time: Feb 01 13:14:00 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG013122.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Jan 31 18:07:10 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.247	152	31892	20.000	ng/u1	0.01
20) Naphthalene-d8	11.084	136	131209	20.000	ng/u1	# 0.00
38) Acenaphthene-d10	14.885	164	92571	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.628	188	219116	20.000	ng/u1	# 0.00
79) Chrysene-d12	21.946	240	223876	20.000	ng/u1	# 0.00
88) Perylene-d12	25.423	264	216777	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.565	96	4387	5.550	ng/uL	0.00
4) Pyridine-d5	3.994	84	66283	26.957	ng/u1	0.00
7) Phenol-d5	7.389	99	81456	28.080	ng/u1	0.01
9) Bis-(2-Chloroethyl)eth...	7.554	67	50448	28.625	ng/u1	0.01
11) 2-Chlorophenol-d4	7.771	132	58395	28.888	ng/u1	0.00
15) 4-Methylphenol-d8	8.952	113	61108	27.759	ng/u1	0.01
21) Nitrobenzene-d5	9.416	128	30524	28.571	ng/u1	0.00
24) 2-Nitrophenol-d4	10.144	143	33845	28.991	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.696	165	59978	26.354	ng/u1	0.00
31) 4-Chloroaniline-d4	11.207	131	80681	27.948	ng/u1	0.00
46) Dimethylphthalate-d6	14.274	166	221035	30.377	ng/u1	0.00
49) Acenaphthylene-d8	14.579	160	262775	28.973	ng/u1	0.00
54) 4-Nitrophenol-d4	15.067	143	26779	24.300	ng/u1	0.00
60) Fluorene-d10	15.872	176	184980	29.017	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.983	200	30749	22.731	ng/u1	0.00
73) Anthracene-d10	17.728	188	316514	30.640	ng/u1	0.00
81) Pyrene-d10	20.007	212	387159	28.855	ng/u1	0.00
92) Benzo(a)pyrene-d12	25.183	264	362073	31.439	ng/u1	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG020122\
Data File : BG052223.D
Acq On : 1 Feb 2022 10:52
Operator : CG/JU
Sample : PB142372BL
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
SBLK372

Quant Time: Feb 01 13:14:00 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG013122.M
Quant Title : SVOA CALIBRATION
QLast Update : Mon Jan 31 18:07:10 2022
Response via : Initial Calibration

