

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG020122\
 Data File : BG052224.D
 Acq On : 1 Feb 2022 11:30
 Operator : CG/JU
 Sample : PB142408BL
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SBLK408

Quant Time: Feb 01 13:16:01 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG013122.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Jan 31 18:07:10 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.235	152	24107	20.000	ng/u1	0.00
20) Naphthalene-d8	11.072	136	95398	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.879	164	68070	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.622	188	161915	20.000	ng/u1	# 0.00
79) Chrysene-d12	21.940	240	160714	20.000	ng/u1	# 0.00
88) Perylene-d12	25.411	264	158733	20.000	ng/u1	-0.01

System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.565	96	4233	7.084	ng/uL	0.00
4) Pyridine-d5	3.988	84	67506	36.321	ng/u1	0.00
7) Phenol-d5	7.377	99	80038	36.501	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.542	67	51685	38.798	ng/u1	0.00
11) 2-Chlorophenol-d4	7.759	132	58765	38.459	ng/u1	0.00
15) 4-Methylphenol-d8	8.940	113	60182	36.167	ng/u1	0.00
21) Nitrobenzene-d5	9.410	128	30666	39.479	ng/u1	0.00
24) 2-Nitrophenol-d4	10.138	143	33431	39.386	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.684	165	60862	36.781	ng/u1	0.00
31) 4-Chloroaniline-d4	11.195	131	82365	39.241	ng/u1	0.00
46) Dimethylphthalate-d6	14.268	166	219065	40.943	ng/u1	0.00
49) Acenaphthylene-d8	14.573	160	264110	39.602	ng/u1	0.00
54) 4-Nitrophenol-d4	15.061	143	27464	33.891	ng/u1	0.00
60) Fluorene-d10	15.866	176	185250	39.519	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.977	200	31695	31.707	ng/u1	0.00
73) Anthracene-d10	17.722	188	315106	41.280	ng/u1	0.00
81) Pyrene-d10	20.001	212	382797	39.742	ng/u1	0.00
92) Benzo(a)pyrene-d12	25.177	264	352893	41.847	ng/u1	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG020122\
Data File : BG052224.D
Acq On : 1 Feb 2022 11:30
Operator : CG/JU
Sample : PB142408BL
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
SBLK408

Quant Time: Feb 01 13:16:01 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG013122.M
Quant Title : SVOA CALIBRATION
QLast Update : Mon Jan 31 18:07:10 2022
Response via : Initial Calibration

