

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG020123\
 Data File : BG056506.D
 Acq On : 1 Feb 2023 19:05
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Feb 02 05:05:08 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG012723.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Feb 01 01:15:52 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	100	0.00
2	1,4-Dioxane	0.453	0.451	0.4	101	0.00
3	Pyridine	1.349	1.349	0.0	97	0.00
4	n-Nitrosodimethylamine	0.287	0.286	0.3	98	0.00
5 S	2-Fluorophenol	1.087	1.073	1.3	98	0.00
6	Aniline	2.109	2.060	2.3	96	0.00
7 S	Phenol-d6	1.611	1.586	1.6	96	0.00
8	2-Chlorophenol	1.194	1.116	6.5	93	0.00
9	Benzaldehyde	0.801	0.762	4.9	106	0.00
10 C	Phenol	1.637	1.588	3.0	96	0.00
11	bis(2-Chloroethyl)ether	1.125	1.112	1.2	97	0.00
12	1,3-Dichlorobenzene	1.375	1.343	2.3	99	0.00
13 C	1,4-Dichlorobenzene	1.393	1.356	2.7	99	0.00
14	1,2-Dichlorobenzene	1.351	1.295	4.1	95	0.00
15	Benzyl Alcohol	1.105	1.092	1.2	94	0.00
16	2,2'-oxybis(1-Chloropropane	0.238	0.209	12.2	85	0.00
17	2-Methylphenol	1.246	1.217	2.3	96	0.00
18	Hexachloroethane	0.421	0.425	-1.0	101	0.00
19 P	n-Nitroso-di-n-propylamine	0.928	0.923	0.5	96	0.00
20	3+4-Methylphenols	1.746	1.684	3.6	93	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	96	0.00
22	Acetophenone	0.531	0.532	-0.2	94	0.00
23 S	Nitrobenzene-d5	0.352	0.361	-2.6	98	0.00
24	Nitrobenzene	0.340	0.357	-5.0	99	0.00
25	Isophorone	0.727	0.725	0.3	93	0.00
26 C	2-Nitrophenol	0.206	0.207	-0.5	93	0.00
27	2,4-Dimethylphenol	0.345	0.345	0.0	93	0.00
28	bis(2-Chloroethoxy)methane	0.401	0.410	-2.2	97	0.00
29 C	2,4-Dichlorophenol	0.388	0.395	-1.8	94	0.00
30	1,2,4-Trichlorobenzene	0.431	0.445	-3.2	96	0.00
31	Naphthalene	1.056	1.059	-0.3	94	0.00
32	Benzoic acid	0.193	0.152	21.2	72	0.00
33	4-Chloroaniline	0.493	0.483	2.0	90	0.00
34 C	Hexachlorobutadiene	0.303	0.329	-8.6	103	0.00
35	Caprolactam	0.131	0.118	9.9	83	0.00
36 C	4-Chloro-3-methylphenol	0.375	0.375	0.0	92	0.00
37	2-Methylnaphthalene	0.870	0.882	-1.4	94	0.00
38	1-Methylnaphthalene	0.813	0.810	0.4	93	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	95	0.00
40	1,2,4,5-Tetrachlorobenzene	0.731	0.751	-2.7	98	0.00
41 P	Hexachlorocyclopentadiene	0.325	0.344	-5.8	96	0.00
42 S	2,4,6-Tribromophenol	0.266	0.254	4.5	90	0.00
43 C	2,4,6-Trichlorophenol	0.471	0.466	1.1	91	0.00
44	2,4,5-Trichlorophenol	0.551	0.536	2.7	92	0.00
45 S	2-Fluorobiphenyl	1.443	1.431	0.8	94	0.00
46	1,1'-Biphenyl	1.451	1.431	1.4	93	0.00
47	2-Chloronaphthalene	1.134	1.114	1.8	93	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG020123\
 Data File : BG056506.D
 Acq On : 1 Feb 2023 19:05
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Feb 02 05:05:08 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG012723.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Feb 01 01:15:52 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
48	2-Nitroaniline	0.234	0.231	1.3	91	0.00
49	Acenaphthylene	1.845	1.805	2.2	91	0.00
50	Dimethylphthalate	1.618	1.554	4.0	91	0.00
51	2,6-Dinitrotoluene	0.353	0.335	5.1	90	0.00
52 C	Acenaphthene	1.094	1.062	2.9	91	0.00
53	3-Nitroaniline	0.342	0.317	7.3	86	0.00
54 P	2,4-Dinitrophenol	0.221	0.193	12.7	77	0.00
55	Dibenzofuran	1.962	1.922	2.0	92	0.00
56 P	4-Nitrophenol	0.234	0.217	7.3	84	0.00
57	2,4-Dinitrotoluene	0.497	0.473	4.8	88	0.00
58	Fluorene	1.561	1.515	2.9	91	0.00
59	2,3,4,6-Tetrachlorophenol	0.490	0.474	3.3	87	0.00
60	Diethylphthalate	1.519	1.446	4.8	90	0.00
61	4-Chlorophenyl-phenylether	0.953	0.948	0.5	93	0.00
62	4-Nitroaniline	0.342	0.329	3.8	90	0.00
63	Azobenzene	1.045	1.062	-1.6	95	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	94	0.00
65	4,6-Dinitro-2-methylphenol	0.141	0.138	2.1	88	0.00
66 c	n-Nitrosodiphenylamine	0.566	0.568	-0.4	92	0.00
67	4-Bromophenyl-phenylether	0.256	0.260	-1.6	92	0.00
68	Hexachlorobenzene	0.251	0.248	1.2	91	0.00
69	Atrazine	0.218	0.221	-1.4	91	0.00
70 C	Pentachlorophenol	0.152	0.132	13.2	77	0.00
71	Phenanthrene	1.047	1.044	0.3	92	0.00
72	Anthracene	1.075	1.064	1.0	91	0.00
73	Carbazole	0.915	0.888	3.0	90	0.00
74	Di-n-butylphthalate	0.961	0.938	2.4	91	0.00
75 C	Fluoranthene	1.321	1.312	0.7	92	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	97	0.00
77	Benzydine	0.541	0.552	-2.0	97	0.00
78	Pyrene	1.422	1.377	3.2	93	0.00
79 S	Terphenyl-d14	1.139	1.101	3.3	93	0.00
80	Butylbenzylphthalate	0.434	0.416	4.1	91	0.00
81	Benzo(a)anthracene	1.398	1.381	1.2	94	0.00
82	3,3'-Dichlorobenzidine	0.504	0.491	2.6	93	0.00
83	Chrysene	1.314	1.303	0.8	95	0.00
84	Bis(2-ethylhexyl)phthalate	0.623	0.605	2.9	93	0.00
85 c	Di-n-octyl phthalate	1.037	1.025	1.2	94	0.00
86 I	Perylene-d12	1.000	1.000	0.0	97	0.00
87	Indeno(1,2,3-cd)pyrene	1.464	1.479	-1.0	96	0.00
88	Benzo(b)fluoranthene	1.241	1.257	-1.3	96	0.00
89	Benzo(k)fluoranthene	1.253	1.250	0.2	95	0.00
90 C	Benzo(a)pyrene	1.223	1.229	-0.5	96	0.00
91	Dibenzo(a,h)anthracene	1.229	1.238	-0.7	96	0.00
92	Benzo(g,h,i)perylene	1.186	1.200	-1.2	96	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG020123\
Data File : BG056506.D
Acq On : 1 Feb 2023 19:05
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTDCCC040

Quant Time: Feb 02 05:05:08 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG012723.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Feb 01 01:15:52 2023
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
----------	-------	------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0