

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG020223\
 Data File : BG056513.D
 Acq On : 2 Feb 2023 14:01
 Operator : CG/JU
 Sample : N4955-09
 Misc : MDL-MDL-WATER 8PPM
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 MDL-WATER-03-QT4-2022

Quant Time: Feb 03 05:21:40 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG012723.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Feb 01 01:15:52 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.277	152	34597	20.000	ng	0.00
21) Naphthalene-d8	11.109	136	129650	20.000	ng	0.00
39) Acenaphthene-d10	14.892	164	109200	20.000	ng	0.00
64) Phenanthrene-d10	17.630	188	289271	20.000	ng	0.00
76) Chrysene-d12	21.925	240	284145	20.000	ng	# 0.00
86) Perylene-d12	25.339	264	308278	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.791	112	199231	105.934	ng	0.00
7) Phenol-d6	7.419	99	288271	103.450	ng	0.00
23) Nitrobenzene-d5	9.452	82	192328	84.237	ng	0.00
42) 2,4,6-Tribromophenol	16.379	330	159165	109.637	ng	0.00
45) 2-Fluorobiphenyl	13.517	172	658717	83.632	ng	0.00
79) Terphenyl-d14	20.210	244	1271966	78.623	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.600	88	6448	8.234	ng	99
3) Pyridine	4.023	79	16167	6.926	ng	97
4) n-Nitrosodimethylamine	3.929	42	3948	7.953	ng	# 80
6) Aniline	7.589	93	26990	7.398	ng	98
8) 2-Chlorophenol	7.836	128	14588	7.061	ng	93
9) Benzaldehyde	7.401	77	14033	8.560	ng	93
10) Phenol	7.442	94	20772	7.334	ng	94
11) bis(2-Chloroethyl)ether	7.677	93	14774	7.589	ng	92
12) 1,3-Dichlorobenzene	8.165	146	18952	7.967	ng	95
13) 1,4-Dichlorobenzene	8.312	146	20318	8.434	ng	94
14) 1,2-Dichlorobenzene	8.635	146	19121	8.181	ng	93
15) Benzyl Alcohol	8.512	79	13562	7.094	ng	94
16) 2,2'-oxybis(1-Chloropr...	8.788	45	3083	7.474	ng	86
17) 2-Methylphenol	8.711	107	15479	7.184	ng	94
18) Hexachloroethane	9.369	117	6193	8.500	ng	86
19) n-Nitroso-di-n-propyla...	9.076	70	11930	7.429	ng	# 90
20) 3+4-Methylphenols	9.040	107	20097	6.656	ng	93
22) Acetophenone	9.099	105	29913	8.684	ng	# 93
24) Nitrobenzene	9.493	77	17563	7.975	ng	94
25) Isophorone	10.010	82	37449	7.951	ng	97
26) 2-Nitrophenol	10.210	139	10443	7.815	ng	93
27) 2,4-Dimethylphenol	10.257	122	15945	7.130	ng	92
28) bis(2-Chloroethoxy)met...	10.492	93	22125	8.503	ng	# 95
29) 2,4-Dichlorophenol	10.756	162	19278	7.659	ng	98
30) 1,2,4-Trichlorobenzene	10.968	180	23523	8.410	ng	92
31) Naphthalene	11.161	128	54819	8.011	ng	98
32) Benzoic acid	10.380	122	6818	5.450	ng	96
33) 4-Chloroaniline	11.261	127	25106	7.860	ng	# 95
34) Hexachlorobutadiene	11.432	225	17519	8.923	ng	96
35) Caprolactam	12.002	113	6690	7.881	ng	88
36) 4-Chloro-3-methylphenol	12.366	107	18463	7.595	ng	99
37) 2-Methylnaphthalene	12.742	142	45365	8.041	ng	99
38) 1-Methylnaphthalene	12.959	142	42963	8.154	ng	# 91
40) 1,2,4,5-Tetrachloroben...	13.106	216	33639	8.432	ng	99
41) Hexachlorocyclopentadiene	13.083	237	9260	10.401	ng	99
43) 2,4,6-Trichlorophenol	13.341	196	19162	7.453	ng	99

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG020223\
 Data File : BG056513.D
 Acq On : 2 Feb 2023 14:01
 Operator : CG/JU
 Sample : N4955-09
 Misc : MDL-MDL-WATER 8PPM
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 MDL-WATER-03-QT4-2022


Quant Time: Feb 03 05:21:40 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG012723.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Feb 01 01:15:52 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
44) 2,4,5-Trichlorophenol	13.418	196	21535	7.154	ng	97
46) 1,1'-Biphenyl	13.729	154	67289	8.496	ng	98
47) 2-Chloronaphthalene	13.782	162	52145	8.423	ng	97
48) 2-Nitroaniline	13.982	65	9970	7.790	ng	84
49) Acenaphthylene	14.622	152	81398	8.081	ng	99
50) Dimethylphthalate	14.328	163	71227	8.062	ng	96
51) 2,6-Dinitrotoluene	14.458	165	14942	7.756	ng	93
52) Acenaphthene	14.957	154	48461	8.113	ng	95
53) 3-Nitroaniline	14.792	138	14612	7.823	ng	98
54) 2,4-Dinitrophenol	15.010	184	10991	9.100	ng	96
55) Dibenzofuran	15.286	168	88401	8.252	ng	98
56) 4-Nitrophenol	15.104	139	8624	6.757	ng	94
57) 2,4-Dinitrotoluene	15.251	165	21989	8.103	ng	90
58) Fluorene	15.932	166	70293	8.247	ng	97
59) 2,3,4,6-Tetrachlorophenol	15.515	232	19427	7.259	ng	# 97
60) Diethylphthalate	15.680	149	65604	7.908	ng	99
61) 4-Chlorophenyl-phenyle...	15.915	204	42701	8.210	ng	96
62) 4-Nitroaniline	15.950	138	13817	7.393	ng	91
63) Azobenzene	16.208	77	46642	8.171	ng	95
65) 4,6-Dinitro-2-methylph...	16.015	198	11971	5.886	ng	94
66) n-Nitrosodiphenylamine	16.132	169	63273	7.726	ng	98
67) 4-Bromophenyl-phenylether	16.808	248	29272	7.905	ng	96
68) Hexachlorobenzene	16.937	284	30078	8.293	ng	95
69) Atrazine	17.060	200	27970	8.852	ng	97
70) Pentachlorophenol	17.290	266	14805	6.746	ng	97
71) Phenanthrene	17.677	178	121258	8.009	ng	99
72) Anthracene	17.765	178	123310	7.934	ng	99
73) Carbazole	18.030	167	105470	7.968	ng	96
74) Di-n-butylphthalate	18.559	149	105642	7.601	ng	99
75) Fluoranthene	19.669	202	156621	8.196	ng	100
77) Benzidine	19.834	184	80658	10.490	ng	99
78) Pyrene	20.027	202	158170	7.827	ng	98
80) Butylbenzylphthalate	20.885	149	44587	7.224	ng	97
81) Benzo(a)anthracene	21.902	228	156978	7.902	ng	99
82) 3,3'-Dichlorobenzidine	21.802	252	62893	8.792	ng	97
83) Chrysene	21.972	228	146594	7.855	ng	97
84) Bis(2-ethylhexyl)phtha...	21.761	149	63987	7.231	ng	99
85) Di-n-octyl phthalate	23.030	149	107829	7.317	ng	99
87) Indeno(1,2,3-cd)pyrene	29.275	276	172508	7.644	ng	100
88) Benzo(b)fluoranthene	24.246	252	159854	8.354	ng	97
89) Benzo(k)fluoranthene	24.317	252	157232	8.139	ng	99
90) Benzo(a)pyrene	25.180	252	141135	7.488	ng	96
91) Dibenzo(a,h)anthracene	29.322	278	153111	8.083	ng	99
92) Benzo(g,h,i)perylene	30.503	276	148629	8.132	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Feb 03 05:21:40 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG012723.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Feb 01 01:15:52 2023
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 8.277 min Scan# 84
Delta R.T. 0.002 min
Lab File: BG056513.D
Acq: 2 Feb 2023 14:01




Instrument :
BNA_G
ClientSampleId :
MDL-WATER-03-QT4-2022


Tgt Ion:152 Resp: 34597
Ion Ratio Lower Upper
152 100
150 143.1 120.2 180.4
115 50.2 36.3 54.5



#2
1,4-Dioxane
Concen: 8.234 ng
RT: 3.600 min Scan# 53
Delta R.T. 0.001 min
Lab File: BG056513.D
Acq: 2 Feb 2023 14:01

Tgt Ion: 88 Resp: 6448
Ion Ratio Lower Upper
88 100
58 28.3 23.0 34.4
43 9.5 8.0 12.0





#3

Pyridine

Concen: 6.926 ng

RT: 4.023 min Scan# 110

Delta R.T. 0.001 min

Lab File: BG056513.D

Acq: 2 Feb 2023 14:01

Instrument :

BNA_G

ClientSampleId :

MDL-WATER-03-QT4-2022





Tgt Ion: 79 Resp: 16167

Ion Ratio Lower Upper

79 100

52 18.5 16.1 24.1

51 13.0 10.7 16.1



#4

n-Nitrosodimethylamine

Concen: 7.953 ng

RT: 3.929 min Scan# 109

Delta R.T. 0.001 min

Lab File: BG056513.D

Acq: 2 Feb 2023 14:01




Tgt Ion: 42 Resp: 3948


Ion Ratio Lower Upper

42 100

74 235.4 217.9 326.9

44 7.2 3.0 4.6#









#5
2-Fluorophenol
Concen: 105.934 ng
RT: 5.791 min Scan# 41
Delta R.T. -0.005 min
Lab File: BG056513.D
Acq: 2 Feb 2023 14:01




Instrument :
BNA_G
ClientSampleId :
MDL-WATER-03-QT4-2022


Tgt Ion:112 Resp: 199231
Ion Ratio Lower Upper
112 100
64 40.7 31.9 47.9
63 23.1 17.4 26.0



#6
Aniline
Concen: 7.398 ng
RT: 7.589 min Scan# 732
Delta R.T. 0.001 min
Lab File: BG056513.D
Acq: 2 Feb 2023 14:01

Tgt Ion: 93 Resp: 26990
Ion Ratio Lower Upper
93 100
66 28.2 22.2 33.2
65 12.3 11.3 16.9









#7
Phenol-d6
Concen: 103.450 ng
RT: 7.419 min Scan# 703
Delta R.T. 0.001 min
Lab File: BG056513.D
Acq: 2 Feb 2023 14:01




Instrument :
BNA_G
ClientSampleId :
MDL-WATER-03-QT4-2022


Tgt Ion: 99 Resp: 288271
Ion Ratio Lower Upper
99 100
42 7.6 6.6 10.0
71 29.9 23.4 35.0



#8
2-Chlorophenol
Concen: 7.061 ng
RT: 7.836 min Scan# 774
Delta R.T. 0.001 min
Lab File: BG056513.D
Acq: 2 Feb 2023 14:01

Tgt Ion: 128 Resp: 14588
Ion Ratio Lower Upper
128 100
130 39.2 12.8 52.8
64 30.3 11.3 51.3









#9
Benzaldehyde
Concen: 8.560 ng
RT: 7.401 min Scan# 701
Delta R.T. 0.001 min
Lab File: BG056513.D
Acq: 2 Feb 2023 14:01




Instrument :
BNA_G
ClientSampleId :
MDL-WATER-03-QT4-2022


Tgt Ion: 77 Resp: 14033
Ion Ratio Lower Upper
77 100
105 116.0 101.8 141.8
106 102.4 93.0 133.0



#10
Phenol
Concen: 7.334 ng
RT: 7.442 min Scan# 707
Delta R.T. -0.005 min
Lab File: BG056513.D
Acq: 2 Feb 2023 14:01

Tgt Ion: 94 Resp: 20772
Ion Ratio Lower Upper
94 100
65 19.6 2.3 42.3
66 35.0 11.2 51.2





#11

bis(2-Chloroethyl)ether

Concen: 7.589 ng

RT: 7.677 min Scan# 74

Delta R.T. -0.005 min

Lab File: BG056513.D

Acq: 2 Feb 2023 14:01

Instrument :

BNA_G

ClientSampleId :

MDL-WATER-03-QT4-2022





Tgt Ion: 93 Resp: 14774

Ion Ratio Lower Upper

93 100

63 45.5 32.4 72.4

95 29.2 11.8 51.8



#12

1,3-Dichlorobenzene

Concen: 7.967 ng

RT: 8.165 min Scan# 830

Delta R.T. 0.001 min

Lab File: BG056513.D

Acq: 2 Feb 2023 14:01




Tgt Ion:146 Resp: 18952


Ion Ratio Lower Upper

146 100

148 66.9 50.6 75.8

75 26.5 19.6 29.4









#13
1,4-Dichlorobenzene
Concen: 8.434 ng
RT: 8.312 min Scan# 81
Delta R.T. 0.001 min
Lab File: BG056513.D
Acq: 2 Feb 2023 14:01




Instrument :
BNA_G
ClientSampleId :
MDL-WATER-03-QT4-2022


Tgt Ion:146 Resp: 20318
Ion Ratio Lower Upper
146 100
148 58.6 51.8 77.6
111 36.7 28.6 42.8



#14
1,2-Dichlorobenzene
Concen: 8.181 ng
RT: 8.635 min Scan# 910
Delta R.T. 0.001 min
Lab File: BG056513.D
Acq: 2 Feb 2023 14:01

Tgt Ion:146 Resp: 19121
Ion Ratio Lower Upper
146 100
148 65.9 50.1 75.1
111 45.0 30.9 46.3









#15
Benzyl Alcohol
Concen: 7.094 ng
RT: 8.512 min Scan# 890
Delta R.T. 0.001 min
Lab File: BG056513.D
Acq: 2 Feb 2023 14:01




Instrument :
BNA_G
ClientSampleId :
MDL-WATER-03-QT4-2022


Tgt Ion: 79 Resp: 13562
Ion Ratio Lower Upper
79 100
108 81.6 70.1 105.1
77 59.3 50.3 75.5



#16
2,2'-oxybis(1-Chloropropane)
Concen: 7.474 ng
RT: 8.788 min Scan# 936
Delta R.T. -0.005 min
Lab File: BG056513.D
Acq: 2 Feb 2023 14:01

Tgt Ion: 45 Resp: 3083
Ion Ratio Lower Upper
45 100
77 68.1 62.1 102.1
79 71.2 42.7 82.7









#17
2-Methylphenol
Concen: 7.184 ng
RT: 8.711 min Scan# 91
Delta R.T. -0.005 min
Lab File: BG056513.D
Acq: 2 Feb 2023 14:01




Instrument :
BNA_G
ClientSampleId :
MDL-WATER-03-QT4-2022


Tgt Ion:107 Resp: 15479
Ion Ratio Lower Upper
107 100
108 110.7 83.9 125.9
77 37.1 31.6 47.4
79 45.8 32.4 48.6



#18
Hexachloroethane
Concen: 8.500 ng
RT: 9.369 min Scan# 1035
Delta R.T. -0.005 min
Lab File: BG056513.D
Acq: 2 Feb 2023 14:01

Tgt Ion:117 Resp: 6193
Ion Ratio Lower Upper
117 100
119 92.5 79.4 119.0
201 135.4 128.5 192.7









#19
n-Nitroso-di-n-propylamine
Concen: 7.429 ng
RT: 9.076 min Scan# 91
Delta R.T. -0.005 min
Lab File: BG056513.D
Acq: 2 Feb 2023 14:01




Instrument :
BNA_G
ClientSampleId :
MDL-WATER-03-QT4-2022


Tgt Ion: 70 Resp: 11930
Ion Ratio Lower Upper
70 100
42 19.7 20.2 30.4#
101 8.8 9.1 13.7#
130 21.6 21.1 31.7



#20
3+4-Methylphenols
Concen: 6.656 ng
RT: 9.040 min Scan# 979
Delta R.T. -0.005 min
Lab File: BG056513.D
Acq: 2 Feb 2023 14:01

Tgt Ion: 107 Resp: 20097
Ion Ratio Lower Upper
107 100
108 95.3 66.7 106.7
77 27.3 10.6 50.6
79 24.7 4.9 44.9









#21
Naphthalene-d8
Concen: 20.000 ng
RT: 11.109 min Scan# 11
Delta R.T. 0.001 min
Lab File: BG056513.D
Acq: 2 Feb 2023 14:01




Instrument :
BNA_G
ClientSampleId :
MDL-WATER-03-QT4-2022


Tgt Ion:136 Resp: 129650
Ion Ratio Lower Upper
136 100
137 11.6 9.4 14.0
54 2.8 2.2 3.4
68 4.8 3.6 5.4



#22
Acetophenone
Concen: 8.684 ng
RT: 9.099 min Scan# 989
Delta R.T. -0.005 min
Lab File: BG056513.D
Acq: 2 Feb 2023 14:01

Tgt Ion:105 Resp: 29913
Ion Ratio Lower Upper
105 100
71 0.9 0.6 0.8#
51 10.4 9.2 13.8
120 17.1 17.4 26.0#









#23
Nitrobenzene-d5
Concen: 84.237 ng
RT: 9.452 min Scan# 1049
Delta R.T. -0.005 min
Lab File: BG056513.D
Acq: 2 Feb 2023 14:01




Instrument :
BNA_G
ClientSampleId :
MDL-WATER-03-QT4-2022


Tgt Ion: 82 Resp: 192328
Ion Ratio Lower Upper
82 100
128 49.3 39.4 59.2
54 26.8 21.4 32.2



#24
Nitrobenzene
Concen: 7.975 ng
RT: 9.493 min Scan# 1056
Delta R.T. -0.005 min
Lab File: BG056513.D
Acq: 2 Feb 2023 14:01

Tgt Ion: 77 Resp: 17563
Ion Ratio Lower Upper
77 100
123 49.3 44.0 66.0
65 12.0 9.7 14.5









#25
Isophorone
Concen: 7.951 ng
RT: 10.010 min Scan# 1146
Delta R.T. -0.005 min
Lab File: BG056513.D
Acq: 2 Feb 2023 14:01




Instrument :
BNA_G
ClientSampleId :
MDL-WATER-03-QT4-2022


Tgt Ion: 82 Resp: 37449
Ion Ratio Lower Upper
82 100
95 8.4 6.3 9.5
138 22.4 16.6 25.0



#26
2-Nitrophenol
Concen: 7.815 ng
RT: 10.210 min Scan# 1178
Delta R.T. -0.005 min
Lab File: BG056513.D
Acq: 2 Feb 2023 14:01

Tgt Ion:139 Resp: 10443
Ion Ratio Lower Upper
139 100
109 27.0 26.9 40.3
65 38.1 29.0 43.4









#27
2,4-Dimethylphenol
Concen: 7.130 ng
RT: 10.257 min Scan# 1186
Delta R.T. 0.001 min
Lab File: BG056513.D
Acq: 2 Feb 2023 14:01




Instrument :
BNA_G
ClientSampleId :
MDL-WATER-03-QT4-2022


Tgt Ion:122 Resp: 15945
Ion Ratio Lower Upper
122 100
107 108.9 92.9 139.3
121 64.3 45.0 67.4



#28
bis(2-Chloroethoxy)methane
Concen: 8.503 ng
RT: 10.492 min Scan# 1226
Delta R.T. 0.001 min
Lab File: BG056513.D
Acq: 2 Feb 2023 14:01

Tgt Ion: 93 Resp: 22125
Ion Ratio Lower Upper
93 100
95 32.8 27.9 41.9
123 13.5 8.6 12.8#





#29

2,4-Dichlorophenol

Concen: 7.659 ng

RT: 10.756 min Scan# 1

Delta R.T. 0.001 min

Lab File: BG056513.D

Acq: 2 Feb 2023 14:01

Instrument :

BNA_G

ClientSampleId :

MDL-WATER-03-QT4-2022





Tgt Ion:162 Resp: 19278

Ion Ratio Lower Upper

162 100

164 60.7 42.8 82.8

98 25.6 6.5 46.5



#30

1,2,4-Trichlorobenzene

Concen: 8.410 ng

RT: 10.968 min Scan# 1307

Delta R.T. 0.001 min

Lab File: BG056513.D

Acq: 2 Feb 2023 14:01




Tgt Ion:180 Resp: 23523


Ion Ratio Lower Upper

180 100

182 86.2 75.9 113.9

145 22.7 19.4 29.0









#31
Naphthalene
Concen: 8.011 ng
RT: 11.161 min Scan# 11
Delta R.T. 0.001 min
Lab File: BG056513.D
Acq: 2 Feb 2023 14:01




Instrument :
BNA_G
ClientSampleId :
MDL-WATER-03-QT4-2022


Tgt Ion:128 Resp: 54819
Ion Ratio Lower Upper
128 100
129 11.5 8.7 13.1
127 12.3 10.4 15.6



#32
Benzoic acid
Concen: 5.450 ng
RT: 10.380 min Scan# 1207
Delta R.T. -0.022 min
Lab File: BG056513.D
Acq: 2 Feb 2023 14:01

Tgt Ion:122 Resp: 6818
Ion Ratio Lower Upper
122 100
105 117.7 96.5 136.5
77 69.4 57.3 97.3









#33
4-Chloroaniline
Concen: 7.860 ng
RT: 11.261 min Scan# 11
Delta R.T. -0.005 min
Lab File: BG056513.D
Acq: 2 Feb 2023 14:01




Instrument :
BNA_G
ClientSampleId :
MDL-WATER-03-QT4-2022


Tgt Ion:	127	Resp:	25106
Ion	Ratio	Lower	Upper
127	100		
129	29.8	25.0	37.6
65	18.4	18.5	27.7#
92	18.4	15.2	22.8



#34
Hexachlorobutadiene
Concen: 8.923 ng
RT: 11.432 min Scan# 1386
Delta R.T. -0.005 min
Lab File: BG056513.D
Acq: 2 Feb 2023 14:01

Tgt Ion:	225	Resp:	17519
Ion	Ratio	Lower	Upper
225	100		
223	62.0	50.6	75.8
227	62.4	53.7	80.5









#35
Caprolactam
Concen: 7.881 ng
RT: 12.002 min Scan# 14
Delta R.T. -0.028 min
Lab File: BG056513.D
Acq: 2 Feb 2023 14:01




Instrument :
BNA_G
ClientSampleId :
MDL-WATER-03-QT4-2022


Tgt Ion:113 Resp: 6690
Ion Ratio Lower Upper
113 100
55 55.9 33.7 73.7
56 61.6 27.1 67.1



#36
4-Chloro-3-methylphenol
Concen: 7.595 ng
RT: 12.366 min Scan# 1545
Delta R.T. 0.001 min
Lab File: BG056513.D
Acq: 2 Feb 2023 14:01

Tgt Ion:107 Resp: 18463
Ion Ratio Lower Upper
107 100
144 27.6 22.3 33.5
142 85.1 68.9 103.3









#37
2-Methylnaphthalene
Concen: 8.041 ng
RT: 12.742 min Scan# 1609
Delta R.T. -0.005 min
Lab File: BG056513.D
Acq: 2 Feb 2023 14:01




Instrument :
BNA_G
ClientSampleId :
MDL-WATER-03-QT4-2022


Tgt Ion:142 Resp: 45365
Ion Ratio Lower Upper
142 100
141 90.2 71.4 107.0
115 35.8 28.4 42.6



#38
1-Methylnaphthalene
Concen: 8.154 ng
RT: 12.959 min Scan# 1646
Delta R.T. -0.005 min
Lab File: BG056513.D
Acq: 2 Feb 2023 14:01

Tgt Ion:142 Resp: 42963
Ion Ratio Lower Upper
142 100
141 87.9 77.7 116.5
116 2.9 3.4 5.2#









#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 14.892 min Scan# 1975
Delta R.T. -0.005 min
Lab File: BG056513.D
Acq: 2 Feb 2023 14:01




Instrument :
BNA_G
ClientSampleId :
MDL-WATER-03-QT4-2022


Tgt Ion:164 Resp: 109200
Ion Ratio Lower Upper
164 100
162 105.5 79.6 119.4
160 39.8 34.9 52.3



#40
1,2,4,5-Tetrachlorobenzene
Concen: 8.432 ng
RT: 13.106 min Scan# 1671
Delta R.T. 0.001 min
Lab File: BG056513.D
Acq: 2 Feb 2023 14:01

Tgt Ion:216 Resp: 33639
Ion Ratio Lower Upper
216 100
214 78.5 63.0 94.6
179 17.3 14.1 21.1
108 12.4 9.0 13.6









#41
Hexachlorocyclopentadiene
Concen: 10.401 ng
RT: 13.083 min Scan# 1
Delta R.T. 0.001 min
Lab File: BG056513.D
Acq: 2 Feb 2023 14:01




Instrument :
BNA_G
ClientSampleId :
MDL-WATER-03-QT4-2022


Tgt Ion:237 Resp: 9260
Ion Ratio Lower Upper
237 100
235 65.3 45.6 85.6
272 10.3 0.0 32.3



#42
2,4,6-Tribromophenol
Concen: 109.637 ng
RT: 16.379 min Scan# 2228
Delta R.T. 0.001 min
Lab File: BG056513.D
Acq: 2 Feb 2023 14:01

Tgt Ion:330 Resp: 159165
Ion Ratio Lower Upper
330 100
332 98.2 79.0 118.6
141 31.2 24.2 36.4









#43
2,4,6-Trichlorophenol
Concen: 7.453 ng
RT: 13.341 min Scan# 1712
Delta R.T. 0.001 min
Lab File: BG056513.D
Acq: 2 Feb 2023 14:01




Instrument :
BNA_G
ClientSampleId :
MDL-WATER-03-QT4-2022


Tgt Ion:196 Resp: 19162
Ion Ratio Lower Upper
196 100
198 94.8 75.4 113.0
200 32.0 25.0 37.6



#44
2,4,5-Trichlorophenol
Concen: 7.154 ng
RT: 13.418 min Scan# 1724
Delta R.T. 0.001 min
Lab File: BG056513.D
Acq: 2 Feb 2023 14:01

Tgt Ion:196 Resp: 21535
Ion Ratio Lower Upper
196 100
198 93.4 76.9 115.3
97 31.2 25.8 38.6
132 15.0 13.3 19.9









#45
2-Fluorobiphenyl
Concen: 83.632 ng
RT: 13.517 min Scan# 1741
Delta R.T. -0.005 min
Lab File: BG056513.D
Acq: 2 Feb 2023 14:01




Instrument :
BNA_G
ClientSampleId :
MDL-WATER-03-QT4-2022


Tgt Ion:172 Resp: 658717
Ion Ratio Lower Upper
172 100
171 37.5 29.5 44.3
170 23.6 19.2 28.8



#46
1,1'-Biphenyl
Concen: 8.496 ng
RT: 13.729 min Scan# 1777
Delta R.T. -0.005 min
Lab File: BG056513.D
Acq: 2 Feb 2023 14:01

Tgt Ion:154 Resp: 67289
Ion Ratio Lower Upper
154 100
153 40.0 21.5 61.5
76 8.4 0.0 28.8









#47
2-Chloronaphthalene
Concen: 8.423 ng
RT: 13.782 min Scan# 1787
Delta R.T. -0.005 min
Lab File: BG056513.D
Acq: 2 Feb 2023 14:01




Instrument : BNA_G
ClientSampleId : MDL-WATER-03-QT4-2022


Tgt Ion:162 Resp: 52145
Ion Ratio Lower Upper
162 100
127 32.9 25.2 37.8
164 34.7 26.2 39.2



#48
2-Nitroaniline
Concen: 7.790 ng
RT: 13.982 min Scan# 1820
Delta R.T. 0.001 min
Lab File: BG056513.D
Acq: 2 Feb 2023 14:01

Tgt Ion: 65 Resp: 9970
Ion Ratio Lower Upper
65 100
92 85.8 75.8 113.6
138 139.1 133.2 199.8









#49
Acenaphthylene
Concen: 8.081 ng
RT: 14.622 min Scan# 1930
Delta R.T. -0.005 min
Lab File: BG056513.D
Acq: 2 Feb 2023 14:01




Instrument :
BNA_G
ClientSampleId :
MDL-WATER-03-QT4-2022


Tgt Ion:152 Resp: 81398
Ion Ratio Lower Upper
152 100
151 19.6 15.9 23.9
153 14.0 10.5 15.7



#50
Dimethylphthalate
Concen: 8.062 ng
RT: 14.328 min Scan# 1879
Delta R.T. -0.011 min
Lab File: BG056513.D
Acq: 2 Feb 2023 14:01

Tgt Ion:163 Resp: 71227
Ion Ratio Lower Upper
163 100
194 3.7 3.5 5.3
164 11.9 8.1 12.1









#51
2,6-Dinitrotoluene
Concen: 7.756 ng
RT: 14.458 min Scan# 1901
Delta R.T. -0.011 min
Lab File: BG056513.D
Acq: 2 Feb 2023 14:01




Instrument :
BNA_G
ClientSampleId :
MDL-WATER-03-QT4-2022


Tgt Ion:165 Resp: 14942
Ion Ratio Lower Upper
165 100
63 35.6 26.9 40.3
89 51.3 35.6 53.4



#52
Acenaphthene
Concen: 8.113 ng
RT: 14.957 min Scan# 1986
Delta R.T. -0.005 min
Lab File: BG056513.D
Acq: 2 Feb 2023 14:01

Tgt Ion:154 Resp: 48461
Ion Ratio Lower Upper
154 100
153 109.5 93.4 140.0
152 53.5 43.4 65.2





#53

3-Nitroaniline

Concen: 7.823 ng

RT: 14.792 min Scan# 1958

Delta R.T. -0.005 min

Lab File: BG056513.D


Acq: 2 Feb 2023 14:01

Instrument :

BNA_G

ClientSampleId :

MDL-WATER-03-QT4-2022






Tgt Ion:138 Resp: 14612

Ion Ratio Lower Upper

138 100

108 8.3 8.0 12.0

92 103.5 84.1 126.1



#54

2,4-Dinitrophenol

Concen: 9.100 ng

RT: 15.010 min Scan# 1995

Delta R.T. 0.001 min

Lab File: BG056513.D

Acq: 2 Feb 2023 14:01




Tgt Ion:184 Resp: 10991


Ion Ratio Lower Upper

184 100

63 30.5 25.1 37.7

154 51.9 44.8 67.2









#55
Dibenzenofuran
Concen: 8.252 ng
RT: 15.286 min Scan# 2042
Delta R.T. -0.005 min
Lab File: BG056513.D
Acq: 2 Feb 2023 14:01




Instrument :
BNA_G
ClientSampleId :
MDL-WATER-03-QT4-2022


Tgt Ion:168 Resp: 88401
Ion Ratio Lower Upper
168 100
139 31.2 26.2 39.2
169 13.4 11.0 16.6



#56
4-Nitrophenol
Concen: 6.757 ng
RT: 15.104 min Scan# 2011
Delta R.T. 0.007 min
Lab File: BG056513.D
Acq: 2 Feb 2023 14:01

Tgt Ion:139 Resp: 8624
Ion Ratio Lower Upper
139 100
109 85.9 67.5 107.5
65 61.9 50.7 90.7





#57

2,4-Dinitrotoluene

Concen: 8.103 ng

RT: 15.251 min Scan# 2037

Delta R.T. 0.001 min

Lab File: BG056513.D

Acq: 2 Feb 2023 14:01

Instrument :

BNA_G

ClientSampleId :

MDL-WATER-03-QT4-2022

Tgt Ion:165 Resp: 21989





Ion Ratio Lower Upper

165 100

63 21.7 20.6 31.0

89 45.8 43.1 64.7

182 2.7 2.4 3.6



#58

Fluorene

Concen: 8.247 ng

RT: 15.932 min Scan# 2152

Delta R.T. -0.005 min

Lab File: BG056513.D

Acq: 2 Feb 2023 14:01




Tgt Ion:166 Resp: 70293


Ion Ratio Lower Upper

166 100

165 101.2 78.8 118.2

167 14.5 10.6 15.8









#59
2,3,4,6-Tetrachlorophenol
Concen: 7.259 ng
RT: 15.515 min Scan# 2081
Delta R.T. 0.001 min
Lab File: BG056513.D
Acq: 2 Feb 2023 14:01




Instrument :
BNA_G
ClientSampleId :
MDL-WATER-03-QT4-2022


Tgt Ion:232 Resp: 19427
Ion Ratio Lower Upper
232 100
131 26.0 22.8 34.2
130 0.9 1.4 2.0#
166 24.4 18.8 28.2



#60
Diethylphthalate
Concen: 7.908 ng
RT: 15.680 min Scan# 2109
Delta R.T. -0.005 min
Lab File: BG056513.D
Acq: 2 Feb 2023 14:01

Tgt Ion:149 Resp: 65604
Ion Ratio Lower Upper
149 100
177 25.5 21.0 31.4
150 12.3 9.9 14.9









#61
4-Chlorophenyl-phenylether
Concen: 8.210 ng
RT: 15.915 min Scan# 2149
Delta R.T. -0.005 min
Lab File: BG056513.D
Acq: 2 Feb 2023 14:01




Instrument :
BNA_G
ClientSampleId :
MDL-WATER-03-QT4-2022


Tgt Ion:204 Resp: 42701
Ion Ratio Lower Upper
204 100
206 34.2 26.2 39.4
141 44.6 32.7 49.1



#62
4-Nitroaniline
Concen: 7.393 ng
RT: 15.950 min Scan# 2155
Delta R.T. -0.011 min
Lab File: BG056513.D
Acq: 2 Feb 2023 14:01

Tgt Ion:138 Resp: 13817
Ion Ratio Lower Upper
138 100
92 56.3 27.8 67.8
108 85.2 72.6 112.6





#63

Azobenzene

Concen: 8.171 ng

RT: 16.208 min Scan# 21

Delta R.T. -0.005 min

Lab File: BG056513.D

Acq: 2 Feb 2023 14:01

Instrument :

BNA_G

ClientSampleId :

MDL-WATER-03-QT4-2022

Tgt Ion: 77 Resp: 46642





Ion Ratio Lower Upper

77 100

182 38.6 23.1 63.1

105 23.9 4.2 44.2

51 16.6 0.0 39.8



#64

Phenanthrene-d10

Concen: 20.000 ng

RT: 17.630 min Scan# 2441

Delta R.T. -0.005 min

Lab File: BG056513.D

Acq: 2 Feb 2023 14:01




Tgt Ion:188 Resp: 289271


Ion Ratio Lower Upper

188 100

94 5.5 4.1 6.1

80 5.7 5.0 7.4









#65
4,6-Dinitro-2-methylphenol
Concen: 5.886 ng
RT: 16.015 min Scan# 2167
Delta R.T. 0.001 min
Lab File: BG056513.D
Acq: 2 Feb 2023 14:01




Instrument :
BNA_G
ClientSampleId :
MDL-WATER-03-QT4-2022


Tgt Ion:198 Resp: 11971
Ion Ratio Lower Upper
198 100
51 13.2 0.0 35.0
105 31.1 7.4 47.4



#66
n-Nitrosodiphenylamine
Concen: 7.726 ng
RT: 16.132 min Scan# 2186
Delta R.T. 0.001 min
Lab File: BG056513.D
Acq: 2 Feb 2023 14:01

Tgt Ion:169 Resp: 63273
Ion Ratio Lower Upper
169 100
168 71.6 55.8 83.6
167 38.6 29.8 44.8





#67

4-Bromophenyl-phenylether

Concen: 7.905 ng

RT: 16.808 min Scan# 21

Delta R.T. -0.005 min

Lab File: BG056513.D

Acq: 2 Feb 2023 14:01

Instrument :

BNA_G

ClientSampleId :

MDL-WATER-03-QT4-2022





Tgt Ion:248 Resp: 29272

Ion Ratio Lower Upper

248 100

250 92.0 77.6 116.4

141 50.6 39.0 58.4



#68

Hexachlorobenzene

Concen: 8.293 ng

RT: 16.937 min Scan# 2323

Delta R.T. -0.005 min

Lab File: BG056513.D

Acq: 2 Feb 2023 14:01




Tgt Ion:284 Resp: 30078


Ion Ratio Lower Upper

284 100

142 26.9 18.0 27.0

249 34.9 26.7 40.1





#69

Atrazine

Concen: 8.852 ng

RT: 17.060 min Scan# 21

Delta R.T. -0.011 min

Lab File: BG056513.D

Acq: 2 Feb 2023 14:01

Instrument :

BNA_G

ClientSampleId :

MDL-WATER-03-QT4-2022





Tgt Ion:200 Resp: 27970

Ion Ratio Lower Upper

200 100

173 22.4 3.2 43.2

215 48.6 30.7 70.7



#70

Pentachlorophenol

Concen: 6.746 ng

RT: 17.290 min Scan# 2383

Delta R.T. 0.007 min

Lab File: BG056513.D

Acq: 2 Feb 2023 14:01




Tgt Ion:266 Resp: 14805


Ion Ratio Lower Upper

266 100

268 63.7 50.2 75.4

264 61.0 51.4 77.0









#71
Phenanthrene
Concen: 8.009 ng
RT: 17.677 min Scan# 2449
Delta R.T. 0.001 min
Lab File: BG056513.D
Acq: 2 Feb 2023 14:01




Instrument :
BNA_G
ClientSampleId :
MDL-WATER-03-QT4-2022


Tgt Ion:178 Resp: 121258
Ion Ratio Lower Upper
178 100
176 19.2 15.6 23.4
179 16.2 13.3 19.9



#72
Anthracene
Concen: 7.934 ng
RT: 17.765 min Scan# 2464
Delta R.T. -0.005 min
Lab File: BG056513.D
Acq: 2 Feb 2023 14:01

Tgt Ion:178 Resp: 123310
Ion Ratio Lower Upper
178 100
176 18.8 14.5 21.7
179 15.2 12.5 18.7









#73
Carbazole
Concen: 7.968 ng
RT: 18.030 min Scan# 2509
Delta R.T. -0.005 min
Lab File: BG056513.D
Acq: 2 Feb 2023 14:01




Instrument : BNA_G
ClientSampleId : MDL-WATER-03-QT4-2022


Tgt Ion:167 Resp: 105470
Ion Ratio Lower Upper
167 100
166 20.1 17.8 26.6
139 9.4 8.9 13.3



#74
Di-n-butylphthalate
Concen: 7.601 ng
RT: 18.559 min Scan# 2599
Delta R.T. -0.005 min
Lab File: BG056513.D
Acq: 2 Feb 2023 14:01

Tgt Ion:149 Resp: 105642
Ion Ratio Lower Upper
149 100
150 9.4 7.8 11.6
104 4.6 4.0 6.0









#75
Fluoranthene
Concen: 8.196 ng
RT: 19.669 min Scan# 21
Delta R.T. 0.001 min
Lab File: BG056513.D
Acq: 2 Feb 2023 14:01




Instrument :
BNA_G
ClientSampleId :
MDL-WATER-03-QT4-2022


Tgt Ion:202 Resp: 156621
Ion Ratio Lower Upper
202 100
101 6.1 0.0 25.6
203 18.1 0.0 38.1



#76
Chrysene-d12
Concen: 20.000 ng
RT: 21.925 min Scan# 3172
Delta R.T. 0.001 min
Lab File: BG056513.D
Acq: 2 Feb 2023 14:01

Tgt Ion:240 Resp: 284145
Ion Ratio Lower Upper
240 100
120 3.6 3.8 5.8#
236 25.5 20.2 30.2









#77
Benzidine
Concen: 10.490 ng
RT: 19.834 min Scan# 2816
Delta R.T. -0.005 min
Lab File: BG056513.D
Acq: 2 Feb 2023 14:01




Instrument :
BNA_G
ClientSampleId :
MDL-WATER-03-QT4-2022


Tgt Ion:184 Resp: 80658
Ion Ratio Lower Upper
184 100
185 15.7 12.3 18.5
183 12.9 9.8 14.8



#78
Pyrene
Concen: 7.827 ng
RT: 20.027 min Scan# 2849
Delta R.T. -0.005 min
Lab File: BG056513.D
Acq: 2 Feb 2023 14:01

Tgt Ion:202 Resp: 158170
Ion Ratio Lower Upper
202 100
200 20.6 16.7 25.1
203 16.6 14.4 21.6









#79
 Terphenyl-d14
 Concen: 78.623 ng
 RT: 20.210 min Scan# 21
 Delta R.T. 0.001 min
 Lab File: BG056513.D
 Acq: 2 Feb 2023 14:01




Instrument :
 BNA_G
 ClientSampleId :
 MDL-WATER-03-QT4-2022


Tgt Ion:244 Resp: 1271966
 Ion Ratio Lower Upper
 244 100
 212 7.0 5.5 8.3
 122 5.5 4.4 6.6



#80
 Butylbenzylphthalate
 Concen: 7.224 ng
 RT: 20.885 min Scan# 2995
 Delta R.T. 0.001 min
 Lab File: BG056513.D
 Acq: 2 Feb 2023 14:01

Tgt Ion:149 Resp: 44587
 Ion Ratio Lower Upper
 149 100
 91 59.8 49.4 74.0
 206 29.1 22.0 33.0









#81
Benzo(a)anthracene
Concen: 7.902 ng
RT: 21.902 min Scan# 3170
Delta R.T. 0.001 min
Lab File: BG056513.D
Acq: 2 Feb 2023 14:01




Instrument :
BNA_G
ClientSampleId :
MDL-WATER-03-QT4-2022


Tgt Ion:228 Resp: 156978
Ion Ratio Lower Upper
228 100
226 27.9 22.4 33.6
229 19.7 16.5 24.7



#82
3,3'-Dichlorobenzidine
Concen: 8.792 ng
RT: 21.802 min Scan# 3151
Delta R.T. -0.005 min
Lab File: BG056513.D
Acq: 2 Feb 2023 14:01

Tgt Ion:252 Resp: 62893
Ion Ratio Lower Upper
252 100
254 60.0 50.0 75.0
126 6.4 4.3 6.5





#83

Chrysene

Concen: 7.855 ng

RT: 21.972 min Scan# 3180

Delta R.T. -0.005 min

Lab File: BG056513.D

Acq: 2 Feb 2023 14:01

Instrument :

BNA_G

ClientSampleId :

MDL-WATER-03-QT4-2022





Tgt Ion:228 Resp: 146594

Ion Ratio Lower Upper

228 100

226 29.3 25.0 37.4

229 19.6 16.4 24.6



#84

Bis(2-ethylhexyl)phthalate

Concen: 7.231 ng

RT: 21.761 min Scan# 3144

Delta R.T. -0.005 min

Lab File: BG056513.D

Acq: 2 Feb 2023 14:01




Tgt Ion:149 Resp: 63987


Ion Ratio Lower Upper

149 100

167 28.9 23.4 35.2

279 5.4 4.4 6.6





#85

Di-n-octyl phthalate

Concen: 7.317 ng

RT: 23.030 min Scan# 31

Delta R.T. -0.005 min

Lab File: BG056513.D

Acq: 2 Feb 2023 14:01

Instrument :

BNA_G

ClientSampleId :

MDL-WATER-03-QT4-2022





Tgt Ion:149 Resp: 107829

Ion Ratio Lower Upper

149 100

167 1.9 1.4 2.0

43 3.1 2.7 4.1



#86

Perylene-d12

Concen: 20.000 ng

RT: 25.339 min Scan# 3753

Delta R.T. -0.005 min

Lab File: BG056513.D

Acq: 2 Feb 2023 14:01




Tgt Ion:264 Resp: 308278


Ion Ratio Lower Upper

264 100

260 24.3 18.2 27.4

265 21.5 18.1 27.1





#87

Indeno(1,2,3-cd)pyrene

Concen: 7.644 ng

RT: 29.275 min Scan# 4428

Delta R.T. -0.005 min

Lab File: BG056513.D

Acq: 2 Feb 2023 14:01

Instrument :

BNA_G

ClientSampleId :

MDL-WATER-03-QT4-2022





Tgt Ion:276 Resp: 172508

Ion Ratio Lower Upper

276 100

138 11.7 9.4 14.2

277 25.3 20.3 30.5



#88

Benzo(b)fluoranthene

Concen: 8.354 ng

RT: 24.246 min Scan# 3567

Delta R.T. -0.005 min

Lab File: BG056513.D

Acq: 2 Feb 2023 14:01




Tgt Ion:252 Resp: 159854


Ion Ratio Lower Upper

252 100

253 20.9 18.1 27.1

125 4.9 3.9 5.9









#89
Benzo(k)fluoranthene
Concen: 8.139 ng
RT: 24.317 min Scan# 31
Delta R.T. -0.005 min
Lab File: BG056513.D
Acq: 2 Feb 2023 14:01




Instrument :
BNA_G
ClientSampleId :
MDL-WATER-03-QT4-2022


Tgt Ion:252 Resp: 157232
Ion Ratio Lower Upper
252 100
253 21.4 17.5 26.3
125 4.7 3.4 5.0



#90
Benzo(a)pyrene
Concen: 7.488 ng
RT: 25.180 min Scan# 3726
Delta R.T. -0.005 min
Lab File: BG056513.D
Acq: 2 Feb 2023 14:01

Tgt Ion:252 Resp: 141135
Ion Ratio Lower Upper
252 100
253 20.7 18.4 27.6
125 4.9 4.1 6.1





#91

Dibenzo(a,h)anthracene

Concen: 8.083 ng

RT: 29.322 min Scan# 4431

Delta R.T. -0.022 min

Lab File: BG056513.D

Acq: 2 Feb 2023 14:01

Instrument :

BNA_G

ClientSampleId :

MDL-WATER-03-QT4-2022





Tgt Ion:278 Resp: 153111

Ion Ratio Lower Upper

278 100

139 7.5 5.5 8.3

279 23.8 18.8 28.2



#92

Benzo(g,h,i)perylene

Concen: 8.132 ng

RT: 30.503 min Scan# 4632

Delta R.T. -0.022 min

Lab File: BG056513.D

Acq: 2 Feb 2023 14:01

Tgt Ion:276 Resp: 148629

Ion Ratio Lower Upper

276 100

277 22.9 18.5 27.7

138 9.7 7.4 11.0

