

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG020322\
 Data File : BG052265.D
 Acq On : 3 Feb 2022 12:44
 Operator : CG/JU
 Sample : N1349-08
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 C0HP7

Quant Time: Feb 03 23:59:43 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG013122.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Jan 31 18:07:10 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.247	152	30454	20.000	ng/u1	0.01
20) Naphthalene-d8	11.084	136	121909	20.000	ng/u1	# 0.00
38) Acenaphthene-d10	14.885	164	83973	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.628	188	203837	20.000	ng/u1	# 0.00
79) Chrysene-d12	21.946	240	213406	20.000	ng/u1	# 0.00
88) Perylene-d12	25.418	264	215348	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.565	96	3639	4.821	ng/uL	0.00
4) Pyridine-d5	4.023	84	2991	1.274	ng/u1	0.04
7) Phenol-d5	7.389	99	23907	8.630	ng/u1	0.01
9) Bis-(2-Chloroethyl)eth...	7.554	67	67675	40.213	ng/u1	0.01
11) 2-Chlorophenol-d4	7.777	132	60626	31.408	ng/u1	0.01
15) 4-Methylphenol-d8	8.952	113	42863	20.390	ng/u1	0.01
21) Nitrobenzene-d5	9.416	128	41180	41.486	ng/u1	0.00
24) 2-Nitrophenol-d4	10.150	143	41640	38.389	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.696	165	74425	35.197	ng/u1	0.00
31) 4-Chloroaniline-d4	11.207	131	44689	16.661	ng/u1	0.00
46) Dimethylphthalate-d6	14.274	166	271644	41.155	ng/u1	0.00
49) Acenaphthylene-d8	14.579	160	338572	41.153	ng/u1	0.00
54) 4-Nitrophenol-d4	15.067	143	9310	9.313	ng/u1	0.00
60) Fluorene-d10	15.872	176	243111	42.041	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.983	200	45337	36.027	ng/u1	0.00
73) Anthracene-d10	17.728	188	426167	44.348	ng/u1	0.00
81) Pyrene-d10	20.007	212	525221	41.065	ng/u1	0.00
92) Benzo(a)pyrene-d12	25.183	264	503191	43.983	ng/u1	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG020322\
Data File : BG052265.D
Acq On : 3 Feb 2022 12:44
Operator : CG/JU
Sample : N1349-08
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
C0HP7

Quant Time: Feb 03 23:59:43 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG013122.M
Quant Title : SVOA CALIBRATION
QLast Update : Mon Jan 31 18:07:10 2022
Response via : Initial Calibration

