


Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG020619\
 Data File : BG039366.D
 Acq On : 6 Feb 2019 12:03
 Operator : JU/SJ
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\8270-BG012919.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Tue Jan 29 15:41:51 2019



AutoFind: Scans 2527, 2528, 2529; Background Corrected with Scan 2520

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	38.1	20093	PASS
68	69	0.00	2	2.0	333	PASS
69	198	0.00	100	31.7	16721	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	42.9	22610	PASS
197	198	0.00	2	0.6	295	PASS
198	198	100	100	100.0	52760	PASS
199	198	5	9	7.3	3835	PASS
275	198	10	60	27.1	14324	PASS
365	198	1	100	3.6	1910	PASS
441	198	0.01	100	9.7	5115	PASS
442	198	50	100	63.9	33709	PASS
443	442	15	24	18.8	6326	PASS