

Data Path : Z:\SVOASRV\HPCHEM1\BNA G\DATA\BG020720\
 Data File : BG044347.D
 Acq On : 7 Feb 2020 21:33
 Operator : CG/JU
 Sample : L1440-01
 Misc :
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 TW-1

Manual Integrations
 APPROVED

mohammad
 2/11/2020 8:26:16 AM

Quant Time: Feb 08 03:11:09 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG013020.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jan 30 16:05:09 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.91	152	1255	20.00	ng	0.00
21) Naphthalene-d8	10.28	136	113	20.00	ng	-0.44
39) Acenaphthene-d10	14.00	164	3152	20.00	ng	-0.55
64) Phenanthrene-d10	16.49	188	55	20.00	ng	-0.81
76) Chrysene-d12	20.71	240	137	20.00	ng	-0.84
87) Perylene-d12	0.00	264	0	0.00	ng	-24.64
System Monitoring Compounds						
5) 2-Fluorophenol	5.49	112	210708	2963.08	ng	0.00
7) Phenol-d6	7.08	99	165182	1729.81	ng	0.00
23) Nitrobenzene-d5	9.07	82	462177	230909.77	ng	0.00
42) 2,4,6-Tribromophenol	0.00	330	0	0.00	ng	
45) 2-Fluorobiphenyl	0.00	172	0	0.00	ng	
79) Terphenyl-d14	0.00	244	0	0.00	ng	
Target Compounds						
10) Phenol	7.11	94	5460	57.775	ng	Qvalue 79
22) Acetophenone	8.73	105	35936m	13072.671	ng	
31) Naphthalene	10.57	128	501	91.383	ng	# 67
32) Benzoic acid	10.02	122	52214	39048.568	ng	93
74) Di-n-butylphthalate	18.26	149	6317m	2105.886	ng	
84) Bis(2-ethylhexyl)phthalate	21.44	149	4515m	818.136	ng	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA G\DATA\BG020720\
Data File : BG044347.D
Acq On : 7 Feb 2020 21:33
Operator : CG/JU
Sample : L1440-01
Misc :
ALS Vial : 18 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
TW-1

Manual Integrations
APPROVED

mohammad
2/11/2020 8:26:16 AM

Quant Time: Feb 08 03:11:09 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG013020.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jan 30 16:05:09 2020
Response via : Initial Calibration

