

Data Path : Z:\SVOASRV\HPCHEM1\BNA G\DATA\BG020720\
 Data File : BG044349.D
 Acq On : 7 Feb 2020 23:31
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Feb 08 02:35:59 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG013020.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jan 30 16:05:09 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	128	0.00
2	1,4-Dioxane	0.509	0.493	3.1	119	0.00
3	Pyridine	1.357	1.423	-4.9	119	0.00
4	n-Nitrosodimethylamine	0.567	0.572	-0.9	127	0.00
5 S	2-Fluorophenol	1.133	1.149	-1.4	125	0.00
6	Aniline	1.889	1.883	0.3	122	0.00
7 S	Phenol-d6	1.522	1.531	-0.6	124	0.00
8	2-Chlorophenol	1.245	1.264	-1.5	124	0.00
9	Benzaldehyde	0.867	0.830	4.3	124	0.00
10 C	Phenol	1.506	1.531	-1.7	125	0.00
11	bis(2-Chloroethyl)ether	1.288	1.277	0.9	123	0.00
12	1,3-Dichlorobenzene	1.487	1.492	-0.3	126	0.00
13 C	1,4-Dichlorobenzene	1.473	1.478	-0.3	124	0.00
14	1,2-Dichlorobenzene	1.392	1.395	-0.2	123	0.00
15	Benzyl Alcohol	1.077	1.111	-3.2	126	0.00
16	2,2'-oxybis(1-Chloropropane	1.743	1.784	-2.4	126	0.00
17	2-Methylphenol	1.075	1.084	-0.8	124	0.00
18	Hexachloroethane	0.553	0.560	-1.3	127	0.00
19 P	n-Nitroso-di-n-propylamine	1.014	0.987	2.7	119	0.00
20	3+4-Methylphenols	1.481	1.482	-0.1	122	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	121	0.00
22	Acetophenone	0.487	0.475	2.5	119	0.00
23 S	Nitrobenzene-d5	0.354	0.352	0.6	120	0.00
24	Nitrobenzene	0.346	0.345	0.3	122	0.00
25	Isophorone	0.660	0.647	2.0	119	0.00
26 C	2-Nitrophenol	0.179	0.188	-5.0	125	0.00
27	2,4-Dimethylphenol	0.253	0.254	-0.4	121	0.00
28	bis(2-Chloroethoxy)methane	0.440	0.429	2.5	117	0.00
29 C	2,4-Dichlorophenol	0.306	0.308	-0.7	122	0.00
30	1,2,4-Trichlorobenzene	0.351	0.353	-0.6	122	0.00
31	Naphthalene	0.970	0.957	1.3	120	0.00
32	Benzoic acid	0.135	0.147	-8.9	135	0.00
33	4-Chloroaniline	0.441	0.434	1.6	120	0.00
34 C	Hexachlorobutadiene	0.216	0.219	-1.4	122	0.00
35	Caprolactam	0.112	0.106	5.4	117	0.00
36 C	4-Chloro-3-methylphenol	0.321	0.311	3.1	119	0.00
37	2-Methylnaphthalene	0.720	0.701	2.6	119	0.00
38	1-Methylnaphthalene	0.679	0.657	3.2	118	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	118	0.00
40	1,2,4,5-Tetrachlorobenzene	0.598	0.609	-1.8	119	0.00
41 P	Hexachlorocyclopentadiene	0.287	0.336	-17.1	133	0.00
42 S	2,4,6-Tribromophenol	0.235	0.234	0.4	120	0.00
43 C	2,4,6-Trichlorophenol	0.387	0.404	-4.4	122	0.00
44	2,4,5-Trichlorophenol	0.395	0.409	-3.5	120	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA G\DATA\BG020720\
 Data File : BG044349.D
 Acq On : 7 Feb 2020 23:31
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Feb 08 02:35:59 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG013020.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jan 30 16:05:09 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
45 S	2-Fluorobiphenyl	1.194	1.177	1.4	115	0.00
46	1,1'-Biphenyl	1.443	1.443	0.0	118	0.00
47	2-Chloronaphthalene	1.148	1.141	0.6	118	0.00
48	2-Nitroaniline	0.339	0.340	-0.3	119	0.00
49	Acenaphthylene	1.684	1.658	1.5	116	0.00
50	Dimethylphthalate	1.438	1.380	4.0	114	0.00
51	2,6-Dinitrotoluene	0.321	0.313	2.5	118	0.00
52 C	Acenaphthene	1.091	1.074	1.6	117	0.00
53	3-Nitroaniline	0.355	0.344	3.1	116	0.00
54 P	2,4-Dinitrophenol	0.159	0.171	-7.5	123	0.00
55	Dibenzofuran	1.652	1.605	2.8	115	0.00
56 P	4-Nitrophenol	0.260	0.266	-2.3	120	0.00
57	2,4-Dinitrotoluene	0.449	0.434	3.3	117	0.00
58	Fluorene	1.301	1.253	3.7	115	0.00
59	2,3,4,6-Tetrachlorophenol	0.357	0.356	0.3	118	0.00
60	Diethylphthalate	1.433	1.366	4.7	114	0.00
61	4-Chlorophenyl-phenylether	0.706	0.684	3.1	115	0.00
62	4-Nitroaniline	0.354	0.347	2.0	119	0.00
63	Azobenzene	1.256	1.212	3.5	115	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	116	0.00
65	4,6-Dinitro-2-methylphenol	0.110	0.117	-6.4	120	0.00
66 c	n-Nitrosodiphenylamine	0.560	0.565	-0.9	116	0.00
67	4-Bromophenyl-phenylether	0.218	0.221	-1.4	117	0.00
68	Hexachlorobenzene	0.244	0.244	0.0	115	0.00
69	Atrazine	0.192	0.191	0.5	112	0.00
70 C	Pentachlorophenol	0.113	0.129	-14.2	124	0.00
71	Phenanthrene	0.968	0.958	1.0	114	0.00
72	Anthracene	0.957	0.946	1.1	114	0.00
73	Carbazole	0.883	0.873	1.1	114	0.00
74	Di-n-butylphthalate	1.091	1.088	0.3	112	0.00
75 C	Fluoranthene	1.120	1.097	2.1	112	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	115	0.00
77	Benzidine	0.555	0.596	-7.4	108	0.00
78	Pyrene	1.284	1.269	1.2	111	0.00
79 S	Terphenyl-d14	0.972	0.880	9.5	109	0.00
80	Butylbenzylphthalate	0.585	0.603	-3.1	115	0.00
81	Benzo(a)anthracene	1.233	1.251	-1.5	115	0.00
82	3,3'-Dichlorobenzidine	0.478	0.504	-5.4	115	0.00
83	Chrysene	1.191	1.191	0.0	112	0.00
84	Bis(2-ethylhexyl)phthalate	0.806	0.815	-1.1	113	0.00
85 c	Di-n-octyl phthalate	1.394	1.456	-4.4	116	-0.01
86	Indeno(1,2,3-cd)pyrene	1.554	1.577	-1.5	117	-0.01
87 I	Perylene-d12	1.000	1.000	0.0	115	-0.01

Data Path : Z:\SVOASRV\HPCHEM1\BNA G\DATA\BG020720\
Data File : BG044349.D
Acq On : 7 Feb 2020 23:31
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTDCCC040

Quant Time: Feb 08 02:35:59 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG013020.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jan 30 16:05:09 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88	Benzo(b)fluoranthene	1.152	1.169	-1.5	116	-0.01
89	Benzo(k)fluoranthene	1.123	1.110	1.2	111	0.00
90 C	Benzo(a)pyrene	1.090	1.099	-0.8	115	-0.01
91	Dibenzo(a,h)anthracene	1.078	1.085	-0.6	116	-0.02
92	Benzo(q,h,i)perylene	1.080	1.085	-0.5	116	-0.02

(#) = Out of Range

SPCC's out = 0 CCC's out = 0