



Benzaldehyde

Response Ratio



$R = -1.515e-001 A^2 + 1.302e+000 A + 2.071e-002$
Coef of Det (r^2) = 0.992373 Curve Fit: Quadratic
Method Name: Z:\svoasrv\HPCHEM1\BNA G\Methods\8270-BG020722.M
Calibration Table Last Updated: Tue Feb 08 01:21:20 2022


Benzoic acid



$\text{Response} = 2.023\text{e-}001 * \text{Amt} - 6.046\text{e-}002$
 Coef of Det (r^2) = 0.996144 Curve Fit: Linear
 Method Name: Z:\svoasrv\HPCHEM1\BNA G\Methods\8270-BG020722.M
 Calibration Table Last Updated: Tue Feb 08 01:21:20 2022

Hexachlorocyclopentadiene

Response Ratio




$$\text{Response} = 3.392\text{e-}001 * \text{Amt} - 8.919\text{e-}002$$

Coef of Det (r^2) = 0.997177 Curve Fit: Linear

Method Name: Z:\svoasrv\HPCHEM1\BNA G\Methods\8270-BG020722.M

Calibration Table Last Updated: Tue Feb 08 01:21:20 2022

2,4-Dinitrophenol



Response = $2.181e-001 * Amt - 5.346e-002$
 Coef of Det (r^2) = 0.999397 Curve Fit: Linear
 Method Name: Z:\svoasrv\HPCHEM1\BNA G\Methods\8270-BG020722.M
 Calibration Table Last Updated: Tue Feb 08 01:21:20 2022