

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG020822\
 Data File : BG052312.D
 Acq On : 8 Feb 2022 10:30
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Feb 08 11:07:03 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG020722.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Feb 08 01:21:20 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	83	0.00
2	1,4-Dioxane	40.000	43.187	-8.0	95	0.00
3	Pyridine	40.000	44.428	-11.1	91	0.00
4	n-Nitrosodimethylamine	40.000	42.926	-7.3	91	0.00
5 S	2-Fluorophenol	80.000	87.163	-9.0	92	0.01
6	Aniline	40.000	43.846	-9.6	89	0.00
7 S	Phenol-d6	80.000	87.430	-9.3	90	0.00
8	2-Chlorophenol	40.000	42.771	-6.9	92	0.00
9	Benzaldehyde	40.000	47.621	-19.1	91	0.00
10 C	Phenol	40.000	43.326	-8.3	90	0.00
11	bis(2-Chloroethyl)ether	40.000	41.097	-2.7	88	0.00
12	1,3-Dichlorobenzene	40.000	42.268	-5.7	90	0.00
13 C	1,4-Dichlorobenzene	40.000	42.652	-6.6	90	0.00
14	1,2-Dichlorobenzene	40.000	42.170	-5.4	90	0.00
15	Benzyl Alcohol	40.000	45.115	-12.8	91	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	41.602	-4.0	90	0.01
17	2-Methylphenol	40.000	44.729	-11.8	93	0.01
18	Hexachloroethane	40.000	40.345	-0.9	88	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	42.426	-6.1	90	0.01
20	3+4-Methylphenols	40.000	43.782	-9.5	90	0.01
21 I	Naphthalene-d8	20.000	20.000	0.0	84	0.00
22	Acetophenone	40.000	42.494	-6.2	92	0.00
23 S	Nitrobenzene-d5	80.000	84.876	-6.1	91	0.00
24	Nitrobenzene	40.000	42.013	-5.0	88	0.00
25	Isophorone	40.000	43.480	-8.7	92	0.00
26 C	2-Nitrophenol	40.000	44.584	-11.5	91	0.00
27	2,4-Dimethylphenol	40.000	44.276	-10.7	93	0.00
28	bis(2-Chloroethoxy)methane	40.000	42.732	-6.8	91	0.00
29 C	2,4-Dichlorophenol	40.000	42.960	-7.4	91	0.00
30	1,2,4-Trichlorobenzene	40.000	42.042	-5.1	91	0.00
31	Naphthalene	40.000	42.765	-6.9	92	0.00
32	Benzoic acid	40.000	39.807	0.5	89	0.00
33	4-Chloroaniline	40.000	43.883	-9.7	91	0.00
34 C	Hexachlorobutadiene	40.000	40.587	-1.5	89	0.00
35	Caprolactam	40.000	45.551	-13.9	95	0.00
36 C	4-Chloro-3-methylphenol	40.000	45.508	-13.8	94	0.00
37	2-Methylnaphthalene	40.000	42.828	-7.1	92	0.00
38	1-Methylnaphthalene	40.000	43.019	-7.5	91	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	84	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	42.271	-5.7	92	0.00
41 P	Hexachlorocyclopentadiene	40.000	41.830	-4.6	94	0.00
42 S	2,4,6-Tribromophenol	80.000	89.448	-11.8	95	0.00
43 C	2,4,6-Trichlorophenol	40.000	44.963	-12.4	95	0.00
44	2,4,5-Trichlorophenol	40.000	44.280	-10.7	94	0.00
45 S	2-Fluorobiphenyl	80.000	85.675	-7.1	92	0.00
46	1,1'-Biphenyl	40.000	43.180	-7.9	93	0.00
47	2-Chloronaphthalene	40.000	43.194	-8.0	93	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG020822\
 Data File : BG052312.D
 Acq On : 8 Feb 2022 10:30
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Feb 08 11:07:03 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG020722.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Feb 08 01:21:20 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	45.172	-12.9	96	0.00
49	Acenaphthylene	40.000	43.661	-9.2	94	0.00
50	Dimethylphthalate	40.000	43.448	-8.6	94	0.00
51	2,6-Dinitrotoluene	40.000	44.628	-11.6	94	0.00
52 C	Acenaphthene	40.000	44.396	-11.0	95	0.00
53	3-Nitroaniline	40.000	45.520	-13.8	95	0.00
54 P	2,4-Dinitrophenol	40.000	43.378	-8.4	96	0.00
55	Dibenzofuran	40.000	42.928	-7.3	94	0.00
56 P	4-Nitrophenol	40.000	47.155	-17.9	97	0.00
57	2,4-Dinitrotoluene	40.000	45.510	-13.8	95	0.00
58	Fluorene	40.000	44.724	-11.8	96	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	44.532	-11.3	95	0.00
60	Diethylphthalate	40.000	44.308	-10.8	96	0.00
61	4-Chlorophenyl-phenylether	40.000	43.289	-8.2	95	0.00
62	4-Nitroaniline	40.000	47.359	-18.4	100	0.00
63	Azobenzene	40.000	43.564	-8.9	95	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	88	0.00
65	4,6-Dinitro-2-methylphenol	40.000	46.203	-15.5	96	0.00
66 c	n-Nitrosodiphenylamine	40.000	43.042	-7.6	94	0.00
67	4-Bromophenyl-phenylether	40.000	42.800	-7.0	95	0.00
68	Hexachlorobenzene	40.000	42.782	-7.0	95	0.00
69	Atrazine	40.000	43.699	-9.2	95	0.00
70 C	Pentachlorophenol	40.000	44.715	-11.8	95	0.00
71	Phenanthrene	40.000	43.138	-7.8	96	0.00
72	Anthracene	40.000	43.140	-7.9	96	0.00
73	Carbazole	40.000	43.427	-8.6	97	0.00
74	Di-n-butylphthalate	40.000	44.040	-10.1	97	0.00
75 C	Fluoranthene	40.000	42.869	-7.2	97	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	90	0.00
77	Benzydine	40.000	49.565	-23.9	97	0.00
78	Pyrene	40.000	44.012	-10.0	97	0.00
79 S	Terphenyl-d14	80.000	86.059	-7.6	97	0.00
80	Butylbenzylphthalate	40.000	45.891	-14.7	100	0.00
81	Benzo(a)anthracene	40.000	44.002	-10.0	98	0.00
82	3,3'-Dichlorobenzidine	40.000	44.981	-12.5	98	0.00
83	Chrysene	40.000	42.973	-7.4	97	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	45.510	-13.8	97	0.00
85 c	Di-n-octyl phthalate	40.000	45.774	-14.4	99	0.00
86 I	Perylene-d12	20.000	20.000	0.0	90	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	43.772	-9.4	99	0.02
88	Benzo(b)fluoranthene	40.000	43.506	-8.8	98	0.00
89	Benzo(k)fluoranthene	40.000	43.904	-9.8	98	0.01
90 C	Benzo(a)pyrene	40.000	43.703	-9.3	98	0.00
91	Dibenzo(a,h)anthracene	40.000	43.149	-7.9	98	0.01
92	Benzo(g,h,i)perylene	40.000	43.259	-8.1	98	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG020822\
Data File : BG052312.D
Acq On : 8 Feb 2022 10:30
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTDCCC040

Quant Time: Feb 08 11:07:03 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG020722.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Feb 08 01:21:20 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0