

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG020921\
 Data File : BG048121.D
 Acq On : 9 Feb 2021 14:09
 Operator : CG/JU
 Sample : PB134329BL
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 PB134329BL

Quant Time: Feb 09 15:45:44 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\8270-BG012521.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Feb 09 14:29:17 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.109	152	50511	20.00	ng	0.00
21) Naphthalene-d8	10.924	136	195244	20.00	ng	0.00
39) Acenaphthene-d10	14.737	164	139405	20.00	ng	0.00
64) Phenanthrene-d10	17.475	188	356041	20.00	ng	# 0.00
76) Chrysene-d12	21.746	240	395139	20.00	ng	# 0.00
86) Perylene-d12	25.013	264	378892	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.665	112	442567	134.65	ng	0.00
7) Phenol-d6	7.257	99	586359	117.34	ng	0.00
23) Nitrobenzene-d5	9.273	82	402552	81.58	ng	0.00
42) 2,4,6-Tribromophenol	16.217	330	285819	123.10	ng	0.00
45) 2-Fluorobiphenyl	13.362	172	883522	81.94	ng	0.00
79) Terphenyl-d14	20.078	244	1598133	70.12	ng	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG020921\
Data File : BG048121.D
Acq On : 9 Feb 2021 14:09
Operator : CG/JU
Sample : PB134329BL
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
PB134329BL

Quant Time: Feb 09 15:45:44 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\8270-BG012521.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Feb 09 14:29:17 2021
Response via : Initial Calibration

