

Data Path : Z:\svoasrv\HPCHEM1\BNA G\Data\BG021020\
 Data File : BG044356.D
 Acq On : 10 Feb 2020 13:20
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Feb 10 16:13:50 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG013020.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Feb 10 16:11:50 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	77	0.00
2	1,4-Dioxane	0.509	0.472	7.3	69	0.00
3	Pyridine	1.357	1.480	-9.1	75	0.00
4	n-Nitrosodimethylamine	0.567	0.559	1.4	75	0.00
5 S	2-Fluorophenol	1.133	1.157	-2.1	76	0.00
6	Aniline	1.889	2.037	-7.8	79	0.00
7 S	Phenol-d6	1.522	1.648	-8.3	81	0.00
8	2-Chlorophenol	1.245	1.323	-6.3	79	0.00
9	Benzaldehyde	0.867	0.835	3.7	76	0.00
10 C	Phenol	1.506	1.636	-8.6	81	0.00
11	bis(2-Chloroethyl)ether	1.288	1.374	-6.7	80	0.00
12	1,3-Dichlorobenzene	1.487	1.521	-2.3	77	0.00
13 C	1,4-Dichlorobenzene	1.473	1.532	-4.0	78	0.00
14	1,2-Dichlorobenzene	1.392	1.460	-4.9	78	0.00
15	Benzyl Alcohol	1.077	1.185	-10.0	81	0.00
16	2,2'-oxybis(1-Chloropropane	1.743	1.852	-6.3	79	0.00
17	2-Methylphenol	1.075	1.173	-9.1	81	0.00
18	Hexachloroethane	0.553	0.563	-1.8	77	0.00
19 P	n-Nitroso-di-n-propylamine	1.014	1.140	-12.4	83	0.00
20	3+4-Methylphenols	1.481	1.637	-10.5	81	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	83	0.00
22	Acetophenone	0.487	0.483	0.8	82	0.00
23 S	Nitrobenzene-d5	0.354	0.351	0.8	82	0.00
24	Nitrobenzene	0.346	0.339	2.0	82	0.00
25	Isophorone	0.660	0.679	-2.9	85	0.00
26 C	2-Nitrophenol	0.179	0.183	-2.2	83	0.00
27	2,4-Dimethylphenol	0.253	0.255	-0.8	83	0.00
28	bis(2-Chloroethoxy)methane	0.440	0.445	-1.1	83	0.00
29 C	2,4-Dichlorophenol	0.306	0.310	-1.3	84	0.00
30	1,2,4-Trichlorobenzene	0.351	0.345	1.7	82	0.00
31	Naphthalene	0.970	0.975	-0.5	83	0.00
32	Benzoic acid	0.135	0.122	9.6	76	0.00
33	4-Chloroaniline	0.441	0.457	-3.6	86	0.00
34 C	Hexachlorobutadiene	0.216	0.206	4.6	79	0.00
35	Caprolactam	0.112	0.121	-8.0	91	0.00
36 C	4-Chloro-3-methylphenol	0.321	0.337	-5.0	88	0.00
37	2-Methylnaphthalene	0.720	0.743	-3.2	86	0.00
38	1-Methylnaphthalene	0.679	0.706	-4.0	86	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	90	0.00
40	1,2,4,5-Tetrachlorobenzene	0.598	0.578	3.3	86	0.00
41 P	Hexachlorocyclopentadiene	0.287	0.252	12.2	76	0.00
42 S	2,4,6-Tribromophenol	0.235	0.245	-4.3	95	0.00
43 C	2,4,6-Trichlorophenol	0.387	0.389	-0.5	89	0.00
44	2,4,5-Trichlorophenol	0.395	0.401	-1.5	90	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA G\Data\BG021020\
 Data File : BG044356.D
 Acq On : 10 Feb 2020 13:20
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Feb 10 16:13:50 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG013020.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Feb 10 16:11:50 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
45 S	2-Fluorobiphenyl	1.194	1.188	0.5	88	0.00
46	1,1'-Biphenyl	1.443	1.430	0.9	89	0.00
47	2-Chloronaphthalene	1.148	1.123	2.2	88	0.00
48	2-Nitroaniline	0.339	0.339	0.0	91	0.00
49	Acenaphthylene	1.684	1.688	-0.2	90	0.00
50	Dimethylphthalate	1.438	1.461	-1.6	92	0.00
51	2,6-Dinitrotoluene	0.321	0.326	-1.6	93	0.00
52 C	Acenaphthene	1.091	1.091	0.0	91	0.00
53	3-Nitroaniline	0.355	0.363	-2.3	93	0.00
54 P	2,4-Dinitrophenol	0.159	0.178	-11.9	97	0.00
55	Dibenzofuran	1.652	1.673	-1.3	91	0.00
56 P	4-Nitrophenol	0.260	0.281	-8.1	96	0.00
57	2,4-Dinitrotoluene	0.449	0.460	-2.4	94	0.00
58	Fluorene	1.301	1.333	-2.5	93	0.00
59	2,3,4,6-Tetrachlorophenol	0.357	0.373	-4.5	94	0.00
60	Diethylphthalate	1.433	1.473	-2.8	94	0.00
61	4-Chlorophenyl-phenylether	0.706	0.710	-0.6	91	0.00
62	4-Nitroaniline	0.354	0.370	-4.5	97	0.00
63	Azobenzene	1.256	1.269	-1.0	92	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	95	0.00
65	4,6-Dinitro-2-methylphenol	0.110	0.115	-4.5	97	0.00
66 c	n-Nitrosodiphenylamine	0.560	0.562	-0.4	94	0.00
67	4-Bromophenyl-phenylether	0.218	0.219	-0.5	95	0.00
68	Hexachlorobenzene	0.244	0.245	-0.4	95	0.00
69	Atrazine	0.192	0.191	0.5	92	0.00
70 C	Pentachlorophenol	0.113	0.124	-9.7	98	0.00
71	Phenanthrene	0.968	0.988	-2.1	96	0.00
72	Anthracene	0.957	0.985	-2.9	97	0.00
73	Carbazole	0.883	0.912	-3.3	97	0.00
74	Di-n-butylphthalate	1.091	1.131	-3.7	96	0.00
75 C	Fluoranthene	1.120	1.151	-2.8	96	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	96	0.00
77	Benzidine	0.555	0.534	3.8	81	0.00
78	Pyrene	1.284	1.312	-2.2	96	0.00
79 S	Terphenyl-d14	0.972	0.958	1.4	100	0.00
80	Butylbenzylphthalate	0.585	0.589	-0.7	94	0.00
81	Benzo(a)anthracene	1.233	1.255	-1.8	96	0.00
82	3,3'-Dichlorobenzidine	0.478	0.495	-3.6	95	0.00
83	Chrysene	1.191	1.211	-1.7	96	0.00
84	Bis(2-ethylhexyl)phthalate	0.806	0.808	-0.2	94	0.00
85 c	Di-n-octyl phthalate	1.394	1.411	-1.2	94	0.00
86	Indeno(1,2,3-cd)pyrene	1.554	1.549	0.3	96	0.00
87 I	Perylene-d12	1.000	1.000	0.0	94	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA G\Data\BG021020\
Data File : BG044356.D
Acq On : 10 Feb 2020 13:20
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTDCCC040

Quant Time: Feb 10 16:13:50 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG013020.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Feb 10 16:11:50 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88	Benzo(b)fluoranthene	1.152	1.161	-0.8	95	0.00
89	Benzo(k)fluoranthene	1.123	1.164	-3.7	96	0.00
90 C	Benzo(a)pyrene	1.090	1.119	-2.7	96	0.00
91	Dibenzo(a,h)anthracene	1.078	1.094	-1.5	96	0.00
92	Benzo(q,h,i)perylene	1.080	1.095	-1.4	96	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0