


Data Path : Z:\SVOASRV\HPCHEM1\BNA G\DATA\BG021119\
 Data File : BG039430.D
 Acq On : 11 Feb 2019 12:20
 Operator : JU/SJ
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\8270-BG012919.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Tue Jan 29 15:41:51 2019



AutoFind: Scans 2526, 2527, 2528; Background Corrected with Scan 2519

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	34.9	23233	PASS
68	69	0.00	2	0.7	144	PASS
69	198	0.00	100	30.8	20487	PASS
70	69	0.00	2	0.5	112	PASS
127	198	10	80	41.5	27637	PASS
197	198	0.00	2	0.2	137	PASS
198	198	100	100	100.0	66578	PASS
199	198	5	9	7.0	4649	PASS
275	198	10	60	26.2	17446	PASS
365	198	1	100	3.4	2254	PASS
441	198	0.01	100	10.2	6773	PASS
442	198	50	100	65.2	43408	PASS
443	442	15	24	19.2	8336	PASS