

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG022222\
 Data File : BG052449.D
 Acq On : 22 Feb 2022 14:06
 Operator : CG/JU
 Sample : SST01008
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SST010408

Quant Time: Feb 22 15:25:23 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG022222.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Feb 22 15:23:39 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.222	152	26287	20.000	ng/u1	0.00
20) Naphthalene-d8	11.060	136	110953	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.866	164	77277	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.616	188	193876	20.000	ng/u1	# 0.00
79) Chrysene-d12	21.933	240	203405	20.000	ng/u1	# 0.00
88) Perylene-d12	25.393	264	204528	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.546	96	2972	4.561	ng/uL	0.00
4) Pyridine-d5	3.981	84	20142	9.938	ng/u1	0.00
7) Phenol-d5	7.371	99	26152	10.937	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.529	67	15470	10.650	ng/u1	0.00
11) 2-Chlorophenol-d4	7.752	132	18801	11.284	ng/u1	0.00
15) 4-Methylphenol-d8	8.927	113	19574	10.788	ng/u1	0.00
21) Nitrobenzene-d5	9.391	128	10410	11.523	ng/u1	0.00
24) 2-Nitrophenol-d4	10.126	143	10794	10.934	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.678	165	20213	10.503	ng/u1	0.00
31) 4-Chloroaniline-d4	11.189	131	27385	11.218	ng/u1	0.00
46) Dimethylphthalate-d6	14.255	166	68888	11.341	ng/u1	0.00
49) Acenaphthylene-d8	14.561	160	84410	11.149	ng/u1	0.00
54) 4-Nitrophenol-d4	15.054	143	9359	10.173	ng/u1	0.00
60) Fluorene-d10	15.853	176	58976	11.082	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.971	200	11211	9.366	ng/u1	0.00
73) Anthracene-d10	17.715	188	102642	11.230	ng/u1	0.00
81) Pyrene-d10	19.995	212	128053	10.504	ng/u1	0.00
92) Benzo(a)pyrene-d12	25.152	264	117836	10.845	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.588	88	3523	4.851	ng/uL#	90
5) Pyridine	3.999	79	20806	9.803	ng/u1#	84
6) Benzaldehyde	7.341	77	18833	13.893	ng/u1#	89
8) Phenol	7.394	94	26970	11.255	ng/u1	93
10) Bis(2-Chloroethyl)ether	7.623	93	19812	10.941	ng/u1	95
12) 2-Chlorophenol	7.782	128	19399	11.279	ng/u1	93
13) 2-Methylphenol	8.663	108	19461	10.791	ng/u1	97
14) 2,2'-oxybis(1-Chloropr...	8.745	45	28162	10.889	ng/u1#	95
16) Acetophenone	9.045	105	31804	11.303	ng/u1	94
17) N-Nitroso-di-n-propyla...	9.021	70	19494	11.549	ng/u1	96
18) 4-Methylphenol	8.992	108	20703	10.897	ng/u1	94
19) Hexachloroethane	9.321	117	7709	10.407	ng/u1#	81
22) Nitrobenzene	9.438	77	26251	10.418	ng/u1	96
23) Isophorone	9.961	82	49700	10.659	ng/u1#	99
25) 2-Nitrophenol	10.161	139	11674	10.817	ng/u1	93
26) 2,4-Dimethylphenol	10.208	107	24328	10.932	ng/u1	94
27) Bis(2-Chloroethoxy)met...	10.437	93	27312	10.922	ng/u1	97
29) 2,4-Dichlorophenol	10.707	162	19398	10.427	ng/u1	94
30) Naphthalene	11.113	128	67681	11.156	ng/u1#	94
32) 4-Chloroaniline	11.213	127	28473	11.313	ng/u1	94
33) Hexachlorobutadiene	11.400	225	15954	10.513	ng/u1	98
34) Caprolactam	11.953	113	6771	11.222	ng/u1	86
35) 4-Chloro-3-methylphenol	12.329	107	20896	10.414	ng/u1	99
36) 2-Methylnaphthalene	12.705	142	48307	11.347	ng/u1	90

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG022222\
 Data File : BG052449.D
 Acq On : 22 Feb 2022 14:06
 Operator : CG/JU
 Sample : SSTD01008
 Misc :
 ALS Vial : 4 Sample Multiplier: 1


Instrument :
 BNA_G
 ClientSampleId :
 SSTD010408


Quant Time: Feb 22 15:25:23 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG022222.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Feb 22 15:23:39 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.928	142	48969	11.181	ng/ul	97
39) 1,2,4,5-Tetrachloroben...	13.075	216	30025	10.794	ng/ul	99
40) Hexachlorocyclopentadiene	13.051	237	10083	6.350	ng/ul#	92
41) 2,4,6-Trichlorophenol	13.310	196	17483	10.527	ng/ul	98
42) 2,4,5-Trichlorophenol	13.386	196	18156	10.150	ng/ul	96
43) 1,1'-Biphenyl	13.697	154	67515	11.144	ng/ul#	94
44) 2-Chloronaphthalene	13.750	162	51428	11.095	ng/ul	96
45) 2-Nitroaniline	13.944	65	15938	10.212	ng/ul#	88
47) Dimethylphthalate	14.302	163	66885	11.360	ng/ul#	98
48) 2,6-Dinitrotoluene	14.426	165	14962	11.780	ng/ul#	91
50) Acenaphthylene	14.590	152	85949	11.326	ng/ul	96
51) 3-Nitroaniline	14.761	138	13336	12.269	ng/ul	96
52) Acenaphthene	14.931	153	55986	11.254	ng/ul	98
53) 2,4-Dinitrophenol	14.978	184	3561	5.399	ng/ul	88
55) 4-Nitrophenol	15.072	109	9510	9.863	ng/ul	86
56) Dibenzofuran	15.266	168	80348	11.155	ng/ul	100
57) 2,4-Dinitrotoluene	15.213	165	20842	11.447	ng/ul#	84
58) 2,3,4,6-Tetrachlorophenol	15.495	232	16579	10.588	ng/ul#	86
59) Diethylphthalate	15.659	149	67571	11.318	ng/ul	99
61) Fluorene	15.912	166	68556	11.469	ng/ul	98
62) 4-Chlorophenyl-phenyle...	15.894	204	37358	11.364	ng/ul	89
63) 4-Nitroaniline	15.918	138	14481	13.730	ng/ul	96
66) 4,6-Dinitro-2-methylph...	15.988	198	10979	9.631	ng/ul#	84
67) N-Nitrosodiphenylamine	16.106	169	59808	11.252	ng/ul	93
68) 4-Bromophenyl-phenylether	16.793	248	24357	10.407	ng/ul#	86
69) Hexachlorobenzene	16.922	284	27937	10.621	ng/ul#	88
70) Atrazine	17.046	200	26659	11.369	ng/ul	96
71) Pentachlorophenol	17.269	266	6323	5.014	ng/ul	99
72) Phenanthrene	17.657	178	113979	11.159	ng/ul	97
74) Anthracene	17.751	178	115126	11.420	ng/ul	95
75) 1,2,3,4-Tetrachloroben...	13.674	216	32984	10.190	ng/ul	96
76) Pentachlorobenzene	15.189	250	32273	10.808	ng/ul	95
77) Carbazole	18.015	167	103633	11.795	ng/ul	98
78) Di-n-butylphthalate	18.555	149	121349	11.515	ng/ul	99
80) Fluoranthene	19.660	202	150239	10.465	ng/ul#	90
82) Pyrene	20.024	202	152265	10.834	ng/ul#	91
83) Butylbenzylphthalate	20.894	149	54303	11.046	ng/ul#	90
84) 3,3'-Dichlorobenzidine	21.810	252	55850	12.823	ng/ul#	93
85) Benzo(a)anthracene	21.910	228	150835	11.025	ng/ul	97
86) Bis(2-ethylhexyl)phtha...	21.786	149	72063	10.226	ng/ul#	97
87) Chrysene	21.980	228	146330	11.320	ng/ul	97
89) Di-n-octyl phthalate	23.079	149	126355	10.889	ng/ul	100
90) Benzo(b)fluoranthene	24.289	252	157163	11.081	ng/ul#	95
91) Benzo(k)fluoranthene	24.359	252	142255	10.963	ng/ul#	98
93) Benzo(a)pyrene	25.223	252	149249	11.283	ng/ul#	94
94) Indeno(1,2,3-cd)pyrene	29.376	276	166157	11.277	ng/ul#	92
95) Dibenzo(a,h)anthracene	29.453	278	142678	11.678	ng/ul#	94
96) Benzo(g,h,i)perylene	30.633	276	138336	11.029	ng/ul#	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Feb 22 15:25:23 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG022222.M
Quant Title : SVOA CALIBRATION
QLast Update : Tue Feb 22 15:23:39 2022
Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 4.851 ng/uL
RT: 3.588 min Scan# 11
Delta R.T. 0.005 min
Lab File: BG052449.D
Acq: 22 Feb 2022 14:06




Instrument :
BNA_G
ClientSampleId :
SSTD010408


Tgt Ion: 88 Resp: 3523
Ion Ratio Lower Upper
88 100
43 37.6 23.0 34.4#
58 84.1 62.4 93.6



#5
Pyridine
Concen: 9.803 ng/uL
RT: 3.999 min Scan# 87
Delta R.T. 0.005 min
Lab File: BG052449.D
Acq: 22 Feb 2022 14:06

Tgt Ion: 79 Resp: 20806
Ion Ratio Lower Upper
79 100
52 92.4 60.8 91.2#
51 32.2 30.5 45.7









#6
Benzaldehyde
Concen: 13.893 ng/ul
RT: 7.341 min Scan# 61
Delta R.T. -0.001 min
Lab File: BG052449.D
Acq: 22 Feb 2022 14:06




Instrument :
BNA_G
ClientSampleId :
SSTD010408


Tgt Ion: 77 Resp: 18833
Ion Ratio Lower Upper
77 100
105 92.8 70.4 105.6
106 92.4 61.2 91.8#



#8
Phenol
Concen: 11.255 ng/ul
RT: 7.394 min Scan# 665
Delta R.T. -0.001 min
Lab File: BG052449.D
Acq: 22 Feb 2022 14:06

Tgt Ion: 94 Resp: 26970
Ion Ratio Lower Upper
94 100
65 33.2 25.1 37.7
66 45.8 31.6 47.4









#10
Bis(2-Chloroethyl)ether
Concen: 10.941 ng/ul
RT: 7.623 min Scan# 704
Delta R.T. -0.001 min
Lab File: BG052449.D
Acq: 22 Feb 2022 14:06




Instrument :
BNA_G
ClientSampleId :
SSTD010408


Tgt Ion: 93	Resp:	19812	
Ion	Ratio	Lower	Upper
93	100		
63	82.0	68.6	103.0
95	35.8	26.1	39.1



#12
2-Chlorophenol
Concen: 11.279 ng/ul
RT: 7.782 min Scan# 731
Delta R.T. -0.001 min
Lab File: BG052449.D
Acq: 22 Feb 2022 14:06

Tgt Ion:128	Resp:	19399	
Ion Ratio	Lower	Upper	
128 100			
64 49.6	45.3	67.9	
130 34.9	27.1	40.7	









#13
2-Methylphenol
Concen: 10.791 ng/ul
RT: 8.663 min Scan# 881
Delta R.T. -0.001 min
Lab File: BG052449.D
Acq: 22 Feb 2022 14:06




Instrument :
BNA_G
ClientSampleId :
SSTD010408


Tgt Ion:108 Resp: 19461
Ion Ratio Lower Upper
108 100
107 86.9 71.8 107.6



#14
2,2'-oxybis(1-Chloropropane)
Concen: 10.889 ng/ul
RT: 8.745 min Scan# 895
Delta R.T. -0.007 min
Lab File: BG052449.D
Acq: 22 Feb 2022 14:06

Tgt Ion: 45 Resp: 28162
Ion Ratio Lower Upper
45 100
77 17.1 10.7 16.1#
79 10.3 8.5 12.7









#16
Acetophenone
Concen: 11.303 ng/ul
RT: 9.045 min Scan# 946
Delta R.T. -0.001 min
Lab File: BG052449.D
Acq: 22 Feb 2022 14:06




Instrument :
BNA_G
ClientSampleId :
SSTD010408


Tgt Ion:105 Resp: 31804
Ion Ratio Lower Upper
105 100
77 89.4 67.3 100.9
51 33.2 24.0 36.0



#17
N-Nitroso-di-n-propylamine
Concen: 11.549 ng/ul
RT: 9.021 min Scan# 942
Delta R.T. -0.007 min
Lab File: BG052449.D
Acq: 22 Feb 2022 14:06

Tgt Ion: 70 Resp: 19494
Ion Ratio Lower Upper
70 100
42 53.9 46.0 69.0
101 10.4 7.0 10.4
130 15.7 13.4 20.2









#18
4-Methylphenol
Concen: 10.897 ng/ul
RT: 8.992 min Scan# 91
Delta R.T. -0.007 min
Lab File: BG052449.D
Acq: 22 Feb 2022 14:06




Instrument :
BNA_G
ClientSampleId :
SSTD010408


Tgt Ion:108 Resp: 20703
Ion Ratio Lower Upper
108 100
107 111.6 95.0 142.4



#19
Hexachloroethane
Concen: 10.407 ng/ul
RT: 9.321 min Scan# 993
Delta R.T. -0.007 min
Lab File: BG052449.D
Acq: 22 Feb 2022 14:06

Tgt Ion:117 Resp: 7709
Ion Ratio Lower Upper
117 100
201 120.6 77.2 115.8#
199 71.4 49.9 74.9





#22

Nitrobenzene

Concen: 10.418 ng/ul

RT: 9.438 min Scan# 1013

Delta R.T. -0.001 min

Lab File: BG052449.D


Acq: 22 Feb 2022 14:06

Instrument :

BNA_G

ClientSampleId :

SSTD010408






Tgt Ion: 77 Resp: 26251

Ion Ratio Lower Upper

77 100

123 35.2 30.1 45.1

65 13.7 9.6 14.4



#23

Isophorone

Concen: 10.659 ng/ul

RT: 9.961 min Scan# 1102

Delta R.T. -0.001 min

Lab File: BG052449.D

Acq: 22 Feb 2022 14:06




Tgt Ion: 82 Resp: 49700


Ion Ratio Lower Upper

82 100

95 8.4 5.6 8.4#

138 14.8 11.8 17.8









#25
2-Nitrophenol
Concen: 10.817 ng/ul
RT: 10.161 min Scan# 1136
Delta R.T. -0.001 min
Lab File: BG052449.D
Acq: 22 Feb 2022 14:06




Instrument :
BNA_G
ClientSampleId :
SSTD010408


Tgt Ion:139 Resp: 11674
Ion Ratio Lower Upper
139 100
65 54.6 48.0 72.0
109 32.1 28.6 42.8



#26
2,4-Dimethylphenol
Concen: 10.932 ng/ul
RT: 10.208 min Scan# 1144
Delta R.T. -0.007 min
Lab File: BG052449.D
Acq: 22 Feb 2022 14:06

Tgt Ion:107 Resp: 24328
Ion Ratio Lower Upper
107 100
121 44.3 39.3 58.9
122 84.0 63.7 95.5





#27

Bis(2-Chloroethoxy)methane

Concen: 10.922 ng/ul

RT: 10.437 min Scan# 1183

Delta R.T. -0.007 min

Lab File: BG052449.D

Acq: 22 Feb 2022 14:06

Instrument :

BNA_G

ClientSampleId :

SSTD010408





Tgt Ion: 93 Resp: 27312

Ion Ratio Lower Upper

93 100

95 31.7 26.5 39.7

123 9.1 8.1 12.1



#29

2,4-Dichlorophenol

Concen: 10.427 ng/ul

RT: 10.707 min Scan# 1229

Delta R.T. 0.005 min

Lab File: BG052449.D

Acq: 22 Feb 2022 14:06




Tgt Ion:162 Resp: 19398


Ion Ratio Lower Upper

162 100

164 66.6 49.4 74.0

98 34.1 30.4 45.6









#30
Naphthalene
Concen: 11.156 ng/ul
RT: 11.113 min Scan# 11
Delta R.T. -0.001 min
Lab File: BG052449.D
Acq: 22 Feb 2022 14:06




Instrument :
BNA_G
ClientSampleId :
SSTD010408


Tgt Ion:128 Resp: 67681
Ion Ratio Lower Upper
128 100
129 11.4 8.5 12.7
127 11.2 11.8 17.6#



#32
4-Chloroaniline
Concen: 11.313 ng/ul
RT: 11.213 min Scan# 1315
Delta R.T. -0.001 min
Lab File: BG052449.D
Acq: 22 Feb 2022 14:06

Tgt Ion:127 Resp: 28473
Ion Ratio Lower Upper
127 100
129 29.1 26.2 39.2









#33
Hexachlorobutadiene
Concen: 10.513 ng/ul
RT: 11.400 min Scan# 11
Delta R.T. 0.005 min
Lab File: BG052449.D
Acq: 22 Feb 2022 14:06




Instrument :
BNA_G
ClientSampleId :
SSTD010408


Tgt Ion:225 Resp: 15954
Ion Ratio Lower Upper
225 100
223 61.9 50.6 75.8
227 63.5 49.8 74.8



#34
Caprolactam
Concen: 11.222 ng/ul
RT: 11.953 min Scan# 1441
Delta R.T. -0.007 min
Lab File: BG052449.D
Acq: 22 Feb 2022 14:06

Tgt Ion:113 Resp: 6771
Ion Ratio Lower Upper
113 100
55 165.7 147.0 220.6
56 117.2 109.2 163.8









#35
 4-Chloro-3-methylphenol
 Concen: 10.414 ng/ul
 RT: 12.329 min Scan# 1505
 Delta R.T. -0.001 min
 Lab File: BG052449.D
 Acq: 22 Feb 2022 14:06




Instrument :
 BNA_G
 ClientSampleId :
 SSTD010408


Tgt Ion:	107	Resp:	20896
Ion	Ratio	Lower	Upper
107	100		
144	24.3	19.8	29.6
142	71.4	56.2	84.4



#36
 2-Methylnaphthalene
 Concen: 11.347 ng/ul
 RT: 12.705 min Scan# 1569
 Delta R.T. -0.001 min
 Lab File: BG052449.D
 Acq: 22 Feb 2022 14:06

Tgt Ion:	142	Resp:	48307
Ion	Ratio	Lower	Upper
142	100		
141	82.1	73.3	109.9









#37
1-Methylnaphthalene
Concen: 11.181 ng/ul
RT: 12.928 min Scan# 1
Delta R.T. 0.005 min
Lab File: BG052449.D
Acq: 22 Feb 2022 14:06




Instrument :
BNA_G
ClientSampleId :
SSTD010408


Tgt Ion:142 Resp: 48969
Ion Ratio Lower Upper
142 100
141 93.9 77.8 116.8
116 4.6 3.3 4.9



#39
1,2,4,5-Tetrachlorobenzene
Concen: 10.794 ng/ul
RT: 13.075 min Scan# 1632
Delta R.T. 0.005 min
Lab File: BG052449.D
Acq: 22 Feb 2022 14:06

Tgt Ion:216 Resp: 30025
Ion Ratio Lower Upper
216 100
214 75.3 61.0 91.4
179 20.6 16.2 24.2
108 14.5 13.1 19.7









#40
Hexachlorocyclopentadiene
Concen: 6.350 ng/ul
RT: 13.051 min Scan# 1628
Delta R.T. -0.001 min
Lab File: BG052449.D
Acq: 22 Feb 2022 14:06




Instrument :
BNA_G
ClientSampleId :
SSTD010408


Tgt Ion:237 Resp: 10083
Ion Ratio Lower Upper
237 100
235 56.6 49.4 74.0
272 8.6 11.1 16.7#



#41
2,4,6-Trichlorophenol
Concen: 10.527 ng/ul
RT: 13.310 min Scan# 1672
Delta R.T. -0.001 min
Lab File: BG052449.D
Acq: 22 Feb 2022 14:06

Tgt Ion:196 Resp: 17483
Ion Ratio Lower Upper
196 100
198 91.2 74.3 111.5
200 28.7 23.8 35.6









#42
2,4,5-Trichlorophenol
Concen: 10.150 ng/ul
RT: 13.386 min Scan# 1
Delta R.T. -0.001 min
Lab File: BG052449.D
Acq: 22 Feb 2022 14:06




Instrument :
BNA_G
ClientSampleId :
SSTD010408


Tgt Ion:196 Resp: 18156
Ion Ratio Lower Upper
196 100
198 99.4 77.0 115.4
200 29.2 25.4 38.2



#43
1,1'-Biphenyl
Concen: 11.144 ng/ul
RT: 13.697 min Scan# 1738
Delta R.T. -0.001 min
Lab File: BG052449.D
Acq: 22 Feb 2022 14:06

Tgt Ion:154 Resp: 67515
Ion Ratio Lower Upper
154 100
153 40.0 34.5 51.7
76 12.1 13.1 19.7#









#44
2-Chloronaphthalene
Concen: 11.095 ng/ul
RT: 13.750 min Scan# 1746
Delta R.T. 0.005 min
Lab File: BG052449.D
Acq: 22 Feb 2022 14:06




Instrument :
BNA_G
ClientSampleId :
SSTD010408


Tgt Ion:162 Resp: 51428
Ion Ratio Lower Upper
162 100
164 31.5 26.1 39.1
127 35.6 31.0 46.4



#45
2-Nitroaniline
Concen: 10.212 ng/ul
RT: 13.944 min Scan# 1780
Delta R.T. -0.001 min
Lab File: BG052449.D
Acq: 22 Feb 2022 14:06

Tgt Ion: 65 Resp: 15938
Ion Ratio Lower Upper
65 100
92 67.3 53.0 79.4
138 103.0 67.4 101.0#





#47

Dimethylphthalate

Concen: 11.360 ng/ul

RT: 14.302 min Scan# 1841

Delta R.T. -0.001 min

Lab File: BG052449.D

Acq: 22 Feb 2022 14:06

Instrument :

BNA_G

ClientSampleId :

SSTD010408





Tgt Ion:163 Resp: 66885

Ion Ratio Lower Upper

163 100

194 4.6 3.0 4.4#

164 9.5 8.2 12.4



#48

2,6-Dinitrotoluene

Concen: 11.780 ng/ul

RT: 14.426 min Scan# 1862

Delta R.T. -0.007 min

Lab File: BG052449.D

Acq: 22 Feb 2022 14:06




Tgt Ion:165 Resp: 14962


Ion Ratio Lower Upper

165 100

89 65.1 57.4 86.0

121 18.2 18.8 28.2#









#50
Acenaphthylene
Concen: 11.326 ng/ul
RT: 14.590 min Scan# 11
Delta R.T. -0.001 min
Lab File: BG052449.D
Acq: 22 Feb 2022 14:06




Instrument :
BNA_G
ClientSampleId :
SSTD010408


Tgt Ion:152 Resp: 85949
Ion Ratio Lower Upper
152 100
151 19.1 17.4 26.0
153 13.0 10.3 15.5



#51
3-Nitroaniline
Concen: 12.269 ng/ul
RT: 14.761 min Scan# 1919
Delta R.T. -0.001 min
Lab File: BG052449.D
Acq: 22 Feb 2022 14:06

Tgt Ion:138 Resp: 13336
Ion Ratio Lower Upper
138 100
108 10.8 8.8 13.2
92 128.0 98.0 147.0









#52
Acenaphthene
Concen: 11.254 ng/ul
RT: 14.931 min Scan# 1948
Delta R.T. -0.001 min
Lab File: BG052449.D
Acq: 22 Feb 2022 14:06




Instrument :
BNA_G
ClientSampleId :
SSTD010408


Tgt Ion:153 Resp: 55986
Ion Ratio Lower Upper
153 100
152 45.8 36.5 54.7
154 84.5 65.8 98.6



#53
2,4-Dinitrophenol
Concen: 5.399 ng/ul
RT: 14.978 min Scan# 1956
Delta R.T. 0.005 min
Lab File: BG052449.D
Acq: 22 Feb 2022 14:06

Tgt Ion:184 Resp: 3561
Ion Ratio Lower Upper
184 100
63 68.9 66.2 99.2
154 61.2 53.6 80.4









#55
4-Nitrophenol
Concen: 9.863 ng/ul
RT: 15.072 min Scan# 1972
Delta R.T. -0.001 min
Lab File: BG052449.D
Acq: 22 Feb 2022 14:06




Instrument :
BNA_G
ClientSampleId :
SSTD010408


Tgt Ion:109 Resp: 9510
Ion Ratio Lower Upper
109 100
139 102.0 88.7 133.1
65 119.7 113.6 170.4



#56
Dibenzofuran
Concen: 11.155 ng/ul
RT: 15.266 min Scan# 2005
Delta R.T. 0.005 min
Lab File: BG052449.D
Acq: 22 Feb 2022 14:06

Tgt Ion:168 Resp: 80348
Ion Ratio Lower Upper
168 100
139 39.1 31.2 46.8









#57
2,4-Dinitrotoluene
Concen: 11.447 ng/ul
RT: 15.213 min Scan# 1996
Delta R.T. -0.001 min
Lab File: BG052449.D
Acq: 22 Feb 2022 14:06




Instrument :
BNA_G
ClientSampleId :
SSTD010408


Tgt Ion:165 Resp: 20842
Ion Ratio Lower Upper
165 100
63 40.9 42.1 63.1#
182 3.3 1.8 2.8#



#58
2,3,4,6-Tetrachlorophenol
Concen: 10.588 ng/ul
RT: 15.495 min Scan# 2044
Delta R.T. -0.001 min
Lab File: BG052449.D
Acq: 22 Feb 2022 14:06

Tgt Ion:232 Resp: 16579
Ion Ratio Lower Upper
232 100
131 38.1 38.9 58.3#
130 2.2 2.4 3.6#
166 24.5 25.8 38.8#









#59
Diethylphthalate
Concen: 11.318 ng/ul
RT: 15.659 min Scan# 2072
Delta R.T. -0.001 min
Lab File: BG052449.D
Acq: 22 Feb 2022 14:06




Instrument :
BNA_G
ClientSampleId :
SSTD010408


Tgt Ion:149 Resp: 67571
Ion Ratio Lower Upper
149 100
177 21.2 16.8 25.2
150 12.2 10.1 15.1



#61
Fluorene
Concen: 11.469 ng/ul
RT: 15.912 min Scan# 2115
Delta R.T. -0.001 min
Lab File: BG052449.D
Acq: 22 Feb 2022 14:06

Tgt Ion:166 Resp: 68556
Ion Ratio Lower Upper
166 100
165 99.5 81.0 121.6
167 12.9 11.0 16.4





#62

4-Chlorophenyl-phenylether

Concen: 11.364 ng/ul

RT: 15.894 min Scan# 2112

Delta R.T. -0.001 min

Lab File: BG052449.D

Acq: 22 Feb 2022 14:06

Instrument :

BNA_G

ClientSampleId :

SSTD010408





Tgt Ion:204 Resp: 37358

Ion Ratio Lower Upper

204 100

206 32.3 27.4 41.2

141 49.3 48.6 73.0



#63

4-Nitroaniline

Concen: 13.730 ng/ul

RT: 15.918 min Scan# 2116

Delta R.T. -0.001 min

Lab File: BG052449.D

Acq: 22 Feb 2022 14:06




Tgt Ion:138 Resp: 14481


Ion Ratio Lower Upper

138 100

92 55.8 49.3 73.9

108 88.9 72.6 108.8









#66
 4,6-Dinitro-2-methylphenol
 Concen: 9.631 ng/ul
 RT: 15.988 min Scan# 2128
 Delta R.T. -0.001 min
 Lab File: BG052449.D
 Acq: 22 Feb 2022 14:06




Instrument :
 BNA_G
 ClientSampleId :
 SST010408


Tgt Ion:198 Resp: 10979
 Ion Ratio Lower Upper
 198 100
 182 2.3 3.8 5.8#
 77 19.5 23.2 34.8#



#67
 N-Nitrosodiphenylamine
 Concen: 11.252 ng/ul
 RT: 16.106 min Scan# 2148
 Delta R.T. -0.001 min
 Lab File: BG052449.D
 Acq: 22 Feb 2022 14:06

Tgt Ion:169 Resp: 59808
 Ion Ratio Lower Upper
 169 100
 168 65.2 58.5 87.7
 167 37.8 31.3 46.9









#68
4-Bromophenyl-phenylether
Concen: 10.407 ng/ul
RT: 16.793 min Scan# 21
Delta R.T. -0.001 min
Lab File: BG052449.D
Acq: 22 Feb 2022 14:06




Instrument :
BNA_G
ClientSampleId :
SSTD010408


Tgt Ion:248 Resp: 24357
Ion Ratio Lower Upper
248 100
250 100.6 75.5 113.3
141 64.4 67.6 101.4#



#69
Hexachlorobenzene
Concen: 10.621 ng/ul
RT: 16.922 min Scan# 2287
Delta R.T. -0.001 min
Lab File: BG052449.D
Acq: 22 Feb 2022 14:06

Tgt Ion:284 Resp: 27937
Ion Ratio Lower Upper
284 100
142 27.8 31.9 47.9#
249 31.0 25.8 38.6





#70

Atrazine

Concen: 11.369 ng/ul

RT: 17.046 min Scan# 21

Delta R.T. -0.001 min

Lab File: BG052449.D

Acq: 22 Feb 2022 14:06

Instrument :

BNA_G

ClientSampleId :

SSTD010408





Tgt Ion:200 Resp: 26659

Ion Ratio Lower Upper

200 100

173 23.1 20.2 30.2

215 45.0 38.1 57.1



#71

Pentachlorophenol

Concen: 5.014 ng/ul

RT: 17.269 min Scan# 2346

Delta R.T. -0.001 min

Lab File: BG052449.D

Acq: 22 Feb 2022 14:06




Tgt Ion:266 Resp: 6323


Ion Ratio Lower Upper

266 100

264 69.5 54.3 81.5

268 63.9 51.0 76.6









#72
Phenanthrene
Concen: 11.159 ng/ul
RT: 17.657 min Scan# 2412
Delta R.T. -0.001 min
Lab File: BG052449.D
Acq: 22 Feb 2022 14:06




Instrument :
BNA_G
ClientSampleId :
SSTD010408


Tgt Ion:178 Resp: 113979
Ion Ratio Lower Upper
178 100
179 14.7 12.3 18.5
176 17.9 15.5 23.3



#74
Anthracene
Concen: 11.420 ng/ul
RT: 17.751 min Scan# 2428
Delta R.T. -0.001 min
Lab File: BG052449.D
Acq: 22 Feb 2022 14:06

Tgt Ion:178 Resp: 115126
Ion Ratio Lower Upper
178 100
179 15.5 13.0 19.6
176 16.4 15.6 23.4









#75
1,2,3,4-Tetrachlorobenzene
Concen: 10.190 ng/uL
RT: 13.674 min Scan# 11
Delta R.T. -0.001 min
Lab File: BG052449.D
Acq: 22 Feb 2022 14:06




Instrument :
BNA_G
ClientSampleId :
SSTD010408


Tgt Ion:216 Resp: 32984
Ion Ratio Lower Upper
216 100
214 79.4 61.3 91.9
218 49.5 37.0 55.6



#76
Pentachlorobenzene
Concen: 10.808 ng/uL
RT: 15.189 min Scan# 1992
Delta R.T. -0.001 min
Lab File: BG052449.D
Acq: 22 Feb 2022 14:06

Tgt Ion:250 Resp: 32273
Ion Ratio Lower Upper
250 100
248 61.6 50.5 75.7
252 57.1 51.2 76.8









#77
 Carbazole
 Concen: 11.795 ng/ul
 RT: 18.015 min Scan# 2473
 Delta R.T. -0.001 min
 Lab File: BG052449.D
 Acq: 22 Feb 2022 14:06




Instrument :
 BNA_G
 ClientSampleId :
 SSTD010408


Tgt Ion:167 Resp: 103633
 Ion Ratio Lower Upper
 167 100
 166 22.9 17.3 25.9
 139 13.3 10.6 16.0



#78
 Di-n-butylphthalate
 Concen: 11.515 ng/ul
 RT: 18.555 min Scan# 2565
 Delta R.T. -0.001 min
 Lab File: BG052449.D
 Acq: 22 Feb 2022 14:06

Tgt Ion:149 Resp: 121349
 Ion Ratio Lower Upper
 149 100
 150 9.3 7.5 11.3
 104 5.7 5.0 7.6









#80
Fluoranthene
Concen: 10.465 ng/ul
RT: 19.660 min Scan# 21
Delta R.T. -0.001 min
Lab File: BG052449.D
Acq: 22 Feb 2022 14:06




Instrument :
BNA_G
ClientSampleId :
SSTD010408


Tgt Ion:202 Resp: 150239
Ion Ratio Lower Upper
202 100
101 7.1 9.3 13.9#
100 6.4 7.0 10.6#



#82
Pyrene
Concen: 10.834 ng/ul
RT: 20.024 min Scan# 2815
Delta R.T. -0.001 min
Lab File: BG052449.D
Acq: 22 Feb 2022 14:06

Tgt Ion:202 Resp: 152265
Ion Ratio Lower Upper
202 100
101 8.8 10.7 16.1#
100 8.5 9.0 13.6#





#83

Butylbenzylphthalate

Concen: 11.046 ng/ul

RT: 20.894 min Scan# 2963

Delta R.T. 0.005 min

Lab File: BG052449.D

Acq: 22 Feb 2022 14:06

Instrument :

BNA_G

ClientSampleId :

SSTD010408





Tgt Ion:149 Resp: 54303

Ion Ratio Lower Upper

149 100

91 73.1 65.2 97.8

206 26.3 16.2 24.2#



#84

3,3'-Dichlorobenzidine

Concen: 12.823 ng/ul

RT: 21.810 min Scan# 3119

Delta R.T. -0.001 min

Lab File: BG052449.D

Acq: 22 Feb 2022 14:06




Tgt Ion:252 Resp: 55850


Ion Ratio Lower Upper

252 100

254 68.3 50.8 76.2

126 7.8 9.0 13.4#









#85
Benzo(a)anthracene
Concen: 11.025 ng/ul
RT: 21.910 min Scan# 3136
Delta R.T. -0.001 min
Lab File: BG052449.D
Acq: 22 Feb 2022 14:06




Instrument :
BNA_G
ClientSampleId :
SSTD010408


Tgt Ion:228 Resp: 150835
Ion Ratio Lower Upper
228 100
229 20.5 15.7 23.5
226 29.6 22.3 33.5



#86
Bis(2-ethylhexyl)phthalate
Concen: 10.226 ng/ul
RT: 21.786 min Scan# 3115
Delta R.T. -0.001 min
Lab File: BG052449.D
Acq: 22 Feb 2022 14:06

Tgt Ion:149 Resp: 72063
Ion Ratio Lower Upper
149 100
167 27.5 21.1 31.7
279 5.1 2.7 4.1#





#87

Chrysene

Concen: 11.320 ng/ul

RT: 21.980 min Scan# 3148

Delta R.T. -0.001 min

Lab File: BG052449.D

Acq: 22 Feb 2022 14:06

Instrument :

BNA_G

ClientSampleId :

SSTD010408





Tgt Ion:228 Resp: 146330

Ion Ratio Lower Upper

228 100

226 28.4 24.8 37.2

229 20.2 15.8 23.6



#89

Di-n-octyl phthalate

Concen: 10.889 ng/ul




RT: 23.079 min Scan# 3335


Delta R.T. -0.001 min

Lab File: BG052449.D

Acq: 22 Feb 2022 14:06

Tgt Ion:149 Resp: 126355









#90
Benzo(b)fluoranthene
Concen: 11.081 ng/ul
RT: 24.289 min Scan# 3540
Delta R.T. 0.005 min
Lab File: BG052449.D
Acq: 22 Feb 2022 14:06




Instrument :
BNA_G
ClientSampleId :
SSTD010408


Tgt Ion:252 Resp: 157163
Ion Ratio Lower Upper
252 100
253 21.6 17.8 26.6
125 5.3 8.5 12.7#



#91
Benzo(k)fluoranthene
Concen: 10.963 ng/ul
RT: 24.359 min Scan# 3553
Delta R.T. -0.001 min
Lab File: BG052449.D
Acq: 22 Feb 2022 14:06

Tgt Ion:252 Resp: 142255
Ion Ratio Lower Upper
252 100
253 21.5 17.3 25.9
125 7.2 7.8 11.6#









#93
Benzo(a)pyrene
Concen: 11.283 ng/ul
RT: 25.223 min Scan# 31
Delta R.T. -0.013 min
Lab File: BG052449.D
Acq: 22 Feb 2022 14:06




Instrument :
BNA_G
ClientSampleId :
SSTD010408


Tgt Ion:252 Resp: 149249
Ion Ratio Lower Upper
252 100
253 20.7 17.4 26.2
125 6.7 9.1 13.7#



#94
Indeno(1,2,3-cd)pyrene
Concen: 11.277 ng/ul
RT: 29.376 min Scan# 4407
Delta R.T. -0.013 min
Lab File: BG052449.D
Acq: 22 Feb 2022 14:06

Tgt Ion:276 Resp: 166157
Ion Ratio Lower Upper
276 100
138 11.7 15.5 23.3#
277 25.2 20.5 30.7









#95
 Dibenzo(a,h)anthracene
 Concen: 11.678 ng/ul
 RT: 29.453 min Scan# 4421
 Delta R.T. -0.007 min
 Lab File: BG052449.D
 Acq: 22 Feb 2022 14:06

Instrument :
 BNA_G
 ClientSampleId :
 SST010408

Tgt Ion:278 Resp: 142678
 Ion Ratio Lower Upper
 278 100
 139 9.9 12.2 18.4#
 279 25.5 19.5 29.3



#96
 Benzo(g,h,i)perylene
 Concen: 11.029 ng/ul
 RT: 30.633 min Scan# 4621
 Delta R.T. -0.007 min
 Lab File: BG052449.D
 Acq: 22 Feb 2022 14:06

Tgt Ion:276 Resp: 138336
 Ion Ratio Lower Upper
 276 100
 138 11.8 16.6 24.8#
 277 27.7 17.6 26.4#

