

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG022222\
 Data File : BG052450.D
 Acq On : 22 Feb 2022 14:44
 Operator : CG/JU
 Sample : SST02009
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SST020409

Quant Time: Feb 22 15:25:58 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG022222.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Feb 22 15:23:39 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.223	152	25910	20.000	ng/u1	0.00
20) Naphthalene-d8	11.061	136	109230	20.000	ng/u1	# 0.00
38) Acenaphthene-d10	14.867	164	78892	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.616	188	200265	20.000	ng/u1	# 0.00
79) Chrysene-d12	21.934	240	197389	20.000	ng/u1	# 0.00
88) Perylene-d12	25.394	264	199452	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.541	96	5402	8.411	ng/uL	0.00
4) Pyridine-d5	3.976	84	39568	19.808	ng/u1	0.00
7) Phenol-d5	7.366	99	49863	21.157	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.530	67	30859	21.553	ng/u1	0.00
11) 2-Chlorophenol-d4	7.747	132	36275	22.088	ng/u1	0.00
15) 4-Methylphenol-d8	8.934	113	39352	22.003	ng/u1	0.00
21) Nitrobenzene-d5	9.398	128	20129	22.632	ng/u1	0.00
24) 2-Nitrophenol-d4	10.127	143	22117	22.757	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.673	165	41471	21.889	ng/u1	0.00
31) 4-Chloroaniline-d4	11.190	131	59789	24.878	ng/u1	0.00
46) Dimethylphthalate-d6	14.256	166	142754	23.021	ng/u1	0.00
49) Acenaphthylene-d8	14.562	160	172802	22.357	ng/u1	0.00
54) 4-Nitrophenol-d4	15.055	143	21760	23.169	ng/u1	0.00
60) Fluorene-d10	15.854	176	123911	22.808	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.972	200	26178	21.173	ng/u1	0.00
73) Anthracene-d10	17.716	188	208167	22.049	ng/u1	0.00
81) Pyrene-d10	19.996	212	254628	21.524	ng/u1	0.00
92) Benzo(a)pyrene-d12	25.159	264	232871	21.977	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.582	88	6023	8.414	ng/uL	95
5) Pyridine	3.994	79	40532	19.375	ng/u1	99
6) Benzaldehyde	7.342	77	31669	23.702	ng/u1#	89
8) Phenol	7.395	94	51223	21.687	ng/u1	98
10) Bis(2-Chloroethyl)ether	7.624	93	39492	22.126	ng/u1	94
12) 2-Chlorophenol	7.783	128	37771	22.280	ng/u1	96
13) 2-Methylphenol	8.664	108	39160	22.031	ng/u1	94
14) 2,2'-oxybis(1-Chloropr...	8.752	45	54398	21.339	ng/u1	96
16) Acetophenone	9.046	105	63236	22.800	ng/u1	96
17) N-Nitroso-di-n-propyla...	9.028	70	38140	22.925	ng/u1	99
18) 4-Methylphenol	8.999	108	42149	22.507	ng/u1	88
19) Hexachloroethane	9.328	117	15244	20.879	ng/u1	86
22) Nitrobenzene	9.439	77	51769	20.869	ng/u1	94
23) Isophorone	9.962	82	102262	22.278	ng/u1	96
25) 2-Nitrophenol	10.162	139	23961	22.552	ng/u1	92
26) 2,4-Dimethylphenol	10.215	107	48995	22.365	ng/u1	98
27) Bis(2-Chloroethoxy)met...	10.444	93	55506	22.547	ng/u1	97
29) 2,4-Dichlorophenol	10.702	162	39528	21.582	ng/u1	96
30) Naphthalene	11.113	128	132198	22.134	ng/u1	95
32) 4-Chloroaniline	11.213	127	61869	24.969	ng/u1	99
33) Hexachlorobutadiene	11.395	225	30278	20.266	ng/u1	97
34) Caprolactam	11.959	113	13592	22.882	ng/u1	88
35) 4-Chloro-3-methylphenol	12.329	107	45763	23.166	ng/u1	87
36) 2-Methylnaphthalene	12.705	142	95614	22.814	ng/u1	96


Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG022222\
 Data File : BG052450.D
 Acq On : 22 Feb 2022 14:44
 Operator : CG/JU
 Sample : SSTD02009
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SSTD020409

Quant Time: Feb 22 15:25:58 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG022222.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Feb 22 15:23:39 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.923	142	97570	22.629	ng/ul	100
39) 1,2,4,5-Tetrachloroben...	13.070	216	60776	21.402	ng/ul	96
40) Hexachlorocyclopentadiene	13.052	237	24859	15.334	ng/ul	99
41) 2,4,6-Trichlorophenol	13.310	196	37436	22.079	ng/ul	98
42) 2,4,5-Trichlorophenol	13.387	196	38680	21.181	ng/ul	97
43) 1,1'-Biphenyl	13.698	154	137663	22.258	ng/ul#	94
44) 2-Chloronaphthalene	13.745	162	104372	22.056	ng/ul	97
45) 2-Nitroaniline	13.945	65	34613	21.723	ng/ul	92
47) Dimethylphthalate	14.303	163	141657	23.568	ng/ul	99
48) 2,6-Dinitrotoluene	14.432	165	30304	23.370	ng/ul#	85
50) Acenaphthylene	14.591	152	173991	22.458	ng/ul	97
51) 3-Nitroaniline	14.761	138	20672	18.629	ng/ul	93
52) Acenaphthene	14.932	153	112842	22.218	ng/ul	96
53) 2,4-Dinitrophenol	14.973	184	12175	18.081	ng/ul#	82
55) 4-Nitrophenol	15.073	109	21369	21.707	ng/ul	89
56) Dibenzofuran	15.261	168	165735	22.538	ng/ul	100
57) 2,4-Dinitrotoluene	15.214	165	45145	24.287	ng/ul	87
58) 2,3,4,6-Tetrachlorophenol	15.496	232	35656	22.305	ng/ul	90
59) Diethylphthalate	15.660	149	140952	23.127	ng/ul	95
61) Fluorene	15.913	166	137702	22.566	ng/ul	99
62) 4-Chlorophenyl-phenyle...	15.895	204	75459	22.485	ng/ul	93
63) 4-Nitroaniline	15.919	138	16998	15.787	ng/ul	94
66) 4,6-Dinitro-2-methylph...	15.989	198	25355	21.533	ng/ul#	90
67) N-Nitrosodiphenylamine	16.107	169	121775	22.180	ng/ul	94
68) 4-Bromophenyl-phenylether	16.794	248	52147	21.570	ng/ul#	88
69) Hexachlorobenzene	16.923	284	59562	21.922	ng/ul#	87
70) Atrazine	17.047	200	53445	22.065	ng/ul	97
71) Pentachlorophenol	17.270	266	23505	18.043	ng/ul	97
72) Phenanthrene	17.657	178	232253	22.014	ng/ul	98
74) Anthracene	17.752	178	234146	22.486	ng/ul	98
75) 1,2,3,4-Tetrachloroben...	13.675	216	67357	20.146	ng/ul	99
76) Pentachlorobenzene	15.190	250	64074	20.774	ng/ul	95
77) Carbazole	18.016	167	211446	23.298	ng/ul	98
78) Di-n-butylphthalate	18.556	149	247776	22.762	ng/ul	99
80) Fluoranthene	19.661	202	303222	21.766	ng/ul#	92
82) Pyrene	20.025	202	297512	21.815	ng/ul#	90
83) Butylbenzylphthalate	20.888	149	104990	22.006	ng/ul	93
84) 3,3'-Dichlorobenzidine	21.811	252	86724	20.518	ng/ul#	98
85) Benzo(a)anthracene	21.911	228	289203	21.784	ng/ul	97
86) Bis(2-ethylhexyl)phtha...	21.787	149	148093	21.655	ng/ul#	99
87) Chrysene	21.981	228	281140	22.412	ng/ul	97
89) Di-n-octyl phthalate	23.080	149	246522	21.786	ng/ul	100
90) Benzo(b)fluoranthene	24.284	252	299801	21.676	ng/ul#	96
91) Benzo(k)fluoranthene	24.360	252	277684	21.944	ng/ul#	96
93) Benzo(a)pyrene	25.235	252	283704	21.994	ng/ul#	95
94) Indeno(1,2,3-cd)pyrene	29.389	276	322021	22.411	ng/ul#	93
95) Dibenzo(a,h)anthracene	29.459	278	275590	23.131	ng/ul#	93
96) Benzo(g,h,i)perylene	30.628	276	135823	11.105	ng/ul#	91





(#) = qualifier out of range (m) = manual integration (+) = signals summed



#2
1,4-Dioxane
Concen: 8.414 ng/uL
RT: 3.582 min Scan# 1
Delta R.T. 0.000 min
Lab File: BG052450.D
Acq: 22 Feb 2022 14:44




Instrument :
BNA_G
ClientSampleId :
SSTD020409


Tgt Ion: 88 Resp: 6023
Ion Ratio Lower Upper
88 100
43 32.7 23.0 34.4
58 75.0 62.4 93.6



#5
Pyridine
Concen: 19.375 ng/uL
RT: 3.994 min Scan# 86
Delta R.T. 0.000 min
Lab File: BG052450.D
Acq: 22 Feb 2022 14:44

Tgt Ion: 79 Resp: 40532
Ion Ratio Lower Upper
79 100
52 76.0 60.8 91.2
51 36.2 30.5 45.7









#6
Benzaldehyde
Concen: 23.702 ng/ul
RT: 7.342 min Scan# 656
Delta R.T. 0.000 min
Lab File: BG052450.D
Acq: 22 Feb 2022 14:44




Instrument :
BNA_G
ClientSampleId :
SSTD020409


Tgt Ion: 77 Resp: 31669
Ion Ratio Lower Upper
77 100
105 92.4 70.4 105.6
106 92.5 61.2 91.8#



#8
Phenol
Concen: 21.687 ng/ul
RT: 7.395 min Scan# 665
Delta R.T. 0.000 min
Lab File: BG052450.D
Acq: 22 Feb 2022 14:44

Tgt Ion: 94 Resp: 51223
Ion Ratio Lower Upper
94 100
65 30.3 25.1 37.7
66 38.4 31.6 47.4









#10
Bis(2-Chloroethyl)ether
Concen: 22.126 ng/ul
RT: 7.624 min Scan# 704
Delta R.T. 0.000 min
Lab File: BG052450.D
Acq: 22 Feb 2022 14:44




Instrument :
BNA_G
ClientSampleId :
SSTD020409


Tgt Ion: 93 Resp: 39492
Ion Ratio Lower Upper
93 100
63 79.4 68.6 103.0
95 33.7 26.1 39.1



#12
2-Chlorophenol
Concen: 22.280 ng/ul
RT: 7.783 min Scan# 731
Delta R.T. 0.000 min
Lab File: BG052450.D
Acq: 22 Feb 2022 14:44

Tgt Ion: 128 Resp: 37771
Ion Ratio Lower Upper
128 100
64 54.0 45.3 67.9
130 30.7 27.1 40.7









#13
2-Methylphenol
Concen: 22.031 ng/ul
RT: 8.664 min Scan# 881
Delta R.T. 0.000 min
Lab File: BG052450.D
Acq: 22 Feb 2022 14:44




Instrument :
BNA_G
ClientSampleId :
SSTD020409


Tgt Ion:108 Resp: 39160
Ion Ratio Lower Upper
108 100
107 84.3 71.8 107.6



#14
2,2'-oxybis(1-Chloropropane)
Concen: 21.339 ng/ul
RT: 8.752 min Scan# 896
Delta R.T. 0.000 min
Lab File: BG052450.D
Acq: 22 Feb 2022 14:44

Tgt Ion: 45 Resp: 54398
Ion Ratio Lower Upper
45 100
77 14.8 10.7 16.1
79 12.3 8.5 12.7









#16
Acetophenone
Concen: 22.800 ng/ul
RT: 9.046 min Scan# 946
Delta R.T. 0.000 min
Lab File: BG052450.D
Acq: 22 Feb 2022 14:44




Instrument :
BNA_G
ClientSampleId :
SSTD020409


Tgt Ion:105 Resp: 63236
Ion Ratio Lower Upper
105 100
77 87.9 67.3 100.9
51 30.9 24.0 36.0



#17
N-Nitroso-di-n-propylamine
Concen: 22.925 ng/ul
RT: 9.028 min Scan# 943
Delta R.T. 0.000 min
Lab File: BG052450.D
Acq: 22 Feb 2022 14:44

Tgt Ion: 70 Resp: 38140
Ion Ratio Lower Upper
70 100
42 57.1 46.0 69.0
101 9.5 7.0 10.4
130 17.0 13.4 20.2









#18
4-Methylphenol
Concen: 22.507 ng/ul
RT: 8.999 min Scan# 911
Delta R.T. 0.000 min
Lab File: BG052450.D
Acq: 22 Feb 2022 14:44




Instrument :
BNA_G
ClientSampleId :
SSTD020409


Tgt Ion:108 Resp: 42149
Ion Ratio Lower Upper
108 100
107 105.5 95.0 142.4



#19
Hexachloroethane
Concen: 20.879 ng/ul
RT: 9.328 min Scan# 994
Delta R.T. 0.000 min
Lab File: BG052450.D
Acq: 22 Feb 2022 14:44

Tgt Ion:117 Resp: 15244
Ion Ratio Lower Upper
117 100
201 110.2 77.2 115.8
199 74.1 49.9 74.9





#22

Nitrobenzene

Concen: 20.869 ng/ul

RT: 9.439 min Scan# 1013

Delta R.T. 0.000 min

Lab File: BG052450.D

Acq: 22 Feb 2022 14:44

Instrument :

BNA_G

ClientSampleId :

SSTD020409





Tgt Ion: 77 Resp: 51769

Ion Ratio Lower Upper

77 100

123 41.4 30.1 45.1

65 13.9 9.6 14.4



#23

Isophorone

Concen: 22.278 ng/ul

RT: 9.962 min Scan# 1102

Delta R.T. 0.000 min

Lab File: BG052450.D

Acq: 22 Feb 2022 14:44




Tgt Ion: 82 Resp: 102262


Ion Ratio Lower Upper

82 100

95 8.1 5.6 8.4

138 16.7 11.8 17.8









#25
2-Nitrophenol
Concen: 22.552 ng/ul
RT: 10.162 min Scan# 1136
Delta R.T. 0.000 min
Lab File: BG052450.D
Acq: 22 Feb 2022 14:44




Instrument :
BNA_G
ClientSampleId :
SSTD020409


Tgt Ion:139	Resp:	23961
Ion Ratio	Lower	Upper
139	100	
65	53.8	48.0 72.0
109	32.0	28.6 42.8



#26
2,4-Dimethylphenol
Concen: 22.365 ng/ul
RT: 10.215 min Scan# 1145
Delta R.T. 0.000 min
Lab File: BG052450.D
Acq: 22 Feb 2022 14:44

Tgt Ion:107	Resp:	48995
Ion Ratio	Lower	Upper
107	100	
121	47.4	39.3 58.9
122	78.7	63.7 95.5









#27
 Bis(2-Chloroethoxy)methane
 Concen: 22.547 ng/ul
 RT: 10.444 min Scan# 1184
 Delta R.T. 0.000 min
 Lab File: BG052450.D
 Acq: 22 Feb 2022 14:44




Instrument :
 BNA_G
 ClientSampleId :
 SSTD020409


Tgt Ion: 93 Resp: 55506
 Ion Ratio Lower Upper
 93 100
 95 31.7 26.5 39.7
 123 11.8 8.1 12.1



#29
 2,4-Dichlorophenol
 Concen: 21.582 ng/ul
 RT: 10.702 min Scan# 1228
 Delta R.T. 0.000 min
 Lab File: BG052450.D
 Acq: 22 Feb 2022 14:44

Tgt Ion: 162 Resp: 39528
 Ion Ratio Lower Upper
 162 100
 164 65.9 49.4 74.0
 98 36.4 30.4 45.6









#30
Naphthalene
Concen: 22.134 ng/ul
RT: 11.113 min Scan# 1298
Delta R.T. 0.000 min
Lab File: BG052450.D
Acq: 22 Feb 2022 14:44




Instrument :
BNA_G
ClientSampleId :
SSTD020409


Tgt Ion:128 Resp: 132198
Ion Ratio Lower Upper
128 100
129 12.1 8.5 12.7
127 12.6 11.8 17.6



#32
4-Chloroaniline
Concen: 24.969 ng/ul
RT: 11.213 min Scan# 1315
Delta R.T. 0.000 min
Lab File: BG052450.D
Acq: 22 Feb 2022 14:44

Tgt Ion:127 Resp: 61869
Ion Ratio Lower Upper
127 100
129 32.2 26.2 39.2









#33
Hexachlorobutadiene
Concen: 20.266 ng/ul
RT: 11.395 min Scan# 1134
Delta R.T. 0.000 min
Lab File: BG052450.D
Acq: 22 Feb 2022 14:44




Instrument :
BNA_G
ClientSampleId :
SSTD020409


Tgt Ion:225 Resp: 30278
Ion Ratio Lower Upper
225 100
223 65.5 50.6 75.8
227 64.2 49.8 74.8



#34
Caprolactam
Concen: 22.882 ng/ul
RT: 11.959 min Scan# 1442
Delta R.T. 0.000 min
Lab File: BG052450.D
Acq: 22 Feb 2022 14:44

Tgt Ion:113 Resp: 13592
Ion Ratio Lower Upper
113 100
55 201.0 147.0 220.6
56 150.0 109.2 163.8









#35
 4-Chloro-3-methylphenol
 Concen: 23.166 ng/ul
 RT: 12.329 min Scan# 1505
 Delta R.T. 0.000 min
 Lab File: BG052450.D
 Acq: 22 Feb 2022 14:44




Instrument :
 BNA_G
 ClientSampleId :
 SST020409


Tgt Ion:	107	Resp:	45763
Ion	Ratio	Lower	Upper
107	100		
144	27.3	19.8	29.6
142	83.8	56.2	84.4



#36
 2-Methylnaphthalene
 Concen: 22.814 ng/ul
 RT: 12.705 min Scan# 1569
 Delta R.T. 0.000 min
 Lab File: BG052450.D
 Acq: 22 Feb 2022 14:44

Tgt Ion:	142	Resp:	95614
Ion	Ratio	Lower	Upper
142	100		
141	88.1	73.3	109.9









#37
1-Methylnaphthalene
Concen: 22.629 ng/ul
RT: 12.923 min Scan# 1606
Delta R.T. 0.000 min
Lab File: BG052450.D
Acq: 22 Feb 2022 14:44




Instrument :
BNA_G
ClientSampleId :
SSTD020409


Tgt Ion:142 Resp: 97570
Ion Ratio Lower Upper
142 100
141 96.9 77.8 116.8
116 3.9 3.3 4.9



#39
1,2,4,5-Tetrachlorobenzene
Concen: 21.402 ng/ul
RT: 13.070 min Scan# 1631
Delta R.T. 0.000 min
Lab File: BG052450.D
Acq: 22 Feb 2022 14:44

Tgt Ion:216 Resp: 60776
Ion Ratio Lower Upper
216 100
214 80.1 61.0 91.4
179 18.7 16.2 24.2
108 15.0 13.1 19.7









#40
Hexachlorocyclopentadiene
Concen: 15.334 ng/ul
RT: 13.052 min Scan# 1628
Delta R.T. 0.000 min
Lab File: BG052450.D
Acq: 22 Feb 2022 14:44




Instrument :
BNA_G
ClientSampleId :
SSTD020409


Tgt Ion:	237	Resp:	24859
Ion Ratio	Lower	Upper	
237	100		
235	62.9	49.4	74.0
272	13.5	11.1	16.7



#41
2,4,6-Trichlorophenol
Concen: 22.079 ng/ul
RT: 13.310 min Scan# 1672
Delta R.T. 0.000 min
Lab File: BG052450.D
Acq: 22 Feb 2022 14:44

Tgt Ion:	196	Resp:	37436
Ion Ratio	Lower	Upper	
196	100		
198	91.3	74.3	111.5
200	28.9	23.8	35.6









#42
2,4,5-Trichlorophenol
Concen: 21.181 ng/ul
RT: 13.387 min Scan# 1685
Delta R.T. 0.000 min
Lab File: BG052450.D
Acq: 22 Feb 2022 14:44




Instrument :
BNA_G
ClientSampleId :
SSTD020409


Tgt Ion:196 Resp: 38680
Ion Ratio Lower Upper
196 100
198 93.9 77.0 115.4
200 33.8 25.4 38.2



#43
1,1'-Biphenyl
Concen: 22.258 ng/ul
RT: 13.698 min Scan# 1738
Delta R.T. 0.000 min
Lab File: BG052450.D
Acq: 22 Feb 2022 14:44

Tgt Ion:154 Resp: 137663
Ion Ratio Lower Upper
154 100
153 40.0 34.5 51.7
76 11.9 13.1 19.7#









#44
2-Chloronaphthalene
Concen: 22.056 ng/ul
RT: 13.745 min Scan# 1746
Delta R.T. 0.000 min
Lab File: BG052450.D
Acq: 22 Feb 2022 14:44




Instrument :
BNA_G
ClientSampleId :
SSTD020409


Tgt Ion:162 Resp: 104372
Ion Ratio Lower Upper
162 100
164 32.0 26.1 39.1
127 36.4 31.0 46.4



#45
2-Nitroaniline
Concen: 21.723 ng/ul
RT: 13.945 min Scan# 1780
Delta R.T. 0.000 min
Lab File: BG052450.D
Acq: 22 Feb 2022 14:44

Tgt Ion: 65 Resp: 34613
Ion Ratio Lower Upper
65 100
92 62.9 53.0 79.4
138 95.0 67.4 101.0





#47

Dimethylphthalate

Concen: 23.568 ng/ul

RT: 14.303 min Scan# 1841

Delta R.T. 0.000 min

Lab File: BG052450.D

Acq: 22 Feb 2022 14:44

Instrument :

BNA_G

ClientSampleId :

SSTD020409





Tgt Ion:163 Resp: 141657

Ion Ratio Lower Upper

163 100

194 3.6 3.0 4.4

164 9.7 8.2 12.4



#48

2,6-Dinitrotoluene

Concen: 23.370 ng/ul

RT: 14.432 min Scan# 1863

Delta R.T. 0.000 min

Lab File: BG052450.D

Acq: 22 Feb 2022 14:44




Tgt Ion:165 Resp: 30304


Ion Ratio Lower Upper

165 100

89 55.8 57.4 86.0#

121 22.2 18.8 28.2









#50
Acenaphthylene
Concen: 22.458 ng/ul
RT: 14.591 min Scan# 1890
Delta R.T. 0.000 min
Lab File: BG052450.D
Acq: 22 Feb 2022 14:44




Instrument :
BNA_G
ClientSampleId :
SSTD020409


Tgt Ion:152 Resp: 173991
Ion Ratio Lower Upper
152 100
151 19.8 17.4 26.0
153 13.4 10.3 15.5



#51
3-Nitroaniline
Concen: 18.629 ng/ul
RT: 14.761 min Scan# 1919
Delta R.T. 0.000 min
Lab File: BG052450.D
Acq: 22 Feb 2022 14:44

Tgt Ion:138 Resp: 20672
Ion Ratio Lower Upper
138 100
108 11.2 8.8 13.2
92 114.1 98.0 147.0









#52
Acenaphthene
Concen: 22.218 ng/ul
RT: 14.932 min Scan# 1948
Delta R.T. 0.000 min
Lab File: BG052450.D
Acq: 22 Feb 2022 14:44




Instrument :
BNA_G
ClientSampleId :
SSTD020409


Tgt Ion:153 Resp: 112842
Ion Ratio Lower Upper
153 100
152 47.1 36.5 54.7
154 86.0 65.8 98.6



#53
2,4-Dinitrophenol
Concen: 18.081 ng/ul
RT: 14.973 min Scan# 1955
Delta R.T. 0.000 min
Lab File: BG052450.D
Acq: 22 Feb 2022 14:44

Tgt Ion:184 Resp: 12175
Ion Ratio Lower Upper
184 100
63 60.8 66.2 99.2#
154 59.5 53.6 80.4









#55
4-Nitrophenol
Concen: 21.707 ng/ul
RT: 15.073 min Scan# 1972
Delta R.T. 0.000 min
Lab File: BG052450.D
Acq: 22 Feb 2022 14:44




Instrument :
BNA_G
ClientSampleId :
SSTD020409


Tgt Ion:109 Resp: 21369
Ion Ratio Lower Upper
109 100
139 113.7 88.7 133.1
65 121.1 113.6 170.4



#56
Dibenzofuran
Concen: 22.538 ng/ul
RT: 15.261 min Scan# 2004
Delta R.T. 0.000 min
Lab File: BG052450.D
Acq: 22 Feb 2022 14:44

Tgt Ion:168 Resp: 165735
Ion Ratio Lower Upper
168 100
139 38.7 31.2 46.8









#57
2,4-Dinitrotoluene
Concen: 24.287 ng/ul
RT: 15.214 min Scan# 1996
Delta R.T. 0.000 min
Lab File: BG052450.D
Acq: 22 Feb 2022 14:44




Instrument :
BNA_G
ClientSampleId :
SSTD020409


Tgt Ion:	165	Resp:	45145
Ion Ratio	Lower	Upper	
165	100		
63	43.3	42.1	63.1
182	2.4	1.8	2.8



#58
2,3,4,6-Tetrachlorophenol
Concen: 22.305 ng/ul
RT: 15.496 min Scan# 2044
Delta R.T. 0.000 min
Lab File: BG052450.D
Acq: 22 Feb 2022 14:44

Tgt Ion:	232	Resp:	35656
Ion Ratio	Lower	Upper	
232	100		
131	42.0	38.9	58.3
130	2.8	2.4	3.6
166	26.1	25.8	38.8









#59
Diethylphthalate
Concen: 23.127 ng/ul
RT: 15.660 min Scan# 2072
Delta R.T. 0.000 min
Lab File: BG052450.D
Acq: 22 Feb 2022 14:44




Instrument :
BNA_G
ClientSampleId :
SSTD020409


Tgt Ion:149 Resp: 140952
Ion Ratio Lower Upper
149 100
177 23.8 16.8 25.2
150 11.3 10.1 15.1



#61
Fluorene
Concen: 22.566 ng/ul
RT: 15.913 min Scan# 2115
Delta R.T. 0.000 min
Lab File: BG052450.D
Acq: 22 Feb 2022 14:44

Tgt Ion:166 Resp: 137702
Ion Ratio Lower Upper
166 100
165 100.2 81.0 121.6
167 13.8 11.0 16.4









#62
4-Chlorophenyl-phenylether
Concen: 22.485 ng/ul
RT: 15.895 min Scan# 2112
Delta R.T. 0.000 min
Lab File: BG052450.D
Acq: 22 Feb 2022 14:44




Instrument :
BNA_G
ClientSampleId :
SSTD020409


Tgt Ion:204 Resp: 75459
Ion Ratio Lower Upper
204 100
206 33.9 27.4 41.2
141 52.8 48.6 73.0



#63
4-Nitroaniline
Concen: 15.787 ng/ul
RT: 15.919 min Scan# 2116
Delta R.T. 0.000 min
Lab File: BG052450.D
Acq: 22 Feb 2022 14:44

Tgt Ion:138 Resp: 16998
Ion Ratio Lower Upper
138 100
92 54.1 49.3 73.9
108 93.3 72.6 108.8









#66
 4,6-Dinitro-2-methylphenol
 Concen: 21.533 ng/ul
 RT: 15.989 min Scan# 2128
 Delta R.T. 0.000 min
 Lab File: BG052450.D
 Acq: 22 Feb 2022 14:44




Instrument :
 BNA_G
 ClientSampleId :
 SST020409


Tgt Ion:198 Resp: 25355
 Ion Ratio Lower Upper
 198 100
 182 3.7 3.8 5.8#
 77 23.1 23.2 34.8#



#67
 N-Nitrosodiphenylamine
 Concen: 22.180 ng/ul
 RT: 16.107 min Scan# 2148
 Delta R.T. 0.000 min
 Lab File: BG052450.D
 Acq: 22 Feb 2022 14:44

Tgt Ion:169 Resp: 121775
 Ion Ratio Lower Upper
 169 100
 168 68.0 58.5 87.7
 167 35.3 31.3 46.9









#68
4-Bromophenyl-phenylether
Concen: 21.570 ng/ul
RT: 16.794 min Scan# 21
Delta R.T. 0.000 min
Lab File: BG052450.D
Acq: 22 Feb 2022 14:44




Instrument :
BNA_G
ClientSampleId :
SSTD020409


Tgt Ion:248 Resp: 52147
Ion Ratio Lower Upper
248 100
250 93.1 75.5 113.3
141 63.7 67.6 101.4#



#69
Hexachlorobenzene
Concen: 21.922 ng/ul
RT: 16.923 min Scan# 2287
Delta R.T. 0.000 min
Lab File: BG052450.D
Acq: 22 Feb 2022 14:44

Tgt Ion:284 Resp: 59562
Ion Ratio Lower Upper
284 100
142 26.9 31.9 47.9#
249 30.1 25.8 38.6





#70

Atrazine

Concen: 22.065 ng/ul

RT: 17.047 min Scan# 21

Delta R.T. 0.000 min

Lab File: BG052450.D

Acq: 22 Feb 2022 14:44

Instrument :

BNA_G

ClientSampleId :

SSTD020409





Tgt Ion:200 Resp: 53445

Ion Ratio Lower Upper

200 100

173 22.3 20.2 30.2

215 49.1 38.1 57.1



#71

Pentachlorophenol

Concen: 18.043 ng/ul

RT: 17.270 min Scan# 2346

Delta R.T. 0.000 min

Lab File: BG052450.D

Acq: 22 Feb 2022 14:44




Tgt Ion:266 Resp: 23505


Ion Ratio Lower Upper

266 100

264 65.9 54.3 81.5

268 66.5 51.0 76.6





#72

Phenanthrene

Concen: 22.014 ng/ul

RT: 17.657 min Scan# 2412

Delta R.T. 0.000 min

Lab File: BG052450.D

Acq: 22 Feb 2022 14:44

Instrument :

BNA_G

ClientSampleId :

SSTD020409





Tgt Ion:178 Resp: 232253

Ion Ratio Lower Upper

178 100

179 14.9 12.3 18.5

176 18.6 15.5 23.3



#74

Anthracene

Concen: 22.486 ng/ul

RT: 17.752 min Scan# 2428

Delta R.T. 0.000 min

Lab File: BG052450.D

Acq: 22 Feb 2022 14:44




Tgt Ion:178 Resp: 234146


Ion Ratio Lower Upper

178 100

179 15.5 13.0 19.6

176 18.8 15.6 23.4









#75
1,2,3,4-Tetrachlorobenzene
Concen: 20.146 ng/uL
RT: 13.675 min Scan# 1
Delta R.T. 0.000 min
Lab File: BG052450.D
Acq: 22 Feb 2022 14:44




Instrument :
BNA_G
ClientSampleId :
SSTD020409


Tgt Ion:216 Resp: 67357
Ion Ratio Lower Upper
216 100
214 75.8 61.3 91.9
218 46.5 37.0 55.6



#76
Pentachlorobenzene
Concen: 20.774 ng/uL
RT: 15.190 min Scan# 1992
Delta R.T. 0.000 min
Lab File: BG052450.D
Acq: 22 Feb 2022 14:44

Tgt Ion:250 Resp: 64074
Ion Ratio Lower Upper
250 100
248 64.9 50.5 75.7
252 69.3 51.2 76.8









#77
 Carbazole
 Concen: 23.298 ng/ul
 RT: 18.016 min Scan# 2473
 Delta R.T. 0.000 min
 Lab File: BG052450.D
 Acq: 22 Feb 2022 14:44




Instrument :
 BNA_G
 ClientSampleId :
 SSTD020409


Tgt Ion:167 Resp: 211446
 Ion Ratio Lower Upper
 167 100
 166 22.0 17.3 25.9
 139 12.2 10.6 16.0



#78
 Di-n-butylphthalate
 Concen: 22.762 ng/ul
 RT: 18.556 min Scan# 2565
 Delta R.T. 0.000 min
 Lab File: BG052450.D
 Acq: 22 Feb 2022 14:44

Tgt Ion:149 Resp: 247776
 Ion Ratio Lower Upper
 149 100
 150 9.2 7.5 11.3
 104 5.8 5.0 7.6









#80
Fluoranthene
Concen: 21.766 ng/u1
RT: 19.661 min Scan# 21
Delta R.T. 0.000 min
Lab File: BG052450.D
Acq: 22 Feb 2022 14:44




Instrument :
BNA_G
ClientSampleId :
SSTD020409


Tgt Ion:202 Resp: 303222
Ion Ratio Lower Upper
202 100
101 7.9 9.3 13.9#
100 6.6 7.0 10.6#



#82
Pyrene
Concen: 21.815 ng/u1
RT: 20.025 min Scan# 2815
Delta R.T. 0.000 min
Lab File: BG052450.D
Acq: 22 Feb 2022 14:44

Tgt Ion:202 Resp: 297512
Ion Ratio Lower Upper
202 100
101 8.7 10.7 16.1#
100 8.0 9.0 13.6#









#83
 Butylbenzylphthalate
 Concen: 22.006 ng/ul
 RT: 20.888 min Scan# 2962
 Delta R.T. 0.000 min
 Lab File: BG052450.D
 Acq: 22 Feb 2022 14:44




Instrument :
 BNA_G
 ClientSampleId :
 SSTD020409


Tgt Ion:149 Resp: 104990
 Ion Ratio Lower Upper
 149 100
 91 74.6 65.2 97.8
 206 22.7 16.2 24.2



#84
 3,3'-Dichlorobenzidine
 Concen: 20.518 ng/ul
 RT: 21.811 min Scan# 3119
 Delta R.T. 0.000 min
 Lab File: BG052450.D
 Acq: 22 Feb 2022 14:44

Tgt Ion:252 Resp: 86724
 Ion Ratio Lower Upper
 252 100
 254 62.6 50.8 76.2
 126 8.1 9.0 13.4#









#85
Benzo(a)anthracene
Concen: 21.784 ng/ul
RT: 21.911 min Scan# 3136
Delta R.T. 0.000 min
Lab File: BG052450.D
Acq: 22 Feb 2022 14:44




Instrument :
BNA_G
ClientSampleId :
SSTD020409


Tgt Ion:228 Resp: 289203
Ion Ratio Lower Upper
228 100
229 21.0 15.7 23.5
226 26.6 22.3 33.5



#86
Bis(2-ethylhexyl)phthalate
Concen: 21.655 ng/ul
RT: 21.787 min Scan# 3115
Delta R.T. 0.000 min
Lab File: BG052450.D
Acq: 22 Feb 2022 14:44

Tgt Ion:149 Resp: 148093
Ion Ratio Lower Upper
149 100
167 26.6 21.1 31.7
279 5.0 2.7 4.1#





#87

Chrysene

Concen: 22.412 ng/ul

RT: 21.981 min Scan# 3148

Delta R.T. 0.000 min

Lab File: BG052450.D

Acq: 22 Feb 2022 14:44

Instrument :

BNA_G

ClientSampleId :

SSTD020409





Tgt Ion:228 Resp: 281140

Ion Ratio Lower Upper

228 100

226 28.2 24.8 37.2

229 19.8 15.8 23.6



#89

Di-n-octyl phthalate

Concen: 21.786 ng/ul




RT: 23.080 min Scan# 3335


Delta R.T. 0.000 min

Lab File: BG052450.D

Acq: 22 Feb 2022 14:44

Tgt Ion:149 Resp: 246522









#90
Benzo(b)fluoranthene
Concen: 21.676 ng/ul
RT: 24.284 min Scan# 3540
Delta R.T. 0.000 min
Lab File: BG052450.D
Acq: 22 Feb 2022 14:44




Instrument :
BNA_G
ClientSampleId :
SSTD020409


Tgt Ion:252 Resp: 299801
Ion Ratio Lower Upper
252 100
253 22.4 17.8 26.6
125 6.0 8.5 12.7#



#91
Benzo(k)fluoranthene
Concen: 21.944 ng/ul
RT: 24.360 min Scan# 3553
Delta R.T. 0.000 min
Lab File: BG052450.D
Acq: 22 Feb 2022 14:44

Tgt Ion:252 Resp: 277684
Ion Ratio Lower Upper
252 100
253 22.0 17.3 25.9
125 5.9 7.8 11.6#





#93

Benzo(a)pyrene

Concen: 21.994 ng/ul

RT: 25.235 min Scan# 3702

Delta R.T. 0.000 min

Lab File: BG052450.D

Acq: 22 Feb 2022 14:44

Instrument :

BNA_G

ClientSampleId :

SSTD020409





Tgt Ion:252 Resp: 283704

Ion Ratio Lower Upper

252 100

253 21.3 17.4 26.2

125 6.9 9.1 13.7#



#94

Indeno(1,2,3-cd)pyrene

Concen: 22.411 ng/ul

RT: 29.389 min Scan# 4409

Delta R.T. 0.000 min

Lab File: BG052450.D

Acq: 22 Feb 2022 14:44




Tgt Ion:276 Resp: 322021


Ion Ratio Lower Upper

276 100

138 13.6 15.5 23.3#

277 24.0 20.5 30.7









#95
 Dibenzo(a,h)anthracene
 Concen: 23.131 ng/ul
 RT: 29.459 min Scan# 4421
 Delta R.T. 0.000 min
 Lab File: BG052450.D
 Acq: 22 Feb 2022 14:44

Instrument :
 BNA_G
 ClientSampleId :
 SSTD020409

Tgt Ion:278 Resp: 275590
 Ion Ratio Lower Upper
 278 100
 139 9.1 12.2 18.4#
 279 23.0 19.5 29.3



#96
 Benzo(g,h,i)perylene
 Concen: 11.105 ng/ul
 RT: 30.628 min Scan# 4620
 Delta R.T. -0.012 min
 Lab File: BG052450.D
 Acq: 22 Feb 2022 14:44

Tgt Ion:276 Resp: 135823
 Ion Ratio Lower Upper
 276 100
 138 13.7 16.6 24.8#
 277 23.2 17.6 26.4

