

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG022222\
 Data File : BG052451.D
 Acq On : 22 Feb 2022 15:23
 Operator : CG/JU
 Sample : SSTD04010
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SSTD040410

Quant Time: Feb 22 16:28:39 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG022222.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Feb 22 15:23:39 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.225	152	29094	20.000	ng/ul	0.00
20) Naphthalene-d8	11.062	136	120290	20.000	ng/ul	0.00
38) Acenaphthene-d10	14.869	164	85744	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.618	188	205713	20.000	ng/ul	# 0.00
79) Chrysene-d12	21.936	240	204386	20.000	ng/ul	# 0.00
88) Perylene-d12	25.396	264	213988	20.000	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.543	96	11516	15.969	ng/uL	0.00
4) Pyridine-d5	3.978	84	89298	39.810	ng/ul	0.00
7) Phenol-d5	7.367	99	104538	39.502	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.526	67	61911	38.508	ng/ul	0.00
11) 2-Chlorophenol-d4	7.755	132	75669	41.033	ng/ul	0.00
15) 4-Methylphenol-d8	8.930	113	81814	40.739	ng/ul	0.00
21) Nitrobenzene-d5	9.400	128	39078	39.898	ng/ul	0.00
24) 2-Nitrophenol-d4	10.128	143	44607	41.678	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.675	165	84901	40.691	ng/ul	0.00
31) 4-Chloroaniline-d4	11.191	131	107783	40.725	ng/ul	0.00
46) Dimethylphthalate-d6	14.258	166	262603	38.963	ng/ul	0.00
49) Acenaphthylene-d8	14.563	160	328477	39.101	ng/ul	0.00
54) 4-Nitrophenol-d4	15.057	143	44309	43.408	ng/ul	0.00
60) Fluorene-d10	15.856	176	231252	39.164	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.973	200	51157	40.281	ng/ul	0.00
73) Anthracene-d10	17.718	188	376154	38.786	ng/ul	0.00
81) Pyrene-d10	19.997	212	466927	38.119	ng/ul	0.00
92) Benzo(a)pyrene-d12	25.161	264	439720	38.679	ng/ul	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.584	88	12971	16.136	ng/uL	97
5) Pyridine	3.995	79	88786	37.796	ng/ul	99
6) Benzaldehyde	7.344	77	64493	42.986	ng/ul	93
8) Phenol	7.397	94	105888	39.925	ng/ul	96
10) Bis(2-Chloroethyl)ether	7.626	93	79301	39.568	ng/ul	96
12) 2-Chlorophenol	7.784	128	77978	40.962	ng/ul	93
13) 2-Methylphenol	8.665	108	79039	39.600	ng/ul	99
14) 2,2'-oxybis(1-Chloropr...	8.748	45	105854	36.979	ng/ul	97
16) Acetophenone	9.047	105	122889	39.459	ng/ul	98
17) N-Nitroso-di-n-propyla...	9.030	70	71828	38.450	ng/ul	94
18) 4-Methylphenol	8.994	108	84851	40.351	ng/ul	93
19) Hexachloroethane	9.323	117	31273	38.146	ng/ul#	81
22) Nitrobenzene	9.441	77	103524	37.895	ng/ul	96
23) Isophorone	9.964	82	199229	39.413	ng/ul	98
25) 2-Nitrophenol	10.158	139	46677	39.893	ng/ul	97
26) 2,4-Dimethylphenol	10.210	107	96023	39.801	ng/ul	98
27) Bis(2-Chloroethoxy)met...	10.445	93	106153	39.156	ng/ul	97
29) 2,4-Dichlorophenol	10.704	162	79242	39.288	ng/ul	96
30) Naphthalene	11.109	128	257467	39.145	ng/ul	97
32) 4-Chloroaniline	11.215	127	110733	40.581	ng/ul	99
33) Hexachlorobutadiene	11.397	225	60167	36.570	ng/ul	96
34) Caprolactam	11.961	113	19719	30.144	ng/ul	93
35) 4-Chloro-3-methylphenol	12.331	107	88781	40.810	ng/ul	92
36) 2-Methylnaphthalene	12.707	142	182380	39.515	ng/ul	99


Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG022222\
 Data File : BG052451.D
 Acq On : 22 Feb 2022 15:23
 Operator : CG/JU
 Sample : SST04010
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SST040410

Quant Time: Feb 22 16:28:39 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG022222.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Feb 22 15:23:39 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.924	142	184689	38.896	ng/ul	97
39) 1,2,4,5-Tetrachloroben...	13.077	216	114710	37.167	ng/ul	95
40) Hexachlorocyclopentadiene	13.054	237	56628	32.140	ng/ul	97
41) 2,4,6-Trichlorophenol	13.306	196	72694	39.447	ng/ul	98
42) 2,4,5-Trichlorophenol	13.383	196	76909	38.749	ng/ul	98
43) 1,1'-Biphenyl	13.700	154	255359	37.989	ng/ul#	95
44) 2-Chloronaphthalene	13.753	162	196524	38.210	ng/ul	98
45) 2-Nitroaniline	13.947	65	68438	39.520	ng/ul	90
47) Dimethylphthalate	14.305	163	256282	39.231	ng/ul	99
48) 2,6-Dinitrotoluene	14.428	165	57275	40.641	ng/ul	87
50) Acenaphthylene	14.593	152	322182	38.263	ng/ul	97
51) 3-Nitroaniline	14.763	138	54485	45.178	ng/ul	96
52) Acenaphthene	14.933	153	213386	38.658	ng/ul	96
53) 2,4-Dinitrophenol	14.969	184	29187	39.883	ng/ul#	83
55) 4-Nitrophenol	15.074	109	41355	38.653	ng/ul	86
56) Dibenzofuran	15.262	168	304468	38.095	ng/ul	98
57) 2,4-Dinitrotoluene	15.221	165	82933	41.050	ng/ul#	86
58) 2,3,4,6-Tetrachlorophenol	15.492	232	68824	39.612	ng/ul#	87
59) Diethylphthalate	15.662	149	261820	39.525	ng/ul	98
61) Fluorene	15.914	166	255699	38.554	ng/ul	97
62) 4-Chlorophenyl-phenyle...	15.897	204	138203	37.890	ng/ul	92
63) 4-Nitroaniline	15.926	138	54039	46.177	ng/ul	97
66) 4,6-Dinitro-2-methylph...	15.991	198	50310	41.594	ng/ul#	92
67) N-Nitrosodiphenylamine	16.108	169	224709	39.844	ng/ul	97
68) 4-Bromophenyl-phenylether	16.796	248	97070	39.089	ng/ul#	85
69) Hexachlorobenzene	16.925	284	106978	38.332	ng/ul#	88
70) Atrazine	17.054	200	100001	40.192	ng/ul	95
71) Pentachlorophenol	17.271	266	46525	34.768	ng/ul	98
72) Phenanthrene	17.659	178	426280	39.335	ng/ul	98
74) Anthracene	17.753	178	422273	39.479	ng/ul	100
75) 1,2,3,4-Tetrachloroben...	13.676	216	129166	37.610	ng/uL	99
76) Pentachlorobenzene	15.192	250	119698	37.780	ng/uL	98
77) Carbazole	18.017	167	388027	41.622	ng/ul	96
78) Di-n-butylphthalate	18.558	149	457262	40.894	ng/ul	99
80) Fluoranthene	19.662	202	553406	38.365	ng/ul#	92
82) Pyrene	20.027	202	534053	37.818	ng/ul#	90
83) Butylbenzylphthalate	20.890	149	201778	40.846	ng/ul#	90
84) 3,3'-Dichlorobenzidine	21.812	252	192995	44.098	ng/ul#	97
85) Benzo(a)anthracene	21.912	228	531890	38.692	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.789	149	281000	39.683	ng/ul#	96
87) Chrysene	21.983	228	507279	39.056	ng/ul	98
89) Di-n-octyl phthalate	23.081	149	469237	38.651	ng/ul	100
90) Benzo(b)fluoranthene	24.291	252	560668	37.784	ng/ul#	96
91) Benzo(k)fluoranthene	24.362	252	520137	38.313	ng/ul#	95
93) Benzo(a)pyrene	25.237	252	525331	37.960	ng/ul#	95
94) Indeno(1,2,3-cd)pyrene	29.396	276	610622	39.610	ng/ul#	91
95) Dibenzo(a,h)anthracene	29.461	278	516928	40.441	ng/ul#	94
96) Benzo(g,h,i)perylene	30.647	276	512450	39.051	ng/ul#	91





(#) = qualifier out of range (m) = manual integration (+) = signals summed



#2
1,4-Dioxane
Concen: 16.136 ng/uL
RT: 3.584 min Scan# 1
Delta R.T. 0.002 min
Lab File: BG052451.D
Acq: 22 Feb 2022 15:23




Instrument :
BNA_G
ClientSampleId :
SSTD040410


Tgt Ion: 88 Resp: 12971
Ion Ratio Lower Upper
88 100
43 33.0 23.0 34.4
58 79.4 62.4 93.6



#5
Pyridine
Concen: 37.796 ng/uL
RT: 3.995 min Scan# 86
Delta R.T. 0.002 min
Lab File: BG052451.D
Acq: 22 Feb 2022 15:23

Tgt Ion: 79 Resp: 88786
Ion Ratio Lower Upper
79 100
52 75.0 60.8 91.2
51 38.1 30.5 45.7









#6
Benzaldehyde
Concen: 42.986 ng/ul
RT: 7.344 min Scan# 611
Delta R.T. 0.002 min
Lab File: BG052451.D
Acq: 22 Feb 2022 15:23




Instrument :
BNA_G
ClientSampleId :
SSTD040410


Tgt Ion: 77 Resp: 64493
Ion Ratio Lower Upper
77 100
105 94.6 70.4 105.6
106 82.8 61.2 91.8



#8
Phenol
Concen: 39.925 ng/ul
RT: 7.397 min Scan# 665
Delta R.T. 0.002 min
Lab File: BG052451.D
Acq: 22 Feb 2022 15:23

Tgt Ion: 94 Resp: 105888
Ion Ratio Lower Upper
94 100
65 33.3 25.1 37.7
66 42.5 31.6 47.4









#10
Bis(2-Chloroethyl)ether
Concen: 39.568 ng/ul
RT: 7.626 min Scan# 704
Delta R.T. 0.002 min
Lab File: BG052451.D
Acq: 22 Feb 2022 15:23




Instrument :
BNA_G
ClientSampleId :
SSTD040410


Tgt Ion: 93	Resp: 79301
Ion Ratio	Lower Upper
93 100	
63 81.4	68.6 103.0
95 33.4	26.1 39.1



#12
2-Chlorophenol
Concen: 40.962 ng/ul
RT: 7.784 min Scan# 731
Delta R.T. 0.002 min
Lab File: BG052451.D
Acq: 22 Feb 2022 15:23

Tgt Ion:128	Resp: 77978
Ion Ratio	Lower Upper
128 100	
64 48.9	45.3 67.9
130 33.1	27.1 40.7









#13
2-Methylphenol
Concen: 39.600 ng/ul
RT: 8.665 min Scan# 881
Delta R.T. 0.002 min
Lab File: BG052451.D
Acq: 22 Feb 2022 15:23




Instrument :
BNA_G
ClientSampleId :
SSTD040410


Tgt Ion:108 Resp: 79039
Ion Ratio Lower Upper
108 100
107 88.9 71.8 107.6



#14
2,2'-oxybis(1-Chloropropane)
Concen: 36.979 ng/ul
RT: 8.748 min Scan# 895
Delta R.T. -0.004 min
Lab File: BG052451.D
Acq: 22 Feb 2022 15:23

Tgt Ion: 45 Resp: 105854
Ion Ratio Lower Upper
45 100
77 14.9 10.7 16.1
79 11.2 8.5 12.7









#16
Acetophenone
Concen: 39.459 ng/ul
RT: 9.047 min Scan# 946
Delta R.T. 0.002 min
Lab File: BG052451.D
Acq: 22 Feb 2022 15:23




Instrument :
BNA_G
ClientSampleId :
SSTD040410


Tgt Ion:105 Resp: 122889
Ion Ratio Lower Upper
105 100
77 82.4 67.3 100.9
51 28.7 24.0 36.0



#17
N-Nitroso-di-n-propylamine
Concen: 38.450 ng/ul
RT: 9.030 min Scan# 943
Delta R.T. 0.002 min
Lab File: BG052451.D
Acq: 22 Feb 2022 15:23

Tgt Ion: 70 Resp: 71828
Ion Ratio Lower Upper
70 100
42 52.9 46.0 69.0
101 8.0 7.0 10.4
130 19.1 13.4 20.2









#18
4-Methylphenol
Concen: 40.351 ng/ul
RT: 8.994 min Scan# 91
Delta R.T. -0.004 min
Lab File: BG052451.D
Acq: 22 Feb 2022 15:23




Instrument :
BNA_G
ClientSampleId :
SSTD040410


Tgt Ion:108 Resp: 84851
Ion Ratio Lower Upper
108 100
107 110.5 95.0 142.4



#19
Hexachloroethane
Concen: 38.146 ng/ul
RT: 9.323 min Scan# 993
Delta R.T. -0.004 min
Lab File: BG052451.D
Acq: 22 Feb 2022 15:23

Tgt Ion:117 Resp: 31273
Ion Ratio Lower Upper
117 100
201 117.8 77.2 115.8#
199 72.9 49.9 74.9









#22
Nitrobenzene
Concen: 37.895 ng/ul
RT: 9.441 min Scan# 1013
Delta R.T. 0.002 min
Lab File: BG052451.D
Acq: 22 Feb 2022 15:23




Instrument :
BNA_G
ClientSampleId :
SSTD040410


Tgt Ion: 77 Resp: 103524
Ion Ratio Lower Upper
77 100
123 39.4 30.1 45.1
65 14.2 9.6 14.4



#23
Isophorone
Concen: 39.413 ng/ul
RT: 9.964 min Scan# 1102
Delta R.T. 0.002 min
Lab File: BG052451.D
Acq: 22 Feb 2022 15:23

Tgt Ion: 82 Resp: 199229
Ion Ratio Lower Upper
82 100
95 7.1 5.6 8.4
138 15.9 11.8 17.8









#25
2-Nitrophenol
Concen: 39.893 ng/ul
RT: 10.158 min Scan# 1135
Delta R.T. -0.004 min
Lab File: BG052451.D
Acq: 22 Feb 2022 15:23




Instrument :
BNA_G
ClientSampleId :
SSTD040410


Tgt Ion:139 Resp: 46677
Ion Ratio Lower Upper
139 100
65 58.6 48.0 72.0
109 39.2 28.6 42.8



#26
2,4-Dimethylphenol
Concen: 39.801 ng/ul
RT: 10.210 min Scan# 1144
Delta R.T. -0.004 min
Lab File: BG052451.D
Acq: 22 Feb 2022 15:23

Tgt Ion:107 Resp: 96023
Ion Ratio Lower Upper
107 100
121 48.0 39.3 58.9
122 81.0 63.7 95.5









#27
 Bis(2-Chloroethoxy)methane
 Concen: 39.156 ng/ul
 RT: 10.445 min Scan# 1184
 Delta R.T. 0.002 min
 Lab File: BG052451.D
 Acq: 22 Feb 2022 15:23




Instrument :
 BNA_G
 ClientSampleId :
 SSTD040410


Tgt Ion: 93 Resp: 106153
 Ion Ratio Lower Upper
 93 100
 95 30.9 26.5 39.7
 123 10.7 8.1 12.1



#29
 2,4-Dichlorophenol
 Concen: 39.288 ng/ul
 RT: 10.704 min Scan# 1228
 Delta R.T. 0.002 min
 Lab File: BG052451.D
 Acq: 22 Feb 2022 15:23

Tgt Ion:162 Resp: 79242
 Ion Ratio Lower Upper
 162 100
 164 59.1 49.4 74.0
 98 35.3 30.4 45.6









#30
Naphthalene
Concen: 39.145 ng/ul
RT: 11.109 min Scan# 11
Delta R.T. -0.004 min
Lab File: BG052451.D
Acq: 22 Feb 2022 15:23




Instrument :
BNA_G
ClientSampleId :
SSTD040410


Tgt Ion:128 Resp: 257467
Ion Ratio Lower Upper
128 100
129 10.9 8.5 12.7
127 12.8 11.8 17.6



#32
4-Chloroaniline
Concen: 40.581 ng/ul
RT: 11.215 min Scan# 1315
Delta R.T. 0.002 min
Lab File: BG052451.D
Acq: 22 Feb 2022 15:23

Tgt Ion:127 Resp: 110733
Ion Ratio Lower Upper
127 100
129 33.5 26.2 39.2





#33

Hexachlorobutadiene

Concen: 36.570 ng/ul

RT: 11.397 min Scan# 11

Delta R.T. 0.002 min

Lab File: BG052451.D

Acq: 22 Feb 2022 15:23

Instrument :

BNA_G

ClientSampleId :

SSTD040410





Tgt Ion:225 Resp: 60167

Ion Ratio Lower Upper

225 100

223 60.0 50.6 75.8

227 65.4 49.8 74.8



#34

Caprolactam

Concen: 30.144 ng/ul

RT: 11.961 min Scan# 1442

Delta R.T. 0.002 min

Lab File: BG052451.D

Acq: 22 Feb 2022 15:23




Tgt Ion:113 Resp: 19719


Ion Ratio Lower Upper

113 100

55 169.2 147.0 220.6

56 132.1 109.2 163.8





#35

4-Chloro-3-methylphenol

Concen: 40.810 ng/ul

RT: 12.331 min Scan# 1

Delta R.T. 0.002 min

Lab File: BG052451.D

Acq: 22 Feb 2022 15:23

Instrument :

BNA_G

ClientSampleId :

SSTD040410





Tgt Ion:107 Resp: 88781

Ion Ratio Lower Upper

107 100

144 23.7 19.8 29.6

142 78.9 56.2 84.4



#36

2-Methylnaphthalene

Concen: 39.515 ng/ul

RT: 12.707 min Scan# 1569

Delta R.T. 0.002 min

Lab File: BG052451.D




Acq: 22 Feb 2022 15:23


Tgt Ion:142 Resp: 182380

Ion Ratio Lower Upper

142 100

141 92.2 73.3 109.9









#37
1-Methylnaphthalene
Concen: 38.896 ng/ul
RT: 12.924 min Scan# 1
Delta R.T. 0.002 min
Lab File: BG052451.D
Acq: 22 Feb 2022 15:23




Instrument :
BNA_G
ClientSampleId :
SSTD040410


Tgt Ion:142 Resp: 184689
Ion Ratio Lower Upper
142 100
141 94.7 77.8 116.8
116 4.1 3.3 4.9



#39
1,2,4,5-Tetrachlorobenzene
Concen: 37.167 ng/ul
RT: 13.077 min Scan# 1632
Delta R.T. 0.008 min
Lab File: BG052451.D
Acq: 22 Feb 2022 15:23

Tgt Ion:216 Resp: 114710
Ion Ratio Lower Upper
216 100
214 80.1 61.0 91.4
179 17.8 16.2 24.2
108 13.3 13.1 19.7









#40
Hexachlorocyclopentadiene
Concen: 32.140 ng/ul
RT: 13.054 min Scan# 1628
Delta R.T. 0.002 min
Lab File: BG052451.D
Acq: 22 Feb 2022 15:23




Instrument :
BNA_G
ClientSampleId :
SSTD040410


Tgt Ion:237 Resp: 56628
Ion Ratio Lower Upper
237 100
235 59.5 49.4 74.0
272 11.8 11.1 16.7



#41
2,4,6-Trichlorophenol
Concen: 39.447 ng/ul
RT: 13.306 min Scan# 1671
Delta R.T. -0.004 min
Lab File: BG052451.D
Acq: 22 Feb 2022 15:23

Tgt Ion:196 Resp: 72694
Ion Ratio Lower Upper
196 100
198 93.6 74.3 111.5
200 32.1 23.8 35.6









#42
2,4,5-Trichlorophenol
Concen: 38.749 ng/ul
RT: 13.383 min Scan# 1684
Delta R.T. -0.004 min
Lab File: BG052451.D
Acq: 22 Feb 2022 15:23




Instrument :
BNA_G
ClientSampleId :
SSTD040410


Tgt Ion:196 Resp: 76909
Ion Ratio Lower Upper
196 100
198 94.6 77.0 115.4
200 29.0 25.4 38.2



#43
1,1'-Biphenyl
Concen: 37.989 ng/ul
RT: 13.700 min Scan# 1738
Delta R.T. 0.002 min
Lab File: BG052451.D
Acq: 22 Feb 2022 15:23

Tgt Ion:154 Resp: 255359
Ion Ratio Lower Upper
154 100
153 41.4 34.5 51.7
76 12.0 13.1 19.7#









#44
2-Chloronaphthalene
Concen: 38.210 ng/ul
RT: 13.753 min Scan# 1746
Delta R.T. 0.008 min
Lab File: BG052451.D
Acq: 22 Feb 2022 15:23




Instrument :
BNA_G
ClientSampleId :
SSTD040410


Tgt Ion:162 Resp: 196524
Ion Ratio Lower Upper
162 100
164 32.1 26.1 39.1
127 36.9 31.0 46.4



#45
2-Nitroaniline
Concen: 39.520 ng/ul
RT: 13.947 min Scan# 1780
Delta R.T. 0.002 min
Lab File: BG052451.D
Acq: 22 Feb 2022 15:23

Tgt Ion: 65 Resp: 68438
Ion Ratio Lower Upper
65 100
92 66.7 53.0 79.4
138 99.4 67.4 101.0









#47
Dimethylphthalate
Concen: 39.231 ng/ul
RT: 14.305 min Scan# 1841
Delta R.T. 0.002 min
Lab File: BG052451.D
Acq: 22 Feb 2022 15:23




Instrument :
BNA_G
ClientSampleId :
SSTD040410


Tgt Ion:163 Resp: 256282
Ion Ratio Lower Upper
163 100
194 4.1 3.0 4.4
164 10.7 8.2 12.4



#48
2,6-Dinitrotoluene
Concen: 40.641 ng/ul
RT: 14.428 min Scan# 1862
Delta R.T. -0.004 min
Lab File: BG052451.D
Acq: 22 Feb 2022 15:23

Tgt Ion:165 Resp: 57275
Ion Ratio Lower Upper
165 100
89 58.8 57.4 86.0
121 20.9 18.8 28.2









#50
Acenaphthylene
Concen: 38.263 ng/ul
RT: 14.593 min Scan# 1890
Delta R.T. 0.002 min
Lab File: BG052451.D
Acq: 22 Feb 2022 15:23




Instrument :
BNA_G
ClientSampleId :
SSTD040410


Tgt Ion:152 Resp: 322182
Ion Ratio Lower Upper
152 100
151 19.7 17.4 26.0
153 13.4 10.3 15.5



#51
3-Nitroaniline
Concen: 45.178 ng/ul
RT: 14.763 min Scan# 1919
Delta R.T. 0.002 min
Lab File: BG052451.D
Acq: 22 Feb 2022 15:23

Tgt Ion:138 Resp: 54485
Ion Ratio Lower Upper
138 100
108 9.4 8.8 13.2
92 118.2 98.0 147.0









#52
Acenaphthene
Concen: 38.658 ng/ul
RT: 14.933 min Scan# 1948
Delta R.T. 0.002 min
Lab File: BG052451.D
Acq: 22 Feb 2022 15:23




Instrument :
BNA_G
ClientSampleId :
SSTD040410


Tgt Ion:153 Resp: 213386
Ion Ratio Lower Upper
153 100
152 48.0 36.5 54.7
154 86.3 65.8 98.6



#53
2,4-Dinitrophenol
Concen: 39.883 ng/ul
RT: 14.969 min Scan# 1954
Delta R.T. -0.004 min
Lab File: BG052451.D
Acq: 22 Feb 2022 15:23

Tgt Ion:184 Resp: 29187
Ion Ratio Lower Upper
184 100
63 59.6 66.2 99.2#
154 62.1 53.6 80.4









#55
4-Nitrophenol
Concen: 38.653 ng/ul
RT: 15.074 min Scan# 1972
Delta R.T. 0.002 min
Lab File: BG052451.D
Acq: 22 Feb 2022 15:23




Instrument :
BNA_G
ClientSampleId :
SSTD040410


Tgt Ion:109 Resp: 41355
Ion Ratio Lower Upper
109 100
139 107.4 88.7 133.1
65 114.3 113.6 170.4



#56
Dibenzofuran
Concen: 38.095 ng/ul
RT: 15.262 min Scan# 2004
Delta R.T. 0.002 min
Lab File: BG052451.D
Acq: 22 Feb 2022 15:23

Tgt Ion:168 Resp: 304468
Ion Ratio Lower Upper
168 100
139 37.9 31.2 46.8









#57
2,4-Dinitrotoluene
Concen: 41.050 ng/ul
RT: 15.221 min Scan# 1996
Delta R.T. 0.008 min
Lab File: BG052451.D
Acq: 22 Feb 2022 15:23




Instrument :
BNA_G
ClientSampleId :
SSTD040410


Tgt Ion:165 Resp: 82933
Ion Ratio Lower Upper
165 100
63 42.1 42.1 63.1
182 3.0 1.8 2.8#



#58
2,3,4,6-Tetrachlorophenol
Concen: 39.612 ng/ul
RT: 15.492 min Scan# 2043
Delta R.T. -0.004 min
Lab File: BG052451.D
Acq: 22 Feb 2022 15:23

Tgt Ion:232 Resp: 68824
Ion Ratio Lower Upper
232 100
131 38.0 38.9 58.3#
130 2.5 2.4 3.6
166 26.9 25.8 38.8









#59
Diethylphthalate
Concen: 39.525 ng/ul
RT: 15.662 min Scan# 2072
Delta R.T. 0.002 min
Lab File: BG052451.D
Acq: 22 Feb 2022 15:23




Instrument :
BNA_G
ClientSampleId :
SSTD040410


Tgt Ion:149 Resp: 261820
Ion Ratio Lower Upper
149 100
177 22.8 16.8 25.2
150 12.5 10.1 15.1



#61
Fluorene
Concen: 38.554 ng/ul
RT: 15.914 min Scan# 2115
Delta R.T. 0.002 min
Lab File: BG052451.D
Acq: 22 Feb 2022 15:23

Tgt Ion:166 Resp: 255699
Ion Ratio Lower Upper
166 100
165 98.3 81.0 121.6
167 13.8 11.0 16.4









#62
4-Chlorophenyl-phenylether
Concen: 37.890 ng/ul
RT: 15.897 min Scan# 2112
Delta R.T. 0.002 min
Lab File: BG052451.D
Acq: 22 Feb 2022 15:23




Instrument :
BNA_G
ClientSampleId :
SSTD040410


Tgt Ion:204 Resp: 138203
Ion Ratio Lower Upper
204 100
206 33.0 27.4 41.2
141 51.7 48.6 73.0



#63
4-Nitroaniline
Concen: 46.177 ng/ul
RT: 15.926 min Scan# 2117
Delta R.T. 0.008 min
Lab File: BG052451.D
Acq: 22 Feb 2022 15:23

Tgt Ion:138 Resp: 54039
Ion Ratio Lower Upper
138 100
92 56.0 49.3 73.9
108 91.0 72.6 108.8









#66
4,6-Dinitro-2-methylphenol
Concen: 41.594 ng/ul
RT: 15.991 min Scan# 2128
Delta R.T. 0.002 min
Lab File: BG052451.D
Acq: 22 Feb 2022 15:23




Instrument :
BNA_G
ClientSampleId :
SSTD040410


Tgt Ion:198 Resp: 50310
Ion Ratio Lower Upper
198 100
182 2.7 3.8 5.8#
77 24.4 23.2 34.8



#67
N-Nitrosodiphenylamine
Concen: 39.844 ng/ul
RT: 16.108 min Scan# 2148
Delta R.T. 0.002 min
Lab File: BG052451.D
Acq: 22 Feb 2022 15:23

Tgt Ion:169 Resp: 224709
Ion Ratio Lower Upper
169 100
168 69.8 58.5 87.7
167 39.0 31.3 46.9









#68
4-Bromophenyl-phenylether
Concen: 39.089 ng/ul
RT: 16.796 min Scan# 21
Delta R.T. 0.002 min
Lab File: BG052451.D
Acq: 22 Feb 2022 15:23




Instrument :
BNA_G
ClientSampleId :
SSTD040410


Tgt Ion:248 Resp: 97070
Ion Ratio Lower Upper
248 100
250 100.4 75.5 113.3
141 61.2 67.6 101.4#



#69
Hexachlorobenzene
Concen: 38.332 ng/ul
RT: 16.925 min Scan# 2287
Delta R.T. 0.002 min
Lab File: BG052451.D
Acq: 22 Feb 2022 15:23

Tgt Ion:284 Resp: 106978
Ion Ratio Lower Upper
284 100
142 26.9 31.9 47.9#
249 32.3 25.8 38.6





#70

Atrazine

Concen: 40.192 ng/ul

RT: 17.054 min Scan# 21

Delta R.T. 0.008 min

Lab File: BG052451.D

Acq: 22 Feb 2022 15:23

Instrument :

BNA_G

ClientSampleId :

SSTD040410





Tgt Ion:200 Resp: 100001

Ion Ratio Lower Upper

200 100

173 21.4 20.2 30.2

215 50.0 38.1 57.1



#71

Pentachlorophenol

Concen: 34.768 ng/ul

RT: 17.271 min Scan# 2346

Delta R.T. 0.002 min

Lab File: BG052451.D

Acq: 22 Feb 2022 15:23




Tgt Ion:266 Resp: 46525


Ion Ratio Lower Upper

266 100

264 66.1 54.3 81.5

268 64.4 51.0 76.6









#72
Phenanthrene
Concen: 39.335 ng/u1
RT: 17.659 min Scan# 2412
Delta R.T. 0.002 min
Lab File: BG052451.D
Acq: 22 Feb 2022 15:23




Instrument :
BNA_G
ClientSampleId :
SSTD040410


Tgt Ion:178 Resp: 426280
Ion Ratio Lower Upper
178 100
179 16.7 12.3 18.5
176 18.6 15.5 23.3



#74
Anthracene
Concen: 39.479 ng/u1
RT: 17.753 min Scan# 2428
Delta R.T. 0.002 min
Lab File: BG052451.D
Acq: 22 Feb 2022 15:23

Tgt Ion:178 Resp: 422273
Ion Ratio Lower Upper
178 100
179 16.1 13.0 19.6
176 19.3 15.6 23.4









#75
1,2,3,4-Tetrachlorobenzene
Concen: 37.610 ng/uL
RT: 13.676 min Scan# 1
Delta R.T. 0.002 min
Lab File: BG052451.D
Acq: 22 Feb 2022 15:23




Instrument :
BNA_G
ClientSampleId :
SSTD040410


Tgt Ion:216 Resp: 129166
Ion Ratio Lower Upper
216 100
214 76.0 61.3 91.9
218 46.7 37.0 55.6



#76
Pentachlorobenzene
Concen: 37.780 ng/uL
RT: 15.192 min Scan# 1992
Delta R.T. 0.002 min
Lab File: BG052451.D
Acq: 22 Feb 2022 15:23

Tgt Ion:250 Resp: 119698
Ion Ratio Lower Upper
250 100
248 65.8 50.5 75.7
252 65.1 51.2 76.8









#77
 Carbazole
 Concen: 41.622 ng/ul
 RT: 18.017 min Scan# 2473
 Delta R.T. 0.002 min
 Lab File: BG052451.D
 Acq: 22 Feb 2022 15:23




Instrument :
 BNA_G
 ClientSampleId :
 SST040410


Tgt Ion:167 Resp: 388027
 Ion Ratio Lower Upper
 167 100
 166 23.5 17.3 25.9
 139 12.1 10.6 16.0



#78
 Di-n-butylphthalate
 Concen: 40.894 ng/ul
 RT: 18.558 min Scan# 2565
 Delta R.T. 0.002 min
 Lab File: BG052451.D
 Acq: 22 Feb 2022 15:23

Tgt Ion:149 Resp: 457262
 Ion Ratio Lower Upper
 149 100
 150 9.4 7.5 11.3
 104 5.7 5.0 7.6









#80
 Fluoranthene
 Concen: 38.365 ng/u1
 RT: 19.662 min Scan# 21
 Delta R.T. 0.002 min
 Lab File: BG052451.D
 Acq: 22 Feb 2022 15:23




Instrument :
 BNA_G
 ClientSampleId :
 SSTD040410


Tgt Ion:202 Resp: 553406
 Ion Ratio Lower Upper
 202 100
 101 7.7 9.3 13.9#
 100 6.6 7.0 10.6#



#82
 Pyrene
 Concen: 37.818 ng/u1
 RT: 20.027 min Scan# 2815
 Delta R.T. 0.002 min
 Lab File: BG052451.D
 Acq: 22 Feb 2022 15:23

Tgt Ion:202 Resp: 534053
 Ion Ratio Lower Upper
 202 100
 101 9.0 10.7 16.1#
 100 7.7 9.0 13.6#









#83
Butylbenzylphthalate
Concen: 40.846 ng/ul
RT: 20.890 min Scan# 2962
Delta R.T. 0.002 min
Lab File: BG052451.D
Acq: 22 Feb 2022 15:23




Instrument :
BNA_G
ClientSampleId :
SSTD040410


Tgt Ion:149 Resp: 201778
Ion Ratio Lower Upper
149 100
91 72.4 65.2 97.8
206 24.9 16.2 24.2#



#84
3,3'-Dichlorobenzidine
Concen: 44.098 ng/ul
RT: 21.812 min Scan# 3119
Delta R.T. 0.002 min
Lab File: BG052451.D
Acq: 22 Feb 2022 15:23

Tgt Ion:252 Resp: 192995
Ion Ratio Lower Upper
252 100
254 65.1 50.8 76.2
126 7.4 9.0 13.4#









#85
Benzo(a)anthracene
Concen: 38.692 ng/ul
RT: 21.912 min Scan# 3136
Delta R.T. 0.002 min
Lab File: BG052451.D
Acq: 22 Feb 2022 15:23




Instrument :
BNA_G
ClientSampleId :
SSTD040410


Tgt Ion:228 Resp: 531890
Ion Ratio Lower Upper
228 100
229 20.6 15.7 23.5
226 28.0 22.3 33.5



#86
Bis(2-ethylhexyl)phthalate
Concen: 39.683 ng/ul
RT: 21.789 min Scan# 3115
Delta R.T. 0.002 min
Lab File: BG052451.D
Acq: 22 Feb 2022 15:23

Tgt Ion:149 Resp: 281000
Ion Ratio Lower Upper
149 100
167 28.6 21.1 31.7
279 5.0 2.7 4.1#









#87
Chrysene
Concen: 39.056 ng/ul
RT: 21.983 min Scan# 3148
Delta R.T. 0.002 min
Lab File: BG052451.D
Acq: 22 Feb 2022 15:23




Instrument :
BNA_G
ClientSampleId :
SSTD040410


Tgt Ion:228 Resp: 507279
Ion Ratio Lower Upper
228 100
226 29.2 24.8 37.2
229 19.5 15.8 23.6



#89
Di-n-octyl phthalate
Concen: 38.651 ng/ul
RT: 23.081 min Scan# 3335
Delta R.T. 0.002 min
Lab File: BG052451.D
Acq: 22 Feb 2022 15:23

Tgt Ion:149 Resp: 469237









#90
Benzo(b)fluoranthene
Concen: 37.784 ng/ul
RT: 24.291 min Scan# 3541
Delta R.T. 0.008 min
Lab File: BG052451.D
Acq: 22 Feb 2022 15:23




Instrument :
BNA_G
ClientSampleId :
SSTD040410


Tgt Ion:252 Resp: 560668
Ion Ratio Lower Upper
252 100
253 22.4 17.8 26.6
125 6.6 8.5 12.7#



#91
Benzo(k)fluoranthene
Concen: 38.313 ng/ul
RT: 24.362 min Scan# 3553
Delta R.T. 0.002 min
Lab File: BG052451.D
Acq: 22 Feb 2022 15:23

Tgt Ion:252 Resp: 520137
Ion Ratio Lower Upper
252 100
253 22.8 17.3 25.9
125 6.0 7.8 11.6#









#93
Benzo(a)pyrene
Concen: 37.960 ng/u1
RT: 25.237 min Scan# 31
Delta R.T. 0.002 min
Lab File: BG052451.D
Acq: 22 Feb 2022 15:23




Instrument :
BNA_G
ClientSampleId :
SSTD040410


Tgt Ion:252 Resp: 525331
Ion Ratio Lower Upper
252 100
253 21.5 17.4 26.2
125 6.7 9.1 13.7#



#94
Indeno(1,2,3-cd)pyrene
Concen: 39.610 ng/u1
RT: 29.396 min Scan# 4410
Delta R.T. 0.008 min
Lab File: BG052451.D
Acq: 22 Feb 2022 15:23

Tgt Ion:276 Resp: 610622
Ion Ratio Lower Upper
276 100
138 12.1 15.5 23.3#
277 24.2 20.5 30.7









#95
 Dibenzo(a,h)anthracene
 Concen: 40.441 ng/ul
 RT: 29.461 min Scan# 4421
 Delta R.T. 0.002 min
 Lab File: BG052451.D
 Acq: 22 Feb 2022 15:23

Instrument :
 BNA_G
 ClientSampleId :
 SST040410

Tgt Ion:278 Resp: 516928
 Ion Ratio Lower Upper
 278 100
 139 8.6 12.2 18.4#
 279 24.7 19.5 29.3



#96
 Benzo(g,h,i)perylene
 Concen: 39.051 ng/ul
 RT: 30.647 min Scan# 4623
 Delta R.T. 0.008 min
 Lab File: BG052451.D
 Acq: 22 Feb 2022 15:23

Tgt Ion:276 Resp: 512450
 Ion Ratio Lower Upper
 276 100
 138 13.3 16.6 24.8#
 277 23.5 17.6 26.4

