

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG022222\
 Data File : BG052454.D
 Acq On : 22 Feb 2022 18:39
 Operator : CG/JU
 Sample : SSTDICV020
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SICV413

Quant Time: Feb 23 03:36:14 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG022222.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Feb 22 17:32:12 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.222	152	25010	20.000	ng/u1	0.00
20) Naphthalene-d8	11.059	136	103114	20.000	ng/u1	# 0.00
38) Acenaphthene-d10	14.866	164	76973	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.615	188	192308	20.000	ng/u1	# 0.00
79) Chrysene-d12	21.933	240	190138	20.000	ng/u1	# 0.00
88) Perylene-d12	25.399	264	193029	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.540	96	4244	6.657	ng/uL	0.00
4) Pyridine-d5	3.975	84	36519	18.987	ng/u1	0.00
7) Phenol-d5	7.370	99	41294	17.412	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.529	67	28880	20.838	ng/u1	0.00
11) 2-Chlorophenol-d4	7.752	132	30430	18.268	ng/u1	0.00
15) 4-Methylphenol-d8	8.927	113	33963	18.406	ng/u1	0.00
21) Nitrobenzene-d5	9.397	128	16122	18.340	ng/u1	0.00
24) 2-Nitrophenol-d4	10.125	143	18699	19.420	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.677	165	34255	19.078	ng/u1	0.00
31) 4-Chloroaniline-d4	11.188	131	45807	18.409	ng/u1	0.00
46) Dimethylphthalate-d6	14.261	166	120219	19.212	ng/u1	0.00
49) Acenaphthylene-d8	14.566	160	150008	19.419	ng/u1	0.00
54) 4-Nitrophenol-d4	15.054	143	18069	18.165	ng/u1	0.00
60) Fluorene-d10	15.859	176	106106	19.303	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.970	200	21281	17.453	ng/u1	0.00
73) Anthracene-d10	17.715	188	175904	18.934	ng/u1	0.00
81) Pyrene-d10	19.994	212	214262	19.258	ng/u1	0.00
92) Benzo(a)pyrene-d12	25.152	264	195287	18.858	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.581	88	4886	6.587	ng/uL#	85
5) Pyridine	3.998	79	30326	15.617	ng/u1	94
6) Benzaldehyde	7.341	77	30482	22.634	ng/u1	94
8) Phenol	7.399	94	42699	17.802	ng/u1	95
10) Bis(2-Chloroethyl)ether	7.623	93	31300	17.642	ng/u1	99
12) 2-Chlorophenol	7.781	128	30056	17.319	ng/u1	93
13) 2-Methylphenol	8.662	108	30949	17.045	ng/u1	96
14) 2,2'-oxybis(1-Chloropr...	8.745	45	42735	17.600	ng/u1	95
16) Acetophenone	9.050	105	49293	17.339	ng/u1	96
17) N-Nitroso-di-n-propyla...	9.027	70	29340	17.530	ng/u1	97
18) 4-Methylphenol	8.991	108	34319	17.773	ng/u1	90
19) Hexachloroethane	9.326	117	12080	17.804	ng/u1#	73
22) Nitrobenzene	9.438	77	43049	19.104	ng/u1#	97
23) Isophorone	9.961	82	84466	19.429	ng/u1	97
25) 2-Nitrophenol	10.160	139	18564	18.347	ng/u1	97
26) 2,4-Dimethylphenol	10.213	107	39953	19.198	ng/u1	97
27) Bis(2-Chloroethoxy)met...	10.442	93	44752	19.020	ng/u1	98
29) 2,4-Dichlorophenol	10.707	162	32010	18.845	ng/u1	99
30) Naphthalene	11.112	128	106343	18.537	ng/u1	99
32) 4-Chloroaniline	11.212	127	47692	18.684	ng/u1	100
33) Hexachlorobutadiene	11.400	225	24541	18.310	ng/u1	96
34) Caprolactam	11.970	113	12610	19.955	ng/u1#	75
35) 4-Chloro-3-methylphenol	12.322	107	37712	19.878	ng/u1	94
36) 2-Methylnaphthalene	12.710	142	80542	19.594	ng/u1	95


Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG022222\
 Data File : BG052454.D
 Acq On : 22 Feb 2022 18:39
 Operator : CG/JU
 Sample : SSTDICV020
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SICV413

Quant Time: Feb 23 03:36:14 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG022222.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Feb 22 17:32:12 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.927	142	75449	18.036	ng/ul	100
39) 1,2,4,5-Tetrachloroben...	13.074	216	47964	17.756	ng/ul	97
40) Hexachlorocyclopentadiene	13.057	237	20218	15.461	ng/ul	98
41) 2,4,6-Trichlorophenol	13.309	196	30526	18.842	ng/ul	95
42) 2,4,5-Trichlorophenol	13.385	196	31331	18.144	ng/ul	96
43) 1,1'-Biphenyl	13.703	154	108180	17.676	ng/ul#	91
44) 2-Chloronaphthalene	13.750	162	88361	18.984	ng/ul	99
45) 2-Nitroaniline	13.944	65	29292	19.269	ng/ul	92
47) Dimethylphthalate	14.308	163	116823	19.099	ng/ul	99
48) 2,6-Dinitrotoluene	14.425	165	26050	19.448	ng/ul	86
50) Acenaphthylene	14.590	152	147580	19.118	ng/ul	97
51) 3-Nitroaniline	14.760	138	27040	24.788	ng/ul#	93
52) Acenaphthene	14.930	153	94505	18.616	ng/ul	95
53) 2,4-Dinitrophenol	14.977	184	10353	15.631	ng/ul#	82
55) 4-Nitrophenol	15.071	109	17254	17.885	ng/ul	94
56) Dibenzofuran	15.265	168	137426	18.831	ng/ul	99
57) 2,4-Dinitrotoluene	15.218	165	37818	19.491	ng/ul#	84
58) 2,3,4,6-Tetrachlorophenol	15.494	232	31203	19.938	ng/ul#	84
59) Diethylphthalate	15.659	149	118968	19.241	ng/ul	97
61) Fluorene	15.911	166	116055	18.905	ng/ul	98
62) 4-Chlorophenyl-phenyle...	15.894	204	62753	18.783	ng/ul	97
63) 4-Nitroaniline	15.923	138	27759	26.504	ng/ul	92
66) 4,6-Dinitro-2-methylph...	15.988	198	20897	17.683	ng/ul#	88
67) N-Nitrosodiphenylamine	16.105	169	103582	19.117	ng/ul	94
68) 4-Bromophenyl-phenylether	16.793	248	43829	18.989	ng/ul#	87
69) Hexachlorobenzene	16.928	284	49864	18.828	ng/ul#	81
70) Atrazine	17.045	200	43739	18.316	ng/ul	97
71) Pentachlorophenol	17.274	266	16016	13.427	ng/ul	95
72) Phenanthrene	17.656	178	195925	18.916	ng/ul	97
74) Anthracene	17.750	178	195405	18.745	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.673	216	48368	15.899	ng/ul	98
76) Pentachlorobenzene	15.189	250	54652	18.817	ng/ul	99
77) Carbazole	18.015	167	180116	19.453	ng/ul	97
78) Di-n-butylphthalate	18.555	149	210378	19.105	ng/ul	99
80) Fluoranthene	19.665	202	246318	18.752	ng/ul#	91
82) Pyrene	20.024	202	247170	18.949	ng/ul#	91
83) Butylbenzylphthalate	20.893	149	86749	18.396	ng/ul#	89
84) 3,3'-Dichlorobenzidine	21.809	252	94496	22.498	ng/ul#	97
85) Benzo(a)anthracene	21.909	228	249019	19.328	ng/ul	98
86) Bis(2-ethylhexyl)phtha...	21.786	149	124412	18.822	ng/ul#	99
87) Chrysene	21.980	228	229488	18.580	ng/ul	98
89) Di-n-octyl phthalate	23.078	149	206963	18.542	ng/ul	100
90) Benzo(b)fluoranthene	24.288	252	252208	18.740	ng/ul#	96
91) Benzo(k)fluoranthene	24.359	252	239737	19.313	ng/ul#	97
93) Benzo(a)pyrene	25.228	252	233053	18.408	ng/ul#	95
94) Indeno(1,2,3-cd)pyrene	29.381	276	129671	9.007	ng/ul#	91
95) Dibenzo(a,h)anthracene	29.446	278	223938	18.222	ng/ul#	94
96) Benzo(g,h,i)perylene	30.633	276	219005	18.118	ng/ul#	89





(#) = qualifier out of range (m) = manual integration (+) = signals summed



#2
 1,4-Dioxane
 Concen: 6.587 ng/uL
 RT: 3.581 min Scan# 1
 Delta R.T. -0.001 min
 Lab File: BG052454.D
 Acq: 22 Feb 2022 18:39




Instrument :
 BNA_G
 ClientSampleId :
 SICV413


Tgt Ion: 88 Resp: 4886
 Ion Ratio Lower Upper
 88 100
 43 38.5 23.0 34.4#
 58 65.5 62.4 93.6



#5
 Pyridine
 Concen: 15.617 ng/uL
 RT: 3.998 min Scan# 87
 Delta R.T. 0.005 min
 Lab File: BG052454.D
 Acq: 22 Feb 2022 18:39

Tgt Ion: 79 Resp: 30326
 Ion Ratio Lower Upper
 79 100
 52 68.8 60.8 91.2
 51 38.1 30.5 45.7









#6
Benzaldehyde
Concen: 22.634 ng/ul
RT: 7.341 min Scan# 656
Delta R.T. -0.001 min
Lab File: BG052454.D
Acq: 22 Feb 2022 18:39




Instrument :
BNA_G
ClientSampleId :
SICV413


Tgt Ion: 77 Resp: 30482
Ion Ratio Lower Upper
77 100
105 91.5 70.4 105.6
106 84.4 61.2 91.8



#8
Phenol
Concen: 17.802 ng/ul
RT: 7.399 min Scan# 666
Delta R.T. 0.005 min
Lab File: BG052454.D
Acq: 22 Feb 2022 18:39

Tgt Ion: 94 Resp: 42699
Ion Ratio Lower Upper
94 100
65 28.0 25.1 37.7
66 42.0 31.6 47.4









#10
Bis(2-Chloroethyl)ether
Concen: 17.642 ng/ul
RT: 7.623 min Scan# 704
Delta R.T. -0.001 min
Lab File: BG052454.D
Acq: 22 Feb 2022 18:39




Instrument :
BNA_G
ClientSampleId :
SICV413


Tgt Ion: 93 Resp: 31300
Ion Ratio Lower Upper
93 100
63 85.7 68.6 103.0
95 34.0 26.1 39.1



#12
2-Chlorophenol
Concen: 17.319 ng/ul
RT: 7.781 min Scan# 731
Delta R.T. -0.001 min
Lab File: BG052454.D
Acq: 22 Feb 2022 18:39

Tgt Ion:128 Resp: 30056
Ion Ratio Lower Upper
128 100
64 50.9 45.3 67.9
130 30.4 27.1 40.7









#13
2-Methylphenol
Concen: 17.045 ng/ul
RT: 8.662 min Scan# 881
Delta R.T. -0.001 min
Lab File: BG052454.D
Acq: 22 Feb 2022 18:39




Instrument :
BNA_G
ClientSampleId :
SICV413


Tgt Ion:108 Resp: 30949
Ion Ratio Lower Upper
108 100
107 93.3 71.8 107.6



#14
2,2'-oxybis(1-Chloropropane)
Concen: 17.600 ng/ul
RT: 8.745 min Scan# 895
Delta R.T. -0.007 min
Lab File: BG052454.D
Acq: 22 Feb 2022 18:39

Tgt Ion: 45 Resp: 42735
Ion Ratio Lower Upper
45 100
77 15.2 10.7 16.1
79 12.3 8.5 12.7









#16
Acetophenone
Concen: 17.339 ng/ul
RT: 9.050 min Scan# 946
Delta R.T. 0.005 min
Lab File: BG052454.D
Acq: 22 Feb 2022 18:39




Instrument :
BNA_G
ClientSampleId :
SICV413


Tgt Ion:105 Resp: 49293
Ion Ratio Lower Upper
105 100
77 88.7 67.3 100.9
51 29.4 24.0 36.0



#17
N-Nitroso-di-n-propylamine
Concen: 17.530 ng/ul
RT: 9.027 min Scan# 943
Delta R.T. -0.001 min
Lab File: BG052454.D
Acq: 22 Feb 2022 18:39

Tgt Ion: 70 Resp: 29340
Ion Ratio Lower Upper
70 100
42 55.4 46.0 69.0
101 9.2 7.0 10.4
130 19.0 13.4 20.2









#18
4-Methylphenol
Concen: 17.773 ng/ul
RT: 8.991 min Scan# 91
Delta R.T. -0.007 min
Lab File: BG052454.D
Acq: 22 Feb 2022 18:39




Instrument :
BNA_G
ClientSampleId :
SICV413


Tgt Ion:108 Resp: 34319
Ion Ratio Lower Upper
108 100
107 107.6 95.0 142.4



#19
Hexachloroethane
Concen: 17.804 ng/ul
RT: 9.326 min Scan# 994
Delta R.T. -0.001 min
Lab File: BG052454.D
Acq: 22 Feb 2022 18:39

Tgt Ion:117 Resp: 12080
Ion Ratio Lower Upper
117 100
201 123.4 77.2 115.8#
199 83.7 49.9 74.9#









#22
Nitrobenzene
Concen: 19.104 ng/ul
RT: 9.438 min Scan# 1013
Delta R.T. -0.001 min
Lab File: BG052454.D
Acq: 22 Feb 2022 18:39




Instrument :
BNA_G
ClientSampleId :
SICV413


Tgt Ion: 77 Resp: 43049
Ion Ratio Lower Upper
77 100
123 36.7 30.1 45.1
65 14.5 9.6 14.4#



#23
Isophorone
Concen: 19.429 ng/ul
RT: 9.961 min Scan# 1102
Delta R.T. -0.001 min
Lab File: BG052454.D
Acq: 22 Feb 2022 18:39

Tgt Ion: 82 Resp: 84466
Ion Ratio Lower Upper
82 100
95 7.1 5.6 8.4
138 16.3 11.8 17.8









#25
2-Nitrophenol
Concen: 18.347 ng/ul
RT: 10.160 min Scan# 1136
Delta R.T. -0.001 min
Lab File: BG052454.D
Acq: 22 Feb 2022 18:39




Instrument :
BNA_G
ClientSampleId :
SICV413


Tgt Ion:139 Resp: 18564
Ion Ratio Lower Upper
139 100
65 57.0 48.0 72.0
109 36.2 28.6 42.8



#26
2,4-Dimethylphenol
Concen: 19.198 ng/ul
RT: 10.213 min Scan# 1145
Delta R.T. -0.001 min
Lab File: BG052454.D
Acq: 22 Feb 2022 18:39

Tgt Ion:107 Resp: 39953
Ion Ratio Lower Upper
107 100
121 49.0 39.3 58.9
122 83.1 63.7 95.5









#27
 Bis(2-Chloroethoxy)methane
 Concen: 19.020 ng/ul
 RT: 10.442 min Scan# 1184
 Delta R.T. -0.001 min
 Lab File: BG052454.D
 Acq: 22 Feb 2022 18:39




Instrument :
 BNA_G
 ClientSampleId :
 SICV413


Tgt Ion: 93 Resp: 44752
 Ion Ratio Lower Upper
 93 100
 95 34.2 26.5 39.7
 123 10.1 8.1 12.1



#29
 2,4-Dichlorophenol
 Concen: 18.845 ng/ul
 RT: 10.707 min Scan# 1229
 Delta R.T. 0.005 min
 Lab File: BG052454.D
 Acq: 22 Feb 2022 18:39

Tgt Ion:162 Resp: 32010
 Ion Ratio Lower Upper
 162 100
 164 61.7 49.4 74.0
 98 36.6 30.4 45.6









#30
Naphthalene
Concen: 18.537 ng/ul
RT: 11.112 min Scan# 11
Delta R.T. -0.001 min
Lab File: BG052454.D
Acq: 22 Feb 2022 18:39




Instrument :
BNA_G
ClientSampleId :
SICV413


Tgt Ion:128 Resp: 106343
Ion Ratio Lower Upper
128 100
129 10.1 8.5 12.7
127 14.4 11.8 17.6



#32
4-Chloroaniline
Concen: 18.684 ng/ul
RT: 11.212 min Scan# 1315
Delta R.T. -0.001 min
Lab File: BG052454.D
Acq: 22 Feb 2022 18:39

Tgt Ion:127 Resp: 47692
Ion Ratio Lower Upper
127 100
129 32.5 26.2 39.2





#33

Hexachlorobutadiene

Concen: 18.310 ng/ul

RT: 11.400 min Scan# 11

Delta R.T. 0.005 min

Lab File: BG052454.D

Acq: 22 Feb 2022 18:39

Instrument :

BNA_G

ClientSampleId :

SICV413





Tgt Ion:225 Resp: 24541

Ion Ratio Lower Upper

225 100

223 66.2 50.6 75.8

227 65.1 49.8 74.8



#34

Caprolactam

Concen: 19.955 ng/ul

RT: 11.970 min Scan# 1444

Delta R.T. 0.010 min

Lab File: BG052454.D

Acq: 22 Feb 2022 18:39




Tgt Ion:113 Resp: 12610


Ion Ratio Lower Upper

113 100

55 147.7 147.0 220.6

56 106.8 109.2 163.8#





#35

4-Chloro-3-methylphenol

Concen: 19.878 ng/ul

RT: 12.322 min Scan# 11

Delta R.T. -0.007 min

Lab File: BG052454.D

Acq: 22 Feb 2022 18:39

Instrument :

BNA_G

ClientSampleId :

SICV413





Tgt Ion:107 Resp: 37712

Ion Ratio Lower Upper

107 100

144 24.3 19.8 29.6

142 76.4 56.2 84.4



#36

2-Methylnaphthalene

Concen: 19.594 ng/ul

RT: 12.710 min Scan# 1570

Delta R.T. 0.005 min

Lab File: BG052454.D




Acq: 22 Feb 2022 18:39


Tgt Ion:142 Resp: 80542

Ion Ratio Lower Upper

142 100

141 87.2 73.3 109.9









#37
1-Methylnaphthalene
Concen: 18.036 ng/ul
RT: 12.927 min Scan# 1
Delta R.T. 0.005 min
Lab File: BG052454.D
Acq: 22 Feb 2022 18:39




Instrument :
BNA_G
ClientSampleId :
SICV413


Tgt Ion:142 Resp: 75449
Ion Ratio Lower Upper
142 100
141 97.0 77.8 116.8
116 4.2 3.3 4.9



#39
1,2,4,5-Tetrachlorobenzene
Concen: 17.756 ng/ul
RT: 13.074 min Scan# 1632
Delta R.T. 0.005 min
Lab File: BG052454.D
Acq: 22 Feb 2022 18:39

Tgt Ion:216 Resp: 47964
Ion Ratio Lower Upper
216 100
214 78.7 61.0 91.4
179 19.4 16.2 24.2
108 15.0 13.1 19.7









#40
 Hexachlorocyclopentadiene
 Concen: 15.461 ng/ul
 RT: 13.057 min Scan# 1629
 Delta R.T. 0.005 min
 Lab File: BG052454.D
 Acq: 22 Feb 2022 18:39




Instrument :
 BNA_G
 ClientSampleId :
 SICV413


Tgt Ion:237 Resp: 20218
 Ion Ratio Lower Upper
 237 100
 235 63.2 49.4 74.0
 272 15.4 11.1 16.7



#41
 2,4,6-Trichlorophenol
 Concen: 18.842 ng/ul
 RT: 13.309 min Scan# 1672
 Delta R.T. -0.001 min
 Lab File: BG052454.D
 Acq: 22 Feb 2022 18:39

Tgt Ion:196 Resp: 30526
 Ion Ratio Lower Upper
 196 100
 198 98.8 74.3 111.5
 200 31.5 23.8 35.6









#42
 2,4,5-Trichlorophenol
 Concen: 18.144 ng/ul
 RT: 13.385 min Scan# 1
 Delta R.T. -0.001 min
 Lab File: BG052454.D
 Acq: 22 Feb 2022 18:39




Instrument :
 BNA_G
 ClientSampleId :
 SICV413


Tgt Ion:196 Resp: 31331
 Ion Ratio Lower Upper
 196 100
 198 100.6 77.0 115.4
 200 30.6 25.4 38.2



#43
 1,1'-Biphenyl
 Concen: 17.676 ng/ul
 RT: 13.703 min Scan# 1739
 Delta R.T. 0.005 min
 Lab File: BG052454.D
 Acq: 22 Feb 2022 18:39

Tgt Ion:154 Resp: 108180
 Ion Ratio Lower Upper
 154 100
 153 38.5 34.5 51.7
 76 10.0 13.1 19.7#









#44
2-Chloronaphthalene
Concen: 18.984 ng/ul
RT: 13.750 min Scan# 1746
Delta R.T. 0.005 min
Lab File: BG052454.D
Acq: 22 Feb 2022 18:39




Instrument :
BNA_G
ClientSampleId :
SICV413


Tgt Ion:162 Resp: 88361
Ion Ratio Lower Upper
162 100
164 32.1 26.1 39.1
127 39.0 31.0 46.4



#45
2-Nitroaniline
Concen: 19.269 ng/ul
RT: 13.944 min Scan# 1780
Delta R.T. -0.001 min
Lab File: BG052454.D
Acq: 22 Feb 2022 18:39

Tgt Ion: 65 Resp: 29292
Ion Ratio Lower Upper
65 100
92 65.2 53.0 79.4
138 96.0 67.4 101.0





#47

Dimethylphthalate

Concen: 19.099 ng/ul

RT: 14.308 min Scan# 11

Delta R.T. 0.005 min

Lab File: BG052454.D

Acq: 22 Feb 2022 18:39

Instrument :

BNA_G

ClientSampleId :

SICV413





Tgt Ion:163 Resp: 116823

Ion Ratio Lower Upper

163 100

194 4.3 3.0 4.4

164 10.0 8.2 12.4



#48

2,6-Dinitrotoluene

Concen: 19.448 ng/ul

RT: 14.425 min Scan# 1862

Delta R.T. -0.007 min

Lab File: BG052454.D

Acq: 22 Feb 2022 18:39




Tgt Ion:165 Resp: 26050


Ion Ratio Lower Upper

165 100

89 58.2 57.4 86.0

121 20.3 18.8 28.2









#50
Acenaphthylene
Concen: 19.118 ng/ul
RT: 14.590 min Scan# 1890
Delta R.T. -0.001 min
Lab File: BG052454.D
Acq: 22 Feb 2022 18:39




Instrument :
BNA_G
ClientSampleId :
SICV413


Tgt Ion:152 Resp: 147580
Ion Ratio Lower Upper
152 100
151 20.2 17.4 26.0
153 12.3 10.3 15.5



#51
3-Nitroaniline
Concen: 24.788 ng/ul
RT: 14.760 min Scan# 1919
Delta R.T. -0.001 min
Lab File: BG052454.D
Acq: 22 Feb 2022 18:39

Tgt Ion:138 Resp: 27040
Ion Ratio Lower Upper
138 100
108 8.0 8.8 13.2#
92 115.2 98.0 147.0









#52
Acenaphthene
Concen: 18.616 ng/ul
RT: 14.930 min Scan# 1948
Delta R.T. -0.001 min
Lab File: BG052454.D
Acq: 22 Feb 2022 18:39




Instrument :
BNA_G
ClientSampleId :
SICV413


Tgt Ion:153 Resp: 94505
Ion Ratio Lower Upper
153 100
152 47.6 36.5 54.7
154 88.2 65.8 98.6



#53
2,4-Dinitrophenol
Concen: 15.631 ng/ul
RT: 14.977 min Scan# 1956
Delta R.T. 0.005 min
Lab File: BG052454.D
Acq: 22 Feb 2022 18:39

Tgt Ion:184 Resp: 10353
Ion Ratio Lower Upper
184 100
63 57.9 66.2 99.2#
154 62.7 53.6 80.4









#55
4-Nitrophenol
Concen: 17.885 ng/ul
RT: 15.071 min Scan# 1972
Delta R.T. -0.001 min
Lab File: BG052454.D
Acq: 22 Feb 2022 18:39




Instrument :
BNA_G
ClientSampleId :
SICV413


Tgt Ion:109 Resp: 17254
Ion Ratio Lower Upper
109 100
139 113.1 88.7 133.1
65 130.6 113.6 170.4



#56
Dibenzofuran
Concen: 18.831 ng/ul
RT: 15.265 min Scan# 2005
Delta R.T. 0.005 min
Lab File: BG052454.D
Acq: 22 Feb 2022 18:39

Tgt Ion:168 Resp: 137426
Ion Ratio Lower Upper
168 100
139 38.2 31.2 46.8









#57
2,4-Dinitrotoluene
Concen: 19.491 ng/ul
RT: 15.218 min Scan# 1996
Delta R.T. 0.005 min
Lab File: BG052454.D
Acq: 22 Feb 2022 18:39




Instrument :
BNA_G
ClientSampleId :
SICV413


Tgt Ion:165 Resp: 37818
Ion Ratio Lower Upper
165 100
63 40.7 42.1 63.1#
182 2.4 1.8 2.8



#58
2,3,4,6-Tetrachlorophenol
Concen: 19.938 ng/ul
RT: 15.494 min Scan# 2044
Delta R.T. -0.001 min
Lab File: BG052454.D
Acq: 22 Feb 2022 18:39

Tgt Ion:232 Resp: 31203
Ion Ratio Lower Upper
232 100
131 35.0 38.9 58.3#
130 2.5 2.4 3.6
166 26.5 25.8 38.8









#59
Diethylphthalate
Concen: 19.241 ng/ul
RT: 15.659 min Scan# 2072
Delta R.T. -0.001 min
Lab File: BG052454.D
Acq: 22 Feb 2022 18:39




Instrument :
BNA_G
ClientSampleId :
SICV413


Tgt Ion:149 Resp: 118968
Ion Ratio Lower Upper
149 100
177 23.0 16.8 25.2
150 12.0 10.1 15.1



#61
Fluorene
Concen: 18.905 ng/ul
RT: 15.911 min Scan# 2115
Delta R.T. -0.001 min
Lab File: BG052454.D
Acq: 22 Feb 2022 18:39

Tgt Ion:166 Resp: 116055
Ion Ratio Lower Upper
166 100
165 102.8 81.0 121.6
167 15.0 11.0 16.4









#62
4-Chlorophenyl-phenylether
Concen: 18.783 ng/ul
RT: 15.894 min Scan# 2112
Delta R.T. -0.001 min
Lab File: BG052454.D
Acq: 22 Feb 2022 18:39




Instrument :
BNA_G
ClientSampleId :
SICV413


Tgt Ion:204 Resp: 62753
Ion Ratio Lower Upper
204 100
206 35.2 27.4 41.2
141 57.6 48.6 73.0



#63
4-Nitroaniline
Concen: 26.504 ng/ul
RT: 15.923 min Scan# 2117
Delta R.T. 0.005 min
Lab File: BG052454.D
Acq: 22 Feb 2022 18:39

Tgt Ion:138 Resp: 27759
Ion Ratio Lower Upper
138 100
92 50.3 49.3 73.9
108 93.9 72.6 108.8









#66
 4,6-Dinitro-2-methylphenol
 Concen: 17.683 ng/ul
 RT: 15.988 min Scan# 2128
 Delta R.T. -0.001 min
 Lab File: BG052454.D
 Acq: 22 Feb 2022 18:39




Instrument :
 BNA_G
 ClientSampleId :
 SICV413


Tgt Ion:198 Resp: 20897
 Ion Ratio Lower Upper
 198 100
 182 4.8 3.8 5.8
 77 21.5 23.2 34.8#



#67
 N-Nitrosodiphenylamine
 Concen: 19.117 ng/ul
 RT: 16.105 min Scan# 2148
 Delta R.T. -0.001 min
 Lab File: BG052454.D
 Acq: 22 Feb 2022 18:39

Tgt Ion:169 Resp: 103582
 Ion Ratio Lower Upper
 169 100
 168 67.5 58.5 87.7
 167 36.5 31.3 46.9









#68
4-Bromophenyl-phenylether
Concen: 18.989 ng/ul
RT: 16.793 min Scan# 21
Delta R.T. -0.001 min
Lab File: BG052454.D
Acq: 22 Feb 2022 18:39




Instrument :
BNA_G
ClientSampleId :
SICV413


Tgt Ion:248 Resp: 43829
Ion Ratio Lower Upper
248 100
250 95.6 75.5 113.3
141 59.9 67.6 101.4#



#69
Hexachlorobenzene
Concen: 18.828 ng/ul
RT: 16.928 min Scan# 2288
Delta R.T. 0.005 min
Lab File: BG052454.D
Acq: 22 Feb 2022 18:39

Tgt Ion:284 Resp: 49864
Ion Ratio Lower Upper
284 100
142 22.6 31.9 47.9#
249 28.3 25.8 38.6









#70
 Atrazine
 Concen: 18.316 ng/ul
 RT: 17.045 min Scan# 21
 Delta R.T. -0.001 min
 Lab File: BG052454.D
 Acq: 22 Feb 2022 18:39




Instrument :
 BNA_G
 ClientSampleId :
 SICV413


Tgt Ion:200 Resp: 43739
 Ion Ratio Lower Upper
 200 100
 173 25.8 20.2 30.2
 215 49.9 38.1 57.1



#71
 Pentachlorophenol
 Concen: 13.427 ng/ul
 RT: 17.274 min Scan# 2347
 Delta R.T. 0.005 min
 Lab File: BG052454.D
 Acq: 22 Feb 2022 18:39

Tgt Ion:266 Resp: 16016
 Ion Ratio Lower Upper
 266 100
 264 63.6 54.3 81.5
 268 60.4 51.0 76.6









#72
Phenanthrene
Concen: 18.916 ng/ul
RT: 17.656 min Scan# 2412
Delta R.T. -0.001 min
Lab File: BG052454.D
Acq: 22 Feb 2022 18:39




Instrument :
BNA_G
ClientSampleId :
SICV413


Tgt Ion:178 Resp: 195925
Ion Ratio Lower Upper
178 100
179 16.9 12.3 18.5
176 18.7 15.5 23.3



#74
Anthracene
Concen: 18.745 ng/ul
RT: 17.750 min Scan# 2428
Delta R.T. -0.001 min
Lab File: BG052454.D
Acq: 22 Feb 2022 18:39

Tgt Ion:178 Resp: 195405
Ion Ratio Lower Upper
178 100
179 16.0 13.0 19.6
176 18.5 15.6 23.4









#75
1,2,3,4-Tetrachlorobenzene
Concen: 15.899 ng/uL
RT: 13.673 min Scan# 1
Delta R.T. -0.001 min
Lab File: BG052454.D
Acq: 22 Feb 2022 18:39




Instrument :
BNA_G
ClientSampleId :
SICV413


Tgt Ion:216 Resp: 48368
Ion Ratio Lower Upper
216 100
214 77.7 61.3 91.9
218 48.2 37.0 55.6



#76
Pentachlorobenzene
Concen: 18.817 ng/uL
RT: 15.189 min Scan# 1992
Delta R.T. -0.001 min
Lab File: BG052454.D
Acq: 22 Feb 2022 18:39

Tgt Ion:250 Resp: 54652
Ion Ratio Lower Upper
250 100
248 62.8 50.5 75.7
252 63.3 51.2 76.8









#77
 Carbazole
 Concen: 19.453 ng/ul
 RT: 18.015 min Scan# 2473
 Delta R.T. -0.001 min
 Lab File: BG052454.D
 Acq: 22 Feb 2022 18:39




Instrument :
 BNA_G
 ClientSampleId :
 SICV413


Tgt Ion:167 Resp: 180116
 Ion Ratio Lower Upper
 167 100
 166 22.1 17.3 25.9
 139 11.3 10.6 16.0



#78
 Di-n-butylphthalate
 Concen: 19.105 ng/ul
 RT: 18.555 min Scan# 2565
 Delta R.T. -0.001 min
 Lab File: BG052454.D
 Acq: 22 Feb 2022 18:39

Tgt Ion:149 Resp: 210378
 Ion Ratio Lower Upper
 149 100
 150 9.6 7.5 11.3
 104 6.1 5.0 7.6









#80
Fluoranthene
Concen: 18.752 ng/u1
RT: 19.665 min Scan# 21
Delta R.T. 0.005 min
Lab File: BG052454.D
Acq: 22 Feb 2022 18:39




Instrument :
BNA_G
ClientSampleId :
SICV413


Tgt Ion:202 Resp: 246318
Ion Ratio Lower Upper
202 100
101 7.6 9.3 13.9#
100 6.2 7.0 10.6#



#82
Pyrene
Concen: 18.949 ng/u1
RT: 20.024 min Scan# 2815
Delta R.T. -0.001 min
Lab File: BG052454.D
Acq: 22 Feb 2022 18:39

Tgt Ion:202 Resp: 247170
Ion Ratio Lower Upper
202 100
101 9.6 10.7 16.1#
100 8.4 9.0 13.6#









#83
 Butylbenzylphthalate
 Concen: 18.396 ng/ul
 RT: 20.893 min Scan# 2963
 Delta R.T. 0.005 min
 Lab File: BG052454.D
 Acq: 22 Feb 2022 18:39




Instrument :
 BNA_G
 ClientSampleId :
 SICV413


Tgt Ion:149 Resp: 86749
 Ion Ratio Lower Upper
 149 100
 91 71.6 65.2 97.8
 206 25.3 16.2 24.2#



#84
 3,3'-Dichlorobenzidine
 Concen: 22.498 ng/ul
 RT: 21.809 min Scan# 3119
 Delta R.T. -0.001 min
 Lab File: BG052454.D
 Acq: 22 Feb 2022 18:39

Tgt Ion:252 Resp: 94496
 Ion Ratio Lower Upper
 252 100
 254 64.6 50.8 76.2
 126 7.0 9.0 13.4#









#85
Benzo(a)anthracene
Concen: 19.328 ng/ul
RT: 21.909 min Scan# 3136
Delta R.T. -0.001 min
Lab File: BG052454.D
Acq: 22 Feb 2022 18:39




Instrument :
BNA_G
ClientSampleId :
SICV413


Tgt Ion:228 Resp: 249019
Ion Ratio Lower Upper
228 100
229 20.7 15.7 23.5
226 26.9 22.3 33.5



#86
Bis(2-ethylhexyl)phthalate
Concen: 18.822 ng/ul
RT: 21.786 min Scan# 3115
Delta R.T. -0.001 min
Lab File: BG052454.D
Acq: 22 Feb 2022 18:39

Tgt Ion:149 Resp: 124412
Ion Ratio Lower Upper
149 100
167 26.9 21.1 31.7
279 4.4 2.7 4.1#









#87
Chrysene
Concen: 18.580 ng/ul
RT: 21.980 min Scan# 3148
Delta R.T. -0.001 min
Lab File: BG052454.D
Acq: 22 Feb 2022 18:39




Instrument :
BNA_G
ClientSampleId :
SICV413


Tgt Ion:228 Resp: 229488
Ion Ratio Lower Upper
228 100
226 29.8 24.8 37.2
229 19.0 15.8 23.6



#89
Di-n-octyl phthalate
Concen: 18.542 ng/ul
RT: 23.078 min Scan# 3335
Delta R.T. -0.001 min
Lab File: BG052454.D
Acq: 22 Feb 2022 18:39

Tgt Ion:149 Resp: 206963









#90
Benzo(b)fluoranthene
Concen: 18.740 ng/ul
RT: 24.288 min Scan# 3540
Delta R.T. 0.005 min
Lab File: BG052454.D
Acq: 22 Feb 2022 18:39




Instrument :
BNA_G
ClientSampleId :
SICV413


Tgt Ion:252 Resp: 252208
Ion Ratio Lower Upper
252 100
253 22.1 17.8 26.6
125 6.1 8.5 12.7#



#91
Benzo(k)fluoranthene
Concen: 19.313 ng/ul
RT: 24.359 min Scan# 3553
Delta R.T. -0.001 min
Lab File: BG052454.D
Acq: 22 Feb 2022 18:39

Tgt Ion:252 Resp: 239737
Ion Ratio Lower Upper
252 100
253 21.9 17.3 25.9
125 7.0 7.8 11.6#









#93
Benzo(a)pyrene
Concen: 18.408 ng/u1
RT: 25.228 min Scan# 31
Delta R.T. -0.007 min
Lab File: BG052454.D
Acq: 22 Feb 2022 18:39




Instrument :
BNA_G
ClientSampleId :
SICV413


Tgt Ion:252 Resp: 233053
Ion Ratio Lower Upper
252 100
253 22.8 17.4 26.2
125 7.4 9.1 13.7#



#94
Indeno(1,2,3-cd)pyrene
Concen: 9.007 ng/u1
RT: 29.381 min Scan# 4408
Delta R.T. -0.007 min
Lab File: BG052454.D
Acq: 22 Feb 2022 18:39

Tgt Ion:276 Resp: 129671
Ion Ratio Lower Upper
276 100
138 11.8 15.5 23.3#
277 24.0 20.5 30.7









#95
 Dibenzo(a,h)anthracene
 Concen: 18.222 ng/ul
 RT: 29.446 min Scan# 4419
 Delta R.T. -0.013 min
 Lab File: BG052454.D
 Acq: 22 Feb 2022 18:39

Instrument :
 BNA_G
 ClientSampleId :
 SICV413

Tgt Ion:278 Resp: 223938
 Ion Ratio Lower Upper
 278 100
 139 8.6 12.2 18.4#
 279 24.3 19.5 29.3



#96
 Benzo(g,h,i)perylene
 Concen: 18.118 ng/ul
 RT: 30.633 min Scan# 4621
 Delta R.T. -0.007 min
 Lab File: BG052454.D
 Acq: 22 Feb 2022 18:39

Tgt Ion:276 Resp: 219005
 Ion Ratio Lower Upper
 276 100
 138 12.0 16.6 24.8#
 277 23.8 17.6 26.4

