

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG022323\
 Data File : BG056707.D
 Acq On : 23 Feb 2023 10:50
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SSTD020591

Quant Time: Feb 23 13:52:41 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG022323.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Feb 23 13:51:30 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.428	152	18755	20.000	ng/u1	0.00
20) Naphthalene-d8	11.266	136	83751	20.000	ng/u1	0.00
38) Acenaphthene-d10	15.032	164	69043	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.764	188	176003	20.000	ng/u1	0.00
79) Chrysene-d12	22.077	240	132537	20.000	ng/u1	0.00
88) Perylene-d12	25.614	264	144256	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.710	96	2921	7.168	ng/uL	0.00
4) Pyridine-d5	4.145	84	24407	16.493	ng/u1	0.00
7) Phenol-d5	7.547	99	33612	18.680	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.729	67	20853	19.370	ng/u1	0.00
11) 2-Chlorophenol-d4	7.952	132	23153	21.360	ng/u1	0.00
15) 4-Methylphenol-d8	9.115	113	26368	19.721	ng/u1	0.00
21) Nitrobenzene-d5	9.597	128	13005	24.900	ng/u1	0.00
24) 2-Nitrophenol-d4	10.332	143	15120	24.726	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.866	165	29459	21.029	ng/u1	0.00
31) 4-Chloroaniline-d4	11.383	131	39291	19.453	ng/u1	0.00
46) Dimethylphthalate-d6	14.415	166	100690	19.596	ng/u1	0.00
49) Acenaphthylene-d8	14.726	160	110285	18.575	ng/u1	0.00
54) 4-Nitrophenol-d4	15.185	143	17000	21.218	ng/u1	0.00
60) Fluorene-d10	16.013	176	95342	18.849	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	16.113	200	19236	22.784	ng/u1	0.00
73) Anthracene-d10	17.864	188	150036	18.575	ng/u1	0.00
81) Pyrene-d10	20.120	212	159519	20.966	ng/u1	0.00
92) Benzo(a)pyrene-d12	25.367	264	140065	18.916	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.739	88	2876	5.753	ng/uL#	83
5) Pyridine	4.162	79	25370	16.732	ng/u1	96
6) Benzaldehyde	7.552	77	16922	20.375	ng/u1	98
8) Phenol	7.576	94	36021	19.701	ng/u1	97
10) Bis(2-Chloroethyl)ether	7.823	93	25032	19.028	ng/u1	96
12) 2-Chlorophenol	7.981	128	23376	21.157	ng/u1	95
13) 2-Methylphenol	8.851	108	25494	18.337	ng/u1	100
14) 2,2'-oxybis(1-Chloropr...	8.945	45	34424	19.703	ng/u1	100
16) Acetophenone	9.250	105	46380	19.422	ng/u1	96
17) N-Nitroso-di-n-propyla...	9.227	70	25768	20.028	ng/u1	95
18) 4-Methylphenol	9.180	108	28863	19.492	ng/u1	88
19) Hexachloroethane	9.527	117	9759	22.530	ng/u1	92
22) Nitrobenzene	9.644	77	37893	25.972	ng/u1	95
23) Isophorone	10.167	82	72804	19.435	ng/u1	98
25) 2-Nitrophenol	10.361	139	15237	24.401	ng/u1	99
26) 2,4-Dimethylphenol	10.402	107	34145	20.162	ng/u1	96
27) Bis(2-Chloroethoxy)met...	10.637	93	36825	19.002	ng/u1	100
29) 2,4-Dichlorophenol	10.896	162	28365	20.568	ng/u1	99
30) Naphthalene	11.319	128	85960	19.477	ng/u1	99
32) 4-Chloroaniline	11.413	127	39461	19.823	ng/u1	99
33) Hexachlorobutadiene	11.595	225	22113	18.396	ng/u1	90
34) Caprolactam	12.147	113	10373	21.898	ng/u1	82
35) 4-Chloro-3-methylphenol	12.494	107	32781	21.465	ng/u1	97
36) 2-Methylnaphthalene	12.887	142	61455	18.903	ng/u1	100

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG022323\
 Data File : BG056707.D
 Acq On : 23 Feb 2023 10:50
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1


Instrument :
 BNA_G
 ClientSampleId :
 SSTD020591


Quant Time: Feb 23 13:52:41 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG022323.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Feb 23 13:51:30 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	13.105	142	62847	18.855	ng/ul	96
39) 1,2,4,5-Tetrachloroben...	13.246	216	44590	18.639	ng/ul	93
40) Hexachlorocyclopentadiene	13.222	237	28487	19.542	ng/ul	99
41) 2,4,6-Trichlorophenol	13.475	196	29082	21.313	ng/ul	98
42) 2,4,5-Trichlorophenol	13.545	196	30869	19.549	ng/ul	98
43) 1,1'-Biphenyl	13.869	154	90694	18.835	ng/ul	98
44) 2-Chloronaphthalene	13.921	162	74550	18.567	ng/ul	94
45) 2-Nitroaniline	14.104	65	26409	30.870	ng/ul	93
47) Dimethylphthalate	14.462	163	101348	19.448	ng/ul#	98
48) 2,6-Dinitrotoluene	14.585	165	21138	25.130	ng/ul	96
50) Acenaphthylene	14.756	152	115697	18.965	ng/ul	98
51) 3-Nitroaniline	14.920	138	17863	21.602	ng/ul#	90
52) Acenaphthene	15.091	153	79503	18.783	ng/ul	96
53) 2,4-Dinitrophenol	15.120	184	11833	19.383	ng/ul	98
55) 4-Nitrophenol	15.202	109	19777	25.704	ng/ul	97
56) Dibenzofuran	15.426	168	118049	18.373	ng/ul	93
57) 2,4-Dinitrotoluene	15.367	165	30810	25.376	ng/ul	94
58) 2,3,4,6-Tetrachlorophenol	15.643	232	29248	21.402	ng/ul#	98
59) Diethylphthalate	15.807	149	100696	20.067	ng/ul	99
61) Fluorene	16.066	166	96177	18.584	ng/ul	99
62) 4-Chlorophenyl-phenyle...	16.048	204	56217	18.232	ng/ul	98
63) 4-Nitroaniline	16.072	138	17327	21.389	ng/ul	97
66) 4,6-Dinitro-2-methylph...	16.131	198	19399	23.296	ng/ul#	95
67) N-Nitrosodiphenylamine	16.260	169	86217	19.052	ng/ul	99
68) 4-Bromophenyl-phenylether	16.941	248	39466	19.892	ng/ul	94
69) Hexachlorobenzene	17.071	284	41725	18.467	ng/ul	97
70) Atrazine	17.188	200	38496	20.438	ng/ul	95
71) Pentachlorophenol	17.406	266	23766	19.537	ng/ul	95
72) Phenanthrene	17.805	178	169265	18.825	ng/ul	99
74) Anthracene	17.899	178	169442	18.654	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.845	216	44394	18.550	ng/ul	98
76) Pentachlorobenzene	15.343	250	45665	18.183	ng/ul	95
77) Carbazole	18.158	167	144066	18.701	ng/ul	100
78) Di-n-butylphthalate	18.686	149	162878	21.058	ng/ul	99
80) Fluoranthene	19.791	202	195133	21.744	ng/ul	98
82) Pyrene	20.149	202	192706	21.551	ng/ul	99
83) Butylbenzylphthalate	21.007	149	61672	25.011	ng/ul	96
84) 3,3'-Dichlorobenzidine	21.947	252	60212	20.975	ng/ul	94
85) Benzo(a)anthracene	22.053	228	170538	18.670	ng/ul	98
86) Bis(2-ethylhexyl)phtha...	21.918	149	81966	21.997	ng/ul	97
87) Chrysene	22.129	228	158787	18.898	ng/ul	99
89) Di-n-octyl phthalate	23.234	149	142251	21.315	ng/ul	100
90) Benzo(b)fluoranthene	24.480	252	172497	18.369	ng/ul	100
91) Benzo(k)fluoranthene	24.550	252	166342	18.403	ng/ul	97
93) Benzo(a)pyrene	25.443	252	150283	18.782	ng/ul	99
94) Indeno(1,2,3-cd)pyrene	29.685	276	205117	18.613	ng/ul	97
95) Dibenzo(a,h)anthracene	29.756	278	171281	18.797	ng/ul	92
96) Benzo(g,h,i)perylene	30.972	276	98057	11.162	ng/ul	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Feb 23 13:52:41 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG022323.M
Quant Title : SVOA CALIBRATION
QLast Update : Thu Feb 23 13:51:30 2023
Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 5.753 ng/uL
RT: 3.739 min Scan# 34
Delta R.T. -0.002 min
Lab File: BG056707.D
Acq: 23 Feb 2023 10:50




Instrument :
BNA_G
ClientSampleId :
SSTD020591


Tgt Ion: 88 Resp: 2876
Ion Ratio Lower Upper
88 100
43 28.4 28.3 42.5
58 79.0 50.6 75.8#



#5
Pyridine
Concen: 16.732 ng/uL
RT: 4.162 min Scan# 106
Delta R.T. -0.002 min
Lab File: BG056707.D
Acq: 23 Feb 2023 10:50

Tgt Ion: 79 Resp: 25370
Ion Ratio Lower Upper
79 100
52 73.5 60.6 91.0
51 36.5 31.9 47.9









#6
Benzaldehyde
Concen: 20.375 ng/ul
RT: 7.552 min Scan# 682
Delta R.T. 0.004 min
Lab File: BG056707.D
Acq: 23 Feb 2023 10:50




Instrument :
BNA_G
ClientSampleId :
SSTD020591


Tgt Ion: 77 Resp: 16922
Ion Ratio Lower Upper
77 100
105 90.2 72.2 108.4
106 81.1 67.4 101.2



#8
Phenol
Concen: 19.701 ng/ul
RT: 7.576 min Scan# 687
Delta R.T. 0.004 min
Lab File: BG056707.D
Acq: 23 Feb 2023 10:50

Tgt Ion: 94 Resp: 36021
Ion Ratio Lower Upper
94 100
65 34.6 25.7 38.5
66 49.5 38.9 58.3









#10
 Bis(2-Chloroethyl)ether
 Concen: 19.028 ng/ul
 RT: 7.823 min Scan# 710
 Delta R.T. -0.002 min
 Lab File: BG056707.D
 Acq: 23 Feb 2023 10:50




Instrument :
 BNA_G
 ClientSampleId :
 SST020591


Tgt Ion: 93	Resp:	25032
Ion	Ratio	Lower Upper
93	100	
63	87.1	66.6 100.0
95	32.5	25.0 37.4



#12
 2-Chlorophenol
 Concen: 21.157 ng/ul
 RT: 7.981 min Scan# 756
 Delta R.T. 0.004 min
 Lab File: BG056707.D
 Acq: 23 Feb 2023 10:50

Tgt Ion: 128	Resp:	23376
Ion	Ratio	Lower Upper
128	100	
64	55.0	45.5 68.3
130	29.1	27.7 41.5









#13
2-Methylphenol
Concen: 18.337 ng/ul
RT: 8.851 min Scan# 90
Delta R.T. -0.002 min
Lab File: BG056707.D
Acq: 23 Feb 2023 10:50




Instrument :
BNA_G
ClientSampleId :
SSTD020591


Tgt Ion:108 Resp: 25494
Ion Ratio Lower Upper
108 100
107 92.1 73.4 110.0



#14
2,2'-oxybis(1-Chloropropane)
Concen: 19.703 ng/ul
RT: 8.945 min Scan# 920
Delta R.T. -0.002 min
Lab File: BG056707.D
Acq: 23 Feb 2023 10:50

Tgt Ion: 45 Resp: 34424
Ion Ratio Lower Upper
45 100
77 15.7 12.4 18.6
79 12.1 9.8 14.6









#16
Acetophenone
Concen: 19.422 ng/ul
RT: 9.250 min Scan# 911
Delta R.T. -0.002 min
Lab File: BG056707.D
Acq: 23 Feb 2023 10:50




Instrument :
BNA_G
ClientSampleId :
SSTD020591


Tgt Ion:105 Resp: 46380
Ion Ratio Lower Upper
105 100
77 84.2 71.4 107.0
51 29.0 23.6 35.4



#17
N-Nitroso-di-n-propylamine
Concen: 20.028 ng/ul
RT: 9.227 min Scan# 968
Delta R.T. 0.004 min
Lab File: BG056707.D
Acq: 23 Feb 2023 10:50

Tgt Ion: 70 Resp: 25768
Ion Ratio Lower Upper
70 100
42 54.5 46.6 70.0
101 7.8 7.3 10.9
130 19.5 13.7 20.5









#18
4-Methylphenol
Concen: 19.492 ng/ul
RT: 9.180 min Scan# 918
Delta R.T. 0.004 min
Lab File: BG056707.D
Acq: 23 Feb 2023 10:50




Instrument :
BNA_G
ClientSampleId :
SSTD020591


Tgt Ion:108 Resp: 28863
Ion Ratio Lower Upper
108 100
107 102.6 92.1 138.1



#19
Hexachloroethane
Concen: 22.530 ng/ul
RT: 9.527 min Scan# 1019
Delta R.T. -0.008 min
Lab File: BG056707.D
Acq: 23 Feb 2023 10:50

Tgt Ion:117 Resp: 9759
Ion Ratio Lower Upper
117 100
201 131.9 95.2 142.8
199 71.5 57.9 86.9









#22
Nitrobenzene
Concen: 25.972 ng/ul
RT: 9.644 min Scan# 1038
Delta R.T. 0.004 min
Lab File: BG056707.D
Acq: 23 Feb 2023 10:50




Instrument :
BNA_G
ClientSampleId :
SSTD020591


Tgt Ion: 77 Resp: 37893
Ion Ratio Lower Upper
77 100
123 37.6 27.8 41.6
65 15.3 10.6 16.0



#23
Isophorone
Concen: 19.435 ng/ul
RT: 10.167 min Scan# 1128
Delta R.T. 0.004 min
Lab File: BG056707.D
Acq: 23 Feb 2023 10:50

Tgt Ion: 82 Resp: 72804
Ion Ratio Lower Upper
82 100
95 8.4 7.7 11.5
138 14.5 12.1 18.1









#25
2-Nitrophenol
Concen: 24.401 ng/ul
RT: 10.361 min Scan# 1161
Delta R.T. -0.002 min
Lab File: BG056707.D
Acq: 23 Feb 2023 10:50




Instrument :
BNA_G
ClientSampleId :
SSTD020591


Tgt Ion:139 Resp: 15237
Ion Ratio Lower Upper
139 100
65 62.3 49.4 74.2
109 40.5 31.6 47.4



#26
2,4-Dimethylphenol
Concen: 20.162 ng/ul
RT: 10.402 min Scan# 1168
Delta R.T. 0.004 min
Lab File: BG056707.D
Acq: 23 Feb 2023 10:50

Tgt Ion:107 Resp: 34145
Ion Ratio Lower Upper
107 100
121 43.9 40.2 60.4
122 78.9 63.1 94.7





#27

Bis(2-Chloroethoxy)methane

Concen: 19.002 ng/ul

RT: 10.637 min Scan# 1208

Delta R.T. -0.002 min

Lab File: BG056707.D

Acq: 23 Feb 2023 10:50

Instrument :

BNA_G

ClientSampleId :

SSTD020591





Tgt Ion: 93 Resp: 36825

Ion Ratio Lower Upper

93 100

95 30.8 24.6 37.0

123 7.9 6.6 10.0



#29

2,4-Dichlorophenol

Concen: 20.568 ng/ul

RT: 10.896 min Scan# 1252

Delta R.T. -0.002 min

Lab File: BG056707.D

Acq: 23 Feb 2023 10:50




Tgt Ion:162 Resp: 28365


Ion Ratio Lower Upper

162 100

164 65.4 53.3 79.9

98 31.3 25.5 38.3









#30
Naphthalene
Concen: 19.477 ng/ul
RT: 11.319 min Scan# 11319
Delta R.T. 0.004 min
Lab File: BG056707.D
Acq: 23 Feb 2023 10:50




Instrument :
BNA_G
ClientSampleId :
SSTD020591


Tgt Ion:128 Resp: 85960
Ion Ratio Lower Upper
128 100
129 10.7 8.6 12.8
127 13.3 9.8 14.8



#32
4-Chloroaniline
Concen: 19.823 ng/ul
RT: 11.413 min Scan# 1340
Delta R.T. 0.004 min
Lab File: BG056707.D
Acq: 23 Feb 2023 10:50

Tgt Ion:127 Resp: 39461
Ion Ratio Lower Upper
127 100
129 28.7 23.4 35.0









#33
Hexachlorobutadiene
Concen: 18.396 ng/ul
RT: 11.595 min Scan# 11
Delta R.T. 0.004 min
Lab File: BG056707.D
Acq: 23 Feb 2023 10:50




Instrument :
BNA_G
ClientSampleId :
SSTD020591


Tgt Ion:225 Resp: 22113
Ion Ratio Lower Upper
225 100
223 64.4 53.5 80.3
227 59.4 58.2 87.2



#34
Caprolactam
Concen: 21.898 ng/ul
RT: 12.147 min Scan# 1465
Delta R.T. -0.002 min
Lab File: BG056707.D
Acq: 23 Feb 2023 10:50

Tgt Ion:113 Resp: 10373
Ion Ratio Lower Upper
113 100
55 178.4 167.0 250.6
56 123.4 113.5 170.3









#35
4-Chloro-3-methylphenol
Concen: 21.465 ng/ul
RT: 12.494 min Scan# 1523
Delta R.T. 0.004 min
Lab File: BG056707.D
Acq: 23 Feb 2023 10:50




Instrument :
BNA_G
ClientSampleId :
SSTD020591


Tgt Ion:107 Resp: 32781
Ion Ratio Lower Upper
107 100
144 23.7 20.5 30.7
142 71.7 55.8 83.6



#36
2-Methylnaphthalene
Concen: 18.903 ng/ul
RT: 12.887 min Scan# 1591
Delta R.T. 0.004 min
Lab File: BG056707.D
Acq: 23 Feb 2023 10:50

Tgt Ion:142 Resp: 61455
Ion Ratio Lower Upper
142 100
141 88.8 71.0 106.6





#37

1-Methylnaphthalene

Concen: 18.855 ng/ul

RT: 13.105 min Scan# 1

Delta R.T. 0.004 min

Lab File: BG056707.D

Acq: 23 Feb 2023 10:50

Instrument :

BNA_G

ClientSampleId :

SSTD020591





Tgt Ion:142 Resp: 62847

Ion Ratio Lower Upper

142 100

141 95.7 73.5 110.3

116 4.2 4.0 6.0



#39

1,2,4,5-Tetrachlorobenzene

Concen: 18.639 ng/ul

RT: 13.246 min Scan# 1652

Delta R.T. 0.004 min

Lab File: BG056707.D

Acq: 23 Feb 2023 10:50

Tgt Ion:216 Resp: 44590




Ion Ratio Lower Upper


216 100

214 73.8 64.4 96.6

179 17.8 16.9 25.3

108 12.5 11.6 17.4









#40
Hexachlorocyclopentadiene
Concen: 19.542 ng/ul
RT: 13.222 min Scan# 1648
Delta R.T. -0.002 min
Lab File: BG056707.D
Acq: 23 Feb 2023 10:50




Instrument :
BNA_G
ClientSampleId :
SSTD020591


Tgt Ion:237	Resp:	28487
Ion Ratio	Lower	Upper
237	100	
235	60.5	47.8 71.8
272	12.3	9.9 14.9



#41
2,4,6-Trichlorophenol
Concen: 21.313 ng/ul
RT: 13.475 min Scan# 1691
Delta R.T. 0.004 min
Lab File: BG056707.D
Acq: 23 Feb 2023 10:50

Tgt Ion:196	Resp:	29082
Ion Ratio	Lower	Upper
196	100	
198	95.9	75.0 112.6
200	26.1	22.1 33.1









#42
2,4,5-Trichlorophenol
Concen: 19.549 ng/ul
RT: 13.545 min Scan# 1702
Delta R.T. 0.004 min
Lab File: BG056707.D
Acq: 23 Feb 2023 10:50




Instrument :
BNA_G
ClientSampleId :
SSTD020591


Tgt Ion:196 Resp: 30869
Ion Ratio Lower Upper
196 100
198 98.7 80.1 120.1
200 32.2 26.7 40.1



#43
1,1'-Biphenyl
Concen: 18.835 ng/ul
RT: 13.869 min Scan# 1758
Delta R.T. -0.002 min
Lab File: BG056707.D
Acq: 23 Feb 2023 10:50

Tgt Ion:154 Resp: 90694
Ion Ratio Lower Upper
154 100
153 38.3 32.0 48.0
76 10.8 9.0 13.4









#44
2-Chloronaphthalene
Concen: 18.567 ng/ul
RT: 13.921 min Scan# 1767
Delta R.T. 0.004 min
Lab File: BG056707.D
Acq: 23 Feb 2023 10:50




Instrument :
BNA_G
ClientSampleId :
SSTD020591


Tgt Ion:162 Resp: 74550
Ion Ratio Lower Upper
162 100
164 33.9 23.3 34.9
127 34.5 25.8 38.8



#45
2-Nitroaniline
Concen: 30.870 ng/ul
RT: 14.104 min Scan# 1798
Delta R.T. -0.002 min
Lab File: BG056707.D
Acq: 23 Feb 2023 10:50

Tgt Ion: 65 Resp: 26409
Ion Ratio Lower Upper
65 100
92 61.0 53.9 80.9
138 86.7 73.8 110.6









#47
 Dimethylphthalate
 Concen: 19.448 ng/ul
 RT: 14.462 min Scan# 1859
 Delta R.T. -0.002 min
 Lab File: BG056707.D
 Acq: 23 Feb 2023 10:50




Instrument :
 BNA_G
 ClientSampleId :
 SSTD020591


Tgt Ion:163 Resp: 101348
 Ion Ratio Lower Upper
 163 100
 194 4.1 2.5 3.7#
 164 10.9 8.2 12.4



#48
 2,6-Dinitrotoluene
 Concen: 25.130 ng/ul
 RT: 14.585 min Scan# 1880
 Delta R.T. -0.002 min
 Lab File: BG056707.D
 Acq: 23 Feb 2023 10:50

Tgt Ion:165 Resp: 21138
 Ion Ratio Lower Upper
 165 100
 89 64.0 48.4 72.6
 121 20.0 15.5 23.3









#50
Acenaphthylene
Concen: 18.965 ng/ul
RT: 14.756 min Scan# 1909
Delta R.T. -0.002 min
Lab File: BG056707.D
Acq: 23 Feb 2023 10:50




Instrument :
BNA_G
ClientSampleId :
SSTD020591


Tgt Ion:152 Resp: 115697
Ion Ratio Lower Upper
152 100
151 21.0 15.8 23.6
153 13.0 10.6 16.0



#51
3-Nitroaniline
Concen: 21.602 ng/ul
RT: 14.920 min Scan# 1937
Delta R.T. 0.004 min
Lab File: BG056707.D
Acq: 23 Feb 2023 10:50

Tgt Ion:138 Resp: 17863
Ion Ratio Lower Upper
138 100
108 11.8 6.5 9.7#
92 116.7 102.2 153.2









#52
Acenaphthene
Concen: 18.783 ng/ul
RT: 15.091 min Scan# 1966
Delta R.T. -0.002 min
Lab File: BG056707.D
Acq: 23 Feb 2023 10:50




Instrument :
BNA_G
ClientSampleId :
SSTD020591


Tgt Ion:153 Resp: 79503
Ion Ratio Lower Upper
153 100
152 46.4 38.6 58.0
154 84.6 70.9 106.3



#53
2,4-Dinitrophenol
Concen: 19.383 ng/ul
RT: 15.120 min Scan# 1971
Delta R.T. 0.004 min
Lab File: BG056707.D
Acq: 23 Feb 2023 10:50

Tgt Ion:184 Resp: 11833
Ion Ratio Lower Upper
184 100
63 55.9 43.8 65.6
154 80.9 66.8 100.2









#55
4-Nitrophenol
Concen: 25.704 ng/ul
RT: 15.202 min Scan# 1985
Delta R.T. 0.004 min
Lab File: BG056707.D
Acq: 23 Feb 2023 10:50




Instrument :
BNA_G
ClientSampleId :
SSTD020591


Tgt Ion:109 Resp: 19777
Ion Ratio Lower Upper
109 100
139 85.0 71.2 106.8
65 105.1 85.1 127.7



#56
Dibenzofuran
Concen: 18.373 ng/ul
RT: 15.426 min Scan# 2023
Delta R.T. 0.004 min
Lab File: BG056707.D
Acq: 23 Feb 2023 10:50

Tgt Ion:168 Resp: 118049
Ion Ratio Lower Upper
168 100
139 39.2 28.2 42.2









#57
2,4-Dinitrotoluene
Concen: 25.376 ng/ul
RT: 15.367 min Scan# 2013
Delta R.T. -0.002 min
Lab File: BG056707.D
Acq: 23 Feb 2023 10:50




Instrument :
BNA_G
ClientSampleId :
SSTD020591


Tgt Ion:165 Resp: 30810
Ion Ratio Lower Upper
165 100
63 52.5 38.6 58.0
182 2.7 1.8 2.8



#58
2,3,4,6-Tetrachlorophenol
Concen: 21.402 ng/ul
RT: 15.643 min Scan# 2060
Delta R.T. 0.004 min
Lab File: BG056707.D
Acq: 23 Feb 2023 10:50

Tgt Ion:232 Resp: 29248
Ion Ratio Lower Upper
232 100
131 34.6 27.8 41.8
130 1.7 1.8 2.6#
166 23.0 20.2 30.4









#59
Diethylphthalate
Concen: 20.067 ng/ul
RT: 15.807 min Scan# 2088
Delta R.T. -0.002 min
Lab File: BG056707.D
Acq: 23 Feb 2023 10:50




Instrument :
BNA_G
ClientSampleId :
SSTD020591


Tgt Ion:149 Resp: 100696
Ion Ratio Lower Upper
149 100
177 23.5 18.6 28.0
150 13.1 10.7 16.1



#61
Fluorene
Concen: 18.584 ng/ul
RT: 16.066 min Scan# 2132
Delta R.T. -0.002 min
Lab File: BG056707.D
Acq: 23 Feb 2023 10:50

Tgt Ion:166 Resp: 96177
Ion Ratio Lower Upper
166 100
165 99.0 78.4 117.6
167 14.1 10.5 15.7









#62
4-Chlorophenyl-phenylether
Concen: 18.232 ng/ul
RT: 16.048 min Scan# 2128
Delta R.T. 0.004 min
Lab File: BG056707.D
Acq: 23 Feb 2023 10:50




Instrument :
BNA_G
ClientSampleId :
SSTD020591


Tgt Ion:204 Resp: 56217
Ion Ratio Lower Upper
204 100
206 34.4 26.1 39.1
141 48.7 39.7 59.5



#63
4-Nitroaniline
Concen: 21.389 ng/ul
RT: 16.072 min Scan# 2133
Delta R.T. 0.004 min
Lab File: BG056707.D
Acq: 23 Feb 2023 10:50

Tgt Ion:138 Resp: 17327
Ion Ratio Lower Upper
138 100
92 57.2 48.6 73.0
108 113.6 93.0 139.6









#66
 4,6-Dinitro-2-methylphenol
 Concen: 23.296 ng/ul
 RT: 16.131 min Scan# 2143
 Delta R.T. 0.004 min
 Lab File: BG056707.D
 Acq: 23 Feb 2023 10:50




Instrument :
 BNA_G
 ClientSampleId :
 SSTD020591


Tgt Ion:198 Resp: 19399
 Ion Ratio Lower Upper
 198 100
 182 4.1 2.4 3.6#
 77 23.7 21.0 31.6



#67
 N-Nitrosodiphenylamine
 Concen: 19.052 ng/ul
 RT: 16.260 min Scan# 2165
 Delta R.T. 0.004 min
 Lab File: BG056707.D
 Acq: 23 Feb 2023 10:50

Tgt Ion:169 Resp: 86217
 Ion Ratio Lower Upper
 169 100
 168 67.5 55.1 82.7
 167 35.7 28.3 42.5









#68
4-Bromophenyl-phenylether
Concen: 19.892 ng/ul
RT: 16.941 min Scan# 21
Delta R.T. -0.002 min
Lab File: BG056707.D
Acq: 23 Feb 2023 10:50




Instrument :
BNA_G
ClientSampleId :
SSTD020591


Tgt Ion:248 Resp: 39466
Ion Ratio Lower Upper
248 100
250 93.6 80.4 120.6
141 56.9 47.6 71.4



#69
Hexachlorobenzene
Concen: 18.467 ng/ul
RT: 17.071 min Scan# 2303
Delta R.T. 0.004 min
Lab File: BG056707.D
Acq: 23 Feb 2023 10:50

Tgt Ion:284 Resp: 41725
Ion Ratio Lower Upper
284 100
142 28.1 21.8 32.6
249 34.4 25.9 38.9









#70
 Atrazine
 Concen: 20.438 ng/ul
 RT: 17.188 min Scan# 21
 Delta R.T. -0.002 min
 Lab File: BG056707.D
 Acq: 23 Feb 2023 10:50




Instrument :
 BNA_G
 ClientSampleId :
 SST020591


Tgt Ion:200 Resp: 38496
 Ion Ratio Lower Upper
 200 100
 173 24.7 17.4 26.0
 215 49.0 41.7 62.5



#71
 Pentachlorophenol
 Concen: 19.537 ng/ul
 RT: 17.406 min Scan# 2360
 Delta R.T. 0.004 min
 Lab File: BG056707.D
 Acq: 23 Feb 2023 10:50

Tgt Ion:266 Resp: 23766
 Ion Ratio Lower Upper
 266 100
 264 60.3 49.9 74.9
 268 62.2 54.5 81.7









#72
Phenanthrene
Concen: 18.825 ng/ul
RT: 17.805 min Scan# 2428
Delta R.T. -0.002 min
Lab File: BG056707.D
Acq: 23 Feb 2023 10:50




Instrument :
BNA_G
ClientSampleId :
SSTD020591


Tgt Ion:178 Resp: 169265
Ion Ratio Lower Upper
178 100
179 15.8 12.1 18.1
176 17.7 14.4 21.6



#74
Anthracene
Concen: 18.654 ng/ul
RT: 17.899 min Scan# 2444
Delta R.T. 0.004 min
Lab File: BG056707.D
Acq: 23 Feb 2023 10:50

Tgt Ion:178 Resp: 169442
Ion Ratio Lower Upper
178 100
179 15.1 12.1 18.1
176 18.2 14.1 21.1









#75
1,2,3,4-Tetrachlorobenzene
Concen: 18.550 ng/uL
RT: 13.845 min Scan# 11
Delta R.T. 0.004 min
Lab File: BG056707.D
Acq: 23 Feb 2023 10:50




Instrument :
BNA_G
ClientSampleId :
SSTD020591


Tgt Ion:216 Resp: 44394
Ion Ratio Lower Upper
216 100
214 80.0 64.2 96.2
218 48.6 36.6 54.8



#76
Pentachlorobenzene
Concen: 18.183 ng/uL
RT: 15.343 min Scan# 2009
Delta R.T. 0.004 min
Lab File: BG056707.D
Acq: 23 Feb 2023 10:50

Tgt Ion:250 Resp: 45665
Ion Ratio Lower Upper
250 100
248 65.6 47.1 70.7
252 64.2 52.4 78.6









#77
 Carbazole
 Concen: 18.701 ng/ul
 RT: 18.158 min Scan# 2487
 Delta R.T. 0.004 min
 Lab File: BG056707.D
 Acq: 23 Feb 2023 10:50




Instrument :
 BNA_G
 ClientSampleId :
 SSTD020591


Tgt Ion:167 Resp: 144066
 Ion Ratio Lower Upper
 167 100
 166 22.8 18.2 27.2
 139 12.1 9.8 14.8



#78
 Di-n-butylphthalate
 Concen: 21.058 ng/ul
 RT: 18.686 min Scan# 2578
 Delta R.T. 0.004 min
 Lab File: BG056707.D
 Acq: 23 Feb 2023 10:50

Tgt Ion:149 Resp: 162878
 Ion Ratio Lower Upper
 149 100
 150 9.3 7.4 11.0
 104 6.3 5.4 8.2









#80
Fluoranthene
Concen: 21.744 ng/ul
RT: 19.791 min Scan# 21
Delta R.T. 0.004 min
Lab File: BG056707.D
Acq: 23 Feb 2023 10:50




Instrument :
BNA_G
ClientSampleId :
SSTD020591


Tgt Ion:202 Resp: 195133
Ion Ratio Lower Upper
202 100
101 5.5 4.7 7.1
100 4.7 4.4 6.6



#82
Pyrene
Concen: 21.551 ng/ul
RT: 20.149 min Scan# 2827
Delta R.T. -0.002 min
Lab File: BG056707.D
Acq: 23 Feb 2023 10:50

Tgt Ion:202 Resp: 192706
Ion Ratio Lower Upper
202 100
101 6.4 5.3 7.9
100 6.0 4.6 7.0





#83

Butylbenzylphthalate

Concen: 25.011 ng/ul

RT: 21.007 min Scan# 2973

Delta R.T. -0.002 min

Lab File: BG056707.D

Acq: 23 Feb 2023 10:50

Instrument :

BNA_G

ClientSampleId :

SSTD020591





Tgt Ion:149 Resp: 61672

Ion Ratio Lower Upper

149 100

91 78.3 59.3 88.9

206 25.0 19.7 29.5



#84

3,3'-Dichlorobenzidine

Concen: 20.975 ng/ul

RT: 21.947 min Scan# 3133

Delta R.T. -0.002 min

Lab File: BG056707.D

Acq: 23 Feb 2023 10:50




Tgt Ion:252 Resp: 60212


Ion Ratio Lower Upper

252 100

254 60.4 52.6 78.8

126 5.8 5.0 7.6





#85

Benzo(a)anthracene

Concen: 18.670 ng/ul

RT: 22.053 min Scan# 3151

Delta R.T. -0.002 min

Lab File: BG056707.D

Acq: 23 Feb 2023 10:50

Instrument :

BNA_G

ClientSampleId :

SSTD020591





Tgt Ion:228 Resp: 170538

Ion Ratio Lower Upper

228 100

229 19.6 16.2 24.4

226 26.7 20.6 31.0



#86

Bis(2-ethylhexyl)phthalate

Concen: 21.997 ng/ul

RT: 21.918 min Scan# 3128

Delta R.T. 0.004 min

Lab File: BG056707.D

Acq: 23 Feb 2023 10:50




Tgt Ion:149 Resp: 81966


Ion Ratio Lower Upper

149 100

167 29.3 21.8 32.8

279 5.4 4.2 6.2









#87
Chrysene
Concen: 18.898 ng/ul
RT: 22.129 min Scan# 3164
Delta R.T. 0.004 min
Lab File: BG056707.D
Acq: 23 Feb 2023 10:50




Instrument :
BNA_G
ClientSampleId :
SSTD020591


Tgt Ion:228 Resp: 158787
Ion Ratio Lower Upper
228 100
226 30.4 23.6 35.4
229 19.8 15.4 23.0



#89
Di-n-octyl phthalate
Concen: 21.315 ng/ul
RT: 23.234 min Scan# 3352
Delta R.T. -0.002 min
Lab File: BG056707.D
Acq: 23 Feb 2023 10:50

Tgt Ion:149 Resp: 142251









#90
Benzo(b)fluoranthene
Concen: 18.369 ng/ul
RT: 24.480 min Scan# 3563
Delta R.T. 0.004 min
Lab File: BG056707.D
Acq: 23 Feb 2023 10:50




Instrument :
BNA_G
ClientSampleId :
SSTD020591


Tgt Ion:252 Resp: 172497
Ion Ratio Lower Upper
252 100
253 21.8 17.4 26.2
125 3.6 3.0 4.6



#91
Benzo(k)fluoranthene
Concen: 18.403 ng/ul
RT: 24.550 min Scan# 3576
Delta R.T. -0.002 min
Lab File: BG056707.D
Acq: 23 Feb 2023 10:50

Tgt Ion:252 Resp: 166342
Ion Ratio Lower Upper
252 100
253 21.2 18.2 27.2
125 3.6 3.4 5.2





#93

Benzo(a)pyrene

Concen: 18.782 ng/ul

RT: 25.443 min Scan# 31

Delta R.T. 0.004 min

Lab File: BG056707.D

Acq: 23 Feb 2023 10:50

Instrument :

BNA_G

ClientSampleId :

SSTD020591





Tgt Ion:252 Resp: 150283

Ion Ratio Lower Upper

252 100

253 21.9 17.4 26.0

125 3.8 3.6 5.4



#94

Indeno(1,2,3-cd)pyrene

Concen: 18.613 ng/ul

RT: 29.685 min Scan# 4450

Delta R.T. 0.004 min

Lab File: BG056707.D

Acq: 23 Feb 2023 10:50




Tgt Ion:276 Resp: 205117


Ion Ratio Lower Upper

276 100

138 8.4 5.7 8.5

277 23.8 18.1 27.1









#95
Dibenzo(a,h)anthracene
Concen: 18.797 ng/ul
RT: 29.756 min Scan# 4462
Delta R.T. -0.002 min
Lab File: BG056707.D
Acq: 23 Feb 2023 10:50

Instrument :
BNA_G
ClientSampleId :
SSTD020591

Tgt Ion:278 Resp: 171281

Ion	Ratio	Lower	Upper
278	100		
139	6.3	4.3	6.5
279	22.3	21.7	32.5



#96
Benzo(g,h,i)perylene
Concen: 11.162 ng/ul
RT: 30.972 min Scan# 4669
Delta R.T. 0.015 min
Lab File: BG056707.D
Acq: 23 Feb 2023 10:50

Tgt Ion:276 Resp: 98057

Ion	Ratio	Lower	Upper
276	100		
138	8.5	6.5	9.7
277	23.8	20.7	31.1

