

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG022723\
 Data File : BG056725.D
 Acq On : 27 Feb 2023 10:40
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SSTD020593

Quant Time: Feb 27 11:25:15 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG022323.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Feb 23 15:40:42 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.423	152	16240	20.000	ng/u1	0.00
20) Naphthalene-d8	11.261	136	69601	20.000	ng/u1	0.00
38) Acenaphthene-d10	15.021	164	55614	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.759	188	137801	20.000	ng/u1	0.00
79) Chrysene-d12	22.066	240	127144	20.000	ng/u1	0.00
88) Perylene-d12	25.597	264	133625	20.000	ng/u1	-0.01
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.699	96	2775	9.479	ng/uL	0.00
4) Pyridine-d5	4.140	84	25093	21.235	ng/u1	0.00
7) Phenol-d5	7.542	99	29065	17.885	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.724	67	19442	19.613	ng/u1	0.00
11) 2-Chlorophenol-d4	7.941	132	19754	18.304	ng/u1	0.00
15) 4-Methylphenol-d8	9.111	113	22352	18.000	ng/u1	0.00
21) Nitrobenzene-d5	9.592	128	11542	19.788	ng/u1	0.00
24) 2-Nitrophenol-d4	10.321	143	11968	17.237	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.862	165	24908	19.043	ng/u1	0.00
31) 4-Chloroaniline-d4	11.379	131	34138	20.010	ng/u1	0.00
46) Dimethylphthalate-d6	14.410	166	87368	19.471	ng/u1	0.00
49) Acenaphthylene-d8	14.722	160	104366	20.465	ng/u1	0.00
54) 4-Nitrophenol-d4	15.180	143	13260	17.140	ng/u1	0.00
60) Fluorene-d10	16.008	176	87257	21.022	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	16.108	200	15251	16.542	ng/u1	0.00
73) Anthracene-d10	17.853	188	137798	20.612	ng/u1	0.00
81) Pyrene-d10	20.115	212	164677	19.141	ng/u1	0.00
92) Benzo(a)pyrene-d12	25.350	264	150510	20.868	ng/u1	-0.01
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.735	88	3389	9.389	ng/uL#	93
5) Pyridine	4.158	79	26834	21.442	ng/u1	95
6) Benzaldehyde	7.542	77	20567	24.157	ng/u1	92
8) Phenol	7.571	94	30647	18.488	ng/u1	98
10) Bis(2-Chloroethyl)ether	7.818	93	24098	20.460	ng/u1	98
12) 2-Chlorophenol	7.977	128	20647	19.319	ng/u1	96
13) 2-Methylphenol	8.840	108	22910	18.323	ng/u1	92
14) 2,2'-oxybis(1-Chloropr...	8.934	45	32799	20.724	ng/u1	96
16) Acetophenone	9.246	105	42956	20.201	ng/u1	98
17) N-Nitroso-di-n-propyla...	9.222	70	24226	19.856	ng/u1	94
18) 4-Methylphenol	9.169	108	25976	19.120	ng/u1	94
19) Hexachloroethane	9.528	117	8889	19.559	ng/u1	91
22) Nitrobenzene	9.634	77	34951	20.664	ng/u1	98
23) Isophorone	10.157	82	68358	20.672	ng/u1#	96
25) 2-Nitrophenol	10.356	139	12058	17.594	ng/u1	99
26) 2,4-Dimethylphenol	10.392	107	29015	18.956	ng/u1	96
27) Bis(2-Chloroethoxy)met...	10.632	93	34732	20.780	ng/u1	99
29) 2,4-Dichlorophenol	10.891	162	24473	19.321	ng/u1	94
30) Naphthalene	11.308	128	77631	20.263	ng/u1	98
32) 4-Chloroaniline	11.402	127	35121	20.782	ng/u1	99
33) Hexachlorobutadiene	11.584	225	20788	20.592	ng/u1	91
34) Caprolactam	12.142	113	7688	17.854	ng/u1	87
35) 4-Chloro-3-methylphenol	12.489	107	27916	19.184	ng/u1	94
36) 2-Methylnaphthalene	12.883	142	56588	20.625	ng/u1	99

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG022723\
 Data File : BG056725.D
 Acq On : 27 Feb 2023 10:40
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SSTD020593

Quant Time: Feb 27 11:25:15 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG022323.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Feb 23 15:40:42 2023
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	13.100	142	56949	21.052	ng/ul	98
39) 1,2,4,5-Tetrachloroben...	13.241	216	40636	20.609	ng/ul	99
40) Hexachlorocyclopentadiene	13.218	237	22714	16.077	ng/ul	99
41) 2,4,6-Trichlorophenol	13.464	196	23903	18.143	ng/ul#	90
42) 2,4,5-Trichlorophenol	13.535	196	27049	18.908	ng/ul	100
43) 1,1'-Biphenyl	13.864	154	83902	20.516	ng/ul	99
44) 2-Chloronaphthalene	13.911	162	68876	20.356	ng/ul	96
45) 2-Nitroaniline	14.099	65	20126	17.136	ng/ul	96
47) Dimethylphthalate	14.457	163	88679	19.434	ng/ul#	96
48) 2,6-Dinitrotoluene	14.581	165	18213	19.471	ng/ul	92
50) Acenaphthylene	14.751	152	104693	20.321	ng/ul	99
51) 3-Nitroaniline	14.910	138	17290	20.653	ng/ul#	92
52) Acenaphthene	15.086	153	72730	20.550	ng/ul	96
53) 2,4-Dinitrophenol	15.110	184	9184	15.326	ng/ul#	85
55) 4-Nitrophenol	15.192	109	14142	15.793	ng/ul	94
56) Dibenzofuran	15.415	168	108468	20.189	ng/ul	95
57) 2,4-Dinitrotoluene	15.362	165	24986	18.632	ng/ul#	99
58) 2,3,4,6-Tetrachlorophenol	15.632	232	23898	17.951	ng/ul#	97
59) Diethylphthalate	15.803	149	86670	18.975	ng/ul	98
61) Fluorene	16.061	166	88529	20.579	ng/ul	99
62) 4-Chlorophenyl-phenyle...	16.044	204	51296	20.131	ng/ul	97
63) 4-Nitroaniline	16.067	138	17254	21.094	ng/ul	92
66) 4,6-Dinitro-2-methylph...	16.120	198	15220	17.148	ng/ul#	86
67) N-Nitrosodiphenylamine	16.249	169	78999	20.616	ng/ul	96
68) 4-Bromophenyl-phenylether	16.937	248	35870	20.540	ng/ul	98
69) Hexachlorobenzene	17.060	284	38349	20.440	ng/ul	97
70) Atrazine	17.184	200	31328	18.824	ng/ul	98
71) Pentachlorophenol	17.401	266	18102	16.992	ng/ul	92
72) Phenanthrene	17.800	178	156130	20.917	ng/ul	99
74) Anthracene	17.889	178	155357	20.595	ng/ul	98
75) 1,2,3,4-Tetrachloroben...	13.835	216	43524	20.646	ng/ul	98
76) Pentachlorobenzene	15.339	250	43733	20.454	ng/ul	94
77) Carbazole	18.153	167	132936	21.062	ng/ul	98
78) Di-n-butylphthalate	18.682	149	138137	18.990	ng/ul	99
80) Fluoranthene	19.781	202	196659	18.904	ng/ul	99
82) Pyrene	20.145	202	195689	19.094	ng/ul	99
83) Butylbenzylphthalate	21.003	149	55337	17.020	ng/ul	97
84) 3,3'-Dichlorobenzidine	21.943	252	65482	20.542	ng/ul	100
85) Benzo(a)anthracene	22.048	228	185721	20.070	ng/ul	98
86) Bis(2-ethylhexyl)phtha...	21.913	149	80886	17.710	ng/ul	99
87) Chrysene	22.119	228	173198	20.096	ng/ul	99
89) Di-n-octyl phthalate	23.224	149	135277	18.621	ng/ul	100
90) Benzo(b)fluoranthene	24.463	252	187549	20.966	ng/ul	98
91) Benzo(k)fluoranthene	24.540	252	178547	20.492	ng/ul	99
93) Benzo(a)pyrene	25.433	252	159247	20.252	ng/ul	96
94) Indeno(1,2,3-cd)pyrene	29.657	276	218617	20.467	ng/ul#	96
95) Dibenzo(a,h)anthracene	29.734	278	177455	19.869	ng/ul	95
96) Benzo(g,h,i)perylene	30.944	276	174801	20.487	ng/ul	94


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG022723\
 Data File : BG056725.D
 Acq On : 27 Feb 2023 10:40
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SSTD020593

Quant Time: Feb 27 11:25:15 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG022323.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Feb 23 15:40:42 2023
 Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 9.389 ng/uL
RT: 3.735 min Scan# 31
Delta R.T. -0.007 min
Lab File: BG056725.D
Acq: 27 Feb 2023 10:40




Instrument :
BNA_G
ClientSampleId :
SSTD020593


Tgt Ion: 88 Resp: 3389
Ion Ratio Lower Upper
88 100
43 24.0 28.3 42.5#
58 62.8 50.6 75.8



#5
Pyridine
Concen: 21.442 ng/uL
RT: 4.158 min Scan# 105
Delta R.T. -0.007 min
Lab File: BG056725.D
Acq: 27 Feb 2023 10:40

Tgt Ion: 79 Resp: 26834
Ion Ratio Lower Upper
79 100
52 74.9 60.6 91.0
51 32.7 31.9 47.9









#6
Benzaldehyde
Concen: 24.157 ng/ul
RT: 7.542 min Scan# 681
Delta R.T. -0.007 min
Lab File: BG056725.D
Acq: 27 Feb 2023 10:40




Instrument :
BNA_G
ClientSampleId :
SSTD020593


Tgt Ion: 77 Resp: 20567
Ion Ratio Lower Upper
77 100
105 96.9 72.2 108.4
106 92.9 67.4 101.2



#8
Phenol
Concen: 18.488 ng/ul
RT: 7.571 min Scan# 686
Delta R.T. -0.001 min
Lab File: BG056725.D
Acq: 27 Feb 2023 10:40

Tgt Ion: 94 Resp: 30647
Ion Ratio Lower Upper
94 100
65 33.7 25.7 38.5
66 47.4 38.9 58.3









#10
Bis(2-Chloroethyl)ether
Concen: 20.460 ng/ul
RT: 7.818 min Scan# 729
Delta R.T. -0.007 min
Lab File: BG056725.D
Acq: 27 Feb 2023 10:40




Instrument :
BNA_G
ClientSampleId :
SSTD020593


Tgt Ion:	93	Resp:	24098
Ion	Ratio	Lower	Upper
93	100		
63	80.9	66.6	100.0
95	31.9	25.0	37.4



#12
2-Chlorophenol
Concen: 19.319 ng/ul
RT: 7.977 min Scan# 755
Delta R.T. -0.001 min
Lab File: BG056725.D
Acq: 27 Feb 2023 10:40

Tgt Ion:	128	Resp:	20647
Ion	Ratio	Lower	Upper
128	100		
64	52.7	45.5	68.3
130	35.8	27.7	41.5





#13

2-Methylphenol

Concen: 18.323 ng/ul

RT: 8.840 min Scan# 904

Delta R.T. -0.013 min

Lab File: BG056725.D

Acq: 27 Feb 2023 10:40

Instrument :

BNA_G

ClientSampleId :





SSTD020593

Tgt Ion:108 Resp: 22910

Ion Ratio Lower Upper

108 100

107 83.7 73.4 110.0



#14

2,2'-oxybis(1-Chloropropane)

Concen: 20.724 ng/ul

RT: 8.934 min Scan# 918

Delta R.T. -0.013 min

Lab File: BG056725.D

Acq: 27 Feb 2023 10:40




Tgt Ion: 45 Resp: 32799


Ion Ratio Lower Upper

45 100

77 18.5 12.4 18.6

79 11.8 9.8 14.6









#16
Acetophenone
Concen: 20.201 ng/ul
RT: 9.246 min Scan# 91
Delta R.T. -0.007 min
Lab File: BG056725.D
Acq: 27 Feb 2023 10:40




Instrument :
BNA_G
ClientSampleId :
SSTD020593


Tgt Ion:105 Resp: 42956
Ion Ratio Lower Upper
105 100
77 88.2 71.4 107.0
51 27.1 23.6 35.4



#17
N-Nitroso-di-n-propylamine
Concen: 19.856 ng/ul
RT: 9.222 min Scan# 967
Delta R.T. -0.001 min
Lab File: BG056725.D
Acq: 27 Feb 2023 10:40

Tgt Ion: 70 Resp: 24226
Ion Ratio Lower Upper
70 100
42 51.5 46.6 70.0
101 9.2 7.3 10.9
130 16.8 13.7 20.5





#18

4-Methylphenol

Concen: 19.120 ng/ul

RT: 9.169 min Scan# 91

Delta R.T. -0.007 min

Lab File: BG056725.D

Acq: 27 Feb 2023 10:40

Instrument :

BNA_G

ClientSampleId :





SSTD020593

Tgt Ion:108 Resp: 25976

Ion Ratio Lower Upper

108 100

107 108.4 92.1 138.1



#19

Hexachloroethane

Concen: 19.559 ng/ul

RT: 9.528 min Scan# 1019

Delta R.T. -0.007 min

Lab File: BG056725.D

Acq: 27 Feb 2023 10:40




Tgt Ion:117 Resp: 8889


Ion Ratio Lower Upper

117 100

201 130.4 95.2 142.8

199 77.9 57.9 86.9





#22

Nitrobenzene

Concen: 20.664 ng/ul

RT: 9.634 min Scan# 110

Delta R.T. -0.007 min

Lab File: BG056725.D

Acq: 27 Feb 2023 10:40

Instrument :

BNA_G

ClientSampleId :

SSTD020593





Tgt Ion: 77 Resp: 34951

Ion Ratio Lower Upper

77 100

123 33.6 27.8 41.6

65 14.5 10.6 16.0



#23

Isophorone

Concen: 20.672 ng/ul

RT: 10.157 min Scan# 1126

Delta R.T. -0.007 min

Lab File: BG056725.D

Acq: 27 Feb 2023 10:40




Tgt Ion: 82 Resp: 68358


Ion Ratio Lower Upper

82 100

95 7.5 7.7 11.5#

138 13.9 12.1 18.1









#25
2-Nitrophenol
Concen: 17.594 ng/ul
RT: 10.356 min Scan# 1160
Delta R.T. -0.007 min
Lab File: BG056725.D
Acq: 27 Feb 2023 10:40




Instrument :
BNA_G
ClientSampleId :
SSTD020593


Tgt Ion:139 Resp: 12058
Ion Ratio Lower Upper
139 100
65 62.7 49.4 74.2
109 38.5 31.6 47.4



#26
2,4-Dimethylphenol
Concen: 18.956 ng/ul
RT: 10.392 min Scan# 1166
Delta R.T. -0.007 min
Lab File: BG056725.D
Acq: 27 Feb 2023 10:40

Tgt Ion:107 Resp: 29015
Ion Ratio Lower Upper
107 100
121 51.0 40.2 60.4
122 84.1 63.1 94.7





#27

Bis(2-Chloroethoxy)methane

Concen: 20.780 ng/ul

RT: 10.632 min Scan# 1207

Delta R.T. -0.007 min

Lab File: BG056725.D

Acq: 27 Feb 2023 10:40

Instrument :

BNA_G

ClientSampleId :

SSTD020593





Tgt Ion: 93 Resp: 34732

Ion Ratio Lower Upper

93 100

95 31.0 24.6 37.0

123 9.5 6.6 10.0



#29

2,4-Dichlorophenol

Concen: 19.321 ng/ul

RT: 10.891 min Scan# 1251

Delta R.T. -0.007 min

Lab File: BG056725.D

Acq: 27 Feb 2023 10:40




Tgt Ion:162 Resp: 24473


Ion Ratio Lower Upper

162 100

164 59.9 53.3 79.9

98 33.2 25.5 38.3









#30
Naphthalene
Concen: 20.263 ng/ul
RT: 11.308 min Scan# 11
Delta R.T. -0.007 min
Lab File: BG056725.D
Acq: 27 Feb 2023 10:40




Instrument :
BNA_G
ClientSampleId :
SSTD020593


Tgt Ion:128 Resp: 77631
Ion Ratio Lower Upper
128 100
129 11.5 8.6 12.8
127 12.7 9.8 14.8



#32
4-Chloroaniline
Concen: 20.782 ng/ul
RT: 11.402 min Scan# 1338
Delta R.T. -0.007 min
Lab File: BG056725.D
Acq: 27 Feb 2023 10:40

Tgt Ion:127 Resp: 35121
Ion Ratio Lower Upper
127 100
129 28.6 23.4 35.0









#33
Hexachlorobutadiene
Concen: 20.592 ng/ul
RT: 11.584 min Scan# 11
Delta R.T. -0.007 min
Lab File: BG056725.D
Acq: 27 Feb 2023 10:40




Instrument :
BNA_G
ClientSampleId :
SSTD020593


Tgt Ion:225 Resp: 20788
Ion Ratio Lower Upper
225 100
223 59.4 53.5 80.3
227 65.3 58.2 87.2



#34
Caprolactam
Concen: 17.854 ng/ul
RT: 12.142 min Scan# 1464
Delta R.T. -0.007 min
Lab File: BG056725.D
Acq: 27 Feb 2023 10:40

Tgt Ion:113 Resp: 7688
Ion Ratio Lower Upper
113 100
55 178.1 167.0 250.6
56 138.6 113.5 170.3









#35
4-Chloro-3-methylphenol
Concen: 19.184 ng/ul
RT: 12.489 min Scan# 11
Delta R.T. -0.001 min
Lab File: BG056725.D
Acq: 27 Feb 2023 10:40




Instrument :
BNA_G
ClientSampleId :
SSTD020593


Tgt Ion:107 Resp: 27916
Ion Ratio Lower Upper
107 100
144 26.5 20.5 30.7
142 76.4 55.8 83.6



#36
2-Methylnaphthalene
Concen: 20.625 ng/ul
RT: 12.883 min Scan# 1590
Delta R.T. -0.001 min
Lab File: BG056725.D
Acq: 27 Feb 2023 10:40

Tgt Ion:142 Resp: 56588
Ion Ratio Lower Upper
142 100
141 89.8 71.0 106.6









#37
1-Methylnaphthalene
Concen: 21.052 ng/ul
RT: 13.100 min Scan# 1
Delta R.T. -0.001 min
Lab File: BG056725.D
Acq: 27 Feb 2023 10:40




Instrument :
BNA_G
ClientSampleId :
SSTD020593


Tgt Ion:142 Resp: 56949
Ion Ratio Lower Upper
142 100
141 90.2 73.5 110.3
116 4.1 4.0 6.0



#39
1,2,4,5-Tetrachlorobenzene
Concen: 20.609 ng/ul
RT: 13.241 min Scan# 1651
Delta R.T. -0.001 min
Lab File: BG056725.D
Acq: 27 Feb 2023 10:40

Tgt Ion:216 Resp: 40636
Ion Ratio Lower Upper
216 100
214 81.3 64.4 96.6
179 19.2 16.9 25.3
108 14.2 11.6 17.4









#40
 Hexachlorocyclopentadiene
 Concen: 16.077 ng/ul
 RT: 13.218 min Scan# 1
 Delta R.T. -0.007 min
 Lab File: BG056725.D
 Acq: 27 Feb 2023 10:40




Instrument :
 BNA_G
 ClientSampleId :
 SST020593


Tgt Ion:237 Resp: 22714
 Ion Ratio Lower Upper
 237 100
 235 60.4 47.8 71.8
 272 13.3 9.9 14.9



#41
 2,4,6-Trichlorophenol
 Concen: 18.143 ng/ul
 RT: 13.464 min Scan# 1689
 Delta R.T. -0.007 min
 Lab File: BG056725.D
 Acq: 27 Feb 2023 10:40

Tgt Ion:196 Resp: 23903
 Ion Ratio Lower Upper
 196 100
 198 101.1 75.0 112.6
 200 37.5 22.1 33.1#









#42
2,4,5-Trichlorophenol
Concen: 18.908 ng/ul
RT: 13.535 min Scan# 1
Delta R.T. -0.007 min
Lab File: BG056725.D
Acq: 27 Feb 2023 10:40




Instrument :
BNA_G
ClientSampleId :
SSTD020593


Tgt Ion:196 Resp: 27049
Ion Ratio Lower Upper
196 100
198 100.7 80.1 120.1
200 33.4 26.7 40.1



#43
1,1'-Biphenyl
Concen: 20.516 ng/ul
RT: 13.864 min Scan# 1757
Delta R.T. -0.007 min
Lab File: BG056725.D
Acq: 27 Feb 2023 10:40

Tgt Ion:154 Resp: 83902
Ion Ratio Lower Upper
154 100
153 40.3 32.0 48.0
76 12.4 9.0 13.4









#44
2-Chloronaphthalene
Concen: 20.356 ng/ul
RT: 13.911 min Scan# 1765
Delta R.T. -0.007 min
Lab File: BG056725.D
Acq: 27 Feb 2023 10:40




Instrument :
BNA_G
ClientSampleId :
SSTD020593


Tgt Ion:162 Resp: 68876
Ion Ratio Lower Upper
162 100
164 30.7 23.3 34.9
127 35.6 25.8 38.8



#45
2-Nitroaniline
Concen: 17.136 ng/ul
RT: 14.099 min Scan# 1797
Delta R.T. -0.007 min
Lab File: BG056725.D
Acq: 27 Feb 2023 10:40

Tgt Ion: 65 Resp: 20126
Ion Ratio Lower Upper
65 100
92 61.1 53.9 80.9
138 92.9 73.8 110.6









#47
Dimethylphthalate
Concen: 19.434 ng/ul
RT: 14.457 min Scan# 1858
Delta R.T. -0.007 min
Lab File: BG056725.D
Acq: 27 Feb 2023 10:40




Instrument :
BNA_G
ClientSampleId :
SSTD020593


Tgt Ion:163 Resp: 88679
Ion Ratio Lower Upper
163 100
194 4.3 2.5 3.7#
164 11.6 8.2 12.4



#48
2,6-Dinitrotoluene
Concen: 19.471 ng/ul
RT: 14.581 min Scan# 1879
Delta R.T. -0.007 min
Lab File: BG056725.D
Acq: 27 Feb 2023 10:40

Tgt Ion:165 Resp: 18213
Ion Ratio Lower Upper
165 100
89 68.5 48.4 72.6
121 20.1 15.5 23.3









#50
Acenaphthylene
Concen: 20.321 ng/ul
RT: 14.751 min Scan# 1908
Delta R.T. -0.007 min
Lab File: BG056725.D
Acq: 27 Feb 2023 10:40




Instrument :
BNA_G
ClientSampleId :
SSTD020593


Tgt Ion:152 Resp: 104693
Ion Ratio Lower Upper
152 100
151 20.3 15.8 23.6
153 13.3 10.6 16.0



#51
3-Nitroaniline
Concen: 20.653 ng/ul
RT: 14.910 min Scan# 1935
Delta R.T. -0.007 min
Lab File: BG056725.D
Acq: 27 Feb 2023 10:40

Tgt Ion:138 Resp: 17290
Ion Ratio Lower Upper
138 100
108 10.1 6.5 9.7#
92 118.9 102.2 153.2









#52
Acenaphthene
Concen: 20.550 ng/ul
RT: 15.086 min Scan# 1966
Delta R.T. -0.007 min
Lab File: BG056725.D
Acq: 27 Feb 2023 10:40




Instrument :
BNA_G
ClientSampleId :
SSTD020593


Tgt Ion:153 Resp: 72730
Ion Ratio Lower Upper
153 100
152 46.8 38.6 58.0
154 83.9 70.9 106.3



#53
2,4-Dinitrophenol
Concen: 15.326 ng/ul
RT: 15.110 min Scan# 1969
Delta R.T. -0.007 min
Lab File: BG056725.D
Acq: 27 Feb 2023 10:40

Tgt Ion:184 Resp: 9184
Ion Ratio Lower Upper
184 100
63 51.5 43.8 65.6
154 103.2 66.8 100.2#









#55
4-Nitrophenol
Concen: 15.793 ng/ul
RT: 15.192 min Scan# 1983
Delta R.T. -0.007 min
Lab File: BG056725.D
Acq: 27 Feb 2023 10:40




Instrument :
BNA_G
ClientSampleId :
SSTD020593


Tgt Ion:109 Resp: 14142
Ion Ratio Lower Upper
109 100
139 95.9 71.2 106.8
65 111.0 85.1 127.7



#56
Dibenzofuran
Concen: 20.189 ng/ul
RT: 15.415 min Scan# 2021
Delta R.T. -0.007 min
Lab File: BG056725.D
Acq: 27 Feb 2023 10:40

Tgt Ion:168 Resp: 108468
Ion Ratio Lower Upper
168 100
139 38.4 28.2 42.2





#57

2,4-Dinitrotoluene

Concen: 18.632 ng/ul

RT: 15.362 min Scan# 2013

Delta R.T. -0.007 min

Lab File: BG056725.D


Acq: 27 Feb 2023 10:40

Instrument :

BNA_G

ClientSampleId :

SSTD020593






Tgt Ion:165 Resp: 24986

Ion Ratio Lower Upper

165 100

63 47.5 38.6 58.0

182 3.8 1.8 2.8#



#58

2,3,4,6-Tetrachlorophenol

Concen: 17.951 ng/ul

RT: 15.632 min Scan# 2058

Delta R.T. -0.007 min

Lab File: BG056725.D

Acq: 27 Feb 2023 10:40

Tgt Ion:232 Resp: 23898




Ion Ratio Lower Upper


232 100

131 33.1 27.8 41.8

130 2.9 1.8 2.6#

166 24.1 20.2 30.4









#59
Diethylphthalate
Concen: 18.975 ng/ul
RT: 15.803 min Scan# 2088
Delta R.T. -0.007 min
Lab File: BG056725.D
Acq: 27 Feb 2023 10:40




Instrument :
BNA_G
ClientSampleId :
SSTD020593


Tgt Ion:149 Resp: 86670
Ion Ratio Lower Upper
149 100
177 23.5 18.6 28.0
150 11.9 10.7 16.1



#61
Fluorene
Concen: 20.579 ng/ul
RT: 16.061 min Scan# 2131
Delta R.T. -0.007 min
Lab File: BG056725.D
Acq: 27 Feb 2023 10:40

Tgt Ion:166 Resp: 88529
Ion Ratio Lower Upper
166 100
165 99.0 78.4 117.6
167 13.5 10.5 15.7









#62
4-Chlorophenyl-phenylether
Concen: 20.131 ng/ul
RT: 16.044 min Scan# 2128
Delta R.T. -0.001 min
Lab File: BG056725.D
Acq: 27 Feb 2023 10:40




Instrument :
BNA_G
ClientSampleId :
SSTD020593


Tgt Ion:204 Resp: 51296
Ion Ratio Lower Upper
204 100
206 33.7 26.1 39.1
141 52.3 39.7 59.5



#63
4-Nitroaniline
Concen: 21.094 ng/ul
RT: 16.067 min Scan# 2132
Delta R.T. -0.001 min
Lab File: BG056725.D
Acq: 27 Feb 2023 10:40

Tgt Ion:138 Resp: 17254
Ion Ratio Lower Upper
138 100
92 53.4 48.6 73.0
108 108.7 93.0 139.6









#66
 4,6-Dinitro-2-methylphenol
 Concen: 17.148 ng/ul
 RT: 16.120 min Scan# 2142
 Delta R.T. -0.007 min
 Lab File: BG056725.D
 Acq: 27 Feb 2023 10:40




Instrument :
 BNA_G
 ClientSampleId :
 SSTD020593


Tgt Ion:198 Resp: 15220
 Ion Ratio Lower Upper
 198 100
 182 1.5 2.4 3.6#
 77 18.6 21.0 31.6#



#67
 N-Nitrosodiphenylamine
 Concen: 20.616 ng/ul
 RT: 16.249 min Scan# 2163
 Delta R.T. -0.007 min
 Lab File: BG056725.D
 Acq: 27 Feb 2023 10:40

Tgt Ion:169 Resp: 78999
 Ion Ratio Lower Upper
 169 100
 168 66.9 55.1 82.7
 167 39.0 28.3 42.5









#68
4-Bromophenyl-phenylether
Concen: 20.540 ng/ul
RT: 16.937 min Scan# 21
Delta R.T. -0.007 min
Lab File: BG056725.D
Acq: 27 Feb 2023 10:40




Instrument :
BNA_G
ClientSampleId :
SSTD020593


Tgt Ion:248 Resp: 35870
Ion Ratio Lower Upper
248 100
250 99.4 80.4 120.6
141 61.6 47.6 71.4



#69
Hexachlorobenzene
Concen: 20.440 ng/ul
RT: 17.060 min Scan# 2301
Delta R.T. -0.007 min
Lab File: BG056725.D
Acq: 27 Feb 2023 10:40

Tgt Ion:284 Resp: 38349
Ion Ratio Lower Upper
284 100
142 29.6 21.8 32.6
249 33.6 25.9 38.9





#70

Atrazine

Concen: 18.824 ng/ul

RT: 17.184 min Scan# 21

Delta R.T. -0.007 min

Lab File: BG056725.D

Acq: 27 Feb 2023 10:40

Instrument :

BNA_G

ClientSampleId :

SSTD020593





Tgt Ion:200 Resp: 31328

Ion Ratio Lower Upper

200 100

173 23.0 17.4 26.0

215 51.3 41.7 62.5



#71

Pentachlorophenol

Concen: 16.992 ng/ul

RT: 17.401 min Scan# 2359

Delta R.T. -0.001 min

Lab File: BG056725.D

Acq: 27 Feb 2023 10:40




Tgt Ion:266 Resp: 18102


Ion Ratio Lower Upper

266 100

264 57.7 49.9 74.9

268 60.1 54.5 81.7









#72
Phenanthrene
Concen: 20.917 ng/ul
RT: 17.800 min Scan# 2428
Delta R.T. -0.007 min
Lab File: BG056725.D
Acq: 27 Feb 2023 10:40




Instrument :
BNA_G
ClientSampleId :
SSTD020593


Tgt Ion:178 Resp: 156130
Ion Ratio Lower Upper
178 100
179 16.2 12.1 18.1
176 17.9 14.4 21.6



#74
Anthracene
Concen: 20.595 ng/ul
RT: 17.889 min Scan# 2442
Delta R.T. -0.007 min
Lab File: BG056725.D
Acq: 27 Feb 2023 10:40

Tgt Ion:178 Resp: 155357
Ion Ratio Lower Upper
178 100
179 15.9 12.1 18.1
176 18.8 14.1 21.1









#75
1,2,3,4-Tetrachlorobenzene
Concen: 20.646 ng/uL
RT: 13.835 min Scan# 11
Delta R.T. -0.007 min
Lab File: BG056725.D
Acq: 27 Feb 2023 10:40




Instrument :
BNA_G
ClientSampleId :
SSTD020593


Tgt Ion:216 Resp: 43524
Ion Ratio Lower Upper
216 100
214 78.7 64.2 96.2
218 46.8 36.6 54.8



#76
Pentachlorobenzene
Concen: 20.454 ng/uL
RT: 15.339 min Scan# 2008
Delta R.T. -0.001 min
Lab File: BG056725.D
Acq: 27 Feb 2023 10:40

Tgt Ion:250 Resp: 43733
Ion Ratio Lower Upper
250 100
248 63.7 47.1 70.7
252 60.5 52.4 78.6









#77
 Carbazole
 Concen: 21.062 ng/ul
 RT: 18.153 min Scan# 2487
 Delta R.T. -0.001 min
 Lab File: BG056725.D
 Acq: 27 Feb 2023 10:40




Instrument :
 BNA_G
 ClientSampleId :
 SST020593


Tgt Ion:167 Resp: 132936
 Ion Ratio Lower Upper
 167 100
 166 21.6 18.2 27.2
 139 11.8 9.8 14.8



#78
 Di-n-butylphthalate
 Concen: 18.990 ng/ul
 RT: 18.682 min Scan# 2577
 Delta R.T. -0.001 min
 Lab File: BG056725.D
 Acq: 27 Feb 2023 10:40

Tgt Ion:149 Resp: 138137
 Ion Ratio Lower Upper
 149 100
 150 8.8 7.4 11.0
 104 6.3 5.4 8.2









#80
Fluoranthene
Concen: 18.904 ng/u1
RT: 19.781 min Scan# 21
Delta R.T. -0.007 min
Lab File: BG056725.D
Acq: 27 Feb 2023 10:40




Instrument :
BNA_G
ClientSampleId :
SSTD020593


Tgt Ion:202 Resp: 196659
Ion Ratio Lower Upper
202 100
101 6.0 4.7 7.1
100 5.2 4.4 6.6



#82
Pyrene
Concen: 19.094 ng/u1
RT: 20.145 min Scan# 2826
Delta R.T. -0.007 min
Lab File: BG056725.D
Acq: 27 Feb 2023 10:40

Tgt Ion:202 Resp: 195689
Ion Ratio Lower Upper
202 100
101 6.7 5.3 7.9
100 6.2 4.6 7.0









#83
Butylbenzylphthalate
Concen: 17.020 ng/ul
RT: 21.003 min Scan# 2972
Delta R.T. -0.007 min
Lab File: BG056725.D
Acq: 27 Feb 2023 10:40




Instrument :
BNA_G
ClientSampleId :
SSTD020593


Tgt Ion:149 Resp: 55337
Ion Ratio Lower Upper
149 100
91 71.4 59.3 88.9
206 22.9 19.7 29.5



#84
3,3'-Dichlorobenzidine
Concen: 20.542 ng/ul
RT: 21.943 min Scan# 3132
Delta R.T. -0.007 min
Lab File: BG056725.D
Acq: 27 Feb 2023 10:40

Tgt Ion:252 Resp: 65482
Ion Ratio Lower Upper
252 100
254 65.9 52.6 78.8
126 6.6 5.0 7.6





#85

Benzo(a)anthracene

Concen: 20.070 ng/ul

RT: 22.048 min Scan# 3150

Delta R.T. -0.007 min

Lab File: BG056725.D

Acq: 27 Feb 2023 10:40

Instrument :

BNA_G

ClientSampleId :

SSTD020593





Tgt Ion:228 Resp: 185721

Ion Ratio Lower Upper

228 100

229 18.6 16.2 24.4

226 26.0 20.6 31.0



#86

Bis(2-ethylhexyl)phthalate

Concen: 17.710 ng/ul

RT: 21.913 min Scan# 3127

Delta R.T. -0.001 min

Lab File: BG056725.D

Acq: 27 Feb 2023 10:40




Tgt Ion:149 Resp: 80886


Ion Ratio Lower Upper

149 100

167 26.8 21.8 32.8

279 4.9 4.2 6.2





#87

Chrysene

Concen: 20.096 ng/ul

RT: 22.119 min Scan# 3162

Delta R.T. -0.007 min

Lab File: BG056725.D

Acq: 27 Feb 2023 10:40

Instrument :

BNA_G

ClientSampleId :

SSTD020593





Tgt Ion:228 Resp: 173198

Ion Ratio Lower Upper

228 100

226 29.6 23.6 35.4

229 19.9 15.4 23.0



#89

Di-n-octyl phthalate

Concen: 18.621 ng/ul




RT: 23.224 min Scan# 3350


Delta R.T. -0.013 min

Lab File: BG056725.D

Acq: 27 Feb 2023 10:40

Tgt Ion:149 Resp: 135277









#90
Benzo(b)fluoranthene
Concen: 20.966 ng/ul
RT: 24.463 min Scan# 3561
Delta R.T. -0.013 min
Lab File: BG056725.D
Acq: 27 Feb 2023 10:40




Instrument :
BNA_G
ClientSampleId :
SSTD020593


Tgt Ion:252 Resp: 187549
Ion Ratio Lower Upper
252 100
253 22.5 17.4 26.2
125 4.4 3.0 4.6



#91
Benzo(k)fluoranthene
Concen: 20.492 ng/ul
RT: 24.540 min Scan# 3574
Delta R.T. -0.013 min
Lab File: BG056725.D
Acq: 27 Feb 2023 10:40

Tgt Ion:252 Resp: 178547
Ion Ratio Lower Upper
252 100
253 22.0 18.2 27.2
125 4.0 3.4 5.2





#93

Benzo(a)pyrene

Concen: 20.252 ng/ul

RT: 25.433 min Scan# 31

Delta R.T. -0.007 min

Lab File: BG056725.D

Acq: 27 Feb 2023 10:40

Instrument :

BNA_G

ClientSampleId :

SSTD020593





Tgt Ion:252 Resp: 159247

Ion Ratio Lower Upper

252 100

253 19.7 17.4 26.0

125 5.2 3.6 5.4



#94

Indeno(1,2,3-cd)pyrene

Concen: 20.467 ng/ul

RT: 29.657 min Scan# 4445

Delta R.T. -0.024 min

Lab File: BG056725.D

Acq: 27 Feb 2023 10:40




Tgt Ion:276 Resp: 218617


Ion Ratio Lower Upper

276 100

138 9.2 5.7 8.5#

277 24.4 18.1 27.1





#95

Dibenzo(a,h)anthracene

Concen: 19.869 ng/ul

RT: 29.734 min Scan# 4458

Delta R.T. -0.024 min

Lab File: BG056725.D

Acq: 27 Feb 2023 10:40

Instrument :

BNA_G

ClientSampleId :

SSTD020593





Tgt Ion:278 Resp: 177455

Ion Ratio Lower Upper

278 100

139 6.2 4.3 6.5

279 24.2 21.7 32.5



#96

Benzo(g,h,i)perylene

Concen: 20.487 ng/ul

RT: 30.944 min Scan# 4664

Delta R.T. -0.013 min

Lab File: BG056725.D

Acq: 27 Feb 2023 10:40

Tgt Ion:276 Resp: 174801

Ion Ratio Lower Upper

276 100

138 8.9 6.5 9.7

277 22.2 20.7 31.1

