

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG022723\
 Data File : BG056745.D
 Acq On : 28 Feb 2023 1:00
 Operator : CG/JU
 Sample : 01648-22
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 DBZW8

Quant Time: Feb 28 03:02:29 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG022323.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Feb 28 00:23:23 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.423	152	15528	20.000	ng/u1	0.00
20) Naphthalene-d8	11.261	136	66563	20.000	ng/u1	0.00
38) Acenaphthene-d10	15.027	164	52606	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.759	188	138017	20.000	ng/u1	0.00
79) Chrysene-d12	22.066	240	126654	20.000	ng/u1	# 0.00
88) Perylene-d12	25.597	264	136770	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.705	96	1998	7.138	ng/uL	0.01
4) Pyridine-d5	4.134	84	29193	25.837	ng/u1	0.00
7) Phenol-d5	7.541	99	49022	31.548	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.724	67	29233	30.842	ng/u1	0.00
11) 2-Chlorophenol-d4	7.941	132	32281	31.283	ng/u1	0.00
15) 4-Methylphenol-d8	9.104	113	36526	30.764	ng/u1	0.00
21) Nitrobenzene-d5	9.592	128	17206	30.845	ng/u1	0.00
24) 2-Nitrophenol-d4	10.321	143	20038	30.178	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.861	165	38026	30.398	ng/u1	0.00
31) 4-Chloroaniline-d4	11.378	131	45464	27.865	ng/u1	0.00
46) Dimethylphthalate-d6	14.410	166	140570	33.119	ng/u1	0.00
49) Acenaphthylene-d8	14.721	160	153057	31.728	ng/u1	0.00
54) 4-Nitrophenol-d4	15.174	143	23475	32.079	ng/u1	0.00
60) Fluorene-d10	16.008	176	125822	32.047	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	16.102	200	26645	28.856	ng/u1	0.00
73) Anthracene-d10	17.853	188	210291	31.407	ng/u1	0.00
81) Pyrene-d10	20.115	212	251872	29.388	ng/u1	0.00
92) Benzo(a)pyrene-d12	25.356	264	234851	31.813	ng/u1	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG022723\
Data File : BG056745.D
Acq On : 28 Feb 2023 1:00
Operator : CG/JU
Sample : 01648-22
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
DBZW8

Quant Time: Feb 28 03:02:29 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG022323.M
Quant Title : SVOA CALIBRATION
QLast Update : Tue Feb 28 00:23:23 2023
Response via : Initial Calibration

