

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG022823\
 Data File : BG056749.D
 Acq On : 28 Feb 2023 9:19
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SSTD020596

Quant Time: Feb 28 09:54:42 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG022323.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Feb 28 00:23:23 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.426	152	13995	20.000	ng/u1	0.00
20) Naphthalene-d8	11.258	136	62197	20.000	ng/u1	0.00
38) Acenaphthene-d10	15.024	164	50448	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.756	188	127667	20.000	ng/u1	0.00
79) Chrysene-d12	22.069	240	125915	20.000	ng/u1	0.00
88) Perylene-d12	25.600	264	137402	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.696	96	2443	9.683	ng/uL	0.00
4) Pyridine-d5	4.137	84	23330	22.910	ng/u1	0.00
7) Phenol-d5	7.539	99	28770	20.543	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.727	67	18294	21.415	ng/u1	0.00
11) 2-Chlorophenol-d4	7.944	132	18927	20.351	ng/u1	0.00
15) 4-Methylphenol-d8	9.107	113	21919	20.483	ng/u1	0.00
21) Nitrobenzene-d5	9.595	128	10336	19.830	ng/u1	0.00
24) 2-Nitrophenol-d4	10.318	143	12849	20.709	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.864	165	23630	20.216	ng/u1	0.00
31) 4-Chloroaniline-d4	11.375	131	31057	20.371	ng/u1	0.00
46) Dimethylphthalate-d6	14.413	166	82809	20.345	ng/u1	0.00
49) Acenaphthylene-d8	14.724	160	95859	20.721	ng/u1	0.00
54) 4-Nitrophenol-d4	15.177	143	14147	20.159	ng/u1	0.00
60) Fluorene-d10	16.005	176	79893	21.219	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	16.105	200	16345	19.136	ng/u1	0.00
73) Anthracene-d10	17.856	188	127778	20.631	ng/u1	0.00
81) Pyrene-d10	20.112	212	157134	18.442	ng/u1	0.00
92) Benzo(a)pyrene-d12	25.353	264	152296	20.535	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.737	88	3130	10.062	ng/uL	94
5) Pyridine	4.160	79	25193	23.360	ng/u1	94
6) Benzaldehyde	7.544	77	18651	25.420	ng/u1	98
8) Phenol	7.568	94	30170	21.120	ng/u1	94
10) Bis(2-Chloroethyl)ether	7.815	93	21717	21.397	ng/u1	97
12) 2-Chlorophenol	7.973	128	19191	20.837	ng/u1	91
13) 2-Methylphenol	8.843	108	21841	20.270	ng/u1	93
14) 2,2'-oxybis(1-Chloropr...	8.937	45	29220	21.424	ng/u1#	93
16) Acetophenone	9.248	105	38092	20.787	ng/u1	98
17) N-Nitroso-di-n-propyla...	9.219	70	21531	20.478	ng/u1	100
18) 4-Methylphenol	9.172	108	24197	20.667	ng/u1	87
19) Hexachloroethane	9.524	117	8520	21.754	ng/u1	94
22) Nitrobenzene	9.636	77	32608	21.574	ng/u1#	89
23) Isophorone	10.159	82	61238	20.723	ng/u1	97
25) 2-Nitrophenol	10.353	139	13047	21.303	ng/u1	96
26) 2,4-Dimethylphenol	10.388	107	28340	20.720	ng/u1	92
27) Bis(2-Chloroethoxy)met...	10.635	93	31835	21.314	ng/u1	96
29) 2,4-Dichlorophenol	10.888	162	23630	20.876	ng/u1	95
30) Naphthalene	11.311	128	70540	20.604	ng/u1	97
32) 4-Chloroaniline	11.399	127	31324	20.742	ng/u1	96
33) Hexachlorobutadiene	11.587	225	19348	21.447	ng/u1	91
34) Caprolactam	12.139	113	7400	19.231	ng/u1	83
35) 4-Chloro-3-methylphenol	12.480	107	26240	20.179	ng/u1	99
36) 2-Methylnaphthalene	12.879	142	50796	20.717	ng/u1	97


Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG022823\
 Data File : BG056749.D
 Acq On : 28 Feb 2023 9:19
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SSTD020596

Quant Time: Feb 28 09:54:42 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG022323.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Feb 28 00:23:23 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	13.097	142	51401	21.263	ng/ul	100
39) 1,2,4,5-Tetrachloroben...	13.238	216	38280	21.402	ng/ul	95
40) Hexachlorocyclopentadiene	13.220	237	23907	18.654	ng/ul	99
41) 2,4,6-Trichlorophenol	13.467	196	23787	19.904	ng/ul	93
42) 2,4,5-Trichlorophenol	13.537	196	26557	20.465	ng/ul	94
43) 1,1'-Biphenyl	13.866	154	76535	20.631	ng/ul	99
44) 2-Chloronaphthalene	13.913	162	63436	20.668	ng/ul	98
45) 2-Nitroaniline	14.101	65	22405	21.030	ng/ul	99
47) Dimethylphthalate	14.454	163	84655	20.452	ng/ul#	98
48) 2,6-Dinitrotoluene	14.583	165	17070	20.118	ng/ul#	83
50) Acenaphthylene	14.748	152	96673	20.686	ng/ul	97
51) 3-Nitroaniline	14.912	138	16434	21.641	ng/ul#	95
52) Acenaphthene	15.089	153	66664	20.765	ng/ul	98
53) 2,4-Dinitrophenol	15.112	184	9345	17.192	ng/ul#	81
55) 4-Nitrophenol	15.194	109	16318	20.089	ng/ul	95
56) Dibenzofuran	15.418	168	100341	20.589	ng/ul	99
57) 2,4-Dinitrotoluene	15.359	165	25966	21.346	ng/ul	93
58) 2,3,4,6-Tetrachlorophenol	15.635	232	24388	20.195	ng/ul	97
59) Diethylphthalate	15.805	149	84946	20.502	ng/ul	99
61) Fluorene	16.058	166	80484	20.625	ng/ul	96
62) 4-Chlorophenyl-phenyle...	16.040	204	47405	20.509	ng/ul	97
63) 4-Nitroaniline	16.064	138	15652	21.095	ng/ul	91
66) 4,6-Dinitro-2-methylph...	16.123	198	16332	19.861	ng/ul#	91
67) N-Nitrosodiphenylamine	16.252	169	73053	20.577	ng/ul	98
68) 4-Bromophenyl-phenylether	16.933	248	33314	20.591	ng/ul	95
69) Hexachlorobenzene	17.063	284	36060	20.746	ng/ul	94
70) Atrazine	17.180	200	31303	20.302	ng/ul	97
71) Pentachlorophenol	17.398	266	16731	16.952	ng/ul	98
72) Phenanthrene	17.797	178	146539	21.190	ng/ul	97
74) Anthracene	17.891	178	145739	20.854	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.837	216	39802	20.379	ng/ul	97
76) Pentachlorobenzene	15.335	250	41670	21.036	ng/ul	96
77) Carbazole	18.150	167	125587	21.477	ng/ul	98
78) Di-n-butylphthalate	18.678	149	141141	20.943	ng/ul	100
80) Fluoranthene	19.783	202	186040	18.058	ng/ul	98
82) Pyrene	20.141	202	187861	18.509	ng/ul	99
83) Butylbenzylphthalate	21.005	149	61196	19.006	ng/ul	98
84) 3,3'-Dichlorobenzidine	21.945	252	68825	21.801	ng/ul	94
85) Benzo(a)anthracene	22.045	228	184006	20.079	ng/ul	98
86) Bis(2-ethylhexyl)phtha...	21.910	149	89764	19.846	ng/ul	99
87) Chrysene	22.121	228	172216	20.177	ng/ul	100
89) Di-n-octyl phthalate	23.226	149	153145	20.501	ng/ul	100
90) Benzo(b)fluoranthene	24.466	252	189208	20.570	ng/ul	100
91) Benzo(k)fluoranthene	24.536	252	181669	20.277	ng/ul	97
93) Benzo(a)pyrene	25.429	252	163784	20.257	ng/ul	99
94) Indeno(1,2,3-cd)pyrene	29.654	276	220270	20.054	ng/ul#	98
95) Dibenzo(a,h)anthracene	29.736	278	181625	19.777	ng/ul	96
96) Benzo(g,h,i)perylene	30.941	276	179116	20.415	ng/ul	96





(#) = qualifier out of range (m) = manual integration (+) = signals summed



#2
1,4-Dioxane
Concen: 10.062 ng/uL
RT: 3.737 min Scan# 34
Delta R.T. 0.003 min
Lab File: BG056749.D
Acq: 28 Feb 2023 9:19




Instrument :
BNA_G
ClientSampleId :
SSTD020596


Tgt Ion: 88 Resp: 3130
Ion Ratio Lower Upper
88 100
43 37.8 28.3 42.5
58 68.4 50.6 75.8



#5
Pyridine
Concen: 23.360 ng/uL
RT: 4.160 min Scan# 106
Delta R.T. 0.003 min
Lab File: BG056749.D
Acq: 28 Feb 2023 9:19

Tgt Ion: 79 Resp: 25193
Ion Ratio Lower Upper
79 100
52 68.7 60.6 91.0
51 38.0 31.9 47.9









#6
Benzaldehyde
Concen: 25.420 ng/ul
RT: 7.544 min Scan# 681
Delta R.T. 0.003 min
Lab File: BG056749.D
Acq: 28 Feb 2023 9:19




Instrument :
BNA_G
ClientSampleId :
SSTD020596


Tgt Ion: 77 Resp: 18651
Ion Ratio Lower Upper
77 100
105 88.3 72.2 108.4
106 86.0 67.4 101.2



#8
Phenol
Concen: 21.120 ng/ul
RT: 7.568 min Scan# 686
Delta R.T. 0.003 min
Lab File: BG056749.D
Acq: 28 Feb 2023 9:19

Tgt Ion: 94 Resp: 30170
Ion Ratio Lower Upper
94 100
65 30.7 25.7 38.5
66 43.0 38.9 58.3









#10
 Bis(2-Chloroethyl)ether
 Concen: 21.397 ng/ul
 RT: 7.815 min Scan# 71
 Delta R.T. -0.003 min
 Lab File: BG056749.D
 Acq: 28 Feb 2023 9:19




Instrument :
 BNA_G
 ClientSampleId :
 SSTD020596


Tgt Ion: 93	Resp:	21717	
Ion	Ratio	Lower	Upper
93	100		
63	86.4	66.6	100.0
95	29.8	25.0	37.4



#12
 2-Chlorophenol
 Concen: 20.837 ng/ul
 RT: 7.973 min Scan# 755
 Delta R.T. 0.003 min
 Lab File: BG056749.D
 Acq: 28 Feb 2023 9:19

Tgt Ion:128	Resp:	19191	
Ion Ratio	Lower	Upper	
128 100			
64 51.1	45.5	68.3	
130 28.5	27.7	41.5	





#13

2-Methylphenol

Concen: 20.270 ng/ul

RT: 8.843 min Scan# 902

Delta R.T. 0.003 min

Lab File: BG056749.D

Acq: 28 Feb 2023 9:19

Instrument :

BNA_G

ClientSampleId :





SSTD020596

Tgt Ion:108 Resp: 21841

Ion Ratio Lower Upper

108 100

107 85.3 73.4 110.0



#14

2,2'-oxybis(1-Chloropropane)

Concen: 21.424 ng/ul

RT: 8.937 min Scan# 919

Delta R.T. -0.003 min

Lab File: BG056749.D

Acq: 28 Feb 2023 9:19




Tgt Ion: 45 Resp: 29220


Ion Ratio Lower Upper

45 100

77 20.2 12.4 18.6#

79 12.5 9.8 14.6









#16
Acetophenone
Concen: 20.787 ng/ul
RT: 9.248 min Scan# 91
Delta R.T. 0.003 min
Lab File: BG056749.D
Acq: 28 Feb 2023 9:19




Instrument :
BNA_G
ClientSampleId :
SSTD020596


Tgt Ion:105 Resp: 38092
Ion Ratio Lower Upper
105 100
77 86.9 71.4 107.0
51 28.7 23.6 35.4



#17
N-Nitroso-di-n-propylamine
Concen: 20.478 ng/ul
RT: 9.219 min Scan# 967
Delta R.T. -0.003 min
Lab File: BG056749.D
Acq: 28 Feb 2023 9:19

Tgt Ion: 70 Resp: 21531
Ion Ratio Lower Upper
70 100
42 58.3 46.6 70.0
101 8.4 7.3 10.9
130 16.6 13.7 20.5









#18
4-Methylphenol
Concen: 20.667 ng/ul
RT: 9.172 min Scan# 91
Delta R.T. 0.003 min
Lab File: BG056749.D
Acq: 28 Feb 2023 9:19




Instrument :
BNA_G
ClientSampleId :
SSTD020596


Tgt Ion:108 Resp: 24197
Ion Ratio Lower Upper
108 100
107 101.3 92.1 138.1



#19
Hexachloroethane
Concen: 21.754 ng/ul
RT: 9.524 min Scan# 1019
Delta R.T. 0.003 min
Lab File: BG056749.D
Acq: 28 Feb 2023 9:19

Tgt Ion:117 Resp: 8520
Ion Ratio Lower Upper
117 100
201 114.4 95.2 142.8
199 79.2 57.9 86.9









#22
Nitrobenzene
Concen: 21.574 ng/ul
RT: 9.636 min Scan# 1037
Delta R.T. 0.003 min
Lab File: BG056749.D
Acq: 28 Feb 2023 9:19




Instrument :
BNA_G
ClientSampleId :
SSTD020596


Tgt Ion: 77 Resp: 32608
Ion Ratio Lower Upper
77 100
123 42.9 27.8 41.6#
65 13.8 10.6 16.0



#23
Isophorone
Concen: 20.723 ng/ul
RT: 10.159 min Scan# 1127
Delta R.T. 0.003 min
Lab File: BG056749.D
Acq: 28 Feb 2023 9:19

Tgt Ion: 82 Resp: 61238
Ion Ratio Lower Upper
82 100
95 9.2 7.7 11.5
138 13.4 12.1 18.1









#25
2-Nitrophenol
Concen: 21.303 ng/ul
RT: 10.353 min Scan# 1160
Delta R.T. -0.003 min
Lab File: BG056749.D
Acq: 28 Feb 2023 9:19




Instrument :
BNA_G
ClientSampleId :
SSTD020596


Tgt Ion:139 Resp: 13047
Ion Ratio Lower Upper
139 100
65 56.8 49.4 74.2
109 39.0 31.6 47.4



#26
2,4-Dimethylphenol
Concen: 20.720 ng/ul
RT: 10.388 min Scan# 1166
Delta R.T. -0.003 min
Lab File: BG056749.D
Acq: 28 Feb 2023 9:19

Tgt Ion:107 Resp: 28340
Ion Ratio Lower Upper
107 100
121 44.0 40.2 60.4
122 72.8 63.1 94.7





#27

Bis(2-Chloroethoxy)methane

Concen: 21.314 ng/ul

RT: 10.635 min Scan# 1

Delta R.T. 0.003 min

Lab File: BG056749.D

Acq: 28 Feb 2023 9:19

Instrument :

BNA_G

ClientSampleId :

SSTD020596





Tgt Ion: 93 Resp: 31835

Ion Ratio Lower Upper

93 100

95 33.4 24.6 37.0

123 9.6 6.6 10.0



#29

2,4-Dichlorophenol

Concen: 20.876 ng/ul

RT: 10.888 min Scan# 1251

Delta R.T. 0.003 min

Lab File: BG056749.D

Acq: 28 Feb 2023 9:19




Tgt Ion:162 Resp: 23630


Ion Ratio Lower Upper

162 100

164 63.3 53.3 79.9

98 35.6 25.5 38.3









#30
Naphthalene
Concen: 20.604 ng/ul
RT: 11.311 min Scan# 11311
Delta R.T. 0.003 min
Lab File: BG056749.D
Acq: 28 Feb 2023 9:19




Instrument :
BNA_G
ClientSampleId :
SSTD020596


Tgt Ion:128 Resp: 70540
Ion Ratio Lower Upper
128 100
129 11.1 8.6 12.8
127 14.0 9.8 14.8



#32
4-Chloroaniline
Concen: 20.742 ng/ul
RT: 11.399 min Scan# 1338
Delta R.T. -0.003 min
Lab File: BG056749.D
Acq: 28 Feb 2023 9:19

Tgt Ion:127 Resp: 31324
Ion Ratio Lower Upper
127 100
129 27.0 23.4 35.0









#33
Hexachlorobutadiene
Concen: 21.447 ng/ul
RT: 11.587 min Scan# 11
Delta R.T. -0.003 min
Lab File: BG056749.D
Acq: 28 Feb 2023 9:19




Instrument :
BNA_G
ClientSampleId :
SSTD020596


Tgt Ion:225 Resp: 19348
Ion Ratio Lower Upper
225 100
223 57.9 53.5 80.3
227 67.1 58.2 87.2



#34
Caprolactam
Concen: 19.231 ng/ul
RT: 12.139 min Scan# 1464
Delta R.T. 0.003 min
Lab File: BG056749.D
Acq: 28 Feb 2023 9:19


Tgt Ion:113 Resp: 7400
Ion Ratio Lower Upper
113 100
55 185.4 167.0 250.6
56 165.3 113.5 170.3








#35
4-Chloro-3-methylphenol
Concen: 20.179 ng/ul
RT: 12.480 min Scan# 11
Delta R.T. -0.003 min
Lab File: BG056749.D
Acq: 28 Feb 2023 9:19

Instrument :
BNA_G
ClientSampleId :
SSTD020596







Tgt Ion:107 Resp: 26240
Ion Ratio Lower Upper
107 100
144 25.3 20.5 30.7
142 70.4 55.8 83.6



#36
2-Methylnaphthalene
Concen: 20.717 ng/ul
RT: 12.879 min Scan# 1590
Delta R.T. -0.003 min
Lab File: BG056749.D
Acq: 28 Feb 2023 9:19

Tgt Ion:142 Resp: 50796
Ion Ratio Lower Upper
142 100
141 91.4 71.0 106.6





#37

1-Methylnaphthalene

Concen: 21.263 ng/ul

RT: 13.097 min Scan# 1

Delta R.T. -0.003 min

Lab File: BG056749.D

Acq: 28 Feb 2023 9:19

Instrument :

BNA_G

ClientSampleId :

SSTD020596





Tgt Ion:142 Resp: 51401

Ion Ratio Lower Upper

142 100

141 91.7 73.5 110.3

116 4.3 4.0 6.0



#39

1,2,4,5-Tetrachlorobenzene

Concen: 21.402 ng/ul

RT: 13.238 min Scan# 1651

Delta R.T. 0.003 min

Lab File: BG056749.D

Acq: 28 Feb 2023 9:19

Tgt Ion:216 Resp: 38280




Ion Ratio Lower Upper


216 100

214 75.7 64.4 96.6

179 19.2 16.9 25.3

108 12.8 11.6 17.4









#40
Hexachlorocyclopentadiene
Concen: 18.654 ng/ul
RT: 13.220 min Scan# 1648
Delta R.T. 0.003 min
Lab File: BG056749.D
Acq: 28 Feb 2023 9:19




Instrument :
BNA_G
ClientSampleId :
SSTD020596


Tgt Ion:237 Resp: 23907
Ion Ratio Lower Upper
237 100
235 59.6 47.8 71.8
272 13.4 9.9 14.9



#41
2,4,6-Trichlorophenol
Concen: 19.904 ng/ul
RT: 13.467 min Scan# 1690
Delta R.T. 0.003 min
Lab File: BG056749.D
Acq: 28 Feb 2023 9:19

Tgt Ion:196 Resp: 23787
Ion Ratio Lower Upper
196 100
198 101.2 75.0 112.6
200 29.8 22.1 33.1





#42

2,4,5-Trichlorophenol

Concen: 20.465 ng/ul

RT: 13.537 min Scan# 1

Delta R.T. 0.003 min

Lab File: BG056749.D

Acq: 28 Feb 2023 9:19

Instrument :

BNA_G

ClientSampleId :

SSTD020596





Tgt Ion:196 Resp: 26557

Ion Ratio Lower Upper

196 100

198 94.3 80.1 120.1

200 30.5 26.7 40.1



#43

1,1'-Biphenyl

Concen: 20.631 ng/ul

RT: 13.866 min Scan# 1758

Delta R.T. 0.003 min

Lab File: BG056749.D

Acq: 28 Feb 2023 9:19




Tgt Ion:154 Resp: 76535


Ion Ratio Lower Upper

154 100

153 39.1 32.0 48.0

76 11.4 9.0 13.4









#44
2-Chloronaphthalene
Concen: 20.668 ng/ul
RT: 13.913 min Scan# 1766
Delta R.T. -0.003 min
Lab File: BG056749.D
Acq: 28 Feb 2023 9:19




Instrument :
BNA_G
ClientSampleId :
SSTD020596


Tgt Ion:162 Resp: 63436
Ion Ratio Lower Upper
162 100
164 30.0 23.3 34.9
127 33.7 25.8 38.8



#45
2-Nitroaniline
Concen: 21.030 ng/ul
RT: 14.101 min Scan# 1798
Delta R.T. 0.003 min
Lab File: BG056749.D
Acq: 28 Feb 2023 9:19

Tgt Ion: 65 Resp: 22405
Ion Ratio Lower Upper
65 100
92 67.0 53.9 80.9
138 94.2 73.8 110.6









#47
Dimethylphthalate
Concen: 20.452 ng/ul
RT: 14.454 min Scan# 1858
Delta R.T. -0.003 min
Lab File: BG056749.D
Acq: 28 Feb 2023 9:19




Instrument :
BNA_G
ClientSampleId :
SSTD020596


Tgt Ion:163 Resp: 84655
Ion Ratio Lower Upper
163 100
194 4.9 2.5 3.7#
164 10.6 8.2 12.4



#48
2,6-Dinitrotoluene
Concen: 20.118 ng/ul
RT: 14.583 min Scan# 1880
Delta R.T. 0.003 min
Lab File: BG056749.D
Acq: 28 Feb 2023 9:19

Tgt Ion:165 Resp: 17070
Ion Ratio Lower Upper
165 100
89 76.2 48.4 72.6#
121 21.1 15.5 23.3









#50
Acenaphthylene
Concen: 20.686 ng/ul
RT: 14.748 min Scan# 1908
Delta R.T. -0.003 min
Lab File: BG056749.D
Acq: 28 Feb 2023 9:19




Instrument :
BNA_G
ClientSampleId :
SSTD020596


Tgt Ion:152 Resp: 96673
Ion Ratio Lower Upper
152 100
151 21.3 15.8 23.6
153 12.6 10.6 16.0



#51
3-Nitroaniline
Concen: 21.641 ng/ul
RT: 14.912 min Scan# 1936
Delta R.T. 0.003 min
Lab File: BG056749.D
Acq: 28 Feb 2023 9:19

Tgt Ion:138 Resp: 16434
Ion Ratio Lower Upper
138 100
108 11.3 6.5 9.7#
92 122.5 102.2 153.2









#52
Acenaphthene
Concen: 20.765 ng/ul
RT: 15.089 min Scan# 1966
Delta R.T. -0.003 min
Lab File: BG056749.D
Acq: 28 Feb 2023 9:19




Instrument :
BNA_G
ClientSampleId :
SSTD020596


Tgt Ion:153 Resp: 66664
Ion Ratio Lower Upper
153 100
152 47.4 38.6 58.0
154 86.6 70.9 106.3



#53
2,4-Dinitrophenol
Concen: 17.192 ng/ul
RT: 15.112 min Scan# 1970
Delta R.T. -0.003 min
Lab File: BG056749.D
Acq: 28 Feb 2023 9:19

Tgt Ion:184 Resp: 9345
Ion Ratio Lower Upper
184 100
63 60.9 43.8 65.6
154 106.4 66.8 100.2#









#55
4-Nitrophenol
Concen: 20.089 ng/ul
RT: 15.194 min Scan# 1983
Delta R.T. 0.003 min
Lab File: BG056749.D
Acq: 28 Feb 2023 9:19




Instrument :
BNA_G
ClientSampleId :
SSTD020596


Tgt Ion:109 Resp: 16318
Ion Ratio Lower Upper
109 100
139 82.3 71.2 106.8
65 109.1 85.1 127.7



#56
Dibenzofuran
Concen: 20.589 ng/ul
RT: 15.418 min Scan# 2022
Delta R.T. -0.003 min
Lab File: BG056749.D
Acq: 28 Feb 2023 9:19

Tgt Ion:168 Resp: 100341
Ion Ratio Lower Upper
168 100
139 35.9 28.2 42.2









#57
2,4-Dinitrotoluene
Concen: 21.346 ng/ul
RT: 15.359 min Scan# 2012
Delta R.T. -0.003 min
Lab File: BG056749.D
Acq: 28 Feb 2023 9:19




Instrument :
BNA_G
ClientSampleId :
SSTD020596


Tgt Ion:165 Resp: 25966
Ion Ratio Lower Upper
165 100
63 53.5 38.6 58.0
182 2.4 1.8 2.8



#58
2,3,4,6-Tetrachlorophenol
Concen: 20.195 ng/ul
RT: 15.635 min Scan# 2059
Delta R.T. 0.003 min
Lab File: BG056749.D
Acq: 28 Feb 2023 9:19

Tgt Ion:232 Resp: 24388
Ion Ratio Lower Upper
232 100
131 34.7 27.8 41.8
130 2.5 1.8 2.6
166 22.3 20.2 30.4









#59
Diethylphthalate
Concen: 20.502 ng/ul
RT: 15.805 min Scan# 2088
Delta R.T. -0.003 min
Lab File: BG056749.D
Acq: 28 Feb 2023 9:19




Instrument :
BNA_G
ClientSampleId :
SSTD020596


Tgt Ion:149 Resp: 84946
Ion Ratio Lower Upper
149 100
177 24.2 18.6 28.0
150 13.1 10.7 16.1



#61
Fluorene
Concen: 20.625 ng/ul
RT: 16.058 min Scan# 2131
Delta R.T. -0.003 min
Lab File: BG056749.D
Acq: 28 Feb 2023 9:19

Tgt Ion:166 Resp: 80484
Ion Ratio Lower Upper
166 100
165 93.9 78.4 117.6
167 13.1 10.5 15.7









#62
4-Chlorophenyl-phenylether
Concen: 20.509 ng/ul
RT: 16.040 min Scan# 2128
Delta R.T. -0.003 min
Lab File: BG056749.D
Acq: 28 Feb 2023 9:19




Instrument :
BNA_G
ClientSampleId :
SSTD020596


Tgt Ion:204 Resp: 47405
Ion Ratio Lower Upper
204 100
206 31.3 26.1 39.1
141 52.2 39.7 59.5



#63
4-Nitroaniline
Concen: 21.095 ng/ul
RT: 16.064 min Scan# 2132
Delta R.T. 0.003 min
Lab File: BG056749.D
Acq: 28 Feb 2023 9:19

Tgt Ion:138 Resp: 15652
Ion Ratio Lower Upper
138 100
92 55.4 48.6 73.0
108 126.7 93.0 139.6









#66
4,6-Dinitro-2-methylphenol
Concen: 19.861 ng/ul
RT: 16.123 min Scan# 2141
Delta R.T. 0.003 min
Lab File: BG056749.D
Acq: 28 Feb 2023 9:19




Instrument :
BNA_G
ClientSampleId :
SSTD020596


Tgt Ion:198 Resp: 16332
Ion Ratio Lower Upper
198 100
182 3.4 2.4 3.6
77 21.0 21.0 31.6#



#67
N-Nitrosodiphenylamine
Concen: 20.577 ng/ul
RT: 16.252 min Scan# 2164
Delta R.T. -0.003 min
Lab File: BG056749.D
Acq: 28 Feb 2023 9:19

Tgt Ion:169 Resp: 73053
Ion Ratio Lower Upper
169 100
168 69.8 55.1 82.7
167 37.2 28.3 42.5









#68
 4-Bromophenyl-phenylether
 Concen: 20.591 ng/ul
 RT: 16.933 min Scan# 21
 Delta R.T. -0.003 min
 Lab File: BG056749.D
 Acq: 28 Feb 2023 9:19




Instrument :
 BNA_G
 ClientSampleId :
 SSTD020596


Tgt Ion:248 Resp: 33314
 Ion Ratio Lower Upper
 248 100
 250 95.9 80.4 120.6
 141 64.1 47.6 71.4



#69
 Hexachlorobenzene
 Concen: 20.746 ng/ul
 RT: 17.063 min Scan# 2302
 Delta R.T. 0.003 min
 Lab File: BG056749.D
 Acq: 28 Feb 2023 9:19

Tgt Ion:284 Resp: 36060
 Ion Ratio Lower Upper
 284 100
 142 28.5 21.8 32.6
 249 37.1 25.9 38.9





#70

Atrazine

Concen: 20.302 ng/ul

RT: 17.180 min Scan# 21

Delta R.T. -0.003 min

Lab File: BG056749.D

Acq: 28 Feb 2023 9:19

Instrument :

BNA_G

ClientSampleId :

SSTD020596





Tgt Ion:200 Resp: 31303

Ion Ratio Lower Upper

200 100

173 23.6 17.4 26.0

215 50.2 41.7 62.5



#71

Pentachlorophenol

Concen: 16.952 ng/ul

RT: 17.398 min Scan# 2359

Delta R.T. -0.003 min

Lab File: BG056749.D

Acq: 28 Feb 2023 9:19




Tgt Ion:266 Resp: 16731


Ion Ratio Lower Upper

266 100

264 64.0 49.9 74.9

268 67.0 54.5 81.7









#72
Phenanthrene
Concen: 21.190 ng/ul
RT: 17.797 min Scan# 2427
Delta R.T. -0.003 min
Lab File: BG056749.D
Acq: 28 Feb 2023 9:19




Instrument :
BNA_G
ClientSampleId :
SSTD020596


Tgt Ion:178 Resp: 146539
Ion Ratio Lower Upper
178 100
179 15.9 12.1 18.1
176 19.6 14.4 21.6



#74
Anthracene
Concen: 20.854 ng/ul
RT: 17.891 min Scan# 2443
Delta R.T. -0.003 min
Lab File: BG056749.D
Acq: 28 Feb 2023 9:19

Tgt Ion:178 Resp: 145739
Ion Ratio Lower Upper
178 100
179 14.2 12.1 18.1
176 17.5 14.1 21.1









#75
1,2,3,4-Tetrachlorobenzene
Concen: 20.379 ng/uL
RT: 13.837 min Scan# 1
Delta R.T. 0.003 min
Lab File: BG056749.D
Acq: 28 Feb 2023 9:19




Instrument :
BNA_G
ClientSampleId :
SSTD020596


Tgt Ion:216 Resp: 39802
Ion Ratio Lower Upper
216 100
214 77.1 64.2 96.2
218 44.6 36.6 54.8



#76
Pentachlorobenzene
Concen: 21.036 ng/uL
RT: 15.335 min Scan# 2008
Delta R.T. -0.003 min
Lab File: BG056749.D
Acq: 28 Feb 2023 9:19

Tgt Ion:250 Resp: 41670
Ion Ratio Lower Upper
250 100
248 62.0 47.1 70.7
252 62.8 52.4 78.6









#77
 Carbazole
 Concen: 21.477 ng/ul
 RT: 18.150 min Scan# 2487
 Delta R.T. -0.003 min
 Lab File: BG056749.D
 Acq: 28 Feb 2023 9:19




Instrument :
 BNA_G
 ClientSampleId :
 SSTD020596


Tgt Ion:167 Resp: 125587
 Ion Ratio Lower Upper
 167 100
 166 21.4 18.2 27.2
 139 12.6 9.8 14.8



#78
 Di-n-butylphthalate
 Concen: 20.943 ng/ul
 RT: 18.678 min Scan# 2577
 Delta R.T. -0.003 min
 Lab File: BG056749.D
 Acq: 28 Feb 2023 9:19

Tgt Ion:149 Resp: 141141
 Ion Ratio Lower Upper
 149 100
 150 9.1 7.4 11.0
 104 6.6 5.4 8.2









#80
 Fluoranthene
 Concen: 18.058 ng/ul
 RT: 19.783 min Scan# 21
 Delta R.T. -0.003 min
 Lab File: BG056749.D
 Acq: 28 Feb 2023 9:19




Instrument :
 BNA_G
 ClientSampleId :
 SSTD020596


Tgt Ion:202 Resp: 186040
 Ion Ratio Lower Upper
 202 100
 101 5.4 4.7 7.1
 100 4.7 4.4 6.6



#82
 Pyrene
 Concen: 18.509 ng/ul
 RT: 20.141 min Scan# 2826
 Delta R.T. -0.003 min
 Lab File: BG056749.D
 Acq: 28 Feb 2023 9:19

Tgt Ion:202 Resp: 187861
 Ion Ratio Lower Upper
 202 100
 101 6.3 5.3 7.9
 100 6.0 4.6 7.0









#83
Butylbenzylphthalate
Concen: 19.006 ng/ul
RT: 21.005 min Scan# 2973
Delta R.T. -0.003 min
Lab File: BG056749.D
Acq: 28 Feb 2023 9:19




Instrument :
BNA_G
ClientSampleId :
SSTD020596


Tgt Ion:149 Resp: 61196
Ion Ratio Lower Upper
149 100
91 72.3 59.3 88.9
206 25.3 19.7 29.5



#84
3,3'-Dichlorobenzidine
Concen: 21.801 ng/ul
RT: 21.945 min Scan# 3133
Delta R.T. 0.003 min
Lab File: BG056749.D
Acq: 28 Feb 2023 9:19

Tgt Ion:252 Resp: 68825
Ion Ratio Lower Upper
252 100
254 60.7 52.6 78.8
126 6.1 5.0 7.6





#85

Benzo(a)anthracene

Concen: 20.079 ng/ul

RT: 22.045 min Scan# 3150

Delta R.T. -0.003 min

Lab File: BG056749.D

Acq: 28 Feb 2023 9:19

Instrument :

BNA_G

ClientSampleId :

SSTD020596





Tgt Ion:228 Resp: 184006

Ion Ratio Lower Upper

228 100

229 19.7 16.2 24.4

226 27.0 20.6 31.0



#86

Bis(2-ethylhexyl)phthalate

Concen: 19.846 ng/ul

RT: 21.910 min Scan# 3127

Delta R.T. -0.003 min

Lab File: BG056749.D

Acq: 28 Feb 2023 9:19




Tgt Ion:149 Resp: 89764


Ion Ratio Lower Upper

149 100

167 26.6 21.8 32.8


279 4.7 4.2 6.2








#87
Chrysene
Concen: 20.177 ng/ul
RT: 22.121 min Scan# 3163
Delta R.T. 0.003 min
Lab File: BG056749.D
Acq: 28 Feb 2023 9:19

Instrument :
BNA_G
ClientSampleId :
SSTD020596







Tgt Ion:228 Resp: 172216
Ion Ratio Lower Upper
228 100
226 29.7 23.6 35.4
229 19.4 15.4 23.0



#89
Di-n-octyl phthalate
Concen: 20.501 ng/ul
RT: 23.226 min Scan# 3351
Delta R.T. -0.003 min
Lab File: BG056749.D
Acq: 28 Feb 2023 9:19

Tgt Ion:149 Resp: 153145









#90
Benzo(b)fluoranthene
Concen: 20.570 ng/ul
RT: 24.466 min Scan# 3561
Delta R.T. 0.003 min
Lab File: BG056749.D
Acq: 28 Feb 2023 9:19




Instrument :
BNA_G
ClientSampleId :
SSTD020596


Tgt Ion:252 Resp: 189208
Ion Ratio Lower Upper
252 100
253 21.7 17.4 26.2
125 4.1 3.0 4.6



#91
Benzo(k)fluoranthene
Concen: 20.277 ng/ul
RT: 24.536 min Scan# 3574
Delta R.T. -0.003 min
Lab File: BG056749.D
Acq: 28 Feb 2023 9:19

Tgt Ion:252 Resp: 181669
Ion Ratio Lower Upper
252 100
253 20.7 18.2 27.2
125 4.4 3.4 5.2





#93

Benzo(a)pyrene

Concen: 20.257 ng/ul

RT: 25.429 min Scan# 31

Delta R.T. 0.003 min

Lab File: BG056749.D

Acq: 28 Feb 2023 9:19

Instrument :

BNA_G

ClientSampleId :

SSTD020596





Tgt Ion:252 Resp: 163784

Ion Ratio Lower Upper

252 100

253 21.6 17.4 26.0

125 5.3 3.6 5.4



#94

Indeno(1,2,3-cd)pyrene

Concen: 20.054 ng/ul

RT: 29.654 min Scan# 4445

Delta R.T. -0.015 min

Lab File: BG056749.D

Acq: 28 Feb 2023 9:19




Tgt Ion:276 Resp: 220270


Ion Ratio Lower Upper

276 100

138 8.8 5.7 8.5#

277 22.2 18.1 27.1









#95
 Dibenzo(a,h)anthracene
 Concen: 19.777 ng/ul
 RT: 29.736 min Scan# 4459
 Delta R.T. 0.003 min
 Lab File: BG056749.D
 Acq: 28 Feb 2023 9:19

Instrument :
 BNA_G
 ClientSampleId :
 SST020596

Tgt Ion:278 Resp: 181625
 Ion Ratio Lower Upper
 278 100
 139 5.9 4.3 6.5
 279 24.8 21.7 32.5



#96
 Benzo(g,h,i)perylene
 Concen: 20.415 ng/ul
 RT: 30.941 min Scan# 4664
 Delta R.T. -0.009 min
 Lab File: BG056749.D
 Acq: 28 Feb 2023 9:19

Tgt Ion:276 Resp: 179116
 Ion Ratio Lower Upper
 276 100
 138 7.3 6.5 9.7
 277 23.7 20.7 31.1

