

Data Path : Z:\SVOASRV\HPCHEM1\BNA G\DATA\BG030119\
 Data File : BG039705.D
 Acq On : 2 Mar 2019 2:52
 Operator : JU/SJ
 Sample : K1675-29
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 Z-3-13-A

Manual Integrations
 APPROVED

Sohil
 3/4/2019 12:19:37 PM

Quant Time: Mar 02 03:40:36 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG012919.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Mar 01 06:15:02 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.16	152	59218	20.00	ng	0.00
21) Naphthalene-d8	10.98	136	265022	20.00	ng	0.00
39) Acenaphthene-d10	14.79	164	177815	20.00	ng	0.00
64) Phenanthrene-d10	17.53	188	423820	20.00	ng	0.00
76) Chrysene-d12	21.82	240	427385	20.00	ng	0.00
87) Perylene-d12	25.20	264	426891	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.71	112	359644	103.94	ng	0.00
7) Phenol-d6	7.31	99	532868	104.63	ng	0.00
23) Nitrobenzene-d5	9.34	82	313645	73.93	ng	0.00
42) 2,4,6-Tribromophenol	16.27	330	182523	95.32	ng	0.00
45) 2-Fluorobiphenyl	13.41	172	748526	60.39	ng	0.00
79) Terphenyl-d14	20.13	244	1125415	55.74	ng	0.00
Target Compounds						
50) Dimethylphthalate	14.23	163	52793	3.676	ng	Qvalue # 97
71) Phenanthrene	17.58	178	66193	3.018	ng	97
75) Fluoranthene	19.57	202	103769	4.076	ng	96
78) Pyrene	19.94	202	92772	3.487	ng	96
88) Benzo(b)fluoranthene	24.12	252	50418m	2.166	ng	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA G\DATA\BG030119\
Data File : BG039705.D
Acq On : 2 Mar 2019 2:52
Operator : JU/SJ
Sample : K1675-29
Misc :
ALS Vial : 17 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
Z-3-13-A

Manual Integrations
APPROVED

Sohil
3/4/2019 12:19:37 PM

Quant Time: Mar 02 03:40:36 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG012919.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Mar 01 06:15:02 2019
Response via : Initial Calibration

