

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG030123\
 Data File : BG056772.D
 Acq On : 1 Mar 2023 20:35
 Operator : CG/JU
 Sample : 01649-07
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 DBZV4

Quant Time: Mar 02 00:16:07 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG022323.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Mar 02 00:12:28 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.420	152	16108	20.000	ng/ul	0.00
20) Naphthalene-d8	11.257	136	71357	20.000	ng/ul	0.00
38) Acenaphthene-d10	15.024	164	55924	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.750	188	137666	20.000	ng/ul	0.00
79) Chrysene-d12	22.057	240	132274	20.000	ng/ul	0.00
88) Perylene-d12	25.582	264	141359	20.000	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.696	96	1587	5.465	ng/uL	0.00
4) Pyridine-d5	4.136	84	22565	19.252	ng/ul	0.00
7) Phenol-d5	7.538	99	32983	20.462	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.720	67	22544	22.928	ng/ul	0.00
11) 2-Chlorophenol-d4	7.938	132	22814	21.313	ng/ul	0.00
15) 4-Methylphenol-d8	9.101	113	25942	21.063	ng/ul	0.00
21) Nitrobenzene-d5	9.589	128	12906	21.582	ng/ul	0.00
24) 2-Nitrophenol-d4	10.317	143	15297	21.490	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.858	165	27850	20.768	ng/ul	0.00
31) 4-Chloroaniline-d4	11.375	131	35932	20.544	ng/ul	0.00
46) Dimethylphthalate-d6	14.407	166	101165	22.421	ng/ul	0.00
49) Acenaphthylene-d8	14.718	160	119413	23.285	ng/ul	0.00
54) 4-Nitrophenol-d4	15.171	143	14493	18.630	ng/ul	0.00
60) Fluorene-d10	16.005	176	95348	22.844	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	16.105	200	16546	17.965	ng/ul	0.00
73) Anthracene-d10	17.850	188	155720	23.316	ng/ul	0.00
81) Pyrene-d10	20.112	212	185466	20.721	ng/ul	0.00
92) Benzo(a)pyrene-d12	25.341	264	174994	22.935	ng/ul	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG030123\
Data File : BG056772.D
Acq On : 1 Mar 2023 20:35
Operator : CG/JU
Sample : 01649-07
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
DBZV4

Quant Time: Mar 02 00:16:07 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG022323.M
Quant Title : SVOA CALIBRATION
QLast Update : Thu Mar 02 00:12:28 2023
Response via : Initial Calibration

