

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG030123\
 Data File : BG056770.D
 Acq On : 1 Mar 2023 19:12
 Operator : CG/JU
 Sample : 01649-11
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 DBZV8

Quant Time: Mar 02 00:15:53 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG022323.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Mar 02 00:12:28 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.417	152	17286	20.000	ng/u1	0.00
20) Naphthalene-d8	11.255	136	72870	20.000	ng/u1	0.00
38) Acenaphthene-d10	15.021	164	59320	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.753	188	144610	20.000	ng/u1	0.00
79) Chrysene-d12	22.060	240	134980	20.000	ng/u1	0.00
88) Perylene-d12	25.585	264	144342	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.699	96	1608	5.160	ng/uL	0.00
4) Pyridine-d5	4.134	84	22874	18.186	ng/u1	0.00
7) Phenol-d5	7.536	99	34250	19.800	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.718	67	22573	21.393	ng/u1	0.00
11) 2-Chlorophenol-d4	7.941	132	23647	20.585	ng/u1	0.00
15) 4-Methylphenol-d8	9.104	113	25854	19.561	ng/u1	0.00
21) Nitrobenzene-d5	9.592	128	13205	21.623	ng/u1	0.00
24) 2-Nitrophenol-d4	10.321	143	15161	20.856	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.855	165	29095	21.246	ng/u1	0.00
31) 4-Chloroaniline-d4	11.378	131	35795	20.040	ng/u1	0.00
46) Dimethylphthalate-d6	14.404	166	101321	21.170	ng/u1	0.00
49) Acenaphthylene-d8	14.721	160	114812	21.107	ng/u1	0.00
54) 4-Nitrophenol-d4	15.174	143	14540	17.621	ng/u1	0.00
60) Fluorene-d10	16.002	176	93977	21.227	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	16.096	200	16622	17.180	ng/u1	0.00
73) Anthracene-d10	17.853	188	150183	21.407	ng/u1	0.00
81) Pyrene-d10	20.109	212	179546	19.657	ng/u1	0.00
92) Benzo(a)pyrene-d12	25.338	264	169413	21.745	ng/u1	-0.01

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG030123\
Data File : BG056770.D
Acq On : 1 Mar 2023 19:12
Operator : CG/JU
Sample : 01649-11
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
DBZV8

Quant Time: Mar 02 00:15:53 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG022323.M
Quant Title : SVOA CALIBRATION
QLast Update : Thu Mar 02 00:12:28 2023
Response via : Initial Calibration

