

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG030123\
 Data File : BG056768.D
 Acq On : 1 Mar 2023 17:49
 Operator : CG/JU
 Sample : 01649-03MSD
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :

BNA_G

ClientSampleId :

DBZV0MSD

Manual Integrations

APPROVED

Reviewed By :Yogesh
 Patel

03/02/2023

Supervised By :Jagrut
 Upadhyay

03/02/2023

Quant Time: Mar 02 00:15:19 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG022323.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Mar 02 00:12:28 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.416	152	13243	20.000	ng/u1	0.00
20) Naphthalene-d8	11.260	136	56578	20.000	ng/u1	0.00
38) Acenaphthene-d10	15.020	164	43547	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.758	188	107046m	20.000	ng/u1	0.00
79) Chrysene-d12	22.065	240	96742	20.000	ng/u1	0.00
88) Perylene-d12	25.578	264	106705	20.000	ng/u1	0.00

System Monitoring Compounds

3) 1,4-Dioxane-d8	3.698	96	1579	6.614	ng/uL	0.00
4) Pyridine-d5	4.133	84	20546	21.322	ng/u1	0.00
7) Phenol-d5	7.541	99	40455	30.527	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.723	67	24024	29.719	ng/u1	0.00
11) 2-Chlorophenol-d4	7.940	132	26564	30.185	ng/u1	0.00
15) 4-Methylphenol-d8	9.109	113	29850	29.479	ng/u1	0.00
21) Nitrobenzene-d5	9.591	128	13713	28.921	ng/u1	0.00
24) 2-Nitrophenol-d4	10.320	143	16755	29.687	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.860	165	34365	32.320	ng/u1	0.00
31) 4-Chloroaniline-d4	11.377	131	34338	24.760	ng/u1	0.00
46) Dimethylphthalate-d6	14.409	166	108937	31.005	ng/u1	0.00
49) Acenaphthylene-d8	14.721	160	122133	30.585	ng/u1	0.00
54) 4-Nitrophenol-d4	15.179	143	17608	29.068	ng/u1	0.00
60) Fluorene-d10	16.001	176	101588	31.257	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	16.107	200	21215	29.623	ng/u1	0.00
73) Anthracene-d10	17.852	188	154385	29.728	ng/u1	0.00
81) Pyrene-d10	20.114	212	186823	28.539	ng/u1	0.00
92) Benzo(a)pyrene-d12	25.343	264	179102	31.097	ng/u1	0.00

Target Compounds

					Qvalue
2) 1,4-Dioxane	3.739	88	5561m	18.892	ng/uL
5) Pyridine	4.156	79	41482	40.649	ng/u1 93
6) Benzaldehyde	7.541	77	35337	50.897	ng/u1 97
8) Phenol	7.570	94	59053	43.687	ng/u1 96
10) Bis(2-Chloroethyl)ether	7.817	93	41547	43.259	ng/u1 94
12) 2-Chlorophenol	7.976	128	38149	43.773	ng/u1 96
13) 2-Methylphenol	8.845	108	42185	41.374	ng/u1 96
14) 2,2'-oxybis(1-Chloropr...	8.939	45	53083	41.131	ng/u1 97
16) Acetophenone	9.245	105	71994	41.519	ng/u1 98
17) N-Nitroso-di-n-propyla...	9.221	70	41991	42.205	ng/u1# 94
18) 4-Methylphenol	9.174	108	46185	41.687	ng/u1 95
19) Hexachloroethane	9.521	117	15594	42.077	ng/u1 91
22) Nitrobenzene	9.632	77	61008	44.372	ng/u1 97
23) Isophorone	10.155	82	114401	42.559	ng/u1 98
25) 2-Nitrophenol	10.355	139	24188	43.416	ng/u1 97
26) 2,4-Dimethylphenol	10.390	107	47107	37.861	ng/u1 96
27) Bis(2-Chloroethoxy)met...	10.631	93	59618	43.879	ng/u1# 98
29) 2,4-Dichlorophenol	10.890	162	45571	44.259	ng/u1 96
30) Naphthalene	11.307	128	134481	43.182	ng/u1 97
32) 4-Chloroaniline	11.401	127	45085	32.819	ng/u1 93
33) Hexachlorobutadiene	11.583	225	37087	45.194	ng/u1 94
34) Caprolactam	12.147	113	14881	42.513	ng/u1# 88
35) 4-Chloro-3-methylphenol	12.482	107	50711	42.871	ng/u1 98
36) 2-Methylnaphthalene	12.882	142	95341	42.747	ng/u1 100

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG030123\
 Data File : BG056768.D
 Acq On : 1 Mar 2023 17:49
 Operator : CG/JU
 Sample : 01649-03MSD
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 DBZV0MSD

Manual Integrations
 APPROVED


Reviewed By :Yogesh
 Patel

03/02/2023
 Supervised By :Jagrut
 Upadhyay

Quant Time: Mar 02 00:15:19 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG022323.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Mar 02 00:12:28 2023
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	13.099	142	99437	45.218	ng/ul	96
39) 1,2,4,5-Tetrachloroben...	13.240	216	71089	46.044	ng/ul	97
40) Hexachlorocyclopentadiene	13.216	237	39895	36.062	ng/ul	96
41) 2,4,6-Trichlorophenol	13.463	196	45706	44.305	ng/ul	96
42) 2,4,5-Trichlorophenol	13.534	196	50772	45.325	ng/ul	93
43) 1,1'-Biphenyl	13.863	154	137938	43.076	ng/ul	99
44) 2-Chloronaphthalene	13.916	162	116049	43.801	ng/ul	95
45) 2-Nitroaniline	14.098	65	39710	43.180	ng/ul	94
47) Dimethylphthalate	14.456	163	151828	42.493	ng/ul#	99
48) 2,6-Dinitrotoluene	14.580	165	32176	43.930	ng/ul	94
50) Acenaphthylene	14.750	152	174617	43.285	ng/ul	99
51) 3-Nitroaniline	14.909	138	28685	43.759	ng/ul#	95
52) Acenaphthene	15.085	153	121255	43.754	ng/ul	97
53) 2,4-Dinitrophenol	15.114	184	20071	42.775	ng/ul	95
55) 4-Nitrophenol	15.196	109	30079	42.898	ng/ul	95
56) Dibenzofuran	15.414	168	180568	42.923	ng/ul	97
57) 2,4-Dinitrotoluene	15.361	165	45280	43.121	ng/ul	95
58) 2,3,4,6-Tetrachlorophenol	15.631	232	44219	42.420	ng/ul	98
59) Diethylphthalate	15.807	149	150573	42.100	ng/ul	98
61) Fluorene	16.060	166	142488	42.300	ng/ul	94
62) 4-Chlorophenyl-phenyle...	16.043	204	87738	43.974	ng/ul	97
63) 4-Nitroaniline	16.066	138	29667	46.320	ng/ul	92
66) 4,6-Dinitro-2-methylph...	16.119	198	29463	42.732	ng/ul	97
67) N-Nitrosodiphenylamine	16.254	169	128744	43.250	ng/ul	98
68) 4-Bromophenyl-phenylether	16.936	248	59344	43.746	ng/ul	96
69) Hexachlorobenzene	17.059	284	64740	44.421	ng/ul	97
70) Atrazine	17.182	200	46386	35.879	ng/ul	97
71) Pentachlorophenol	17.400	266	33384m	40.341	ng/ul	
72) Phenanthrene	17.799	178	246690m	42.544	ng/ul	
74) Anthracene	17.887	178	244087	41.655	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.833	216	71398	43.599	ng/uL	98
76) Pentachlorobenzene	15.337	250	72754	43.804	ng/uL	97
77) Carbazole	18.146	167	206186	42.054	ng/ul	99
78) Di-n-butylphthalate	18.675	149	240010	42.474	ng/ul	99
80) Fluoranthene	19.779	202	309971	39.161	ng/ul	99
82) Pyrene	20.144	202	307094	39.381	ng/ul	99
83) Butylbenzylphthalate	21.001	149	105757	42.751	ng/ul	100
84) 3,3'-Dichlorobenzidine	21.936	252	100007	41.231	ng/ul	99
85) Benzo(a)anthracene	22.041	228	305357	43.368	ng/ul	98
86) Bis(2-ethylhexyl)phtha...	21.900	149	151411	43.570	ng/ul	99
87) Chrysene	22.112	228	284261	43.347	ng/ul	99
89) Di-n-octyl phthalate	23.216	149	256422	44.202	ng/ul	100
90) Benzo(b)fluoranthene	24.456	252	310093	43.410	ng/ul	99
91) Benzo(k)fluoranthene	24.533	252	306161	44.003	ng/ul	97
93) Benzo(a)pyrene	25.420	252	271005	43.160	ng/ul	100
94) Indeno(1,2,3-cd)pyrene	29.650	276	368445	43.195	ng/ul	97
95) Dibenzo(a,h)anthracene	29.721	278	305965	42.901	ng/ul#	94
96) Benzo(g,h,i)perylene	30.925	276	294680	43.250	ng/ul#	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed



#2
1,4-Dioxane
Concen: 18.892 ng/uL m
RT: 3.739 min Scan# 34
Delta R.T. -0.001 min
Lab File: BG056768.D
Acq: 1 Mar 2023 17:49

Instrument :
BNA_G
ClientSampleId :
DBZV0MSD






Tgt Ion: 88 Resp: 556
Ion Ratio Lower Upper
88 100
43 32.6 28.3 42.5
58 91.0 50.6 75.8

Manual Integrations
APPROVED

Reviewed By :Yogesh
Patel




03/02/2023
Supervised By :Jagrut
Upadhyay


03/02/2023



#5
Pyridine
Concen: 40.649 ng/uL
RT: 4.156 min Scan# 105
Delta R.T. -0.001 min
Lab File: BG056768.D
Acq: 1 Mar 2023 17:49


Tgt Ion: 79 Resp: 41482
Ion Ratio Lower Upper
79 100
52 69.5 60.6 91.0
51 36.2 31.9 47.9





#6
Benzaldehyde
Concen: 50.897 ng/ul
RT: 7.541 min Scan# 681
Delta R.T. -0.001 min
Lab File: BG056768.D
Acq: 1 Mar 2023 17:49

Instrument :
BNA_G
ClientSampleId :
DBZV0MSD






Tgt Ion: 77 Resp: 3533
Ion Ratio Lower Upper
77 100
105 85.0 72.2 108.4
106 83.4 67.4 101.2

Manual Integrations
APPROVED

Reviewed By :Yogesh
Patel




03/02/2023
Supervised By :Jagrut
Upadhyay


03/02/2023



#8
Phenol
Concen: 43.687 ng/ul
RT: 7.570 min Scan# 686
Delta R.T. 0.005 min
Lab File: BG056768.D
Acq: 1 Mar 2023 17:49


Tgt Ion: 94 Resp: 59053
Ion Ratio Lower Upper
94 100
65 36.8 25.7 38.5
66 49.2 38.9 58.3





#10
Bis(2-Chloroethyl)ether
Concen: 43.259 ng/ul
RT: 7.817 min Scan# 71
Delta R.T. -0.001 min
Lab File: BG056768.D
Acq: 1 Mar 2023 17:49

Instrument :
BNA_G
ClientSampleId :
DBZV0MSD






Tgt Ion: 93 Resp: 4154
Ion Ratio Lower Upper
93 100
63 77.4 66.6 100.0
95 27.8 25.0 37.4

Manual Integrations
APPROVED


Reviewed By :Yogesh
Patel

03/02/2023
Supervised By :Jagrut
Upadhyay



03/02/2023




#12
2-Chlorophenol
Concen: 43.773 ng/ul
RT: 7.976 min Scan# 755
Delta R.T. 0.005 min
Lab File: BG056768.D
Acq: 1 Mar 2023 17:49



Tgt Ion:128 Resp: 38149
Ion Ratio Lower Upper
128 100
64 58.6 45.5 68.3
130 31.2 27.7 41.5





#13
2-Methylphenol
Concen: 41.374 ng/ul
RT: 8.845 min Scan# 902
Delta R.T. 0.005 min
Lab File: BG056768.D
Acq: 1 Mar 2023 17:49

Instrument :
BNA_G
ClientSampleId :
DBZV0MSD





Tgt Ion:108 Resp: 42185
Ion Ratio Lower Upper
108 100
107 87.9 73.4 110.0

Manual Integrations
APPROVED

Reviewed By :Yogesh
Patel




03/02/2023
Supervised By :Jagrut
Upadhyay


03/02/2023



#14
2,2'-oxybis(1-Chloropropane)
Concen: 41.131 ng/ul
RT: 8.939 min Scan# 919
Delta R.T. 0.005 min
Lab File: BG056768.D
Acq: 1 Mar 2023 17:49


Tgt Ion: 45 Resp: 53083
Ion Ratio Lower Upper
45 100
77 15.5 12.4 18.6
79 9.9 9.8 14.6





#16
Acetophenone
Concen: 41.519 ng/ul
RT: 9.245 min Scan# 91
Delta R.T. 0.005 min
Lab File: BG056768.D
Acq: 1 Mar 2023 17:49

Instrument :
BNA_G
ClientSampleId :
DBZV0MSD






Tgt Ion: 105 Resp: 71994
Ion Ratio Lower Upper
105 100
77 88.1 71.4 107.0
51 31.6 23.6 35.4

Manual Integrations
APPROVED


Reviewed By :Yogesh
Patel

03/02/2023
Supervised By :Jagrut
Upadhyay



03/02/2023




#17
N-Nitroso-di-n-propylamine
Concen: 42.205 ng/ul
RT: 9.221 min Scan# 967
Delta R.T. 0.005 min
Lab File: BG056768.D
Acq: 1 Mar 2023 17:49



Tgt Ion: 70 Resp: 41991
Ion Ratio Lower Upper
70 100
42 54.4 46.6 70.0
101 7.6 7.3 10.9
130 12.9 13.7 20.5#





#18
4-Methylphenol
Concen: 41.687 ng/ul
RT: 9.174 min Scan# 91
Delta R.T. 0.005 min
Lab File: BG056768.D
Acq: 1 Mar 2023 17:49

Instrument :
BNA_G
ClientSampleId :
DBZV0MSD





Tgt Ion:108 Resp: 46185
Ion Ratio Lower Upper
108 100
107 110.1 92.1 138.1

Manual Integrations
APPROVED

Reviewed By :Yogesh
Patel




03/02/2023
Supervised By :Jagrut
Upadhyay


03/02/2023



#19
Hexachloroethane
Concen: 42.077 ng/ul
RT: 9.521 min Scan# 1018
Delta R.T. -0.001 min
Lab File: BG056768.D
Acq: 1 Mar 2023 17:49

Tgt Ion:117 Resp: 15594
Ion Ratio Lower Upper
117 100
201 131.3 95.2 142.8
199 76.0 57.9 86.9





#22

Nitrobenzene

Concen: 44.372 ng/ul

RT: 9.632 min Scan# 1037

Delta R.T. -0.001 min

Lab File: BG056768.D

Acq: 1 Mar 2023 17:49

Instrument :

BNA_G

ClientSampleId :

DBZV0MSD

Tgt Ion: 77	Resp: 61008
Ion Ratio	Lower Upper
77 100	
123 36.4	27.8 41.6
65 14.5	10.6 16.0





Manual Integrations
APPROVED

Reviewed By :Yogesh
Patel

03/02/2023

Supervised By :Jagrut
Upadhyay

03/02/2023



#23

Isophorone

Concen: 42.559 ng/ul




RT: 10.155 min Scan# 1126


Delta R.T. -0.001 min

Lab File: BG056768.D

Acq: 1 Mar 2023 17:49

Tgt Ion: 82	Resp: 114401
Ion Ratio	Lower Upper
82 100	
95 10.0	7.7 11.5
138 16.2	12.1 18.1






#25
2-Nitrophenol
Concen: 43.416 ng/ul
RT: 10.355 min Scan# 1160
Delta R.T. -0.001 min
Lab File: BG056768.D
Acq: 1 Mar 2023 17:49

Instrument :

BNA_G

ClientSampleId :

DBZV0MSD



Tgt Ion:139 Resp: 24188
Ion Ratio Lower Upper
139 100
65 60.8 49.4 74.2
109 43.5 31.6 47.4




Manual Integrations
APPROVED

Reviewed By :Yogesh
Patel

03/02/2023




Supervised By :Jagrut
Upadhyay


03/02/2023



#26
2,4-Dimethylphenol
Concen: 37.861 ng/ul
RT: 10.390 min Scan# 1166
Delta R.T. -0.001 min
Lab File: BG056768.D
Acq: 1 Mar 2023 17:49


Tgt Ion:107 Resp: 47107
Ion Ratio Lower Upper
107 100
121 44.9 40.2 60.4
122 77.2 63.1 94.7





#27
 Bis(2-Chloroethoxy)methane
 Concen: 43.879 ng/ul
 RT: 10.631 min Scan# 1207
 Delta R.T. -0.001 min
 Lab File: BG056768.D
 Acq: 1 Mar 2023 17:49

Instrument :
 BNA_G
 ClientSampleId :
 DBZV0MSD






Tgt Ion: 93 Resp: 59618
 Ion Ratio Lower Upper
 93 100
 95 31.7 24.6 37.0
 123 10.2 6.6 10.0

Manual Integrations
 APPROVED

Reviewed By :Yogesh
 Patel




03/02/2023
 Supervised By :Jagrut
 Upadhyay


03/02/2023



#29
 2,4-Dichlorophenol
 Concen: 44.259 ng/ul
 RT: 10.890 min Scan# 1251
 Delta R.T. 0.005 min
 Lab File: BG056768.D
 Acq: 1 Mar 2023 17:49


Tgt Ion:162 Resp: 45571
 Ion Ratio Lower Upper
 162 100
 164 69.3 53.3 79.9
 98 34.7 25.5 38.3





#30
Naphthalene
Concen: 43.182 ng/ul
RT: 11.307 min Scan# 1322
Delta R.T. -0.001 min
Lab File: BG056768.D
Acq: 1 Mar 2023 17:49

Instrument :
BNA_G
ClientSampleId :
DBZV0MSD






Tgt Ion:128 Resp: 13448
Ion Ratio Lower Upper
128 100
129 12.6 8.6 12.8
127 12.9 9.8 14.8

Manual Integrations
APPROVED

Reviewed By :Yogesh
Patel




03/02/2023
Supervised By :Jagrut
Upadhyay


03/02/2023



#32
4-Chloroaniline
Concen: 32.819 ng/ul
RT: 11.401 min Scan# 1338
Delta R.T. -0.001 min
Lab File: BG056768.D
Acq: 1 Mar 2023 17:49

Tgt Ion:127 Resp: 45085
Ion Ratio Lower Upper
127 100
129 33.2 23.4 35.0





#33
Hexachlorobutadiene
Concen: 45.194 ng/ul
RT: 11.583 min Scan# 1369
Delta R.T. -0.001 min
Lab File: BG056768.D
Acq: 1 Mar 2023 17:49

Instrument :

BNA_G

ClientSampleId :

DBZV0MSD

Tgt Ion:225 Resp: 3708
Ion Ratio Lower Upper
225 100
223 62.0 53.5 80.3
227 68.3 58.2 87.2





Manual Integrations
APPROVED

Reviewed By :Yogesh
Patel

03/02/2023




Supervised By :Jagrut
Upadhyay


03/02/2023



#34
Caprolactam
Concen: 42.513 ng/ul
RT: 12.147 min Scan# 1465
Delta R.T. 0.011 min
Lab File: BG056768.D
Acq: 1 Mar 2023 17:49

Tgt Ion:113 Resp: 14881
Ion Ratio Lower Upper
113 100
55 203.8 167.0 250.6
56 172.0 113.5 170.3#






#35
4-Chloro-3-methylphenol
Concen: 42.871 ng/ul
RT: 12.482 min Scan# 1522
Delta R.T. -0.001 min
Lab File: BG056768.D
Acq: 1 Mar 2023 17:49

Instrument :

BNA_G

ClientSampleId :

DBZV0MSD



Tgt Ion:107 Resp: 50711
Ion Ratio Lower Upper
107 100
144 23.0 20.5 30.7
142 69.5 55.8 83.6




Manual Integrations
APPROVED

Reviewed By :Yogesh
Patel

03/02/2023




Supervised By :Jagrut
Upadhyay


03/02/2023



#36
2-Methylnaphthalene
Concen: 42.747 ng/ul
RT: 12.882 min Scan# 1590
Delta R.T. 0.005 min
Lab File: BG056768.D
Acq: 1 Mar 2023 17:49


Tgt Ion:142 Resp: 95341
Ion Ratio Lower Upper
142 100
141 89.1 71.0 106.6





#37
1-Methylnaphthalene
Concen: 45.218 ng/ul
RT: 13.099 min Scan# 1627
Delta R.T. 0.005 min
Lab File: BG056768.D
Acq: 1 Mar 2023 17:49

Instrument :
BNA_G
ClientSampleId :
DBZV0MSD






Tgt Ion:142 Resp: 9943
Ion Ratio Lower Upper
142 100
141 95.5 73.5 110.3
116 4.4 4.0 6.0

Manual Integrations
APPROVED

Reviewed By :Yogesh
Patel




03/02/2023
Supervised By :Jagrut
Upadhyay


03/02/2023



#39
1,2,4,5-Tetrachlorobenzene
Concen: 46.044 ng/ul
RT: 13.240 min Scan# 1651
Delta R.T. 0.005 min
Lab File: BG056768.D
Acq: 1 Mar 2023 17:49


Tgt Ion:216 Resp: 71089
Ion Ratio Lower Upper
216 100
214 83.5 64.4 96.6
179 19.5 16.9 25.3
108 13.8 11.6 17.4





#40
Hexachlorocyclopentadiene
Concen: 36.062 ng/ul
RT: 13.216 min Scan# 1647
Delta R.T. -0.001 min
Lab File: BG056768.D
Acq: 1 Mar 2023 17:49

Instrument :
BNA_G
ClientSampleId :
DBZV0MSD






Tgt Ion:237 Resp: 39895
Ion Ratio Lower Upper
237 100
235 63.7 47.8 71.8
272 11.9 9.9 14.9

Manual Integrations
APPROVED

Reviewed By :Yogesh
Patel




03/02/2023
Supervised By :Jagrut
Upadhyay


03/02/2023



#41
2,4,6-Trichlorophenol
Concen: 44.305 ng/ul
RT: 13.463 min Scan# 1689
Delta R.T. -0.001 min
Lab File: BG056768.D
Acq: 1 Mar 2023 17:49

Tgt Ion:196 Resp: 45706
Ion Ratio Lower Upper
196 100
198 97.1 75.0 112.6
200 31.2 22.1 33.1





#42
2,4,5-Trichlorophenol
Concen: 45.325 ng/ul
RT: 13.534 min Scan# 1701
Delta R.T. -0.001 min
Lab File: BG056768.D
Acq: 1 Mar 2023 17:49

Instrument :

BNA_G

ClientSampleId :

DBZV0MSD

Tgt Ion:196 Resp: 50772
Ion Ratio Lower Upper
196 100
198 92.4 80.1 120.1
200 31.1 26.7 40.1





Manual Integrations
APPROVED

Reviewed By :Yogesh
Patel

03/02/2023




Supervised By :Jagrut
Upadhyay


03/02/2023



#43
1,1'-Biphenyl
Concen: 43.076 ng/ul
RT: 13.863 min Scan# 1757
Delta R.T. -0.001 min
Lab File: BG056768.D
Acq: 1 Mar 2023 17:49


Tgt Ion:154 Resp: 137938
Ion Ratio Lower Upper
154 100
153 40.5 32.0 48.0
76 12.4 9.0 13.4





#44
2-Chloronaphthalene
Concen: 43.801 ng/ul
RT: 13.916 min Scan# 1766
Delta R.T. 0.005 min
Lab File: BG056768.D
Acq: 1 Mar 2023 17:49

Instrument :
BNA_G
ClientSampleId :
DBZV0MSD






Tgt Ion:162 Resp: 116049
Ion Ratio Lower Upper
162 100
164 33.3 23.3 34.9
127 33.9 25.8 38.8

Manual Integrations
APPROVED

Reviewed By :Yogesh
Patel




03/02/2023
Supervised By :Jagrut
Upadhyay


03/02/2023



#45
2-Nitroaniline
Concen: 43.180 ng/ul
RT: 14.098 min Scan# 1797
Delta R.T. -0.001 min
Lab File: BG056768.D
Acq: 1 Mar 2023 17:49


Tgt Ion: 65 Resp: 39710
Ion Ratio Lower Upper
65 100
92 64.4 53.9 80.9
138 85.0 73.8 110.6





#47
Dimethylphthalate
Concen: 42.493 ng/ul
RT: 14.456 min Scan# 1858
Delta R.T. -0.001 min
Lab File: BG056768.D
Acq: 1 Mar 2023 17:49

Instrument :
BNA_G
ClientSampleId :
DBZV0MSD






Tgt Ion:163 Resp: 151828
Ion Ratio Lower Upper
163 100
194 4.4 2.5 3.7
164 10.5 8.2 12.4

Manual Integrations
APPROVED

Reviewed By :Yogesh
Patel




03/02/2023
Supervised By :Jagrut
Upadhyay


03/02/2023



#48
2,6-Dinitrotoluene
Concen: 43.930 ng/ul
RT: 14.580 min Scan# 1879
Delta R.T. -0.001 min
Lab File: BG056768.D
Acq: 1 Mar 2023 17:49


Tgt Ion:165 Resp: 32176
Ion Ratio Lower Upper
165 100
89 65.7 48.4 72.6
121 21.9 15.5 23.3





#50
Acenaphthylene
Concen: 43.285 ng/ul
RT: 14.750 min Scan# 1908
Delta R.T. -0.001 min
Lab File: BG056768.D
Acq: 1 Mar 2023 17:49

Instrument :
BNA_G
ClientSampleId :
DBZV0MSD





Tgt Ion:152 Resp: 17461
Ion Ratio Lower Upper
152 100
151 20.2 15.8 23.6
153 12.5 10.6 16.0

Manual Integrations
APPROVED

Reviewed By :Yogesh
Patel




03/02/2023
Supervised By :Jagrut
Upadhyay


03/02/2023



#51
3-Nitroaniline
Concen: 43.759 ng/ul
RT: 14.909 min Scan# 1935
Delta R.T. -0.001 min
Lab File: BG056768.D
Acq: 1 Mar 2023 17:49


Tgt Ion:138 Resp: 28685
Ion Ratio Lower Upper
138 100
108 10.5 6.5 9.7#
92 133.4 102.2 153.2





#52
Acenaphthene
Concen: 43.754 ng/ul
RT: 15.085 min Scan# 1965
Delta R.T. -0.001 min
Lab File: BG056768.D
Acq: 1 Mar 2023 17:49

Instrument :
BNA_G
ClientSampleId :
DBZV0MSD





Tgt Ion:153 Resp: 121255
Ion Ratio Lower Upper
153 100
152 44.7 38.6 58.0
154 86.6 70.9 106.3

Manual Integrations
APPROVED

Reviewed By :Yogesh
Patel




03/02/2023
Supervised By :Jagrut
Upadhyay


03/02/2023



#53
2,4-Dinitrophenol
Concen: 42.775 ng/ul
RT: 15.114 min Scan# 1970
Delta R.T. 0.005 min
Lab File: BG056768.D
Acq: 1 Mar 2023 17:49


Tgt Ion:184 Resp: 20071
Ion Ratio Lower Upper
184 100
63 59.7 43.8 65.6
154 80.8 66.8 100.2





#55
4-Nitrophenol
Concen: 42.898 ng/ul
RT: 15.196 min Scan# 1983
Delta R.T. 0.005 min
Lab File: BG056768.D
Acq: 1 Mar 2023 17:49

Instrument :
BNA_G
ClientSampleId :
DBZV0MSD






Tgt Ion:109 Resp: 30079
Ion Ratio Lower Upper
109 100
139 82.3 71.2 106.8
65 109.5 85.1 127.7

Manual Integrations
APPROVED


Reviewed By :Yogesh
Patel

03/02/2023
Supervised By :Jagrut
Upadhyay



03/02/2023




#56
Dibenzofuran
Concen: 42.923 ng/ul
RT: 15.414 min Scan# 2021
Delta R.T. -0.001 min
Lab File: BG056768.D
Acq: 1 Mar 2023 17:49




Tgt Ion:168 Resp: 180568
Ion Ratio Lower Upper
168 100
139 36.9 28.2 42.2





#57
2,4-Dinitrotoluene
Concen: 43.121 ng/ul
RT: 15.361 min Scan# 2012
Delta R.T. -0.001 min
Lab File: BG056768.D
Acq: 1 Mar 2023 17:49

Instrument :
BNA_G
ClientSampleId :
DBZV0MSD






Tgt Ion:165 Resp: 45280
Ion Ratio Lower Upper
165 100
63 51.6 38.6 58.0
182 2.0 1.8 2.8

Manual Integrations
APPROVED

Reviewed By :Yogesh
Patel




03/02/2023
Supervised By :Jagrut
Upadhyay


03/02/2023



#58
2,3,4,6-Tetrachlorophenol
Concen: 42.420 ng/ul
RT: 15.631 min Scan# 2058
Delta R.T. -0.001 min
Lab File: BG056768.D
Acq: 1 Mar 2023 17:49

Tgt Ion:232 Resp: 44219
Ion Ratio Lower Upper
232 100
131 35.8 27.8 41.8
130 1.8 1.8 2.6
166 26.1 20.2 30.4






#59
Diethylphthalate
Concen: 42.100 ng/ul
RT: 15.807 min Scan# 2087
Delta R.T. 0.005 min
Lab File: BG056768.D
Acq: 1 Mar 2023 17:49

Instrument :

BNA_G

ClientSampleId :

DBZV0MSD



Tgt Ion:149 Resp: 150573
Ion Ratio Lower Upper
149 100
177 22.5 18.6 28.0
150 12.5 10.7 16.1




Manual Integrations
APPROVED

Reviewed By :Yogesh
Patel

03/02/2023




Supervised By :Jagrut
Upadhyay


03/02/2023



#61
Fluorene
Concen: 42.300 ng/ul
RT: 16.060 min Scan# 2131
Delta R.T. -0.001 min
Lab File: BG056768.D
Acq: 1 Mar 2023 17:49

Tgt Ion:166 Resp: 142488
Ion Ratio Lower Upper
166 100
165 104.0 78.4 117.6
167 15.4 10.5 15.7





#62
4-Chlorophenyl-phenylether
Concen: 43.974 ng/ul
RT: 16.043 min Scan# 2128
Delta R.T. -0.001 min
Lab File: BG056768.D
Acq: 1 Mar 2023 17:49

Instrument :
BNA_G
ClientSampleId :
DBZV0MSD





Tgt Ion:204 Resp: 87738
Ion Ratio Lower Upper
204 100
206 31.7 26.1 39.1
141 47.0 39.7 59.5

Manual Integrations
APPROVED

Reviewed By :Yogesh
Patel




03/02/2023
Supervised By :Jagrut
Upadhyay


03/02/2023



#63
4-Nitroaniline
Concen: 46.320 ng/ul
RT: 16.066 min Scan# 2132
Delta R.T. 0.005 min
Lab File: BG056768.D
Acq: 1 Mar 2023 17:49

Tgt Ion:138 Resp: 29667
Ion Ratio Lower Upper
138 100
92 55.1 48.6 73.0
108 126.0 93.0 139.6





#66
 4,6-Dinitro-2-methylphenol
 Concen: 42.732 ng/ul
 RT: 16.119 min Scan# 2141
 Delta R.T. -0.001 min
 Lab File: BG056768.D
 Acq: 1 Mar 2023 17:49

Instrument :

BNA_G

ClientSampleId :

DBZV0MSD

Tgt Ion:198 Resp: 29463
 Ion Ratio Lower Upper
 198 100
 182 3.6 2.4 3.6
 77 24.6 21.0 31.6





Manual Integrations
 APPROVED

Reviewed By :Yogesh
 Patel

03/02/2023




Supervised By :Jagrut
 Upadhyay


03/02/2023



#67
 N-Nitrosodiphenylamine
 Concen: 43.250 ng/ul
 RT: 16.254 min Scan# 2164
 Delta R.T. 0.005 min
 Lab File: BG056768.D
 Acq: 1 Mar 2023 17:49


Tgt Ion:169 Resp: 128744
 Ion Ratio Lower Upper
 169 100
 168 68.4 55.1 82.7
 167 37.8 28.3 42.5





#68
4-Bromophenyl-phenylether
Concen: 43.746 ng/ul
RT: 16.936 min Scan# 21
Delta R.T. -0.001 min
Lab File: BG056768.D
Acq: 1 Mar 2023 17:49

Instrument :
BNA_G
ClientSampleId :
DBZV0MSD






Tgt Ion:248 Resp: 59344
Ion Ratio Lower Upper
248 100
250 95.7 80.4 120.6
141 57.9 47.6 71.4

Manual Integrations
APPROVED

Reviewed By :Yogesh
Patel




03/02/2023
Supervised By :Jagrut
Upadhyay


03/02/2023



#69
Hexachlorobenzene
Concen: 44.421 ng/ul
RT: 17.059 min Scan# 2301
Delta R.T. -0.001 min
Lab File: BG056768.D
Acq: 1 Mar 2023 17:49

Tgt Ion:284 Resp: 64740
Ion Ratio Lower Upper
284 100
142 26.5 21.8 32.6
249 29.4 25.9 38.9





#70

Atrazine

Concen: 35.879 ng/ul

RT: 17.182 min Scan# 2321

Delta R.T. 0.005 min

Lab File: BG056768.D

Acq: 1 Mar 2023 17:49

Instrument :

BNA_G

ClientSampleId :

DBZV0MSD

Tgt Ion:200 Resp: 46386
 Ion Ratio Lower Upper
 200 100
 173 20.9 17.4 26.0
 215 49.2 41.7 62.5





Manual Integrations
 APPROVED

Reviewed By :Yogesh
 Patel

03/02/2023

Supervised By :Jagrut
 Upadhyay

03/02/2023



#71

Pentachlorophenol

Concen: 40.341 ng/ul m




RT: 17.400 min Scan# 2359


Delta R.T. 0.005 min

Lab File: BG056768.D

Acq: 1 Mar 2023 17:49


Tgt Ion:266 Resp: 33384
 Ion Ratio Lower Upper
 266 100
 264 62.7 49.9 74.9
 268 63.0 54.5 81.7





#72
Phenanthrene
Concen: 42.544 ng/ul m
RT: 17.799 min Scan# 2427
Delta R.T. -0.001 min
Lab File: BG056768.D
Acq: 1 Mar 2023 17:49

Instrument :
BNA_G
ClientSampleId :
DBZV0MSD






Tgt Ion:178 Resp: 246690
Ion Ratio Lower Upper
178 100
179 15.6 12.1 18.1
176 18.1 14.4 21.6

Manual Integrations
APPROVED

Reviewed By :Yogesh
Patel




03/02/2023
Supervised By :Jagrut
Upadhyay


03/02/2023



#74
Anthracene
Concen: 41.655 ng/ul
RT: 17.887 min Scan# 2442
Delta R.T. -0.001 min
Lab File: BG056768.D
Acq: 1 Mar 2023 17:49


Tgt Ion:178 Resp: 244087
Ion Ratio Lower Upper
178 100
179 15.4 12.1 18.1
176 18.4 14.1 21.1





#75
1,2,3,4-Tetrachlorobenzene
Concen: 43.599 ng/uL
RT: 13.833 min Scan# 1752
Delta R.T. -0.001 min
Lab File: BG056768.D
Acq: 1 Mar 2023 17:49

Instrument :
BNA_G
ClientSampleId :
DBZV0MSD






Tgt Ion:216 Resp: 71398
Ion Ratio Lower Upper
216 100
214 78.5 64.2 96.2
218 47.4 36.6 54.8

Manual Integrations
APPROVED


Reviewed By :Yogesh
Patel

03/02/2023
Supervised By :Jagrut
Upadhyay



03/02/2023




#76
Pentachlorobenzene
Concen: 43.804 ng/uL
RT: 15.337 min Scan# 2008
Delta R.T. -0.001 min
Lab File: BG056768.D
Acq: 1 Mar 2023 17:49




Tgt Ion:250 Resp: 72754
Ion Ratio Lower Upper
250 100
248 64.2 47.1 70.7
252 65.4 52.4 78.6





#77
 Carbazole
 Concen: 42.054 ng/ul
 RT: 18.146 min Scan# 2486
 Delta R.T. -0.001 min
 Lab File: BG056768.D
 Acq: 1 Mar 2023 17:49

Instrument :
 BNA_G
 ClientSampleId :
 DBZV0MSD






Tgt Ion:167 Resp: 206186
 Ion Ratio Lower Upper
 167 100
 166 22.0 18.2 27.2
 139 12.0 9.8 14.8

Manual Integrations
 APPROVED


Reviewed By :Yogesh
 Patel

03/02/2023
 Supervised By :Jagrut
 Upadhyay



03/02/2023




#78
 Di-n-butylphthalate
 Concen: 42.474 ng/ul
 RT: 18.675 min Scan# 2576
 Delta R.T. -0.001 min
 Lab File: BG056768.D
 Acq: 1 Mar 2023 17:49




Tgt Ion:149 Resp: 240010
 Ion Ratio Lower Upper
 149 100
 150 8.8 7.4 11.0
 104 6.6 5.4 8.2





#80
Fluoranthene
Concen: 39.161 ng/u1
RT: 19.779 min Scan# 21
Delta R.T. -0.001 min
Lab File: BG056768.D
Acq: 1 Mar 2023 17:49

Instrument :
BNA_G
ClientSampleId :
DBZV0MSD





Tgt Ion:202 Resp: 30997
Ion Ratio Lower Upper
202 100
101 5.4 4.7 7.1
100 5.1 4.4 6.6

Manual Integrations
APPROVED

Reviewed By :Yogesh
Patel




03/02/2023
Supervised By :Jagrut
Upadhyay


03/02/2023



#82
Pyrene
Concen: 39.381 ng/u1
RT: 20.144 min Scan# 2826
Delta R.T. 0.005 min
Lab File: BG056768.D
Acq: 1 Mar 2023 17:49

Tgt Ion:202 Resp: 307094
Ion Ratio Lower Upper
202 100
101 6.2 5.3 7.9
100 5.7 4.6 7.0






#83
Butylbenzylphthalate
Concen: 42.751 ng/ul
RT: 21.001 min Scan# 2972
Delta R.T. -0.001 min
Lab File: BG056768.D
Acq: 1 Mar 2023 17:49

Instrument :

BNA_G

ClientSampleId :

DBZV0MSD



Tgt Ion:149 Resp: 105752
Ion Ratio Lower Upper
149 100
91 74.6 59.3 88.9
206 24.6 19.7 29.5




Manual Integrations
APPROVED

Reviewed By :Yogesh
Patel

03/02/2023




Supervised By :Jagrut
Upadhyay


03/02/2023



#84
3,3'-Dichlorobenzidine
Concen: 41.231 ng/ul
RT: 21.936 min Scan# 3131
Delta R.T. -0.001 min
Lab File: BG056768.D
Acq: 1 Mar 2023 17:49


Tgt Ion:252 Resp: 100007
Ion Ratio Lower Upper
252 100
254 64.6 52.6 78.8
126 6.0 5.0 7.6





#85
Benzo(a)anthracene
Concen: 43.368 ng/ul
RT: 22.041 min Scan# 3149
Delta R.T. -0.001 min
Lab File: BG056768.D
Acq: 1 Mar 2023 17:49

Instrument :
BNA_G
ClientSampleId :
DBZV0MSD






Tgt Ion:228 Resp: 305352
Ion Ratio Lower Upper
228 100
229 20.4 16.2 24.4
226 27.3 20.6 31.0

Manual Integrations
APPROVED

Reviewed By :Yogesh
Patel




03/02/2023
Supervised By :Jagrut
Upadhyay


03/02/2023



#86
Bis(2-ethylhexyl)phthalate
Concen: 43.570 ng/ul
RT: 21.900 min Scan# 3125
Delta R.T. -0.007 min
Lab File: BG056768.D
Acq: 1 Mar 2023 17:49


Tgt Ion:149 Resp: 151411
Ion Ratio Lower Upper
149 100
167 26.6 21.8 32.8
279 4.8 4.2 6.2





#87
 Chrysene
 Concen: 43.347 ng/ul
 RT: 22.112 min Scan# 3161
 Delta R.T. -0.001 min
 Lab File: BG056768.D
 Acq: 1 Mar 2023 17:49

Instrument :
 BNA_G
 ClientSampleId :
 DBZV0MSD



Tgt Ion:228 Resp: 284263



Ion	Ratio	Lower	Upper
228	100		
226	29.8	23.6	35.4
229	20.2	15.4	23.0

Manual Integrations
 APPROVED

Reviewed By :Yogesh
 Patel




03/02/2023
 Supervised By :Jagrut
 Upadhyay


03/02/2023



#89
 Di-n-octyl phthalate
 Concen: 44.202 ng/ul
 RT: 23.216 min Scan# 3349
 Delta R.T. -0.001 min
 Lab File: BG056768.D
 Acq: 1 Mar 2023 17:49


Tgt Ion:149 Resp: 256422





#90
Benzo(b)fluoranthene
Concen: 43.410 ng/ul
RT: 24.456 min Scan# 3560
Delta R.T. -0.001 min
Lab File: BG056768.D
Acq: 1 Mar 2023 17:49

Instrument :
BNA_G
ClientSampleId :
DBZV0MSD





Tgt Ion:252 Resp: 310093
Ion Ratio Lower Upper
252 100
253 22.0 17.4 26.2
125 4.3 3.0 4.6

Manual Integrations
APPROVED

Reviewed By :Yogesh
Patel




03/02/2023
Supervised By :Jagrut
Upadhyay


03/02/2023



#91
Benzo(k)fluoranthene
Concen: 44.003 ng/ul
RT: 24.533 min Scan# 3573
Delta R.T. -0.001 min
Lab File: BG056768.D
Acq: 1 Mar 2023 17:49


Tgt Ion:252 Resp: 306161
Ion Ratio Lower Upper
252 100
253 21.4 18.2 27.2
125 3.9 3.4 5.2





#93
Benzo(a)pyrene
Concen: 43.160 ng/ul
RT: 25.420 min Scan# 3724
Delta R.T. -0.001 min
Lab File: BG056768.D
Acq: 1 Mar 2023 17:49

Instrument :
BNA_G
ClientSampleId :
DBZV0MSD





Tgt Ion:252 Resp: 271005
Ion Ratio Lower Upper
252 100
253 21.8 17.4 26.0
125 4.1 3.6 5.4

Manual Integrations
APPROVED

Reviewed By :Yogesh
Patel




03/02/2023
Supervised By :Jagrut
Upadhyay


03/02/2023



#94
Indeno(1,2,3-cd)pyrene
Concen: 43.195 ng/ul
RT: 29.650 min Scan# 4444
Delta R.T. 0.011 min
Lab File: BG056768.D
Acq: 1 Mar 2023 17:49


Tgt Ion:276 Resp: 368445
Ion Ratio Lower Upper
276 100
138 7.7 5.7 8.5
277 24.2 18.1 27.1





#95
Dibenzo(a,h)anthracene
Concen: 42.901 ng/ul
RT: 29.721 min Scan# 4455
Delta R.T. 0.005 min
Lab File: BG056768.D
Acq: 1 Mar 2023 17:49

Instrument :
BNA_G
ClientSampleId :
DBZV0MSD






Tgt Ion:278 Resp: 305965
Ion Ratio Lower Upper
278 100
139 6.6 4.3 6.5
279 23.9 21.7 32.5

Manual Integrations
APPROVED

Reviewed By :Yogesh
Patel

03/02/2023
Supervised By :Jagrut
Upadhyay

03/02/2023



#96
Benzo(g,h,i)perylene
Concen: 43.250 ng/ul
RT: 30.925 min Scan# 4661
Delta R.T. -0.001 min
Lab File: BG056768.D
Acq: 1 Mar 2023 17:49

Tgt Ion:276 Resp: 294680
Ion Ratio Lower Upper
276 100
138 10.0 6.5 9.7#
277 24.2 20.7 31.1

