

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG030123\
 Data File : BG056773.D
 Acq On : 1 Mar 2023 21:16
 Operator : CG/JU
 Sample : 01649-10
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 DBZV7

Quant Time: Mar 02 00:16:14 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG022323.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Mar 02 00:12:28 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc Units	Dev(Min)

Internal Standards					
1) 1,4-Dichlorobenzene-d4	8.423	152	14517	20.000 ng/u1	0.00
20) Naphthalene-d8	11.255	136	61681	20.000 ng/u1	0.00
38) Acenaphthene-d10	15.021	164	50541	20.000 ng/u1	0.00
64) Phenanthrene-d10	17.753	188	124857	20.000 ng/u1	0.00
79) Chrysene-d12	22.059	240	120665	20.000 ng/u1	0.00
88) Perylene-d12	25.585	264	128705	20.000 ng/u1	0.00
System Monitoring Compounds					
3) 1,4-Dioxane-d8	3.699	96	1288	4.922 ng/uL	0.00
4) Pyridine-d5	4.139	84	13597	12.872 ng/u1	0.00
7) Phenol-d5	7.541	99	26902	18.519 ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.717	67	18410	20.776 ng/u1	0.00
11) 2-Chlorophenol-d4	7.941	132	18914	19.606 ng/u1	0.00
15) 4-Methylphenol-d8	9.104	113	20875	18.806 ng/u1	0.00
21) Nitrobenzene-d5	9.592	128	10587	20.481 ng/u1	0.00
24) 2-Nitrophenol-d4	10.320	143	12538	20.377 ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.855	165	23488	20.263 ng/u1	0.00
31) 4-Chloroaniline-d4	11.378	131	27953	18.489 ng/u1	0.00
46) Dimethylphthalate-d6	14.404	166	85029	20.852 ng/u1	0.00
49) Acenaphthylene-d8	14.721	160	100074	21.593 ng/u1	0.00
54) 4-Nitrophenol-d4	15.173	143	11581	16.472 ng/u1	0.00
60) Fluorene-d10	16.002	176	81946	21.725 ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	16.096	200	13956	16.707 ng/u1	0.00
73) Anthracene-d10	17.853	188	134601	22.221 ng/u1	0.00
81) Pyrene-d10	20.109	212	159886	19.581 ng/u1	0.00
92) Benzo(a)pyrene-d12	25.338	264	153104	22.039 ng/u1	-0.01

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG030123\
Data File : BG056773.D
Acq On : 1 Mar 2023 21:16
Operator : CG/JU
Sample : 01649-10
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
DBZV7

Quant Time: Mar 02 00:16:14 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG022323.M
Quant Title : SVOA CALIBRATION
QLast Update : Thu Mar 02 00:12:28 2023
Response via : Initial Calibration

