

Data Path : Z:\SVOASRV\HPCHEM1\BNA G\DATA\BG030220\
 Data File : BG044659.D
 Acq On : 3 Mar 2020 10:02
 Operator : CG/JU
 Sample : L1743-16
 Misc :
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 C-1

Quant Time: Mar 03 18:32:09 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG022420.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Feb 24 16:11:23 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.81	152	76197	20.00	ng	-0.07
21) Naphthalene-d8	10.60	136	353583	20.00	ng	-0.07
39) Acenaphthene-d10	14.46	164	226337	20.00	ng	-0.06
64) Phenanthrene-d10	17.20	188	453346	20.00	ng	-0.06
76) Chrysene-d12	21.44	240	391334	20.00	ng	-0.07
87) Perylene-d12	24.46	264	449968	20.00	ng	-0.11
System Monitoring Compounds						
5) 2-Fluorophenol	5.38	112	461529	97.18	ng	-0.07
7) Phenol-d6	6.98	99	626118	96.32	ng	-0.07
23) Nitrobenzene-d5	8.96	82	390319	56.87	ng	-0.07
42) 2,4,6-Tribromophenol	15.95	330	233475	72.75	ng	-0.06
45) 2-Fluorobiphenyl	13.08	172	868664	54.96	ng	-0.06
79) Terphenyl-d14	19.82	244	1144333	55.36	ng	-0.06
Target Compounds						
50) Dimethylphthalate	13.91	163	96790	5.298	ng	98
84) Bis(2-ethylhexyl)phthalate	21.35	149	403682	23.577	ng	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA G\DATA\BG030220\
Data File : BG044659.D
Acq On : 3 Mar 2020 10:02
Operator : CG/JU
Sample : L1743-16
Misc :
ALS Vial : 26 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
C-1

Quant Time: Mar 03 18:32:09 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG022420.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Feb 24 16:11:23 2020
Response via : Initial Calibration

