




Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG030421\
 Data File : BG048322.D
 Acq On : 4 Mar 2021 8:55
 Operator : CG/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Quant Time: Mar 04 09:47:49 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG021021.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Mar 03 16:40:27 2021
 Response via : Initial Calibration






TIC: BG048322.D\data.ms

(77) Benzidine**19.721min (+ 0.003) 267348.27 ng****response 655157**

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	15.50	14.77
183.00	13.20	12.32
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG030421\
Data File : BG048322.D
Acq On : 4 Mar 2021 8:55
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Mar 04 09:47:49 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG021021.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Mar 03 16:40:27 2021
Response via : Initial Calibration



TIC: BG048322.D\data.ms

(70) Pentachlorophenol (C)

17.159min (+ 0.003) 16337.95 ng

response 67219

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	62.20	60.89
264.00	61.60	57.41
0.00	0.00	0.00