


Data Path : Z:\SVOASRV\HPCHEM1\BNA_G\DATA\BG030520\
 Data File : BG044680.D
 Acq On : 5 Mar 2020 12:08
 Operator : CG/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\8270-BG030520.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Thu Mar 05 17:42:35 2020



AutoFind: Scans 2521, 2522, 2523; Background Corrected with Scan 2514

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	41.9	38546	PASS
68	69	0.00	2	0.9	371	PASS
69	198	0.00	100	43.6	40076	PASS
70	69	0.00	2	0.3	123	PASS
127	198	10	80	47.0	43248	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	91981	PASS
199	198	5	9	6.9	6330	PASS
275	198	10	60	29.7	27296	PASS
365	198	1	100	5.2	4759	PASS
441	198	0.01	100	13.5	12424	PASS
442	198	50	100	78.5	72160	PASS
443	442	15	24	18.4	13293	PASS