




Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM030822\
Data File : BM034152.D
Acq On : 08 Mar 2022 16:48
Operator : CG/JU
Sample : N1733-04
Misc :
ALS Vial : 11 Sample Multiplier: 1

Quant Time: Mar 09 00:31:07 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM030822.M
Quant Title : SVOA CALIBRATION
QLast Update : Tue Mar 08 13:09:16 2022
Response via : Initial Calibration



TIC: BM034152.D\data.ms

(3) 1,4-Dioxane-d8 (S)**3.372min (+ 0.000) 7.33 ng/uL****response 24909**

Ion	Exp%	Act%
96.00	100.00	100.00
64.00	74.20	66.25
34.00	0.00	0.00
0.00	0.00	0.00